

计量授权证书附件

序号 No.	开展商品量/商品 包装计量检验的参数 名称 Parameters of Metrological Testing for Quantity/Package of Products	测量范围 Measurement Range	准确度等级或 测量扩展不确定度 Accuracy or Expanded Measurement Uncertainty	依据检定 规程编号 Number of Verification Specification Referred to
1	质量标注定量包装商 品净含量(Q_n)	(0~50) g (50~100) g (100~200) g (200~300) g (300~500) g (500~1000) g (1000~10000) g (10000~15000) g (15000~50000) g	$\leq Q_n \times 1.8\%$ ≤ 0.9 g $\leq Q_n \times 0.9\%$ ≤ 1.8 g $\leq Q_n \times 0.6\%$ ≤ 3 g $\leq Q_n \times 0.3\%$ ≤ 30 g $\leq Q_n \times 0.2\%$	JJF 1070 JJF 1070.1 JJF 1070.2
2	体积标注定量包装商 品净含量(Q_n)	(0~50) mL (50~100) mL (100~200) mL (200~300) mL (300~500) mL (500~1000) mL (1000~10000) mL (10000~15000) mL (15000~50000) mL	$\leq Q_n \times 1.8\%$ ≤ 0.9 mL $\leq Q_n \times 0.9\%$ ≤ 1.8 mL $\leq Q_n \times 0.6\%$ ≤ 3 mL $\leq Q_n \times 0.3\%$ ≤ 30 mL $\leq Q_n \times 0.2\%$	JJF 1070
3	长度标注定量包装商 品净含量(Q_n)	$Q_n \leq 5$ m $Q_n > 5$ m	$\leq Q_n \times 0.2\%$ $\leq Q_n \times 0.4\%$	JJF 1070
4	面积标注定量包装商 品净含量(Q_n)	全部 Q_n	$\leq Q_n \times 0.6\%$	JJF 1070
5	计数标注定量包装商 品净含量(Q_n)	$Q_n \leq 50$ $Q_n > 50$	--- $\leq Q_n \times 0.2\%$	JJF 1070
6	单件大米商品净含量 (Q_n)	(0~50) g (50~100) g (100~200) g (200~300) g (300~500) g (500~1000) g (1000~10000) g	$\leq Q_n \times 1.8\%$ ≤ 0.9 g $\leq Q_n \times 0.9\%$ ≤ 1.8 g $\leq Q_n \times 0.6\%$ ≤ 3 g $\leq Q_n \times 0.3\%$ ≤ 30 g	JJF 1070.3

序号 No.	开展商品量/商品 包装计量检验的参数 名称 Parameters of Metrological Testing for Quantity/Package of Products	测量范围 Measurement Range	准确度等级或 测量扩展不确定度 Accuracy or Expanded Measurement Uncertainty	依据检定 规程编号 Number of Verification Specification Referred to
		(10000~15000) g (15000~50000) g	$\leq Q_n \times 0.2\%$	
7	零售商品称重	(0~50) g (50~100) g (100~1000)g (1000~4000)g (4000~6000)g (6000~10000)g (10000g~15000)g (15000g~25000)g	≤ 0.006 g ≤ 0.012 g ≤ 1.2 g ≤ 2.3 g ≤ 3.5 g ≤ 5.8 g ≤ 8.7 g ≤ 17 g	JJF 1647

-----以下空白-----