

Name: South China National Centre of Metrology/Guangdong Institute of Metrology

Address: No.152, Shipai Section, Dongyuan Avenue, Shipai, Dongguan, Guangdong, China

Registration No. CNAS L0730

Accreditation Criteria: ISO/IEC 17025:2017 and relevant requirements of CNAS

Effective Date: 2025-01-26 Expiry Date: 2029-10-21

## SCHEDULE 5 ACCREDITED CALIBRATION AND MEASUREMENT CAPABILITY SCOPE

Note: The instruments with \* represents onsite calibration can be performed.

| No                              | Instrument                                                    | Measurand | Calibration Method                   | Range                                                                         | Expanded Uncertainty<br>( $k=2$ ) | Note | Effective Date |
|---------------------------------|---------------------------------------------------------------|-----------|--------------------------------------|-------------------------------------------------------------------------------|-----------------------------------|------|----------------|
| 1 Geometry Measuring Instrument |                                                               |           |                                      |                                                                               |                                   |      |                |
| 1                               | *Dial Indicator                                               | Length    | V. R. of Dial Gauges JJG 34          | (0~30)mm/0.001mm                                                              | $U=(1.6\sim3.7)\mu\text{m}$       |      |                |
|                                 |                                                               |           |                                      | (0~50)mm/0.005mm                                                              | $U=(3.0\sim6.0)\mu\text{m}$       |      |                |
|                                 |                                                               |           |                                      | (0~50)mm/0.01mm                                                               | $U=(3.6\sim10)\mu\text{m}$        |      |                |
| 2                               | *Dial Test Indicator                                          | Length    | V. R. of Dial Test Indicator JJG 35  | (0~0.4) mm/0.001mm                                                            | $U=1.2\mu\text{m}$                |      |                |
|                                 |                                                               |           |                                      | (0~1) mm/0.01mm                                                               | $U=2.6\mu\text{m}$                |      |                |
| 3                               | *Bore Dial Indicator with a Protective Orientation Attachment | Length    | C.S. of Bore Dial Indicator JJF 1102 | $\Phi(10\sim160)\text{mm}/0.001\text{mm}$                                     | $U=1.5\mu\text{m}$                |      |                |
|                                 |                                                               |           |                                      | $\Phi(6\sim160)\text{mm}/0.01\text{mm}$ ,<br>Working stroke $\leq 1\text{mm}$ | $U=2.4\mu\text{m}$                |      |                |



No. CNAS L0730

第 1 页 共 163 页

The scope of the accreditation in Chinese remains the definitive version.

| №  | Instrument                                 | Measurand   | Calibration Method                                            | Range                                                                         | Expanded Uncertainty<br>( $k=2$ )                   | Note | Effective Date |
|----|--------------------------------------------|-------------|---------------------------------------------------------------|-------------------------------------------------------------------------------|-----------------------------------------------------|------|----------------|
|    |                                            |             |                                                               | $\Phi(6\sim 160)\text{mm}/0.01\text{mm}$ , Working stroke $\leq 1.6\text{mm}$ | $U=4.2\text{ }\mu\text{m}$                          |      |                |
| 4  | *Current Caliper                           | Length      | V. R. of Current Caliper<br>JJG 30                            | $(0\sim 1000)\text{mm}/0.01\text{mm}$                                         | $U=(0.01\sim 0.02)\text{mm}$                        |      |                |
| 5  | *Height Caliper                            | Length      | V. R. of Height Caliper JJG 31                                | $(0\sim 1000)\text{mm}/0.01\text{mm}$                                         | $U=(0.01\sim 0.02)\text{mm}$                        |      |                |
| 6  | *Micrometer                                | Length      | V. R. of Micrometer JJG 21                                    | $(0\sim 500)\text{mm}/0.001\text{mm}$                                         | $U=(1.0\sim 2.0)\text{ }\mu\text{m}$                |      |                |
|    |                                            |             |                                                               | Calibration measuring rod: $(25\sim 475)\text{mm}$                            | $U=(0.6\sim 1.4)\text{ }\mu\text{m}$                |      |                |
| 7  | Gauge blocks                               | Length      | V. R. of Gauge Blocks<br>JJG 146                              | $(0.5\sim 100)\text{mm}$                                                      | $U=0.20\text{ }\mu\text{m}+2\times 10^{-6}L(k=2.7)$ |      |                |
|    |                                            |             |                                                               | $(>100\sim 300)\text{mm}$                                                     | $U=0.50\text{ }\mu\text{m}+5\times 10^{-6}L(k=2.7)$ |      |                |
| 8  | *Micrometers of Measuring Inside Dimension | Length      | C.S. of Micrometers of Measuring Inside Dimension<br>JJF 1411 | $(5\sim 100)\text{mm}$                                                        | $U=1.5\text{ }\mu\text{m}$                          |      |                |
| 9  | Smooth Limit Gauge                         | Length      | V. R. of Smooth Limit Gauge JJG 343                           | Plug Gauge: $\Phi(0.3\sim 100)\text{mm}$                                      | $U=(0.3\sim 0.6)\text{ }\mu\text{m}$                |      |                |
|    |                                            |             |                                                               | Ring Gauge: $\Phi(6\sim 100)\text{mm}$                                        | $U=(0.7\sim 1.0)\text{ }\mu\text{m}$                |      |                |
| 10 | Optical Flat                               | Flatness    | V. R. of Optical Flat JJG 28                                  | $\Phi(30\sim 100)\text{mm}$                                                   | $U=0.01\text{ }\mu\text{m}$                         |      |                |
|    |                                            | Parallelism |                                                               | $(0\sim 2)\text{ }\mu\text{m}$                                                | $U=0.2\text{ }\mu\text{m}$                          |      |                |
| 11 | Standard Ring gauge                        | Length      | V.R. for standard Ring gauge<br>JJG 894                       | $\Phi(6\sim 100)\text{mm}$                                                    | $U=0.7\text{ }\mu\text{m}+3\times 10^{-6}L$         |      |                |
| 12 | *Wooden Rule                               | Length      | V.R. for Wooden Rule JJG 2                                    | $(0\sim 1000)\text{mm}$                                                       | $U=0.2\text{mm}$                                    |      |                |
| 13 | *Depth Micrometer                          | Length      | V.R. for Depth Micrometer<br>JJG 24                           | $(0\sim 300)\text{mm}$                                                        | $U=(1.4\sim 3.5)\text{ }\mu\text{m}$                |      |                |



在线扫码获取验证

No. CNAS L0730

第 2 页 共 163 页

The scope of the accreditation in Chinese remains the definitive version.

| №  | Instrument                    | Measurand   | Calibration Method                             | Range                                                     | Expanded Uncertainty<br>( $k=2$ ) | Note | Effective Date |
|----|-------------------------------|-------------|------------------------------------------------|-----------------------------------------------------------|-----------------------------------|------|----------------|
|    |                               |             |                                                | Calibration measuring rod: (25~275)mm                     | $U=(0.7\sim2.1)\mu\text{m}$       |      |                |
| 14 | *General Bevel Protractors    | Angle       | C.S. for General Bevel Protractors JJF 1959    | $(0\sim360)^\circ/5'$                                     | $U=3'$                            |      |                |
|    |                               |             |                                                | $(0\sim360)^\circ/2'$                                     | $U=1'$                            |      |                |
| 15 | Fineness of Grind Gauge       | Length      | V.R. for Fineness of Grind Gauge JJG 905       | $H:(0\sim150)\mu\text{m}$                                 | $U=(0.2\sim0.5)\mu\text{m}$       |      |                |
| 16 | Small parts cylinder          | Length      | C.S. of Small parts cylinder ZDCD-007L         | $(0\sim100)\text{mm}$                                     | $U=0.005\text{mm}$                |      |                |
| 17 | Probe C,D                     | Length      | C.S. of Probe C,D ZDCD-008L                    | $\Phi(10\sim50)\text{mm}$                                 | $U=0.02\text{mm}$                 |      |                |
| 18 | Straight Edge                 | Linear      | V.R. for Straight Edge JJG 63                  | $L: 75\text{mm}\sim200\text{mm}$                          | $U=0.3\mu\text{m}$                |      |                |
| 19 | Test template A and B         | Length      | C.S. of Test template A and B ZDCD-009L        | $(0\sim100)\text{mm}$                                     | $U=0.02\text{mm}$                 |      |                |
| 20 | *Ultrasonic Thickness Gauge   | Length      | C.S. of Ultrasonic Thickness Gauge JJF 1126    | $(0.5\sim200)\text{mm}$                                   | $U=0.04\text{mm}$                 |      |                |
| 21 | Squares                       | Verticality | V.R. for Squares JJG 7                         | $H:50\text{mm}\sim500\text{mm}$                           | $U=(0.002\sim0.02)\text{mm}$      |      |                |
|    |                               | Length      |                                                | $H:(50\sim500)\text{mm}$                                  | $U=0.1\text{mm}$                  |      |                |
| 22 | *Straight Edges               | Linear      | C.S. for Straight Edges JJF 1097               | $300\text{mm}\sim2000\text{mm}$                           | $U=(0.6\sim1.0)\mu\text{m}$       |      |                |
| 23 | Screw Thread Gauge            | Length      | V.R. for Screw Thread Gauge JJG 60             | $P:(0.40\sim6.00)\text{mm}, (\alpha: 55^\circ, 60^\circ)$ | $U=2\mu\text{m}$                  |      |                |
| 24 | Frame Levels and Shaft levels | Angle       | C.S. of Frame Levels and Shaft levels JJF 1084 | division $d: 0.02\text{mm/m}\sim 0.1\text{mm/m}$          | $U=d\times 5.8\%$                 |      |                |
| 25 | *Concentricity Tester         | Length      | C.S. of Concentricity Tester JJF 1109          | $(0\sim100)\mu\text{m}$                                   | $U=2\mu\text{m}$                  |      |                |
| 26 | Box Plate                     | Vertical    | V.R. of Box Plates JJG 194                     | Side length $L: 100\text{mm}\sim 400\text{mm}$            | $U=3\mu\text{m}$                  |      |                |



No. CNAS L0730

第 3 页 共 163 页

The scope of the accreditation in Chinese remains the definitive version.

| №  | Instrument                    | Measurand | Calibration Method                            | Range                                | Expanded Uncertainty<br>( $k=2$ )          | Note | Effective Date |
|----|-------------------------------|-----------|-----------------------------------------------|--------------------------------------|--------------------------------------------|------|----------------|
| 27 | Steel Tape                    | Length    | V.R. for Steel Tape JJG 4                     | (0~200) m                            | $U=0.03 \text{ mm} + 1.2 \times 10^{-5} L$ |      |                |
| 28 | Optical Theodolites           | Angle     | V.R. for Optical Theodolites JJG 414          | Level: $0^\circ \sim 360^\circ$      | $U=0.16''$                                 |      |                |
|    |                               |           |                                               | Vertical: $-32^\circ \sim +32^\circ$ | $U=1.4''$                                  |      |                |
| 29 | Levels                        | Angle     | V.R. for Levels JJG 425                       | Blister type: $-25'' \sim +25''$     | $U=1.3''$                                  |      |                |
|    |                               |           |                                               | Anping type: $-25'' \sim +25''$      | $U=1.1''$                                  |      |                |
| 30 | *Coordinate Measuring Machine | Length    | C.S. of Coordinate Measuring Machine JJF 1064 | Contact type: (0~1000)mm             | $U_E=0.5 \mu\text{m} + 3 \times 10^{-6} L$ |      |                |
|    |                               |           |                                               | $\Phi(10 \sim 50)\text{mm}$          | $U_P=0.3 \mu\text{m}$                      |      |                |
| 31 | *Projectors                   | Length    | C.S. of Projectors JJF 1093                   | Shadow screen diameter: (100~600)mm  | $U=1.8 \mu\text{m}$                        |      |                |
| 32 | Inductive Micrometers         | Length    | C.S. for Inductive Micrometers JJF 1331       | (-1000~+1000) $\mu\text{m}$          | $U=(0.3 \sim 3.0) \mu\text{m}$             |      |                |
| 33 | *Tool makes Microscope        | Length    | V.R. for Tool makes Microscope JJG 56         | (0~200)mm                            | $U=(0.4 \sim 1.2) \mu\text{m}$             |      |                |
| 34 | *Steel Rule                   | Length    | V.R. for Steel Rule JJG 1                     | (0~1)m                               | $U=0.05\text{mm}$                          |      |                |
|    |                               |           |                                               | (1~2)m                               | $U=0.10\text{mm}$                          |      |                |
| 35 | Digital high meters           | Length    | C.S. for Digital high meters ZDCD-016L        | (0~100)mm                            | $U=2 \mu\text{m}$                          |      |                |
| 36 | Testers for Dial Gauges       | Length    | V.R. of Testers for Dial Gauges JJG 201       | (0~50)mm                             | $U=(0.2 \sim 0.6) \mu\text{m}$             |      |                |
| 37 | *Welding Measuring Rulers     | Length    | V.R. for Welding Measuring Rulers JJG 704     | (0~60)mm                             | $U=0.01\text{mm}$                          |      |                |
|    |                               | Angle     |                                               | (0~180)°                             | $U=6'$                                     |      |                |
| 38 | *Plates                       | Flatness  | V.R. for Plates JJG 117                       | 630 mm × 630 mm                      | $U=3.2 \mu\text{m}$                        |      |                |



No. CNAS L0730

第 4 页 共 163 页

The scope of the accreditation in Chinese remains the definitive version.

| №  | Instrument                                                            | Measurand | Calibration Method                                                                 | Range                                        | Expanded Uncertainty<br>( $k=2$ )      | Note | Effective Date |
|----|-----------------------------------------------------------------------|-----------|------------------------------------------------------------------------------------|----------------------------------------------|----------------------------------------|------|----------------|
|    |                                                                       |           |                                                                                    | 1000 mm × 1000 mm                            | $U=5.2 \mu\text{m}$                    |      |                |
|    |                                                                       |           |                                                                                    | 1600 mm × 1600 mm                            | $U=7.8 \mu\text{m}$                    |      |                |
|    |                                                                       |           |                                                                                    | 2500 mm × 1600 mm                            | $U=10 \mu\text{m}$                     |      |                |
| 39 | Cylindrical Thread Gauges                                             | Length    | C.S. for Cylindrical Thread Gauges JJF 1345                                        | Thread plug gauge:<br>M(1~100)mm             | $U=3 \mu\text{m}$                      |      |                |
|    |                                                                       |           |                                                                                    | Thread ring gauge:<br>M(3~100)mm             | $U=4 \mu\text{m}$                      |      |                |
| 40 | Feeler Gauge                                                          | Length    | V.R. for Feeler Gauge JJG 62                                                       | (0.01~3)mm                                   | $U=1 \mu\text{m}$                      |      |                |
| 41 | *Snap Gauge Reading in 0.01mm                                         | Length    | V.R. for Snap Gauge Reading in 0.01mm JJG 109                                      | (0~50) mm                                    | $U=3.4 \mu\text{m}$                    |      |                |
| 42 | Sharp point tester                                                    | Length    | C.S. for sharp point tester ZDCD-010L                                              | (1.02 × 1.15)mm                              | $U=0.005\text{mm}$                     |      |                |
| 43 | Accessibility probe                                                   | Length    | C.S for Accessibility probe ZDCD-006L                                              | (2.8~470)mm                                  | $U=0.02\text{mm}$                      |      |                |
| 44 | *Magnetic and Eddy Current Measuring Instrument for Coating thickness | Length    | V.R. for magic and Eddy Current Measuring Instrument for Coating thickness JJG 818 | Thickness sheet: (7~50) $\mu\text{m}$        | $U=0.3 \mu\text{m}$                    |      |                |
|    |                                                                       |           |                                                                                    | Thickness sheet: (50~3000) $\mu\text{m}$     | $U_{\text{rel}}=0.4\%$                 |      |                |
|    |                                                                       |           |                                                                                    | (10~1250) $\mu\text{m}$                      | $U_{\text{rel}}=0.6\%+0.3 \mu\text{m}$ |      |                |
| 45 | Radius Gauge                                                          | Length    | V.R. or Radius Gauge JJG 58                                                        | R (1~6) mm                                   | $U=6 \mu\text{m}$                      |      |                |
|    |                                                                       |           |                                                                                    | R(6~25)mm                                    | $U=9 \mu\text{m}$                      |      |                |
| 46 | lever rule                                                            | Angle     | C.R of lever rule JJF 1085                                                         | L: (100~1500)mm,<br>( $d:0.5 \text{ mm/m}$ ) | $U=d \times 8\%$                       |      |                |
|    |                                                                       | Length    |                                                                                    | (0~1500) mm                                  | $U=1.0 \text{ mm}$                     |      |                |



No. CNAS L0730

第 5 页 共 163 页

The scope of the accreditation in Chinese remains the definitive version.

| №  | Instrument                                          | Measurand     | Calibration Method                                                 | Range                          | Expanded Uncertainty<br>( $k=2$ ) | Note | Effective Date |
|----|-----------------------------------------------------|---------------|--------------------------------------------------------------------|--------------------------------|-----------------------------------|------|----------------|
| 47 | *Reading microscope                                 | Length        | V.R of reading microscope and measuring microscope JJG 571         | (0~8)mm                        | $U=1.5 \mu m$                     |      |                |
| 48 | *optimeter                                          | Length        | V.R of optimeter JJG 45                                            | (-100~+100) $\mu m$            | $U=0.07 \mu m$                    |      |                |
| 49 | *length measuring instrument                        | Length        | C.S of length measuring instrument JJF 1189                        | Outside: (0~500)mm             | $U=(0.3\sim0.7) \mu m$            |      |                |
|    |                                                     |               |                                                                    | Internal dimensions: (6~50) mm | $U=1.5 \mu m$                     |      |                |
| 50 | *height measuring instrument with digital display   | Length        | C.S of height measuring instrument with digital display JJF 1254   | (0~600) mm                     | $U=(0.7\sim0.9) \mu m$            |      |                |
| 51 | *surface roughness measurement by the profile metod | Roughness     | C.S of surface roughness measurement by the profile metod JJF 1105 | Ra:(0.085~4.36) $\mu m$        | $U_{rel}=6.0\%$                   |      |                |
| 52 | *Combined Type Angle Rules                          | Angle         | C.S of combine type angle rule JJF 1132                            | (0~180) ° /1 °                 | $U=4'$                            |      |                |
| 53 | roughness comparison specimens                      | Roughness     | C.R of roughness comparison specimens JJF 1099                     | Ra:(0.012~100) $\mu m$         | $U_{rel}=(8.0\%\sim6.2\%)$        |      |                |
| 54 | *cable length meter                                 | Length        | V.R of cable lengt meter JJG 987                                   | (0.1~2000)m                    | $U=0.01 mm+5\times 10^{-5}L$      |      |                |
| 55 | test sieves                                         | Length        | C.S of test sieves JJF 1175                                        | $\omega$ : (0.020~5) mm        | $U=1 \mu m$                       |      |                |
|    |                                                     |               |                                                                    | $\omega$ : (>5~125) mm         | $U=0.02mm$                        |      |                |
| 56 | *Biological Microscopes                             | Magnification | C.S of biological microscope JJF 1402                              | 4X~100X                        | $U_{rel}=1\%$                     |      |                |
| 57 | fiber tapes                                         | Length        | V.R of fiber tapes and measuring ropes JJG 5                       | (0~5) m                        | $U=0.8mm$                         |      |                |
|    |                                                     |               |                                                                    | (5~10)m                        | $U=1.6mm$                         |      |                |
|    |                                                     |               |                                                                    | (10~15)m                       | $U=2.4mm$                         |      |                |



No. CNAS L0730

第 6 页 共 163 页

The scope of the accreditation in Chinese remains the definitive version.

| №  | Instrument                        | Measurand | Calibration Method                                | Range              | Expanded Uncertainty<br>(k=2)         | Note | Effective Date |
|----|-----------------------------------|-----------|---------------------------------------------------|--------------------|---------------------------------------|------|----------------|
|    |                                   |           |                                                   | (15~20)mm          | U=3.2mm                               |      |                |
|    |                                   |           |                                                   | (20~30)mm          | U=4.8mm                               |      |                |
|    |                                   |           |                                                   | (30~50)mm          | U=8.0mm                               |      |                |
| 58 | cylindrical measuring pin         | Length    | C.S of cylindrical measuring pin JJF 1207         | $\Phi$ (0.1~25) mm | $U=0.14 \mu m + 1.4 \times 10^{-6} L$ |      |                |
| 59 | angle gauge block                 | angle     | V.R of angle gauge block JJG 70                   | 0° ~90°            | U=1.8 "                               |      |                |
| 60 | *imaging probe measuring machines | Length    | C.S for imaging probe measuring machines JJF 1318 | (0~600)mm          | $U=(0.7 \sim 1.9) \mu m$              |      |                |
| 61 | bite tester                       | Length    | C.S for bite tester ZDCD-020L                     | (0~100)mm          | U=0.05mm                              |      |                |
| 62 | *moulds                           | Length    | C.S for moulds JJF 1307                           | (40~600)mm         | $U=(0.03 \sim 0.10) mm$               |      |                |
| 63 | step gauges                       | Length    | C.S for step gauges JJF 1258                      | (0~600)mm          | $U=2.0 \mu m$                         |      |                |
| 64 | curve tester                      | Angle     | C.S for curve tester ZDCD-021L                    | (0~200)mm          | U=0.05mm                              |      |                |
| 65 | *fall plate                       | Length    | C.S for fall plate ZDCD-022L                      | (0~1000)mm         | U=0.10mm                              |      |                |
| 66 | UL finger                         | Length    | C.S for UL finger ZDCD-023L                       | (0~200)mm          | U=0.05mm                              |      |                |
| 67 | safety pole                       | Length    | C.S for safety pole ZDCD-024L                     | (0~100)mm          | U=0.01mm                              |      |                |
| 68 | *sampler                          | Length    | C.S for sampler ZDCD-025L                         | (0~200)mm          | U=0.02mm                              |      |                |
| 69 | level rod                         | Length    | V.R of level rod JJG 8                            | (0~5)m             | U=0.3mm                               |      |                |
| 70 | Pi tapes                          | length    | C.S for Pi tapes JJF1423                          | (9~5000)mm         | $U=(0.01 \sim 0.05) mm$               |      |                |



No. CNAS L0730

第 7 页 共 163 页

The scope of the accreditation in Chinese remains the definitive version.

| №  | Instrument                          | Measurand   | Calibration Method                                      | Range                                                                                                                                                 | Expanded Uncertainty<br>( $k=2$ )                                                                                                           | Note | Effective Date |
|----|-------------------------------------|-------------|---------------------------------------------------------|-------------------------------------------------------------------------------------------------------------------------------------------------------|---------------------------------------------------------------------------------------------------------------------------------------------|------|----------------|
| 71 | square gauge                        | Verticality | V.R of square gauge JJG1046                             | H: 100 mm ~ 500 mm                                                                                                                                    | $U = (1.5 \sim 2.4) \mu\text{m}$                                                                                                            |      |                |
| 72 | dial snap gauge                     | Length      | C.S for dial snap gauge JJF1253                         | (5 ~ 100) mm / 0.005 mm                                                                                                                               | $U = 0.0052 \text{ mm}$                                                                                                                     |      |                |
| 73 | *common normal micrometer           | Length      | V.R of common normal micrometer JJG82                   | (0 ~ 100) mm<br>Calibration bar (25 ~ 75) mm                                                                                                          | $U = 1.6 \mu\text{m}$<br>$U = 1.0 \mu\text{m}$                                                                                              |      |                |
| 74 | *depth dial gauge                   | Length      | V.R of depth dial gauge JJG830                          | (0 ~ 300) mm<br>Calibration bar: (25 ~ 300) mm                                                                                                        | $U = (2.0 \sim 8.0) \mu\text{m}$<br>$U = (1.6 \sim 4.8) \mu\text{m}$                                                                        |      |                |
| 75 | micrometer with dial comparator     | Length      | V.R of micrometer with dial comparator JJG26            | (0 ~ 100) mm<br>Calibration lever: (25 ~ 75) mm                                                                                                       | $U = 1.0 \mu\text{m}$<br>$U = 1.0 \mu\text{m}$                                                                                              |      |                |
| 76 | *length measuring machine           | Length      | C.S for length measuring machine JJF1066                | Micrometer scale: (-100 ~ +100) $\mu\text{m}$<br>Millimeter scale: (0 ~ 100) mm<br>Decimeter scale: (0 ~ 2000) mm<br>Internal dimensions: (6 ~ 50) mm | $U = 0.07 \mu\text{m}$<br>$U = 0.2 \mu\text{m} + 2 \times 10^{-6} L$<br>$U = 0.4 \mu\text{m} + 3 \times 10^{-6} L$<br>$U = 1.5 \mu\text{m}$ |      |                |
| 77 | electronic tachometer total station | Angle       | V.R of electronic tachometer total station JJG100       | level: $0^\circ \sim 360^\circ$<br>Vertical right angle: $-32^\circ \sim +32^\circ$                                                                   | $U = 0.38''$<br>$U = 1.4''$                                                                                                                 |      |                |
| 78 | *Track gauge                        | Length      | V. R. of Track gages for standard gauges railway JJG219 | (1410 ~ 1470) mm                                                                                                                                      | $U = 0.08 \text{ mm}$                                                                                                                       |      |                |



No. CNAS L0730

第 8 页 共 163 页

The scope of the accreditation in Chinese remains the definitive version.


| №  | Instrument                        | Measurand     | Calibration Method                                 | Range                          | Expanded Uncertainty<br>( $k=2$ )   | Note | Effective Date |
|----|-----------------------------------|---------------|----------------------------------------------------|--------------------------------|-------------------------------------|------|----------------|
| 79 | *metallographic microscope        | magnification | V.R.for metallographic microscope JJG(JW)012       | 4 X~1000 X                     | $U_{rel}=1\%$                       |      |                |
| 80 | *Clinometers                      | Angel         | C.S.for Clinometers JJF 1915                       | $0^{\circ} \sim 360^{\circ}$   | $U=0.02^{\circ} \sim 0.2^{\circ}$   |      |                |
| 81 | *Wet Film Thickness Gauges        | Length        | C.S.for Wet Film Thickness Gauges JJF 1484         | comb gauge: (0~3000) $\mu m$   | $U=(2.4 \sim 15) \mu m$             |      |                |
|    |                                   |               |                                                    | wheel gauge(0~1000) $\mu m$    | $U=(0.6 \sim 6) \mu m$              |      |                |
| 82 | *Borehole Clinometers             | Angel         | C.S.for Borehole Clinometers JJF 1550              | $-30^{\circ} \sim +30^{\circ}$ | $U=0.008\%FS$                       |      |                |
| 83 | *Thickness Gauges                 | Length        | C.S.for Thickness Gauges JJF 1255                  | (0~30)mm                       | $U=(2 \sim 4) \mu m$                |      |                |
| 84 | Grating Micrometers               | Length        | C.S.for Grating Micrometers JJF 1682               | (0.01~100)mm                   | $U=(0.05 \sim 2.2) \mu m$           |      |                |
| 85 | *Surface Profile Gauges           | Length        | C.S.for Surface Profile Gauges JJF 1476            | (0.1~6.5)mm                    | $U=(0.9 \sim 1.6) \mu m$            |      |                |
| 86 | Microcator                        | Length        | V.R.of Microcator JJG 118                          | (0~ $\pm 100$ )分度              | $U=(0.054 \sim 0.54) \mu m$         |      |                |
| 87 | Comparators of Machine Type       | Length        | V.R.of Comparators of Machine Type JJG 39          | (0~ $\pm 100$ )分度              | $U=(0.054 \sim 0.54) \mu m$         |      |                |
| 88 | Coordinate Measuring Spheres      | Diameter      | C.S.for Coordinate Measuring Spheres JJF 1422      | $\Phi(10 \sim 100)mm$          | $U=0.35 \mu m$                      |      |                |
|    |                                   | roundness     |                                                    | (0~1) mm                       | $U=0.04 \mu m$                      |      |                |
| 89 | Blocks used in Ultrasonic Testing | Length        | C.S.for Blocks used in Ultrasonic Testing JJF 1487 | (0.1~300)mm                    | $U=(0.004 \sim 0.012)mm$            |      |                |
| 90 | Laser Interferometers             | Length        | V.R.of Laser Interferometers JJG739                | (0~1)m                         | $U=0.20 \mu m+0.20 \times 10^{-6}L$ |      |                |
|    |                                   |               |                                                    | (>1~50)m                       | $U=0.30 \mu m+0.30 \times 10^{-6}L$ |      |                |
|    |                                   | Angle         |                                                    | $0^{\circ} \sim 360^{\circ}$   | $U=0.30''$                          |      |                |
| 91 | High-precision Line Scale         | Length        | V.R.of High-precision Line Scale JJG73             | (0~1)m                         | $U=0.14 \mu m+1 \times 10^{-6}L$    |      |                |



No. CNAS L0730

第 9 页 共 163 页

The scope of the accreditation in Chinese remains the definitive version.

| №  | Instrument                                   | Measurand      | Calibration Method                                                | Range                                                           | Expanded Uncertainty<br>( $k=2$ ) | Note                                                                                 | Effective Date |
|----|----------------------------------------------|----------------|-------------------------------------------------------------------|-----------------------------------------------------------------|-----------------------------------|--------------------------------------------------------------------------------------|----------------|
| 92 | Master Gear                                  | Length         | V.R.of Master Gear JJG1008                                        | Distance of Gear: Pitch Diameter:(20~280)mm, [ $m_n$ (1~10) mm] | $U=(2.5\sim3.7) \mu m$            |                                                                                      |                |
|    |                                              | Length         |                                                                   | Gear Form: Pitch Diameter:(20~280)mm, [ $m_n$ (1~10) mm]        | $U=(2.6\sim3.6) \mu m$            |                                                                                      |                |
|    |                                              | Length         |                                                                   | Gear Helix: Pitch Diameter:(20~280)mm, [ $m_n$ (1~10) mm]       | $U=(2.7\sim4.1) \mu m$            |                                                                                      |                |
| 93 | Laser Tracker 3-Dimensional Measuring System | Length         | C.S.for Laser Tracker 3-Dimensional Measuring System JJF1242      | (0~50)m                                                         | $U=0.5 \mu m+1 \times 10^{-6}L$   |                                                                                      |                |
| 94 | MEMS Gyroscopes                              | Angle Velocity | C.S.for MEMS Gyroscopes JJF1535                                   | Scale Factor: (0.01~15)mV/(°/s)                                 | $U_{rel}=1.0\%$                   |                                                                                      |                |
| 95 | Angular Motion Transducers                   | Angle Velocity | C.S.for Angular Motion Transducers(Primary Angular Shock) JJF1453 | (1~400)°/s                                                      | $U_{rel}=1.5\%$                   |                                                                                      |                |
| 96 | Strapdown Inertial Flight Attitudes          | Angle Velocity | C.S.for Strapdown Inertial Flight Attitudes JJF1536               | (1~400)°/s                                                      | $U_{rel}=1.0\%$                   |                                                                                      |                |
| 97 | Gyroscopes with Dynamic Features             | Phase          | C.S.for Gyroscopes with Dynamic Features JJF1537                  | (0.1~180)°                                                      | $U=1.7^\circ$                     |  |                |
|    |                                              | Sensitivity    |                                                                   | (0.01~15)mV/(°/s)                                               | $U_{rel}=3.0\%$                   |                                                                                      |                |
| 98 | Micro Inertial Measurement Units(MIMU)       |                | C.S.for Micro Inertial Measurement Units(MIMU) JJF2003            | Scale Factor: (0.5~1.5)                                         | $U_{rel}=1.0\%$                   |                                                                                      |                |



No. CNAS L0730

第 10 页 共 163 页

The scope of the accreditation in Chinese remains the definitive version.

| No                                | Instrument                                                                           | Measurand   | Calibration Method                                             | Range                                            | Expanded Uncertainty<br>(k=2)      | Note | Effective Date |
|-----------------------------------|--------------------------------------------------------------------------------------|-------------|----------------------------------------------------------------|--------------------------------------------------|------------------------------------|------|----------------|
| 2 Thermology Measuring Instrument |                                                                                      |             |                                                                |                                                  |                                    |      |                |
| 1                                 | *Digital Temperature Indicators and Controllers                                      | Temperature | V.R.of Digital Temperature Indicators and Controllers JJG 617  | Part with thermocouple: (-200~0)°C               | $U=(0.32\sim0.14)^{\circ}\text{C}$ |      |                |
|                                   |                                                                                      |             |                                                                | Part with thermocouple: (>0~800)°C               | $U=(0.14\sim0.2)^{\circ}\text{C}$  |      |                |
|                                   |                                                                                      |             |                                                                | Part with thermocouple: (800~1800)°C             | $U=(0.2\sim1.1)^{\circ}\text{C}$   |      |                |
|                                   |                                                                                      |             |                                                                | Part with Resistance Thermo- meters: (-200~0)°C  | $U=(0.08\sim0.10)^{\circ}\text{C}$ |      |                |
|                                   |                                                                                      |             |                                                                | Part with Resistance Thermo- meters: (>0~800)°C  | $U=(0.10\sim0.16)^{\circ}\text{C}$ |      |                |
| 2                                 | *Recorders for Industrial-Process Measurements                                       | Temperature | V.R.of the Recorders for Industrial-Process Measurement JJG 74 | (-200~0)°C                                       | $U=(0.5\sim0.3)^{\circ}\text{C}$   |      |                |
|                                   |                                                                                      |             |                                                                | (0~1800)°C                                       | $U=(0.3\sim1.0)^{\circ}\text{C}$   |      |                |
| 3                                 | Industry platinum Resistance Thermo- meters/Industry cuprumResistance Thermo- meters | Temperature | V.R.of Industry Platinum Copper Resistance Thermometer JJG 229 | platinum Resistance Thermo- meters: (-60~+300)°C | $U=(0.02\sim0.03)^{\circ}\text{C}$ |      |                |
|                                   |                                                                                      | Temperature |                                                                | cuprumResistance Thermo- meters : (-60~+300)°C   | $U=(0.02\sim0.03)^{\circ}\text{C}$ |      |                |
| 4                                 | Base Metal Thermocouples                                                             | Temperature | C.S.for Base Metal Thermocouples JJF 1637                      | (-50~+300)°C                                     | $U=(0.18\sim0.15)^{\circ}\text{C}$ |      |                |
|                                   |                                                                                      |             |                                                                | (300~1100)°C                                     | $U=(0.15\sim0.8)^{\circ}\text{C}$  |      |                |
| 5                                 | Liquid-in-Glass Thermome- ters                                                       | Temperature | V.R.of Liquid-in-Glass Thermometers for working JJG 130        | (-60~+300)°C                                     | $U=0.04^{\circ}\text{C}$           |      |                |



No. CNAS L0730

第 11 页 共 163 页

The scope of the accreditation in Chinese remains the definitive version.

| №  | Instrument                                                                           | Measurand            | Calibration Method                                                                                                                  | Range                                  | Expanded Uncertainty<br>( $k=2$ )                   | Note | Effective Date |
|----|--------------------------------------------------------------------------------------|----------------------|-------------------------------------------------------------------------------------------------------------------------------------|----------------------------------------|-----------------------------------------------------|------|----------------|
| 6  | Mobile<br>Detecting Meter of<br>Temperature                                          | Temperature          | C.S.for Temperature and<br>Humidity Itinerant Detecting<br>Instruments JJF 1171                                                     | $(-60\sim+300)^{\circ}\text{C}$        | $U=0.12^{\circ}\text{C}$                            |      |                |
| 7  | Radiation<br>Thermo meters                                                           | Temperature          | V.R.of Radiation<br>Thermometers JJG 856                                                                                            | $(-20\sim+300)^{\circ}\text{C}$        | $U=(0.6\sim1.0)^{\circ}\text{C}$                    |      |                |
|    |                                                                                      |                      |                                                                                                                                     | $(\geq 300\sim 1000)^{\circ}\text{C}$  | $U=(1.0\sim 3.2)^{\circ}\text{C}$                   |      |                |
| 8  | Bimetallic<br>Thermo meters                                                          | Temperature          | C.S.for Bimetallic<br>Thermometers JJF 1908                                                                                         | $(-60\sim+300)^{\circ}\text{C}$        | $U=0.2^{\circ}\text{C}$                             |      |                |
| 9  | Filled System<br>Thermometers                                                        | Temperature          | C.S.for Filled System<br>Thermometers JJF 1909                                                                                      | $(-60\sim+300)^{\circ}\text{C}$        | $U=0.3^{\circ}\text{C}$                             |      |                |
| 10 | *Thermostatic<br>Baths for<br>Temperature<br>Calibration                             | Temperature          | Measurement and Test Norm<br>of Metrological<br>Characteristics of<br>Thermostatic Baths for<br>Temperature Calibration JJF<br>1030 | $(-60\sim+300)^{\circ}\text{C}$        | $U=0.014^{\circ}\text{C}$                           |      |                |
| 11 | mechanical<br>Thermo/hygro<br>meter                                                  | Temperature          | V.R.of mechanical<br>Thermo/hygro meter JJG 205                                                                                     | $(-50\sim+100)^{\circ}\text{C}$        | $U=0.6^{\circ}\text{C}$                             |      |                |
|    |                                                                                      | Relative<br>humidity |                                                                                                                                     | $(20\sim 95)\%$                        | $U=2.2\%$                                           |      |                |
| 12 | *Environmental<br>Testing Equipment<br>for Temperature<br>and Humidity<br>Parameters | Temperature          | C.S.for Environmental<br>Testing Equipment for<br>Temperature and Humidity<br>Parameters JJF 1101                                   | $(-80\sim 300)^{\circ}\text{C}$        | $U=0.3^{\circ}\text{C}$                             |      |                |
|    |                                                                                      | Relative<br>humidity |                                                                                                                                     | $(20\sim 98)\%$                        | $U=2.0\%$                                           |      |                |
| 13 | Humidity<br>Sensors                                                                  | Relative<br>humidity | C.S.for Digital Temperature-<br>hygrometers JJF 1076                                                                                | $(10\sim 95)\%$                        | $U=2.2\%$                                           |      |                |
| 14 | *Temperature and<br>Humidity Standard<br>Chambers                                    | Temperature          | C.S.for Temperature and<br>Humidity Standard Chambers<br>JJF 1564                                                                   | $(5\sim 50)^{\circ}\text{C}$           | uniformity: $U=0.05^{\circ}\text{C}$                |      |                |
|    |                                                                                      |                      |                                                                                                                                     | $(5\sim 50)^{\circ}\text{C}$           | fluctuation: $U=0.02^{\circ}\text{C}$               |      |                |
|    |                                                                                      |                      |                                                                                                                                     | $(0\sim 1)^{\circ}\text{C}/\text{min}$ | variation rate: $U=0.03^{\circ}\text{C}/\text{min}$ |      |                |



No. CNAS L0730

第 12 页 共 163 页

The scope of the accreditation in Chinese remains the definitive version.

| №  | Instrument                                              | Measurand         | Calibration Method                                                | Range           | Expanded Uncertainty<br>( $k=2$ )                       | Note | Effective Date |
|----|---------------------------------------------------------|-------------------|-------------------------------------------------------------------|-----------------|---------------------------------------------------------|------|----------------|
|    |                                                         | Relative humidity |                                                                   | (10~90) %       | Humidity uniformity:: $U=0.5\%$                         |      |                |
|    |                                                         |                   |                                                                   | (10~90) %       | fluctuation: $U=0.1\%$                                  |      |                |
|    |                                                         |                   |                                                                   | (0~2) %/min     | variation rate: $U=0.2\%/min$                           |      |                |
| 15 | *Box-type Resistance Furnace                            | Temperature       | C.S.for Box-type Resistance Furnace JJF 1376                      | (300~1100)°C    | $U=2.0^{\circ}\text{C}$                                 |      |                |
| 16 | *Calibration Specification for the Surface Thermometers | Temperature       | C.S.for the Surface Thermometers JJF 1409                         | (30~400)°C      | $U=0.4^{\circ}\text{C}$                                 |      |                |
| 17 | Digital Thermo-hygro meters                             | Temperature       | C.S.for Digital Temperature-hygrometers JJF 1076                  | (-50~+100)°C    | $U=0.6^{\circ}\text{C}$                                 |      |                |
|    |                                                         | Relative Humidity |                                                                   | (20~95)%        | $U=2.2\%$                                               |      |                |
| 18 | *Temperature Transmitter                                | Temperature       | Calibration Specification of the Temperature Transmitter JJF 1183 | (-60~+300)°C    | $U=0.16^{\circ}\text{C}$                                |      |                |
| 19 | Sheathed Thermocouples                                  | Temperature       | Calibration Specification for Sheathed Thermocouples JJF 1262     | (-40~+1100)°C   | $U=(0.15\sim0.80)^{\circ}\text{C}$                      |      |                |
| 20 | *Temperature Indicators                                 | Temperature       | Calibration Specification for Temperature Indicators JJF 1664     | (-200~+1800) °C | $U=(0.4\sim1.0)^{\circ}\text{C}$ (With thermocouple)    |      |                |
|    |                                                         |                   |                                                                   | (-200~+1800) °C | $U=0.4^{\circ}\text{C}$ (With Resistance Thermo-meters) |      |                |
| 21 | *Dissolution Tester                                     | Temperature       | Verification Regulation of Dissolution Tester JJG (粤) 048         | (0~50)°C        | $U=0.14^{\circ}\text{C}$                                |      |                |
|    |                                                         | Rotate Speed      |                                                                   | (0~300)r/min    | $U_{\text{rel}}=0.8\%$                                  |      |                |
|    |                                                         | Displacement      |                                                                   | (0~3) mm        | $U=0.12\text{ mm}$                                      |      |                |



No. CNAS L0730

第 13 页 共 163 页

The scope of the accreditation in Chinese remains the definitive version.

| №                                | Instrument                                                 | Measurand   | Calibration Method                                                                              | Range         | Expanded Uncertainty<br>( $k=2$ )  | Note | Effective Date |
|----------------------------------|------------------------------------------------------------|-------------|-------------------------------------------------------------------------------------------------|---------------|------------------------------------|------|----------------|
| 22                               | Infrared Ear Thermometers                                  | Temperature | Verification Regulation of Infrared Ear Thermometers JJG1164                                    | (35~42)°C     | $U=0.12^{\circ}\text{C}$           |      |                |
| 23                               | Temperature Data Acquisition Instruments                   | Temperature | Calibration Specification of Temperature Data Acquisition Instruments JJF1366                   | (-80~+300)°C  | $U=(0.03\sim0.04)^{\circ}\text{C}$ |      |                |
| 24                               | Infrared Thermometers for Measurement of Human Temperature | Temperature | Calibration Specification of Infrared Thermometers for Measurement of Human Temperature JJF1107 | (22.0~42.0)°C | $U=0.2^{\circ}\text{C}$            |      |                |
| 25                               | Clinical Electronic Thermometers                           | Temperature | Verification Regulation of Clinical Electronic Thermometers JJG1162                             | (35.0~41.0)°C | $U=0.03^{\circ}\text{C}$           |      |                |
| 26                               | *Textile Dryer of the Rotary Tunble                        | Temperature | Calibration Specification for Textile Dryer of the Rotary Tunble JJF(纺织)072                     | (0~100)°C     | $U=0.3^{\circ}\text{C}$            |      |                |
|                                  |                                                            | time        |                                                                                                 | (0.01s~1h)    | $U=0.2\text{s}$                    |      |                |
| 27                               | Temperature Parameter of Passive Medical Cold Boxes        | Temperature | C.S.for Temperature Parameter of Passive Medical Cold Boxes JJF 1676                            | (-20~+20)°C   | $U=0.3^{\circ}\text{C}$            |      |                |
| 3 Mechanics Measuring Instrument |                                                            |             |                                                                                                 |               |                                    |      |                |
| 1                                | *Working Dynamometers                                      | Force value | V.R.of Working Dynamometers JJG 455                                                             | 0.1cN~2MN     | $U_{\text{rel}}=0.12\%$            |      |                |
|                                  |                                                            |             |                                                                                                 | >2MN~5MN      | $U_{\text{rel}}=0.33\%$            |      |                |
| 2                                | Shore A Durometers                                         | Test force  | V.R.of Shore A Durometers JJG 304                                                               | (50~10000) mN | $U=20\text{ mN}$                   |      |                |
|                                  |                                                            | Length      |                                                                                                 | (0~150) mm    | $U=0.018\text{ mm}$                |      |                |
|                                  |                                                            | Angle       |                                                                                                 | (0~40) °      | $U=2.2\text{ '}$                   |      |                |

No. CNAS L0730

第 14 页 共 163 页



The scope of the accreditation in Chinese remains the definitive version.


| № | Instrument                                         | Measurand                 | Calibration Method                                               | Range          | Expanded Uncertainty<br>( $k=2$ )         | Note | Effective Date |
|---|----------------------------------------------------|---------------------------|------------------------------------------------------------------|----------------|-------------------------------------------|------|----------------|
| 3 | *Weights                                           | Mass                      | V.R.of Weights JJG 99                                            | 1 mg~500 mg    | $U=0.005 \text{ mg} \sim 0.02 \text{ mg}$ |      |                |
|   |                                                    |                           |                                                                  | > 500 mg~100 g | $U=0.02 \text{ mg} \sim 0.10 \text{ mg}$  |      |                |
|   |                                                    |                           |                                                                  | 100 g~1 kg     | $U=0.10 \text{ mg} \sim 0.8 \text{ mg}$   |      |                |
|   |                                                    |                           |                                                                  | >1 kg~10 kg    | $U=0.8 \text{ mg} \sim 9 \text{ mg}$      |      |                |
|   |                                                    |                           |                                                                  | >10 kg~20 kg   | $U=9 \text{ mg} \sim 0.03 \text{ g}$      |      |                |
|   |                                                    |                           |                                                                  | >20 kg~40 kg   | $U=0.03 \text{ g} \sim 0.18 \text{ g}$    |      |                |
|   |                                                    |                           |                                                                  | >40 kg~150 kg  | $U=0.18 \text{ g} \sim 2.2 \text{ g}$     |      |                |
|   |                                                    |                           |                                                                  |                |                                           |      |                |
| 4 | *Electronic Balance                                | Mass                      | C.S.for Electronic Balances JJF 1847                             | (0~220)g       | $U=(0.02 \sim 0.08) \text{ mg}$           |      |                |
|   |                                                    |                           |                                                                  | >220g~320g     | $U=0.7 \text{ mg}$                        |      |                |
|   |                                                    |                           |                                                                  | >320g~1000g    | $U=(3 \sim 7) \text{ mg}$                 |      |                |
|   |                                                    |                           |                                                                  | >1000g~10000g  | $U=13 \text{ mg} \sim 0.06 \text{ g}$     |      |                |
|   |                                                    |                           |                                                                  | >10000g~40000g | $U=0.10 \text{ g} \sim 0.25 \text{ g}$    |      |                |
|   |                                                    |                           |                                                                  | >40000g~100kg  | $U=2.2 \text{ g} \sim 2.8 \text{ g}$      |      |                |
|   |                                                    |                           |                                                                  | >100kg~200kg   | $U=13 \text{ g} \sim 15 \text{ g}$        |      |                |
|   |                                                    |                           |                                                                  | >200kg~800kg   | $U=0.03 \text{ kg} \sim 0.11 \text{ kg}$  |      |                |
| 5 | *Universal Tension and Compression Testing Machine | Force value               | V.R.of Universal Tension and Compression Testing Machine JJG 139 | 0.1cN~2MN      | $U_{\text{rel}}=0.12\%$                   |      |                |
|   |                                                    | Displacement              |                                                                  | >2MN~5MN       | $U_{\text{rel}}=0.33\%$                   |      |                |
|   |                                                    | Extensometer displacement |                                                                  | (0.05~1000) mm | $U_{\text{rel}}=0.15\%$                   |      |                |
|   |                                                    |                           |                                                                  | (0.05~25) mm   | $U_{\text{rel}}=0.28\%$                   |      |                |



No. CNAS L0730

第 15 页 共 163 页

The scope of the accreditation in Chinese remains the definitive version.

| №  | Instrument                                                          | Measurand                 | Calibration Method                                                              | Range                        | Expanded Uncertainty<br>( <i>k</i> =2) | Note                                                                                 | Effective Date |
|----|---------------------------------------------------------------------|---------------------------|---------------------------------------------------------------------------------|------------------------------|----------------------------------------|--------------------------------------------------------------------------------------|----------------|
| 6  | *Metallic Rockwell Hardness Testers                                 | Hardness                  | V.R.of Metallic Rockwell Hardness Testers(Scales A,B,C,D,E,F,G,H,K,N,T) JJG 112 | ( 75～88 ) HRA                | <i>U</i> =0.4 HRA                      |                                                                                      |                |
|    |                                                                     |                           |                                                                                 | ( 80～100 ) HRB               | <i>U</i> =0.5 HRB                      |                                                                                      |                |
|    |                                                                     |                           |                                                                                 | ( 40～70 ) HRC                | <i>U</i> =0.4 HRC                      |                                                                                      |                |
|    |                                                                     |                           |                                                                                 | ( 20～40 ) HRC                | <i>U</i> =0.5 HRC                      |                                                                                      |                |
|    |                                                                     |                           |                                                                                 | ( 0～100 ) HRN                | <i>U</i> =0.4 HRN                      |                                                                                      |                |
|    |                                                                     |                           |                                                                                 | ( 10～100 ) HRTW              | <i>U</i> =1.2HRTW                      |                                                                                      |                |
| 7  | Tachometer                                                          | Rotate speed              | V.R.of Tachometer JJG 105                                                       | (100～999) r/min              | <i>U</i> <sub>rel</sub> =0.08%         |                                                                                      |                |
|    |                                                                     |                           |                                                                                 | (1000～9999) r/min            | <i>U</i> <sub>rel</sub> =0.06%         |                                                                                      |                |
|    |                                                                     |                           |                                                                                 | (10000～20000) r/min          | <i>U</i> <sub>rel</sub> =0.014%        |                                                                                      |                |
| 8  | Torque Wrenches                                                     | Torque force              | V.R.of Torque Wrenches JJG 707                                                  | ( 0.008～2000 ) Nm            | <i>U</i> <sub>rel</sub> =1.3%          |                                                                                      |                |
| 9  | *Mechanical Balance                                                 | Mass                      | V.R.for Mechanical Balance JJG 98                                               | Max: ( 0～200 ) g<br>d=0.1 mg | <i>U</i> =(0.08～0.14) mg               |                                                                                      |                |
|    |                                                                     |                           |                                                                                 | Max: ( 0～20 ) kg<br>d>0.1mg  | <i>U</i> =(0.08～6.5)mg                 |                                                                                      |                |
| 10 | *Table balances                                                     | Mass                      | V.R.for Table Balances JJG 156                                                  | (0～5) kg                     | <i>U</i> =(0.02～1.3) g                 |                                                                                      |                |
| 11 | *Universal Tension and Compression Testing Machine for Non-metallic | Force                     | V.R.of Universal Tension and Compression Testing Machine JJG 139                | 0.1cN～2MN                    | <i>U</i> <sub>rel</sub> =0.12%         |  |                |
|    |                                                                     | Extensometer displacement |                                                                                 | >2MN～5MN                     | <i>U</i> <sub>rel</sub> =0.33%         |                                                                                      |                |
|    |                                                                     |                           |                                                                                 | (0.05～25) mm                 | <i>U</i> <sub>rel</sub> =0.28%         |                                                                                      |                |
| 12 | *Electronic Universal Testing Machine                               | Force                     | V.R.of Electronic Universal Testing Machine JJG 475                             | 0.1cN～2MN                    | <i>U</i> <sub>rel</sub> =0.12%         |                                                                                      |                |
|    |                                                                     |                           |                                                                                 | >2MN～5MN                     | <i>U</i> <sub>rel</sub> =0.33%         |                                                                                      |                |



No. CNAS L0730

第 16 页 共 163 页

The scope of the accreditation in Chinese remains the definitive version.

| №  | Instrument                                                            | Measurand                    | Calibration Method                                                                  | Range                           | Expanded Uncertainty<br>( $k=2$ ) | Note | Effective Date |
|----|-----------------------------------------------------------------------|------------------------------|-------------------------------------------------------------------------------------|---------------------------------|-----------------------------------|------|----------------|
|    |                                                                       | Displacement                 |                                                                                     | (0.05~1000) mm                  | $U_{rel}=0.15\%$                  |      |                |
|    |                                                                       | Extensometer<br>Displacement |                                                                                     | (0.05~25) mm                    | $U_{rel}=0.28\%$                  |      |                |
| 13 | *Building Material<br>Testing Machine<br>of Constant<br>Loading Speed | Force                        | V.R. for Building Material<br>Testing Machine of Constant<br>Loading Speed JJG 1025 | 20N~2MN                         | $U_{rel}=0.12\%$                  |      |                |
|    |                                                                       |                              |                                                                                     | >2MN~5MN                        | $U_{rel}=0.33\%$                  |      |                |
|    |                                                                       | Afterburner<br>speed         |                                                                                     | (100~2000) N/s                  | $U_{rel}=2.5\%$                   |      |                |
| 14 | *Flexure Testing<br>Machine                                           | Force                        | V.R. of Flexure Testing<br>Machine JJG 476                                          | 0.1 cN~100 kN                   | $U_{rel}=0.12\%$                  |      |                |
|    |                                                                       | Test force<br>speed          |                                                                                     | (40~60) N/s                     | $U=1.2\text{N/s}$                 |      |                |
| 15 | *Working Force<br>Measureing<br>Machines for<br>Special Purposes      | Force                        | C.S. for Working Force<br>Measureing Machines for<br>Special Purposes JJF 1134      | 0.1cN~2MN                       | $U_{rel}=0.12\%$                  |      |                |
|    |                                                                       |                              |                                                                                     | >2MN~5MN                        | $U_{rel}=0.33\%$                  |      |                |
| 16 | *Metallic Brinell<br>Hardness Testers                                 | Hardness                     | V.R. of Metallic Brinell<br>Hardness Testers JJG 150                                | 标尺 10/3000: (50~<br>225) HBW    | $U_{rel}=1.2\%$                   |      |                |
|    |                                                                       |                              |                                                                                     | 标尺 10/3000: (225~<br>500) HBW   | $U_{rel}=0.7\%$                   |      |                |
|    |                                                                       |                              |                                                                                     | 标尺 5/750: (50~225)<br>HBW       | $U_{rel}=1.2\%$                   |      |                |
|    |                                                                       |                              |                                                                                     | 标尺 5/750: (225~500)<br>HBW      | $U_{rel}=0.7\%$                   |      |                |
|    |                                                                       |                              |                                                                                     | 标尺 2.5/187.5: (50~<br>225) HBW  | $U_{rel}=1.2\%$                   |      |                |
|    |                                                                       |                              |                                                                                     | 标尺 2.5/187.5: (225~<br>500) HBW | $U_{rel}=0.7\%$                   |      |                |
|    |                                                                       |                              |                                                                                     | 标尺 10/1000: (0~125)<br>HBW      | $U_{rel}=1.2\%$                   |      |                |



No. CNAS L0730

第 17 页 共 163 页

The scope of the accreditation in Chinese remains the definitive version.

| №  | Instrument                                               | Measurand     | Calibration Method                                                    | Range                                               | Expanded Uncertainty<br>(k=2) | Note | Effective Date |
|----|----------------------------------------------------------|---------------|-----------------------------------------------------------------------|-----------------------------------------------------|-------------------------------|------|----------------|
|    |                                                          |               |                                                                       | 标尺 5/250: (50~125) HBW                              | $U_{rel}=1.2\%$               |      |                |
|    |                                                          |               |                                                                       | 标尺 2.5/62.5: (50~125) HBW                           | $U_{rel}=1.2\%$               |      |                |
| 17 | *Metallic Vickers Hardness Testers                       | Hardness      | V.R.of Metallic Vickers Hardness Testers JJG 151                      | (400~600)HV <sub>10</sub>                           | $U_{rel}=1.0\%$               |      |                |
|    |                                                          |               |                                                                       | (175~225)HV <sub>5</sub> , (700~800)HV <sub>5</sub> | $U_{rel}=1.0\%$               |      |                |
|    |                                                          |               |                                                                       | (700~800)HV <sub>1</sub>                            | $U_{rel}=1.3\%$               |      |                |
|    |                                                          |               |                                                                       | (700~800)HV0.5                                      | $U_{rel}=1.7\%$               |      |                |
|    |                                                          |               |                                                                       | (400~800)HV0.2                                      | $U_{rel}=2.3\%$               |      |                |
|    |                                                          |               |                                                                       | (400~600)HV0.1                                      | $U_{rel}=2.5\%$               |      |                |
|    |                                                          |               |                                                                       | (175~225)HV0.05                                     | $U_{rel}=2.9\%$               |      |                |
|    |                                                          |               |                                                                       | (700~800)HV0.3                                      | $U_{rel}=2.0\%$               |      |                |
|    |                                                          |               |                                                                       | (700~800)HV <sub>3</sub>                            | $U_{rel}=1.1\%$               |      |                |
|    |                                                          |               |                                                                       | (400~600)HV <sub>30</sub>                           | $U_{rel}=0.5\%$               |      |                |
| 18 | Shore D Durometer                                        | Testing force | V.R.of Shore D Durometer JJG 1039                                     | (5~50)N                                             | $U=45\text{ mN}$              |      |                |
|    |                                                          | Length        |                                                                       | (0~150) mm                                          | $U=0.04\text{ mm}$            |      |                |
|    |                                                          | Angle         |                                                                       | 0° ~35°                                             | $U=6'$                        |      |                |
| 19 | *Elastic Element Precise Pressure Gauge and Vacuum Gauge | Pressure      | V.R.of Elastic Element Precise Pressure Gauge and Vacuum Gauge JJG 49 | (-0.1~+100) MPa                                     | $U=0.08\%FS$                  |      |                |



No. CNAS L0730

第 18 页 共 163 页

The scope of the accreditation in Chinese remains the definitive version.

| №  | Instrument                                                                                                                                       | Measurand              | Calibration Method                                                                                   | Range                                                                                               | Expanded Uncertainty<br>(k=2)          | Note | Effective Date |
|----|--------------------------------------------------------------------------------------------------------------------------------------------------|------------------------|------------------------------------------------------------------------------------------------------|-----------------------------------------------------------------------------------------------------|----------------------------------------|------|----------------|
| 20 | *Elastic Element Pressure Gauge, Pressure Vacuum Gauge and Vacuum Gauge for General Use/Electro Connecting Pressure Gauge/Capsule Pressure Gauge | Pressure               | V.R.of Elastic Element Pressure Gauge, Pressure Vacuum Gauge and Vacuum Gauge for General Use JJG 52 | Elastic Element Pressure Gauge, Pressure Vacuum Gauge and Vacuum Gauge for General: (-0.1~+100) MPa | $U=0.3\%FS$                            |      |                |
|    |                                                                                                                                                  | Pressure               |                                                                                                      | Electro Connecting Pressure Gauge: (-0.1~+100) MPa                                                  | $U=0.3\%FS$                            |      |                |
|    |                                                                                                                                                  | Pressure               |                                                                                                      | Capsule Pressure Gauge: (-0.1~+0.6)MPa                                                              | $U=0.3\%FS$                            |      |                |
| 21 | *Tyre Pressure Gauge                                                                                                                             | Pressure               | V.R.for For Tyre Pressure Gauge JJG 927                                                              | (0~2.5)MPa                                                                                          | $U=0.4\%FS$                            |      |                |
| 22 | *Taking Blood Electronic Scales                                                                                                                  | Mass                   | V.R. of Taking Blood Electronic Scales JJG 815                                                       | (20~500) g                                                                                          | $U=(0.1 \sim 0.2) \text{ g}$           |      |                |
|    |                                                                                                                                                  |                        |                                                                                                      | (500~1000) g                                                                                        | $U=(0.2 \sim 0.3) \text{ g}$           |      |                |
| 23 | Equotip Hardness Tester                                                                                                                          | Equotip Hardness Value | V.R.of Leeb Hardness Tester JJG 747                                                                  | (750~830) HLD                                                                                       | $U=10 \text{ HLD}$                     |      |                |
|    |                                                                                                                                                  |                        |                                                                                                      | (590~670) HLD                                                                                       | $U=8 \text{ HLD}$                      |      |                |
|    |                                                                                                                                                  |                        |                                                                                                      | (490~570) HLD                                                                                       | $U=7 \text{ HLD}$                      |      |                |
|    |                                                                                                                                                  |                        |                                                                                                      | (550~630) HLG                                                                                       | $U=8 \text{ HLG}$                      |      |                |
|    |                                                                                                                                                  |                        |                                                                                                      | (460~540) HLG                                                                                       | $U=6 \text{ HLG}$                      |      |                |
| 24 | *Thermogravimetric Moisture Meters                                                                                                               | Mass                   | V.R.of Thermogravimetric Moisture Meters JJG 658                                                     | (0.02~6100) g                                                                                       | $U=0.1 \text{ mg} \sim 0.09 \text{ g}$ |      |                |
|    |                                                                                                                                                  | Moisture               |                                                                                                      | NaCl: (0.01~99.994)%                                                                                | $U=0.05\%$                             |      |                |
| 25 | *Hydraulic Tension Jack                                                                                                                          | Force                  | V.R.of Hydraulic Jacks JJG 621                                                                       | 0.1 kN~15 MN                                                                                        | $U_{rel}=0.33\%$                       |      |                |



No. CNAS L0730

第 19 页 共 163 页

The scope of the accreditation in Chinese remains the definitive version.

| №  | Instrument                                        | Measurand        | Calibration Method                                               | Range             | Expanded Uncertainty<br>(k=2) | Note | Effective Date |
|----|---------------------------------------------------|------------------|------------------------------------------------------------------|-------------------|-------------------------------|------|----------------|
| 26 | Calibration<br>Instrument for<br>Torque Wrenches  | Torque force     | V.R. of Calibration<br>Instrument for Torque<br>Wrenches JJG 797 | (0.1~500) N·m     | $U_{rel}=0.3\%$               |      |                |
| 27 | Shore AO<br>Durometers                            | Testing Force    | C.S. for Shore AO<br>Durometers JJF 1312                         | (550~8050) mN     | $U=22$ mN                     |      |                |
|    |                                                   | Length           |                                                                  | (0~150) mm        | $U=0.01$ mm                   |      |                |
| 28 | *Digital Pressure<br>Gauges                       | Pressure         | V.R. of Digital Pressure<br>Gauges JJG 875                       | (-0.1~100) MPa    | $U=0.025\%$ FS                |      |                |
| 29 | *Pressure<br>Transmitter                          | Current          | V.R. of the Pressure<br>Transmitter JJG 882                      | (1~20) mA         | $U_{rel}=0.05\%$              |      |                |
|    |                                                   | Voltage          |                                                                  | (1~5) V           | $U_{rel}=0.05\%$              |      |                |
|    |                                                   | Pressure         |                                                                  | (-0.1~+100) MPa   | $U=0.05\%$                    |      |                |
| 30 | *Metallic Webster<br>Hardness Testing<br>Machines | Hardness         | V.R. of Metallic Webster<br>Hardness Testing Machines<br>JJG 944 | (5~18) HWA        | $U=0.6$ HWA                   |      |                |
|    |                                                   |                  |                                                                  | (5~18) HWB        | $U=0.7$ HWB                   |      |                |
|    |                                                   | Diameter         |                                                                  | (0~2) mm          | $U=0.01$ mm                   |      |                |
|    |                                                   | Angle            |                                                                  | (0~60) °          | $U=0.02$ °                    |      |                |
| 31 | *Spring Hammers                                   | Impact<br>Energy | C.S. for Spring Hammers JJF<br>1475                              | (0.20~2.00) J     | $U=0.004$ J                   |      |                |
| 32 | *Medical<br>Centrifuge                            | Rotate Speed     | C.S. for Medical Centrifuge<br>JJF (浙) 1117                      | (500~20000) r/min | $U_{rel}=(0.4\%\sim 0.1\%)$   |      |                |
| 33 | *Elevator<br>Overspeed<br>Governor Testers        | Speed            | C.S. for Elevator Overspeed<br>Governor Testers JJF 1374         | (0.5~20) m/s      | $U_{rel}=0.3\%$               |      |                |
| 34 | *Aeroid<br>Barometer/Aneroid<br>Barograph         | Pressure         | V.R. of Aneroid Barometer &<br>Aneroid Barograph JJG 272         | (520~1070) hPa    | $U=0.8$ hPa                   |      |                |
| 35 | *Piston Gauge                                     | Pressure         | V.R. of Piston Gauge JJG 59                                      | (0.005~100) MPa   | $U_{rel}=0.014\%$             |      |                |



No. CNAS L0730

第 20 页 共 163 页

The scope of the accreditation in Chinese remains the definitive version.

| №  | Instrument                                          | Measurand        | Calibration Method                                                 | Range                              | Expanded Uncertainty<br>( $k=2$ ) | Note | Effective Date |
|----|-----------------------------------------------------|------------------|--------------------------------------------------------------------|------------------------------------|-----------------------------------|------|----------------|
| 36 | *Piston Pressure Vacuum Gauges                      | Pressure         | V.R. of Piston Pressure Vacuum Gauges JJG236                       | (-0.1~-0.005) MPa                  | $U_{rel}=0.014\%$                 |      |                |
| 37 | *Pendulum Impact Testing Machines                   | Distance         | V.R. of Pendulum Impact Testing Machines JJG 145                   | (0.02~1) m                         | $U_{rel}=0.15\%$                  |      |                |
|    |                                                     | Moment           |                                                                    | (0.01~750) N·m                     | $U_{rel}=0.08\%$                  |      |                |
|    |                                                     | Potential Energy |                                                                    | (1~350) J                          | $U_{rel}=0.22\%$                  |      |                |
| 38 | *Cantivever-Beam (Izod-Type) Impact Testing Machine | Distance         | V.R. of Cantivever-Beam (Izod-Type) Impact Testing Machine JJG 608 | (0.02~1) m                         | $U_{rel}=0.21\%$                  |      |                |
|    |                                                     | Moment           |                                                                    | (0.001~100) N·m                    | $U_{rel}=0.10\%$                  |      |                |
|    |                                                     | Potential Energy |                                                                    | (1~350) J                          | $U_{rel}=0.22\%$                  |      |                |
| 39 | *Taximeters                                         | Rotate Speed     | V.R. of Taximeters JJG 517                                         | (50~1500) r/min                    | $U_{rel}=0.22\%$                  |      |                |
|    |                                                     | Working Time     |                                                                    | 1min~24h                           | $U_{rel}=0.24\%$                  |      |                |
| 40 | *Wire Bending Testing Machine                       | Mass             | V.R. of Wire Bending Testing Machine JJG(粤) 022                    | (50~6200) g                        | $U_{rel}=0.10\% \sim 0.15\%$      |      |                |
|    |                                                     | Angle            |                                                                    | $0^{\circ} \sim 360^{\circ}$       | $U=0.10^{\circ}$                  |      |                |
|    |                                                     | Speed            |                                                                    | (0~150) times per minute           | $U=0.11$ times per minute         |      |                |
| 41 | *Film Impact Testers                                | Mass             | C.S. of Film Impact Testers JJF(石化) 002                            | (0~2000) g                         | $U=0.3$ g                         |      |                |
|    |                                                     |                  |                                                                    | (0~600) mm                         | $U=(0.16 \sim 0.22)$ mm           |      |                |
| 42 | *Tilting Tube Micromanometers                       | Pressure value   | V.R. of Tilting Tube Micromanometers JJG 172                       | Curved tube type:(-2000~2000) Pa   | $U=(0.5 \sim 4)$ Pa               |      |                |
|    |                                                     |                  |                                                                    | Inclined tube type:(-2000~2000) Pa | $U=3.4$ Pa                        |      |                |
| 43 | *Portable Brinell Hardness Testers                  | Force value      | C.S. of Portable Brinell Hardness Testers JJF 1595                 | (3~30) kN                          | $U_{rel}=0.3\%$                   |      |                |



No. CNAS L0730

第 21 页 共 163 页

The scope of the accreditation in Chinese remains the definitive version.

| No             | Instrument                              | Measurand        | Calibration Method                                                         | Range                 | Expanded Uncertainty<br>(k=2)               | Note | Effective Date |
|----------------|-----------------------------------------|------------------|----------------------------------------------------------------------------|-----------------------|---------------------------------------------|------|----------------|
| 44             | *Portable Rockwell Harness Testers      | Hardness         | C.S.for Portable Rockwell Harness Testers JJF 1594                         | ( 100~400 ) HBS10/300 | U <sub>rel</sub> =1.2%                      |      |                |
|                |                                         |                  |                                                                            | ( 175~225 ) HBS5/750  | U <sub>rel</sub> =1.2%                      |      |                |
|                |                                         |                  |                                                                            | (75~650) HBW          | U <sub>rel</sub> =1.2%                      |      |                |
|                |                                         | Hardness         |                                                                            | (0.3~3) kN            | U <sub>rel</sub> =0.3%                      |      |                |
|                |                                         |                  |                                                                            | (22~88) HRA           | U=0.9 HRA                                   |      |                |
|                |                                         |                  |                                                                            | (20~100) HRBW         | U=1.0 HRBW                                  |      |                |
|                |                                         |                  |                                                                            | (20~70) HRC           | U=1.0 HRC                                   |      |                |
|                |                                         |                  |                                                                            | (70~94) HR15N         | U=0.9 HR15N                                 |      |                |
|                |                                         |                  |                                                                            | (42~86) HR30N         | U=0.9 HR30N                                 |      |                |
|                |                                         |                  |                                                                            | (20~77) HR45N         | U=0.9 HR45N                                 |      |                |
|                |                                         |                  |                                                                            | (67~93) HR15TW        | U=0.9 HR15TW                                |      |                |
| (29~82) HR30TW | U=0.9 HR30TW                            |                  |                                                                            |                       |                                             |      |                |
| (10~72) HR45TW | U=0.9 HR45TW                            |                  |                                                                            |                       |                                             |      |                |
| 45             | *Extensometer                           | Scal distnce     | V.R.of Extensometer JJG 762                                                | ( 1~1000 ) mm         | U <sub>rel</sub> =0.13%                     |      |                |
|                |                                         | Deformation      |                                                                            | ( 0~1/3 ) mm          | U=0.42 μ m                                  |      |                |
|                |                                         |                  |                                                                            | ( 1/3~25 ) mm         | U <sub>rel</sub> =0.17%                     |      |                |
| 46             | Standard Equipment for Revolution Speed | Revolution Speed | Verification Regulation of Standard Equipment for Revolution Speed JJG 326 | ( 10~60000 ) r/min    | U <sub>rel</sub> =(3.3~14)×10 <sup>-6</sup> |      |                |



No. CNAS L0730

第 22 页 共 163 页

The scope of the accreditation in Chinese remains the definitive version.

| №  | Instrument                                            | Measurand   | Calibration Method                                                                         | Range                       | Expanded Uncertainty<br>( $k=2$ ) | Note | Effective Date |
|----|-------------------------------------------------------|-------------|--------------------------------------------------------------------------------------------|-----------------------------|-----------------------------------|------|----------------|
| 47 | *Liquid Manometer for working                         | Pressure    | Verification Regulation of Liquid Manometers for working JJG540                            | (0~25)kPa                   | $U=0.3\%FS$                       |      |                |
| 48 | *Plastic Rock Well Hardness Testing Machine           | Hardness    | Verification Regulation of Plastic Rock Well Hardness Testing Machine JJG884               | (60~100)HRE                 | $U=0.6HRE$                        |      |                |
|    |                                                       | Hardness    |                                                                                            | (100~120)HRL                | $U=0.6HRL$                        |      |                |
|    |                                                       | Hardness    |                                                                                            | (80~100)HRM                 | $U=0.5HRM$                        |      |                |
|    |                                                       | Hardness    |                                                                                            | (100~130)HRR                | $U=0.5HRR$                        |      |                |
| 49 | *Pressure Regulators with Bourdon Tube Pressure Gauge | Pressure    | Calibration Specification for Pressure Regulators with Bourdon Tube Pressure Gauge JJF1328 | (0~25) MPa                  | $U=0.8\%FS$                       |      |                |
| 50 | *Electro-hydraulic Servo Universal Testing Machines   | Force value | Verification Regulation of Electro-hydraulic Servo Universal Testing Machines JJG1063      | 0.1cN~2MN                   | $U_{rel}=0.15\%$                  |      |                |
|    |                                                       |             |                                                                                            | 2MN~3MN                     | $U_{rel}=0.33\%$                  |      |                |
| 51 | *Mass Comparators                                     | Mass        | C.S.for Mass Comparators JJF 1326                                                          | Repeatability: 1mg~41kg     | $U=0.06 \mu g \sim 5.8mg$         |      |                |
|    |                                                       |             |                                                                                            | Partial load: 1mg~41kg      | $U=1.0 \mu g \sim 30mg$           |      |                |
|    |                                                       |             |                                                                                            | Local value error: 1mg~41kg | $U=1.6 \mu g \sim 30mg$           |      |                |
| 52 | *Falling Weight Impact Testing Machines               | Mass        | C.S.for Falling Weight Impact Testing Machines JJF 1445                                    | 5g~34kg                     | $U=0.7g$                          |      |                |
|    |                                                       | Height      |                                                                                            | (0.2~5)m                    | $U=(0.25 \sim 0.7)mm$             |      |                |
|    |                                                       | Speed       |                                                                                            | (1.8~10) m/s                | $U=(0.011 \sim 0.6) m/s$          |      |                |
|    |                                                       | Energy loss |                                                                                            | (1.0~50)%                   | $U=(0.4 \sim 2.2)\%$              |      |                |
| 53 | *Oedometers                                           | Force value | C.S.for Oedometers JJF 1311                                                                | 5N~20kN                     | $U_{rel}=0.4\%$                   |      |                |



No. CNAS L0730

第 23 页 共 163 页

The scope of the accreditation in Chinese remains the definitive version.

| No | Instrument                               | Measurand | Calibration Method                                      | Range                                | Expanded Uncertainty<br>( $k=2$ )            | Note | Effective Date |
|----|------------------------------------------|-----------|---------------------------------------------------------|--------------------------------------|----------------------------------------------|------|----------------|
| 54 | *Non-self-indicating weighing instrument | Mass      | V.R. of Non-self-indicating Weighing Instrument JJG 14  | Max:(10~100)kg<br>( $e=5g\sim 50g$ ) | $U=(0.21\sim 6.0)g$ , (0~500)e               |      |                |
|    |                                          |           |                                                         | Max:(10~100)kg<br>( $e=5g\sim 50g$ ) | $U=(0.38\sim 14)g$ , (>500~2000)e            |      |                |
|    |                                          |           |                                                         | Max:(10~100)kg<br>( $e=5g\sim 50g$ ) | $U=(2.1\sim 14)g$ , (>2000~3000)e            |      |                |
|    |                                          |           |                                                         | Max:100kg~1t<br>( $e=50g\sim 500g$ ) | $U=(2.1\sim 62)g$ , (0~500)e                 |      |                |
|    |                                          |           |                                                         | Max:100kg~1t<br>( $e=50g\sim 500g$ ) | $U=(3.8\sim 76)g$ , (>500~2000)e             |      |                |
|    |                                          |           |                                                         | Max:100kg~1t<br>( $e=50g\sim 500g$ ) | $U=(5.0\sim 76)g$ , (>2000~3000)e            |      |                |
|    |                                          |           |                                                         | Max:(1~10)t<br>( $e=0.5kg\sim 5kg$ ) | $U=(0.21\sim 0.38)kg$ , (0~500)e             |      |                |
|    |                                          |           |                                                         | Max:(1~10)t<br>( $e=0.5kg\sim 5kg$ ) | $U=(0.21\sim 0.80)kg$ , (>500~2000)e         |      |                |
|    |                                          |           |                                                         | Max:(1~10)t<br>( $e=0.5kg\sim 5kg$ ) | $U=(0.38\sim 1.8)kg$ , (>2000~3000)e         |      |                |
|    |                                          |           |                                                         | Max:(10~60)t<br>( $e=5kg\sim 20kg$ ) | $U=(1.8\sim 3.8)kg$ , (0~500)e               |      |                |
|    |                                          |           |                                                         | Max:(10~60)t<br>( $e=5kg\sim 20kg$ ) | $U=(1.8\sim 5.0)kg$ , (>500~2000)e           |      |                |
|    |                                          |           |                                                         | Max:(10~60)t<br>( $e=5kg\sim 20kg$ ) | $U=(3.8\sim 7.2)kg$ , (>2000~3000)e          |      |                |
| 55 | *Digital indicating weighing instruments | Mass      | V.R. of Digital indicating weighing instruments JJG 539 | (2~3000)g ( $e=0.1g\sim 0.5g$ )      | $U=(0.02\sim 0.21)g$ , ( $e=0.1g\sim 0.5g$ ) |      |                |
|    |                                          |           |                                                         | Max:(3~100)kg ( $e=1g\sim 50g$ )     | $U=(0.21\sim 6.0)g$ , (0~500)e               |      |                |



No. CNAS L0730

第 24 页 共 163 页

The scope of the accreditation in Chinese remains the definitive version.

| № | Instrument | Measurand | Calibration Method | Range                      | Expanded Uncertainty<br>( $k=2$ )            | Note | Effective Date |
|---|------------|-----------|--------------------|----------------------------|----------------------------------------------|------|----------------|
|   |            |           |                    | Max:(3~100)kg (e=1g~50g)   | $U=(0.38\sim 14)\text{g}$ , (>500~2000)e     |      |                |
|   |            |           |                    | Max:(3~100)kg (e=1g~50g)   | $U=(2.1\sim 14)\text{g}$ , (>2000~3000)e     |      |                |
|   |            |           |                    | Max:100kg~1t (e=50g~500g)  | $U=(2.1\sim 62)\text{g}$ , (0~500)e          |      |                |
|   |            |           |                    | Max:100kg~1t (e=50g~500g)  | $U=(3.8\sim 76)\text{g}$ , (>500~2000)e      |      |                |
|   |            |           |                    | Max:100kg~1t (e=50g~500g)  | $U=(5.0\sim 76)\text{g}$ , (>2000~3000)e     |      |                |
|   |            |           |                    | Max:(1~10)t (e=0.5kg~5kg)  | $U=(0.21\sim 0.38)\text{kg}$ , (0~500)e      |      |                |
|   |            |           |                    | Max:(1~10)t (e=0.5kg~5kg)  | $U=(0.21\sim 0.80)\text{kg}$ , (>500~2000)e  |      |                |
|   |            |           |                    | Max:(1~10)t (e=0.5kg~5kg)  | $U=(0.38\sim 1.8)\text{kg}$ , (>2000~30000)e |      |                |
|   |            |           |                    | Max:(10~100)t (e=5kg~50kg) | $U=(0.8\sim 1.8)\text{kg}$ , (0~500)e        |      |                |
|   |            |           |                    | Max:(10~100)t (e=5kg~50kg) | $U=(0.8\sim 3.8)\text{kg}$ , (>500~2000)e    |      |                |
|   |            |           |                    | Max:(10~100)t (e=5kg~50kg) | $U=(1.8\sim 13)\text{kg}$ , (>2000~3000)e    |      |                |
|   |            |           |                    | Max:(100~150)t (e=50kg)    | $U=(3.8\sim 5.0)\text{kg}$ , (0~500)e        |      |                |
|   |            |           |                    | Max:(100~150)t (e=50kg)    | $U=(5.0\sim 13)\text{kg}$ , (>500~2000)e     |      |                |
|   |            |           |                    | Max:(100~150)t (e=50kg)    | $U=(13\sim 22)\text{kg}$ , (>2000~3000)e     |      |                |



No. CNAS L0730

第 25 页 共 163 页

The scope of the accreditation in Chinese remains the definitive version.

| №  | Instrument                                                           | Measurand | Calibration Method                                                                  | Range                                   | Expanded Uncertainty<br>( $k=2$ )  | Note | Effective Date |
|----|----------------------------------------------------------------------|-----------|-------------------------------------------------------------------------------------|-----------------------------------------|------------------------------------|------|----------------|
| 56 | *Analogue<br>indicating<br>weighing<br>instrument                    | Mass      | V.R. of Analogue indicating<br>Weighing Instrument JJG 13                           | Max:(0~10)kg<br>( $e=10g\sim 50g$ )     | $U=(2.1\sim 6.0)g$ ,(0~50)e        |      |                |
|    |                                                                      |           |                                                                                     | Max:(0~10)kg<br>( $e=10g\sim 50g$ )     | $U=(2.1\sim 6.0)g$ ,(>50~200)e     |      |                |
|    |                                                                      |           |                                                                                     | Max:(0~10)kg<br>( $e=10g\sim 50g$ )     | $U=(2.1\sim 6.0)g$ ,(>200~1000)e   |      |                |
|    |                                                                      |           |                                                                                     | Max:(10~500)kg<br>( $e=100g\sim 500g$ ) | $U=(14\sim 58)g$ ,(0~50)e          |      |                |
|    |                                                                      |           |                                                                                     | Max:(10~500)kg<br>( $e=100g\sim 500g$ ) | $U=(14\sim 62)g$ ,(>50~200)e       |      |                |
|    |                                                                      |           |                                                                                     | Max:(10~500)kg<br>( $e=100g\sim 500g$ ) | $U=(14\sim 76)g$ ,(>200~1000)e     |      |                |
| 57 | *Steelyard Scale                                                     | Mass      | V.R. of Steelyard Scale JJG 17                                                      | Max:(0~800)g( $e=0.1\sim 2g$ )          | $U=(0.01\sim 0.05)g$ ,(0~50)e      |      |                |
|    |                                                                      |           |                                                                                     | Max:(0~800)g( $e=0.1\sim 2g$ )          | $U=(0.05\sim 0.21)g$ ,(>50~200)e   |      |                |
|    |                                                                      |           |                                                                                     | Max:(0~800)g( $e=0.1\sim 2g$ )          | $U=(0.21\sim 0.40)g$ ,(>200~1000)e |      |                |
| 58 | *Automatic<br>gravimetric filling<br>instruments                     | Mass      | V.R. of Automatic<br>Gravimetric Filling<br>Instrument JJG 564                      | 10g~5t                                  | $U_{rel}=0.1\%$                    |      |                |
| 59 | *Automatic<br>instruments for<br>weighing road<br>vehicles in motion | Mass      | V.R. of Automatic<br>Instruments for Weighing<br>Road Vehicles in Motion<br>JJG 907 | Static(0~50)t                           | $U=(1.8\sim 6.1)kg$                |      |                |
|    |                                                                      |           |                                                                                     | Dynamic(0~30)t                          | $U=19kg$ , ( $e=10kg$ )            |      |                |
|    |                                                                      |           |                                                                                     | Dynamic(0~30)t                          | $U=37kg$ , ( $e=20kg$ )            |      |                |
|    |                                                                      |           |                                                                                     | Dynamic(0~30)t                          | $U=91kg$ , ( $e=50kg$ )            |      |                |
|    | *Portable<br>weighing                                                | Mass      | C.S. of Portable Weighing<br>Instruments for Axles of                               | Static(0~50)t                           | $U=(1.8\sim 6.1)kg$                |      |                |



No. CNAS L0730

第 26 页 共 163 页

The scope of the accreditation in Chinese remains the definitive version.

| №  | Instrument                                                                                              | Measurand     | Calibration Method                                                                                             | Range                                                      | Expanded Uncertainty<br>( $k=2$ )     | Note | Effective Date |
|----|---------------------------------------------------------------------------------------------------------|---------------|----------------------------------------------------------------------------------------------------------------|------------------------------------------------------------|---------------------------------------|------|----------------|
|    | instruments for<br>axles of vehicle in<br>motion                                                        |               | Vehicle in Motion JJF 1212                                                                                     | Dynamic(0~30)t                                             | $U=19\text{kg}$ , ( $e=10\text{kg}$ ) |      |                |
|    |                                                                                                         |               |                                                                                                                | Dynamic(0~30)t                                             | $U=37\text{kg}$ , ( $e=20\text{kg}$ ) |      |                |
|    |                                                                                                         |               |                                                                                                                | Dynamic(0~30)t                                             | $U=91\text{kg}$ , ( $e=50\text{kg}$ ) |      |                |
| 61 | *Continuous<br>totalizing<br>automatic<br>weighing<br>instruments(Belt<br>weigher)                      | Mass          | V.R.of Continuous totalizing<br>automatic weighing<br>instruments(Belt weigher)<br>JJG 195                     | (100~8000) t/h                                             | $U_{\text{rel}}=0.08\%\sim 0.32\%$    |      |                |
| 62 | *Concrete<br>Batching Scales                                                                            | mass          | V.R.of Concrete Batching<br>Scales JJG 1171                                                                    | (0.01~15)t                                                 | $U_{\text{rel}}=0.26\%$               |      |                |
| 63 | *Discontinuous<br>Totalizing<br>Automatic<br>Weighing<br>Instruments(Totali<br>zing Hopper<br>Weighers) | mass          | V.R.of Discontinuous<br>Totalizing Automatic<br>Weighing<br>Instruments(Totalizing<br>Hopper Weighers) JJG 648 | (0.01~150)t                                                | $U_{\text{rel}}=0.11\%$               |      |                |
| 64 | Edge test<br>apparatus                                                                                  | Force value   | C.S. for edge test apparatus<br>ZDCD-011L                                                                      | (0.1~10)N                                                  | $U=0.3\text{N}$                       |      |                |
| 65 | *Buoy Type<br>Oxygen Inhalers                                                                           | Pressure      | Verification Regulation of<br>Buoy Type Oxygen Inhalers<br>JJG913                                              | (0~16)MPa                                                  | $U=0.6\text{MPa}$                     |      |                |
|    |                                                                                                         | Flow rate     |                                                                                                                | (0~10)L/min                                                | $U=0.4\text{L/min}$                   |      |                |
| 66 | Working Glass<br>Hydrometers                                                                            | Density value | V.R. of Working Glass<br>Hydrometers JJG 42                                                                    | Working Densitometer:<br>(650~1400)kg/m <sup>3</sup>       | $U=0.4\text{kg/m}^3$                  |      |                |
|    |                                                                                                         | Density value |                                                                                                                | Petroleum<br>Densitometer: (650~<br>1100)kg/m <sup>3</sup> | $U=0.3\text{kg/m}^3$                  |      |                |



No. CNAS L0730

第 27 页 共 163 页

The scope of the accreditation in Chinese remains the definitive version.

| №  | Instrument                                      | Measurand       | Calibration Method                                             | Range                     | Expanded Uncertainty<br>( $k=2$ ) | Note | Effective Date |
|----|-------------------------------------------------|-----------------|----------------------------------------------------------------|---------------------------|-----------------------------------|------|----------------|
|    |                                                 | Baume Degrees   |                                                                | Baume Hydrometer:(0~40)Bh | $U_{rel}=0.2Bh$                   |      |                |
|    |                                                 | Alcohol content |                                                                | Alcohol Meter:q:(0~100)%  | $U=0.2\%$                         |      |                |
| 67 | *Traffic Loop-based speed Meters                | speed           | V.R. of Traffic Loop-based Speed Meters JJG 1122               | (20~180) km/h             | $U_{rel}=1.8\%$ (spot)            |      |                |
|    |                                                 |                 |                                                                | (20~180) km/h             | $U_{rel}=1.0\%$ (simulation)      |      |                |
| 68 | *Fixed Radar Vehicle Speed Measurement Devices  | speed           | V.R. of Fixed Radar Vehicle Speed Measurement Devices JJG 527  | (20~180)km/h              | $U_{rel}=1.8\%$ (spot)            |      |                |
|    |                                                 |                 |                                                                | (20~180) km/h             | $U_{rel}=1.0\%$ (simulation)      |      |                |
|    |                                                 | Frequency       |                                                                | (8~40) GHz                | $U_{rel}=0.013\%$                 |      |                |
| 69 | *Mobile Radar Vehicle Speed Measurement Devices | spot speed      | V.R. of Mobile Radar Vehicle Speed Measurement Devices JJG 528 | (20~180)km/h              | $U_{rel}=1.8\%$ (spot)            |      |                |
|    |                                                 |                 |                                                                | (20~180) km/h             | $U_{rel}=1.0\%$ (simulation)      |      |                |
|    |                                                 | Frequency       |                                                                | (8~40) GHz                | $U_{rel}=0.013\%$                 |      |                |
| 70 | Rotary torque sensor                            | Torque          | Rotary torque sensor FFK1901                                   | (3-300) N·m               | $U_{rel}=1.2\%$                   |      |                |
| 71 | Standard Dynamo- meter                          | Force           | V.R. of Standard Dynamometers JJG144                           | (5~500) kN                | $U_{rel}=0.015\%$                 |      |                |
|    |                                                 |                 |                                                                | (0.5~10) MN               | $U_{rel}=0.08\%$                  |      |                |
|    |                                                 |                 |                                                                | (10~30) MN                | $U_{rel}=0.10\%$                  |      |                |



No. CNAS L0730

第 28 页 共 163 页

The scope of the accreditation in Chinese remains the definitive version.

| No                                     | Instrument                                                                                   | Measurand | Calibration Method                                                                                                                                                                                                                      | Range                                             | Expanded Uncertainty<br>(k=2) | Note | Effective Date |
|----------------------------------------|----------------------------------------------------------------------------------------------|-----------|-----------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|---------------------------------------------------|-------------------------------|------|----------------|
| 72                                     | Working<br>Dynamo- meters                                                                    | Force     | V.R.of Working<br>Dynamometers<br>JJG455,V.R.of Standard<br>Dynamometers ISO376<br>Item1-9,Appendix A-<br>C,Calibration of force -<br>proving instruments used for<br>the verification GB/T13634<br>GB/T13634 Item 1-9,<br>Appendix A,B | 3 MN~30 MN                                        | $U_{rel}=0.3\%$               |      |                |
| 73                                     | Force-proving<br>instruments used<br>for the verification<br>of uniaxial testing<br>machines | Force     | Calibration of force - proving<br>instruments used for the<br>verification GB/T13634<br>GB/T13634,V.R.of Standard<br>Dynamometers ISO376                                                                                                | (5~500) kN                                        | $U_{rel}=0.015\%$             |      |                |
|                                        |                                                                                              |           |                                                                                                                                                                                                                                         | (0.5~10) MN                                       | $U_{rel}=0.08\%$              |      |                |
|                                        |                                                                                              |           |                                                                                                                                                                                                                                         | (10~30) MN                                        | $U_{rel}=0.10\%$              |      |                |
| 74                                     | Standard<br>Torque-meters                                                                    | Torque    | V.R.of Standard Torque-<br>meters JJG557                                                                                                                                                                                                | (2~5000)Nm                                        | $U_{rel}=(0.04\sim0.14)\%$    |      |                |
| 4 Flow & Capacity Measuring Instrument |                                                                                              |           |                                                                                                                                                                                                                                         |                                                   |                               |      |                |
| 1                                      | *Quantitative<br>filling machine for<br>liquid material                                      | Capacity  | V.R.of Quantitative filling<br>machine for liquid material<br>JJG 687                                                                                                                                                                   | Capacity quota filling<br>machine: (0.01~50)<br>L | $U_{rel}=0.33\%$              |      |                |
|                                        |                                                                                              | Mass      |                                                                                                                                                                                                                                         | Mass quota filling<br>machine: (0.01~50)kg        | $U_{rel}=0.10\%$              |      |                |
| 2                                      | *vertical metal<br>tank                                                                      | capacity  | V.R of vertical metal tank<br>capacity JJG 168                                                                                                                                                                                          | (20~100)m <sup>3</sup>                            | $U_{rel}=0.3\%$               |      |                |
|                                        |                                                                                              |           |                                                                                                                                                                                                                                         | (100~700)m <sup>3</sup>                           | $U_{rel}=0.2\%$               |      |                |
|                                        |                                                                                              |           |                                                                                                                                                                                                                                         | (700~200000) m <sup>3</sup>                       | $U_{rel}=0.1\%$               |      |                |



No. CNAS L0730

第 29 页 共 163 页

The scope of the accreditation in Chinese remains the definitive version.

| № | Instrument                 | Measurand             | Calibration Method                                    | Range                                              | Expanded Uncertainty<br>( $k=2$ ) | Note | Effective Date |
|---|----------------------------|-----------------------|-------------------------------------------------------|----------------------------------------------------|-----------------------------------|------|----------------|
| 3 | *horizontal metallic tank  | capacity              | V.R of the volume of horizontal metallic tank JJG 266 | (10~100)m <sup>3</sup>                             | $U_{rel}=0.4\%$                   |      |                |
| 4 | Working Glass Container    | Capacity of container | V.R.of Working Glass Container JJG 196                | (0~1) mL                                           | $U=0.003\text{mL}$                |      |                |
|   |                            |                       |                                                       | (1~25) mL                                          | $U=0.008\text{mL}$                |      |                |
|   |                            |                       |                                                       | (25~100) mL                                        | $U=0.03\text{mL}$                 |      |                |
|   |                            |                       |                                                       | (100~1000) mL                                      | $U=0.11\text{mL}$                 |      |                |
|   |                            |                       |                                                       | (1000~2000) mL                                     | $U=0.24\text{mL}$                 |      |                |
| 5 | Locomotive Pipette         | volume                | V.R.of Locomotive Pipette JJG 646                     | (0.1~5) $\mu\text{L}$                              | $U=0.01\text{ }\mu\text{L}$       |      |                |
|   |                            |                       |                                                       | (>5~30) $\mu\text{L}$                              | $U=0.04\text{ }\mu\text{L}$       |      |                |
|   |                            |                       |                                                       | (>30~50) $\mu\text{L}$                             | $U=0.07\text{ }\mu\text{L}$       |      |                |
|   |                            |                       |                                                       | (>50~100) $\mu\text{L}$                            | $U=0.18\text{ }\mu\text{L}$       |      |                |
|   |                            |                       |                                                       | (>100~1000) $\mu\text{L}$                          | $U=1.3\text{ }\mu\text{L}$        |      |                |
|   |                            |                       |                                                       | (>1000~5000) $\mu\text{L}$                         | $U=1.8\text{ }\mu\text{L}$        |      |                |
|   |                            |                       |                                                       | (>5000~10000) $\mu\text{L}$                        | $U=2.8\text{ }\mu\text{L}$        |      |                |
| 6 | Special Glassware          | volume                | V.R.of Special Glassware JJG 10                       | Micropipette: 20 $\mu\text{L}$                     | $U=0.2\text{ }\mu\text{L}$        |      |                |
|   |                            |                       |                                                       | Volumetric pipette: (0.25~10) mL                   | $U=0.002\text{ mL}$               |      |                |
|   |                            |                       |                                                       | Dissolved oxygenin sea water burette: (0.05~25) mL | $U=0.01\text{ mL}$                |      |                |
| 7 | *Cold Potable Water Meters | Flow                  | V.R.of Cold Potable Water Meters JJG 162              | (0.004~3500) m <sup>3</sup> /h,(DN8~DN800)         | $U_{rel}=0.3\%$                   |      |                |



No. CNAS L0730

第 30 页 共 163 页

The scope of the accreditation in Chinese remains the definitive version.

| №  | Instrument                             | Measurand | Calibration Method                                         | Range                                           | Expanded Uncertainty<br>( $k=2$ ) | Note | Effective Date |
|----|----------------------------------------|-----------|------------------------------------------------------------|-------------------------------------------------|-----------------------------------|------|----------------|
| 8  | Diaphragm Gas Meters                   | Flow      | V.R. of Diaphragm Gas Meters JJG577                        | (0.01~120) m <sup>3</sup> /h, (G1.6~G65)        | $U_{rel}=0.38\%$                  |      |                |
|    |                                        |           |                                                            | (120~192) m <sup>3</sup> /h, (G65~G100)         | $U_{rel}=0.5\%$                   |      |                |
| 9  | *Fuel Dispensers                       | Volume    | V.R. of Fuel Dispensers (for Trial Implementation) JJG 443 | (1~500) L/min                                   | $U_{rel}=0.07\%$                  |      |                |
| 10 | *Compressed Natural Gas Dispensers     | Flow      | V.R. of Compressed Natural Gas Dispensers JJG996           | (1~25) kg/min                                   | $U_{rel}=0.4\%$                   |      |                |
| 11 | *Weirs and Flumes for Flow Measurement | Flow      | V.R. of Weirs and Flumes for Flow Measurement JJG711       | (0.001~93) m <sup>3</sup> /s                    | $U_{rel}=1.5\%$                   |      |                |
| 12 | *Standard Metal Tank                   | Volume    | V.R. of Standard Metal Tank JJG259                         | (10~50) L                                       | $U_{rel}=0.06\%$                  |      |                |
| 13 | Gas Displacement Meters                | Flow      | V.R. of Gas Displacement Meters JJG633                     | (0.01~120) m <sup>3</sup> /h, (DN1~DN80)        | $U_{rel}=0.25\%$                  |      |                |
|    |                                        |           |                                                            | (120~7000) m <sup>3</sup> /h, (DN15~DN400)      | $U_{rel}=0.3\%$                   |      |                |
| 14 | Float Meter                            | Flow      | V.R. of Float Meter JJG257                                 | (0.01~120) m <sup>3</sup> /h, (DN1~DN80, gas)   | $U=0.3\%FS$                       |      |                |
|    |                                        |           |                                                            | (120~7000) m <sup>3</sup> /h, (DN15~DN400, gas) | $U=0.4\%FS$                       |      |                |
|    |                                        |           |                                                            | (1~100) mL/min, (DN1~DN10, water)               | $U=0.3\%FS$                       |      |                |
|    |                                        |           |                                                            | (0.004~20) m <sup>3</sup> /h, (DN6~DN40, water) | $U=0.2\%FS$                       |      |                |



在线扫码获取验证

No. CNAS L0730

第 31 页 共 163 页

The scope of the accreditation in Chinese remains the definitive version.


| No | Instrument                   | Measurand | Calibration Method                          | Range                                                          | Expanded Uncertainty<br>(k=2) | Note | Effective Date |
|----|------------------------------|-----------|---------------------------------------------|----------------------------------------------------------------|-------------------------------|------|----------------|
|    |                              |           |                                             | (1~120)m <sup>3</sup> /h, (DN15~<br>DN100,water)               | $U=0.3\%FS$                   |      |                |
| 15 | Goriolis Mass<br>Flow Meters | Flow      | V.R.of Goriolis Mass Flow<br>Meters JJG1038 | (0.01~120) m <sup>3</sup> /h,<br>(DN15~DN80, gas)              | $U_{rel}=0.25\%$              |      |                |
|    |                              |           |                                             | (120~<br>7000)m <sup>3</sup> /h,(DN15~<br>DN400,gas)           | $U_{rel}=0.3\%$               |      |                |
|    |                              |           |                                             | (0.004~1600) t/h,<br>(DN2~DN300,<br>water)                     | $U_{rel}=0.05\%$              |      |                |
|    |                              |           |                                             | (1600~3500) t/h,<br>(DN300~DN800,<br>water)                    | $U_{rel}=0.15\%$              |      |                |
|    |                              |           |                                             | (0.5~98)m <sup>3</sup> /h,(0.42~<br>80)t/h,(DN15~<br>DN80,oil) | $U_{rel}=0.06\%$              |      |                |
| 16 | *Ultrasonic<br>Flowmeters    | Flow      | V.R.of Ultrasonic Flowmeters<br>JJG1030     | (0.004~<br>1600)m <sup>3</sup> /h,(DN4~<br>DN300,water)        | $U_{rel}=0.06\%$              |      |                |
|    |                              |           |                                             | (1600~<br>3500)m <sup>3</sup> /h,(DN300~<br>DN2400,water)      | $U_{rel}=0.15\%$              |      |                |
|    |                              |           |                                             | (0.01~<br>120)m <sup>3</sup> /h,(DN15~<br>DN80,gas)            | $U_{rel}=0.25\%$              |      |                |
|    |                              |           |                                             | (120~<br>7000)m <sup>3</sup> /h,(DN15~<br>DN400,gas)           | $U_{rel}=0.3\%$               |      |                |



No. CNAS L0730

第 32 页 共 163 页

The scope of the accreditation in Chinese remains the definitive version.

| №  | Instrument                              | Measurand | Calibration Method                                     | Range                                                  | Expanded Uncertainty<br>( $k=2$ ) | Note                                                                                 | Effective Date |
|----|-----------------------------------------|-----------|--------------------------------------------------------|--------------------------------------------------------|-----------------------------------|--------------------------------------------------------------------------------------|----------------|
| 17 | Turbine Flowmeter                       | flow      | V.R. of Turbine Flowmeter JJG1037                      | (0.01~120)m <sup>3</sup> /h,(DN15~DN80,gas)            | $U_{rel}=0.25\%$                  |                                                                                      |                |
|    |                                         |           |                                                        | (120~7000)m <sup>3</sup> /h,(DN15~DN400,gas)           | $U_{rel}=0.3\%$                   |                                                                                      |                |
|    |                                         |           |                                                        | (0.004~1600)m <sup>3</sup> /h,(DN4~DN300,water)        | $U_{rel}=0.06\%$                  |                                                                                      |                |
|    |                                         |           |                                                        | (1600~3500)m <sup>3</sup> /h,(DN300~DN800,water)       | $U_{rel}=0.15\%$                  |                                                                                      |                |
|    |                                         |           |                                                        | (0.5~98)m <sup>3</sup> /h,(0.42~80)t/h,(DN15~DN80,oil) | $U_{rel}=0.06\%$                  |                                                                                      |                |
| 18 | *Liquefied Natural Gas Dispensers       | Flow      | V.R. of Liquefied Natural Gas Dispensers JJG 1114      | (1~80) kg/min                                          | $U_{rel}=0.48\%$                  |                                                                                      |                |
| 19 | *Electromagnetic Flowmeters (on line )  | Flow      | C.S. for Electromagnetic Flowmeters JJF(Su) 228        | (0.1~50000) m <sup>3</sup> /h,(DN10~DN3000)            | $U_{rel}=1.3\%$                   |                                                                                      |                |
| 20 | *Liquid Positive Displacement Flowmeter | Flow      | V.R. of Liquid Positive Displacement flow meter JJG667 | (0.004~1600)m <sup>3</sup> /h,(DN4~DN300,water)        | $U_{rel}=0.06\%$                  |  |                |
|    |                                         |           |                                                        | (1600~3500)m <sup>3</sup> /h,(DN300~DN800,water)       | $U_{rel}=0.15\%$                  |                                                                                      |                |



No. CNAS L0730

第 33 页 共 163 页

The scope of the accreditation in Chinese remains the definitive version.

| №  | Instrument                           | Measurand | Calibration Method                                  | Range                                                  | Expanded Uncertainty<br>( $k=2$ ) | Note | Effective Date |
|----|--------------------------------------|-----------|-----------------------------------------------------|--------------------------------------------------------|-----------------------------------|------|----------------|
|    |                                      |           |                                                     | (0.5~98)m <sup>3</sup> /h,(0.42~80)t/h,(DN15~DN80,oil) | $U_{rel}=0.06\%$                  |      |                |
| 21 | *Tank Cars Capacity                  | Volume    | V.R. of Tank cars capacity JJG133                   | (0.1~100)m <sup>3</sup>                                | $U_{rel}=0.08\%$                  |      |                |
| 22 | *Electromagnetic Flowmeters          | Flow      | V.R. of Electromagnetic Flowmeters JJG1033          | (0.004~1600)m <sup>3</sup> /h,(DN4~DN300)              | $U_{rel}=0.06\%$                  |      |                |
|    |                                      |           |                                                     | (1600~3500)m <sup>3</sup> /h,(DN300~DN800)             | $U_{rel}=0.15\%$                  |      |                |
| 23 | Vortex-shedding Flowmeter            | Flow      | V.R. of Vortex-shedding Flowmeter JJG1029           | (0.004~1600)m <sup>3</sup> /h,(DN15~DN300,water)       | $U_{rel}=0.12\%$                  |      |                |
|    |                                      |           |                                                     | (1600~3500)m <sup>3</sup> /h,(DN300~DN800,water)       | $U_{rel}=0.15\%$                  |      |                |
|    |                                      |           |                                                     | (0.01~120)m <sup>3</sup> /h,(DN6~DN80,gas)             | $U_{rel}=0.25\%$                  |      |                |
|    |                                      |           |                                                     | (120~7000)m <sup>3</sup> /h,(DN15~DN400,gas)           | $U_{rel}=0.3\%$                   |      |                |
| 24 | Differential Pressure Type Flowmeter | Flow      | V.R. of Differential Pressure Type Flowmeter JJG640 | (0.004~1600)m <sup>3</sup> /h,(DN15~DN300,water)       | $U_{rel}=0.12\%$                  |      |                |
|    |                                      |           |                                                     | (1600~3500)m <sup>3</sup> /h,(DN300~DN800,water)       | $U_{rel}=0.15\%$                  |      |                |



No. CNAS L0730

第 34 页 共 163 页

The scope of the accreditation in Chinese remains the definitive version.

| №  | Instrument                                               | Measurand | Calibration Method                                            | Range                                                | Expanded Uncertainty<br>( $k=2$ ) | Note | Effective Date |
|----|----------------------------------------------------------|-----------|---------------------------------------------------------------|------------------------------------------------------|-----------------------------------|------|----------------|
|    |                                                          |           |                                                               | (0.01 ~ 7000)m <sup>3</sup> /h,(DN15 ~ DN400,gas)    | $U_{rel}=0.3\%$                   |      |                |
| 25 | Heat Meters                                              | Flow      | V.R.of Heat Meters JJG225                                     | (0.004 ~ 32)m <sup>3</sup> /h,(DN15 ~ DN80)          | $U_{rel}=0.3\%$                   |      |                |
| 26 | Thermal mass gas Flowmeter                               | flow      | V.R.of Thermal mass gas Flowmeter JJG1132                     | (0.01 ~ 120)m <sup>3</sup> /h,(DN1 ~ DN80)           | $U_{rel}=0.25\%$                  |      |                |
|    |                                                          |           |                                                               | (120 ~ 7000)m <sup>3</sup> /h,(DN15 ~ DN400)         | $U_{rel}=0.3\%$                   |      |                |
| 27 | Target Flowmeter                                         | flow      | V.R.of Target Flowmeter JJG461                                | (0.004 ~ 3500)m <sup>3</sup> /h,(DN15 ~ DN800,water) | $U_{rel}=0.15\%$                  |      |                |
| 28 | Vortex precession Flowmeters                             | flow      | V.R.of Vortex Precession Flowmeters JJG1121                   | (0.01 ~ 7000)m <sup>3</sup> /h,(DN15 ~ DN400,gas)    | $U_{rel}=0.3\%$                   |      |                |
|    |                                                          |           |                                                               | (0.004 ~ 3500)m <sup>3</sup> /h,(DN15 ~ DN800,water) | $U_{rel}=0.25\%$                  |      |                |
| 29 | *Pipe Prover                                             | volume    | V.R.of Pipe Prover JJG209                                     | (1 ~ 10000)L                                         | $U_{rel}=0.028\%$                 |      |                |
| 30 | *Standard Facilities for Liquid Flowmeter                | Flow      | V.R.of Standard Facilities for Liquid Flow rate JJG164        | (0.01 ~ 3000)m <sup>3</sup> /h                       | $U_{rel}=0.05\%$                  |      |                |
| 31 | *Flow Standard Facilities by Master Meter Method(Liquid) | Flow      | V.R.of Flow Standard Facilities by Master Meter Method JJG643 | (0.2 ~ 10000)m <sup>3</sup> /h                       | $U_{rel}=0.15\%$                  |      |                |
| 32 | Wet Gas Meters                                           | Flow      | C.S.for Wet Gas Meters JJF1357                                | (0.01 ~ 120)m <sup>3</sup> /h,(DN10 ~ DN80)          | $U_{rel}=0.22\%$                  |      |                |



No. CNAS L0730

第 35 页 共 163 页

The scope of the accreditation in Chinese remains the definitive version.


| №                                           | Instrument                        | Measurand | Calibration Method                                | Range                                    | Expanded Uncertainty<br>( $k=2$ ) | Note | Effective Date |
|---------------------------------------------|-----------------------------------|-----------|---------------------------------------------------|------------------------------------------|-----------------------------------|------|----------------|
| 33                                          | Critical Flow Venturi Nozzle      | Flow      | V.R.of Critical Flow Venturi Nozzle JJG 620       | (0.016~6)m <sup>3</sup> /h               | $U_{rel}=0.08\%$                  |      |                |
| 34                                          | Gas Laminar Flow Transducers      | Flow      | V.R.of Gas Laminar Flow Transducer JJG 736        | (0.01~7000)m <sup>3</sup> /h,(DN1~DN400) | $U_{rel}=0.3\%$                   |      |                |
| 35                                          | *Flow Integration Meters          | Flow      | V.R.of Flow Intergration Meters JJG1003           | (0.1~9999.99)m <sup>3</sup> /h           | $U_{rel}=0.02\%$                  |      |                |
|                                             |                                   |           |                                                   | (0.1~9999.99)t/h                         | $U_{rel}=0.04\%$                  |      |                |
|                                             |                                   | Frequency |                                                   | 10Hz~15MHz                               | $U_{rel}=3 \times 10^{-6}$        |      |                |
|                                             |                                   | DCV       |                                                   | (0.1~10)V                                | $U_{rel}=1 \times 10^{-4}$        |      |                |
|                                             |                                   | DCA       |                                                   | (0.1~20) mA                              | $U_{rel}=1 \times 10^{-4}$        |      |                |
|                                             |                                   | OHM       |                                                   | (0.01~50) k $\Omega$                     | $U_{rel}=1 \times 10^{-4}$        |      |                |
| 36                                          | Ultrasonic Gas Meters             | Flow      | V.R.of Ultrasonic Gas Meters JJG1190              | (0.01~120) m <sup>3</sup> /h,(G1.6~G65)  | $U_{rel}=0.23\%$                  |      |                |
|                                             |                                   |           |                                                   | (120~192) m <sup>3</sup> /h,(G65~G100)   | $U_{rel}=0.39\%$                  |      |                |
| 37                                          | *Aqueous Urea Solution Dispensers | Flow      | V.R.of Aqueous Urea Solution Dispensers JJG1191   | (1~50) L/min                             | $U_{rel}=0.08\%$                  |      |                |
| 38                                          | Gasoline Vapor Recovery Detectors | Flow      | C.S.for Gasoline Vapor Recovery Detectors JJF1948 | (0.01~30)m <sup>3</sup> /h               | $U_{rel}=0.8\%$                   |      |                |
|                                             |                                   | Pressure  |                                                   | (-5.0~5.0)kPa                            | $U=0.06\%FS$                      |      |                |
|                                             |                                   | Time      |                                                   | (0.01~3600)s                             | $U=0.06s$                         |      |                |
| 39                                          | *Fuel Dispensers For Ships        | Flow      | V.R.of Fuel Dispensers For Ships JJG(Yue)068      | (30~500)L/min                            | $U_{rel}=0.06\%$                  |      |                |
| 5 Acoustics & Vibrance Measuring Instrument |                                   |           |                                                   |                                          |                                   |      |                |



No. CNAS L0730

第 36 页 共 163 页

The scope of the accreditation in Chinese remains the definitive version.

| № | Instrument                                                            | Measurand               | Calibration Method                                                                  | Range                        | Expanded Uncertainty<br>( <i>k</i> =2) | Note                                                                                 | Effective Date |
|---|-----------------------------------------------------------------------|-------------------------|-------------------------------------------------------------------------------------|------------------------------|----------------------------------------|--------------------------------------------------------------------------------------|----------------|
| 1 | *Audio-frequency<br>Signal Generator                                  | Frequency               | V.R. of Audio-frequency<br>Signal Generator JJG 607                                 | 10Hz~100Hz                   | <i>U</i> <sub>rel</sub> =0.11%         |                                                                                      |                |
|   |                                                                       |                         |                                                                                     | 100Hz~200kHz                 | <i>U</i> <sub>rel</sub> =0.02%         |                                                                                      |                |
|   |                                                                       | AC Voltage              |                                                                                     | (1~100)mV                    | <i>U</i> <sub>rel</sub> =0.5%          |                                                                                      |                |
|   |                                                                       |                         |                                                                                     | 100mV~100V                   | <i>U</i> <sub>rel</sub> =0.20%         |                                                                                      |                |
|   |                                                                       | Impedance               |                                                                                     | 2 Ω ~10k Ω                   | <i>U</i> <sub>rel</sub> =0.5%          |                                                                                      |                |
|   |                                                                       | Distortion              |                                                                                     | (0.002~100)%                 | <i>U</i> <sub>rel</sub> =15%           |                                                                                      |                |
|   |                                                                       | Sound<br>pressure level |                                                                                     | (10~130)dB, (10Hz~<br>20kHz) | <i>U</i> =0.2dB                        |                                                                                      |                |
| 2 | *Mechanical<br>Vibration<br>Genetator for<br>Testing                  | Acceleration            | V.R. of Mechanical Vibration<br>Genetator for Testing JJG 189                       | (1~100) m/s <sup>2</sup>     | <i>U</i> <sub>rel</sub> =1.7%          |                                                                                      |                |
|   |                                                                       | Displacement            |                                                                                     | (0.01~10) mm                 | <i>U</i> <sub>rel</sub> =2.5%          |                                                                                      |                |
|   |                                                                       | Frequency               |                                                                                     | (5~2000)Hz                   | <i>U</i> <sub>rel</sub> =0.5%          |                                                                                      |                |
|   |                                                                       | Distortion              |                                                                                     | (0.01~30)%                   | <i>U</i> <sub>rel</sub> =10%           |                                                                                      |                |
| 3 | *Electrodynamic<br>Vibration Testing<br>Systems                       | Acceleration            | V.R. of Electrodynamic<br>Vibration Testing Systems<br>JJG 948                      | (1~100) m/s <sup>2</sup>     | <i>U</i> <sub>rel</sub> =1.7%          |                                                                                      |                |
|   |                                                                       | Frequency               |                                                                                     | (5~2000)Hz                   | <i>U</i> <sub>rel</sub> =0.5%          |                                                                                      |                |
|   |                                                                       | Distortion              |                                                                                     | (0.01~30)%                   | <i>U</i> <sub>rel</sub> =10%           |                                                                                      |                |
| 4 | *Electrodynamic<br>Horizontal<br>Vibraton<br>Generator for<br>Testing | Acceleration            | V.R. of Electrodynamic<br>Horizontal Vibration<br>Generator for Testing JJG<br>1000 | (1~100) m/s <sup>2</sup>     | <i>U</i> <sub>rel</sub> =1.7%          |  |                |
|   |                                                                       | Displacement            |                                                                                     | (0.01~10) mm                 | <i>U</i> <sub>rel</sub> =2.5%          |                                                                                      |                |
|   |                                                                       | Frequency               |                                                                                     | (5~2000)Hz                   | <i>U</i> <sub>rel</sub> =0.5%          |                                                                                      |                |
|   |                                                                       | Distortion              |                                                                                     | (0.01~30)%                   | <i>U</i> <sub>rel</sub> =10%           |                                                                                      |                |



No. CNAS L0730

第 37 页 共 163 页

The scope of the accreditation in Chinese remains the definitive version.

| №  | Instrument                       | Measurand            | Calibration Method                                             | Range                                  | Expanded Uncertainty<br>( $k=2$ ) | Note | Effective Date |
|----|----------------------------------|----------------------|----------------------------------------------------------------|----------------------------------------|-----------------------------------|------|----------------|
| 5  | Piezoelectric Accelerometer      | Acceleration         | V.R. of Piezoelectric Accelerometer JJG 233                    | (1~100)m/s <sup>2</sup> , (20~2000)Hz  | $U_{rel}=1.2\%$                   |      |                |
| 6  | Vibration Meters                 | Acceleration         | V.R. of Vibration Meters JJG 676                               | (1~100)m/s <sup>2</sup> , (20~2000)Hz  | $U_{rel}=1.2\%$                   |      |                |
|    |                                  | Speed                |                                                                | (0.1~100)mm/s                          | $U_{rel}=3.0\%$                   |      |                |
|    |                                  | Displacement         |                                                                | (0.1~100)mm                            | $U_{rel}=3.0\%$                   |      |                |
| 7  | *Shock and Bump Testing Machines | Acceleration         | V.R. of Shock and Bump Testing Machines JJG 1174               | (70~10000)m/s <sup>2</sup>             | $U_{rel}=5\%$                     |      |                |
|    |                                  | Pulse Width          |                                                                | (0.5~30)ms                             | $U_{rel}=3.0\%$                   |      |                |
| 8  | *Filters                         | Relative attenuation | V.R. of Octave-Band and Fractional-Octave-Band Filters JJG 449 | (0.1~100)dB                            | $U=0.1\text{dB}$                  |      |                |
|    |                                  | Frequency            |                                                                | 20Hz~100kHz                            | $U_{rel}=0.01\%$                  |      |                |
| 9  | *SINAD Meters                    | SINAD                | C.S. for SINAD Meters JJF 1165                                 | (0~20) dB                              | $U=0.3\text{dB}$                  |      |                |
|    |                                  | Frequency            |                                                                | 100Hz~10kHz                            | $U_{rel}=0.01\%$                  |      |                |
|    |                                  | Level                |                                                                | 10mV~20V                               | $U_{rel}=0.2\%$                   |      |                |
|    |                                  | Distortion           |                                                                | (0.01~30)%                             | $U_{rel}=15\%$                    |      |                |
| 10 | Sound Level Meters               | Sound pressure level | V.R. of Sound Level Meters JJG 188                             | pressure field: (20~130)dB, (10~20)Hz  | $U=0.5\text{dB}$                  |      |                |
|    |                                  |                      |                                                                | pressure field: (20~130)dB, (20~400)Hz | $U=0.4\text{dB}$                  |      |                |
|    |                                  |                      |                                                                | free field: (20~130)dB, (500~6300)Hz   | $U=0.4\text{dB}$                  |      |                |
|    |                                  |                      |                                                                | free field : (20~130)dB, (6.3~10)kHz   | $U=0.7\text{dB}$                  |      |                |



No. CNAS L0730

第 38 页 共 163 页

The scope of the accreditation in Chinese remains the definitive version.


| №  | Instrument                       | Measurand            | Calibration Method                                | Range                                     | Expanded Uncertainty<br>( $k=2$ ) | Note | Effective Date |
|----|----------------------------------|----------------------|---------------------------------------------------|-------------------------------------------|-----------------------------------|------|----------------|
|    |                                  | Time weight          |                                                   | free field: (20~130)dB, (10~20)kHz        | $U=1.0\text{dB}$                  |      |                |
|    |                                  |                      |                                                   | Electrical signal: (20~130)dB, 20Hz~20kHz | $U=0.1\text{dB}$                  |      |                |
|    |                                  |                      |                                                   | F: (0~90) dB/s                            | $U=1.5\text{dB/s}$                |      |                |
|    |                                  |                      |                                                   | S: (0~90) dB/s                            | $U=0.2\text{dB/s}$                |      |                |
|    |                                  |                      |                                                   | (-70~20) dB                               | $U=0.30\text{dB}$                 |      |                |
| 11 | *Electret Microphone Instruments | Frequency            | C.S. for Electret Microphone Instruments JJF 1145 | 10Hz~10kHz                                | $U_{\text{rel}}=0.1\%$            |      |                |
|    |                                  | Distorting           |                                                   | (0.1~30)%                                 | $U=15\%$                          |      |                |
|    |                                  | Voltage              |                                                   | (0.1~10)V                                 | $U_{\text{rel}}=0.1\%$            |      |                |
|    |                                  | Current              |                                                   | 1 $\mu\text{A}$ ~100mA                    | $U_{\text{rel}}=0.5\%$            |      |                |
|    |                                  | Resistance           |                                                   | 1 $\Omega$ ~100k $\Omega$                 | $U_{\text{rel}}=0.2\%$            |      |                |
|    |                                  | Sound pressure level |                                                   | (10~150)dB                                | $U=0.3\text{dB}$                  |      |                |
|    |                                  | Attenuation          |                                                   | (-10~-60)dB                               | $U=0.1\text{dB}$                  |      |                |
| 12 | *Sound Calibrators               | Sound Pressure Level | V.R. of Sound Calibrators JJG 176                 | (74~120)dB, (31.5~160)Hz                  | $U=0.2\text{dB}$                  |      |                |
|    |                                  |                      |                                                   | (74~120)dB, (160~1250)Hz                  | $U=0.1\text{dB}$                  |      |                |
|    |                                  |                      |                                                   | (74~120)dB, (1250~4000)Hz                 | $U=0.25\text{dB}$                 |      |                |
|    |                                  |                      |                                                   | (74~120)dB, (4~8)kHz                      | $U=0.35\text{dB}$                 |      |                |



No. CNAS L0730

第 39 页 共 163 页

The scope of the accreditation in Chinese remains the definitive version.

| №  | Instrument                                                                      | Measurand            | Calibration Method                                                                               | Range                                  | Expanded Uncertainty<br>( <i>k</i> =2) | Note                                                                                 | Effective Date  |
|----|---------------------------------------------------------------------------------|----------------------|--------------------------------------------------------------------------------------------------|----------------------------------------|----------------------------------------|--------------------------------------------------------------------------------------|-----------------|
|    |                                                                                 |                      |                                                                                                  | (74~120)dB, (8~16)kHz                  | <i>U</i> =0.5dB                        |                                                                                      |                 |
|    |                                                                                 | Frequency            |                                                                                                  | 31.5 Hz~16 kHz                         | <i>U</i> <sub>rel</sub> =0.1%          |                                                                                      |                 |
|    |                                                                                 | Distortion           |                                                                                                  | (0.01~20)%                             | <i>U</i> <sub>rel</sub> =2.0%          |                                                                                      |                 |
|    |                                                                                 | Sound pressure level |                                                                                                  | (-60~30)dB, (10Hz~20kHz)               | <i>U</i> =0.3dB                        |                                                                                      |                 |
| 13 | *Electro-acoustic Products(Loudspeakers) Power Life-span Measurement Equipments | Time                 | C.S. for Electro-acoustic Products(Loudspeakers) Power Life-span Measurement Equipments JJF 1203 | (1~120)s                               | <i>U</i> <sub>rel</sub> =0.3%          |                                                                                      |                 |
|    |                                                                                 | Distortion           |                                                                                                  | (0.01~30)%, (20Hz~20kHz)               | <i>U</i> <sub>rel</sub> =10%           |                                                                                      |                 |
|    |                                                                                 | Power                |                                                                                                  | (0.1~200)W                             | <i>U</i> <sub>rel</sub> =4%            |                                                                                      |                 |
|    |                                                                                 | Peak Factor          |                                                                                                  | 1~5                                    | <i>U</i> <sub>rel</sub> =4%            |                                                                                      |                 |
| 14 | Personal Sound Exposure Meters                                                  | Sound exposure       | V.R.of Personal Sound Exposure Meters JJG 980                                                    | (0.01~99.99)Pa²h                       | <i>U</i> =0.90Pa²h                     |                                                                                      |                 |
|    |                                                                                 | Sound pressure level |                                                                                                  | Pressure field: (20~130)dB, (63~400)Hz | <i>U</i> =0.4dB                        |                                                                                      |                 |
|    |                                                                                 |                      |                                                                                                  | Free field: (20~130)dB, 500Hz~4kHz     | <i>U</i> =0.4dB                        |                                                                                      |                 |
|    |                                                                                 |                      |                                                                                                  |                                        | Free field: (20~130)dB, 4kHz~8kHz      |                                                                                      | <i>U</i> =0.7dB |
| 15 | Noise Dosimeters                                                                | Noise dose           | V.R.of Noise Dosimeters JJG 655                                                                  | 0.1%~10000%                            | <i>U</i> =7%                           |  |                 |
|    |                                                                                 | Sound pressure level |                                                                                                  | Pressure field: (20~130)dB, (20~400)Hz | <i>U</i> =0.4dB                        |                                                                                      |                 |
|    |                                                                                 |                      |                                                                                                  | Free field: (20~130)dB, 500Hz~4kHz     | <i>U</i> =0.4dB                        |                                                                                      |                 |
|    |                                                                                 |                      |                                                                                                  |                                        | Free field: (20~130)dB, 4kHz~8kHz      |                                                                                      | <i>U</i> =0.7dB |



No. CNAS L0730

第 40 页 共 163 页

The scope of the accreditation in Chinese remains the definitive version.



在线扫码获取验证

| №  | Instrument                                              | Measurand                            | Calibration Method                                                        | Range                       | Expanded Uncertainty<br>( $k=2$ ) | Note                       | Effective Date |
|----|---------------------------------------------------------|--------------------------------------|---------------------------------------------------------------------------|-----------------------------|-----------------------------------|----------------------------|----------------|
| 16 | *Tone Burst Generators                                  | Time                                 | V.R. of Tone Burst Generators<br>JJG 199                                  | 0.1ms~10ms                  | $U_{rel}=0.31\%$                  |                            |                |
|    |                                                         |                                      |                                                                           | 10ms~200ms                  | $U_{rel}=0.24\%$                  |                            |                |
|    |                                                         |                                      |                                                                           | 0.2s~10s                    | $U_{rel}=0.2\%$                   |                            |                |
|    |                                                         | Period                               |                                                                           | 0.1ms~10s                   | $U_{rel}=0.3\%$                   |                            |                |
|    |                                                         | Distortion                           |                                                                           | 0.1%~30%                    | $U_{rel}=2.5\%$                   |                            |                |
|    |                                                         | Frequency                            |                                                                           | (20~20000)Hz                | $U_{rel}=0.01\%$                  |                            |                |
|    |                                                         | Voltage                              |                                                                           | 100mV~100V                  | $U_{rel}=0.14\%$                  |                            |                |
| 17 | *Audiometry room                                        | Sound pressure level                 | C.S. for Acoustic Performance of Audiometry Rooms JJF 1191                | (-20~150)dB                 | $U=0.8\text{dB}$                  | Except for diffusion field |                |
| 18 | Portable Vibration Calibration                          | Acceleration                         | V.R. of Portable Vibration Calibration JJG 1062                           | (1~100) m/s <sup>2</sup>    | $U_{rel}=1.2\%$                   |                            |                |
|    |                                                         | Frequency                            |                                                                           | 20 Hz~2 kHz                 | $U_{rel}=0.01\%$                  |                            |                |
|    |                                                         | Distortion                           |                                                                           | (0.01~30)%                  | $U=0.6\%$                         |                            |                |
| 19 | *Standard Vibrators                                     | Acceleration                         | V.R. of Standard Standard Vibrators JJG 298                               | (1~100)m/s <sup>2</sup>     | $U_{rel}=2.0\%$                   |                            |                |
|    |                                                         | Frequency                            |                                                                           | 20Hz~2 kHz                  | $U_{rel}=0.01\%$                  |                            |                |
|    |                                                         | Distortion                           |                                                                           | (0.01~100)%                 | $U_{rel}=2.0\%$                   |                            |                |
| 20 | Head and Torso Simulator Used in Acoustical Measurement | Ear sensitivity level                | C.S. for Head and Torso Simulator Used in Acoustical Measurement JJF 1520 | (-40~40)dB, (1 kHz)         | $U=0.8\text{dB}$                  |                            |                |
|    |                                                         | Frequency response of ear free field |                                                                           | (-40~40)dB, (100 Hz~10 kHz) | $U=(0.5~0.8)\text{dB}$            |                            |                |



No. CNAS L0730

第 41 页 共 163 页

The scope of the accreditation in Chinese remains the definitive version.

| №  | Instrument                                 | Measurand                                                         | Calibration Method                                           | Range                        | Expanded Uncertainty<br>( <i>k</i> =2) | Note | Effective Date |
|----|--------------------------------------------|-------------------------------------------------------------------|--------------------------------------------------------------|------------------------------|----------------------------------------|------|----------------|
|    |                                            | Ear directional frequency response                                |                                                              | (-40~40)dB, (100 Hz~10 kHz)  | <i>U</i> =(0.6~1.0)dB                  |      |                |
|    |                                            | Amplitude-frequency characteristics at the nozzle reference point |                                                              | (-40~40)dB, (100 Hz~10 kHz)  | <i>U</i> =(0.50~0.70)dB                |      |                |
|    |                                            | Distortion at mouth reference point                               |                                                              | (0.01~30)%, (100 Hz~10 kHz)  | <i>U</i> <sub>rel</sub> =1.5 %         |      |                |
|    |                                            | Amplitude-frequency characteristics of the nozzle free field      |                                                              | (-40~40)dB, (100 Hz~10 kHz)  | <i>U</i> =(0.4~0.8)dB                  |      |                |
| 21 | *Audio-frequency Power Amplifiers          | Frequency Response                                                | C.S. for Audio-frequency Power Amplifiers JJF 1200           | (-20dB~20dB), (20Hz~100kHz)  | <i>U</i> =0.08dB                       |      |                |
|    |                                            | Distortion                                                        |                                                              | 0.01%~100%                   | <i>U</i> =0.16%                        |      |                |
|    |                                            | Gain                                                              |                                                              | (-60dB~100dB), (20Hz~100kHz) | <i>U</i> =0.05dB                       |      |                |
| 22 | *Electro-acoustical Measurement Instrument | sound pressure level                                              | C.S. for Electro-acoustical Measurement Instruments JJF 1339 | (10~150)dB                   | <i>U</i> =0.30 dB                      |      |                |
|    |                                            | frequency                                                         |                                                              | 20 Hz~20 kHz                 | <i>U</i> <sub>rel</sub> =0.1%          |      |                |
|    |                                            | voltage                                                           |                                                              | AC: (0.01~100)V              | <i>U</i> <sub>rel</sub> =0.1%          |      |                |
|    |                                            |                                                                   |                                                              | DC: (0.01~100)V              | <i>U</i> <sub>rel</sub> =0.01%         |      |                |



No. CNAS L0730

第 42 页 共 163 页

The scope of the accreditation in Chinese remains the definitive version.

| №  | Instrument                                                            | Measurand                      | Calibration Method                              | Range                         | Expanded Uncertainty<br>( <i>k</i> =2) | Note | Effective Date |
|----|-----------------------------------------------------------------------|--------------------------------|-------------------------------------------------|-------------------------------|----------------------------------------|------|----------------|
|    |                                                                       | total distortion               |                                                 | (0.01~30)%                    | <i>U</i> =0.5%                         |      |                |
| 23 | *Temperature/Humidity/Vibration Combined Environmental Testing System | Temperature                    | C.S.for Temperature/Humidity/Vibration JJF 1270 | (-75~+200)°C                  | <i>U</i> =0.3°C                        |      |                |
|    |                                                                       | Temperature variation rate     |                                                 | (0.5~60)°C/min                | <i>U</i> =0.4°C/min                    |      |                |
|    |                                                                       | Noise                          |                                                 | (30~130)dB                    | <i>U</i> =0.7dB                        |      |                |
|    |                                                                       | Wind speed                     |                                                 | (0.2~20)m/s                   | <i>U</i> <sub>rel</sub> =4%            |      |                |
|    |                                                                       | Acceleration                   |                                                 | (1~100)m/s <sup>2</sup>       | <i>U</i> <sub>rel</sub> =2%            |      |                |
|    |                                                                       | Humidity                       |                                                 | (10~98)%RH                    | <i>U</i> =1.5%RH                       |      |                |
| 24 | *Microphone Preamplifiers                                             | frequency response             | C.S.for Microphone Preamplifiers JJF 1137       | (-20~ +20) dB, (10 Hz~50 kHz) | <i>U</i> =0.24 dB                      |      |                |
|    |                                                                       | transmission loss              |                                                 | (-100~0) dB, (10 Hz~50 kHz)   | <i>U</i> =0.22 dB                      |      |                |
|    |                                                                       | noise                          |                                                 | 0.1 μ V~10V                   | <i>U</i> <sub>rel</sub> =7%            |      |                |
| 25 | *Measuring Amplifiers                                                 | sensitivity                    | C.S.for Measuring Amplifiers JJF 1157           | 94dB, 1kHz                    | <i>U</i> =0.05 dB                      |      |                |
|    |                                                                       | Frequency weight               |                                                 | (-80 ~ 25) dB, (10 Hz~20 kHz) | <i>U</i> =0.2 dB                       |      |                |
|    |                                                                       | Attenuation decrement          |                                                 | (-30~50) dB                   | <i>U</i> =0.05 dB                      |      |                |
|    |                                                                       | Time weight                    |                                                 | F: (0.01~ 40)dB/s             | <i>U</i> =2 dB/s                       |      |                |
|    |                                                                       |                                |                                                 | S: (0.01~ 10)dB/s             | <i>U</i> =0.4dB/s                      |      |                |
|    |                                                                       | Linearity of electrical output |                                                 | (-50~+10) dB                  | <i>U</i> =0.1dB                        |      |                |



No. CNAS L0730

第 43 页 共 163 页

The scope of the accreditation in Chinese remains the definitive version.


| №  | Instrument                              | Measurand                          | Calibration Method                                                                         | Range                              | Expanded Uncertainty<br>( $k=2$ ) | Note | Effective Date |
|----|-----------------------------------------|------------------------------------|--------------------------------------------------------------------------------------------|------------------------------------|-----------------------------------|------|----------------|
|    |                                         | Harmonic distortion                |                                                                                            | (0.01~100)%                        | $U_{rel}=6.6\%$                   |      |                |
|    |                                         | Frequency                          |                                                                                            | 1Hz~200 kHz                        | $U_{rel}=0.1\%$                   |      |                |
|    |                                         | Voltage                            |                                                                                            | 10mV~210 V                         | $U_{rel}=0.5\%$                   |      |                |
| 26 | *Anechoic Rooms and Hemi-anechoic Rooms | background noise                   | C.S.for Acoustic Performance of Anechoic Rooms and Hemi-anechoic Rooms JJF1147             | (-2.5~130)dB                       | $U=0.5\text{dB}$                  |      |                |
|    |                                         | SPL                                |                                                                                            | -10dB~150dB                        | $U=0.5\text{ dB}$                 |      |                |
|    |                                         | Frequency                          |                                                                                            | 20 Hz~20 kHz                       | $U_{rel}=0.1\%$                   |      |                |
| 27 | *Audiometers                            | frequency                          | V.R. of Calibration Specification for Audiological Equipment Pure-tone Audiometers JJG 388 | 125Hz~8 kHz                        | $U_{rel}=1.0\%$                   |      |                |
|    |                                         | THD (total harmonic distortion)    |                                                                                            | (0.01~100)%                        | $U_{rel}=2.0\%$                   |      |                |
|    |                                         | Air conduction hearing level zero  |                                                                                            | (-10~120)dB                        | $U=1.2\text{ dB}$                 |      |                |
|    |                                         | Bone conduction hearing level zero |                                                                                            | (-10~70)dB                         | $U=1.5\text{ dB}$                 |      |                |
|    |                                         | ound pressure level                |                                                                                            | (-10~110)dB                        | $U=1.2\text{ dB}$                 |      |                |
| 28 | Elevator Vibration and Noise Analyzer   | Acceleration                       | C.S.for Elevator Vibration and Noise Analyzer JJF (闽) 1058                                 | (2~80)m/s <sup>2</sup> , (1~80) Hz | $U_{rel}=1.5\%$                   |      |                |
|    |                                         | SPL                                |                                                                                            | (20~135) dB, 10Hz~20kHz            | $U=0.3\text{dB}$                  |      |                |
| 29 | *Active Couplers                        | Total distortion                   | C.S.for Active Couplers JJF 1734                                                           | (0.01~30)%                         | $U_{rel}=15\%$                    |      |                |



No. CNAS L0730

第 44 页 共 163 页

The scope of the accreditation in Chinese remains the definitive version.

| №  | Instrument                        | Measurand            | Calibration Method                                | Range                                     | Expanded Uncertainty<br>(k=2) | Note                                                                                 | Effective Date |  |
|----|-----------------------------------|----------------------|---------------------------------------------------|-------------------------------------------|-------------------------------|--------------------------------------------------------------------------------------|----------------|--|
|    |                                   | Phase                |                                                   | (0~360)°, 10Hz~20kHz                      | U=0.3°                        |                                                                                      |                |  |
|    |                                   | Sound pressure level |                                                   | (30~140)dB, 10Hz~20kHz                    | U=0.22dB                      |                                                                                      |                |  |
| 30 | *Constant<br>Bnadwidth Filters    | Relative damping     | C.S.for Constant Bandwith<br>Filters JJF1490      | (-140~+140)dB                             | U=0.1dB                       |                                                                                      |                |  |
|    |                                   | Bandwidth            |                                                   | 10Hz~100kHz                               | U <sub>rel</sub> =0.1%        |                                                                                      |                |  |
|    |                                   | Frequency            |                                                   | 10Hz~25kHz                                | U <sub>rel</sub> =0.1%        |                                                                                      |                |  |
|    |                                   | Linear of order      |                                                   | (-140~+140)dB                             | U=0.1dB                       |                                                                                      |                |  |
|    |                                   | Form factor          |                                                   | 1~50                                      | U <sub>rel</sub> =0.9%        |                                                                                      |                |  |
|    |                                   |                      |                                                   |                                           |                               |                                                                                      |                |  |
| 31 | Human Vibration Meters            | Acceleration         | V.R. of Human Vibration Meters JJG 1178           | (1~100)m/s <sup>2</sup> , (1~1000)Hz      | U <sub>rel</sub> =1.0%        |                                                                                      |                |  |
| 32 | Environment Vibration Instrument  | Acceleration level   | V.R. of Environment Vibration Instruments JJG 921 | (60~150)dB, (1~80)Hz                      | U=0.2dB                       |                                                                                      |                |  |
| 33 | Noise Level Statistical Analyzers | ound pressure level  | V.R. of Noise Level Statistical Analyzers JJG 778 | Pressure field: (20~130)dB, (10~250)Hz    | U=0.5dB                       |  |                |  |
|    |                                   |                      |                                                   | pressure field: (20~130)dB, (250~400)Hz   | U=0.4dB                       |                                                                                      |                |  |
|    |                                   |                      |                                                   | free field: (20~130)dB, (400~1250)Hz      | U=0.4dB                       |                                                                                      |                |  |
|    |                                   |                      |                                                   | free field : (20~130)dB, (1.25~10)kHz     | U=0.7dB                       |                                                                                      |                |  |
|    |                                   |                      |                                                   | free field: (20~130)dB, (10~20)kHz        | U=1.0dB                       |                                                                                      |                |  |
|    |                                   |                      |                                                   | Electrical signal: (20~130)dB, 20Hz~20kHz | U=0.1dB                       |                                                                                      |                |  |



No. CNAS L0730

第 45 页 共 163 页

The scope of the accreditation in Chinese remains the definitive version.

| No                                     | Instrument                   | Measurand                                          | Calibration Method                           | Range                    | Expanded Uncertainty<br>(k=2) | Note | Effective Date |
|----------------------------------------|------------------------------|----------------------------------------------------|----------------------------------------------|--------------------------|-------------------------------|------|----------------|
|                                        |                              | Linear of order                                    |                                              | (-20~20)dB               | U=0.3dB                       |      |                |
|                                        |                              | Time weightings                                    |                                              | F: (0~90)dB              | U=3.5dB/s                     |      |                |
|                                        |                              | Burst response                                     |                                              | S: (0~90)dB              | U=0.4dB/s                     |      |                |
|                                        |                              |                                                    |                                              | (-70~20)dB               | U=0.3dB                       |      |                |
| 34                                     | Ultrasonic Flaw Detectors    | decrement                                          | V.R. for Ultrasonic Flaw Detectors JJG 746   | (0~110)dB                | U=0.02dB                      |      |                |
|                                        |                              | horizontal linearity                               |                                              | 1%~100%                  | U=0.36%                       |      |                |
|                                        |                              | Vertical linearity error of ultrasonic attenuation |                                              | 1%~100%                  | U=2.0%                        |      |                |
| 6 Electromagnetic Measuring Instrument |                              |                                                    |                                              |                          |                               |      |                |
| 1                                      | *Withstanding Voltage Tester | DC Voltage                                         | V.R. of Withstanding Voltage Testers JJG 795 | (0.1~15) kV              | U <sub>rel</sub> =(1.7~0.12)% |      |                |
|                                        |                              | AC Voltage                                         |                                              | (0.1~15) kV, (45~55) Hz  | U <sub>rel</sub> =(1.7~0.20)% |      |                |
|                                        |                              | DC Current                                         |                                              | 0.1mA~10A                | U <sub>rel</sub> =(6~0.2)%    |      |                |
|                                        |                              | AC Current                                         |                                              | 0.1mA~10A, (45~55) Hz    | U <sub>rel</sub> =(6~0.3)%    |      |                |
|                                        |                              | Time                                               |                                              | (1~999)s                 | U <sub>rel</sub> =(2.7~0.24)% |      |                |
|                                        |                              | Distortion                                         |                                              | (0.01~100)%(20Hz~100kHz) | U <sub>rel</sub> =15%         |      |                |
|                                        |                              | ripple factor                                      |                                              | (0.1~10)%                | U=0.2%                        |      |                |
|                                        |                              | capacity                                           |                                              | (1~500)W                 | U <sub>rel</sub> =1.3%        |      |                |



No. CNAS L0730

第 46 页 共 163 页

The scope of the accreditation in Chinese remains the definitive version.

| № | Instrument                                             | Measurand      | Calibration Method                                                 | Range                               | Expanded Uncertainty<br>( <i>k</i> =2) | Note | Effective Date |
|---|--------------------------------------------------------|----------------|--------------------------------------------------------------------|-------------------------------------|----------------------------------------|------|----------------|
| 2 | *Megohm - meter                                        | Resistance     | V.R. of Insulation Resistance Meter<br>(Megohmmeter) JJG 622       | 100 Ω ~ 10M Ω                       | <i>U</i> <sub>rel</sub> =(0.4~0.3)%    |      |                |
|   |                                                        | 10M Ω ~ 200G Ω |                                                                    | <i>U</i> <sub>rel</sub> =(0.6~11)%  |                                        |      |                |
|   |                                                        | Voltage        |                                                                    | 10V~10kV                            | <i>U</i> <sub>rel</sub> =(2.8~1.1)%    |      |                |
| 3 | *Earth Resistance Meter                                | Resistance     | V.R. of Earth Resistance Meter JJG 366                             | 1m Ω ~ 10k Ω                        | <i>U</i> <sub>rel</sub> =(0.9~0.14)%   |      |                |
| 4 | *Amperemeter<br>Voltmeter<br>Wattmeter and<br>Ohmmeter | DC Voltage     | V.R. of Amperemeter<br>Voltmeter Wattmeter and<br>Ohmmeter JJG 124 | (100~330)mV                         | <i>U</i> <sub>rel</sub> = (0.07~0.05)% |      |                |
|   |                                                        |                |                                                                    | 330mV~1000V                         | <i>U</i> <sub>rel</sub> = (0.04~0.07)% |      |                |
|   |                                                        | AC Voltage     |                                                                    | 100mV~3.3V, (45~55) Hz              | <i>U</i> <sub>rel</sub> =(0.16~0.08)%  |      |                |
|   |                                                        |                |                                                                    | (3.3~330)V, (45~55) Hz              | <i>U</i> <sub>rel</sub> =(0.13~0.1)%   |      |                |
|   |                                                        |                |                                                                    | (330~1000)V, (45~55) Hz             | <i>U</i> <sub>rel</sub> =(0.16~0.14)%  |      |                |
|   |                                                        | DC Current     |                                                                    | 100 μ A~2.2A                        | <i>U</i> <sub>rel</sub> = (0.4~0.07)%  |      |                |
|   |                                                        |                |                                                                    | (2.2~10)A                           | <i>U</i> <sub>rel</sub> = (0.15~0.10)% |      |                |
|   |                                                        |                |                                                                    | (10~30)A                            | <i>U</i> <sub>rel</sub> = (0.17~0.10)% |      |                |
|   |                                                        | AC Current     |                                                                    | 30 μ A~10A, (45~55) Hz              | <i>U</i> <sub>rel</sub> =(0.8~0.09)%   |      |                |
|   |                                                        |                |                                                                    | (10~30)A, (45~55) Hz                | <i>U</i> <sub>rel</sub> =(0.16~0.10)%  |      |                |
|   |                                                        | DC Power       |                                                                    | (0.75~1000)V/(0.01~30) A            | <i>U</i> <sub>rel</sub> =0.15%         |      |                |
|   |                                                        | AC Power       |                                                                    | (0.75~750)V/(0.01~30) A, (45~55) Hz | <i>U</i> <sub>rel</sub> =0.15%         |      |                |



No. CNAS L0730

第 47 页 共 163 页

The scope of the accreditation in Chinese remains the definitive version.

| № | Instrument  | Measurand                | Calibration Method                                  | Range                 | Expanded Uncertainty<br>( <i>k</i> =2)     | Note                                  | Effective Date |  |
|---|-------------|--------------------------|-----------------------------------------------------|-----------------------|--------------------------------------------|---------------------------------------|----------------|--|
|   |             | DC Resistor              |                                                     | 10 Ω ~ 3.3M Ω         | <i>U</i> <sub>rel</sub> =0.15%             |                                       |                |  |
|   |             |                          |                                                     | (3.3~10)M Ω           | <i>U</i> <sub>rel</sub> = (0.4~0.27)%      |                                       |                |  |
| 5 | *Calibrator | DC Voltage               | C.S. for Multifunction<br>Standard Sources JJF 1638 | (10~20) mV            | <i>U</i> <sub>rel</sub> =(0.004~0.003)%    |                                       |                |  |
|   |             |                          |                                                     | (20~200)mV            | <i>U</i> <sub>rel</sub> = (0.0020~0.0010)% |                                       |                |  |
|   |             |                          |                                                     | (0.2~20)V             | <i>U</i> <sub>rel</sub> = (0.0010~0.0006)% |                                       |                |  |
|   |             |                          |                                                     | (20~200)V             | <i>U</i> <sub>rel</sub> = (0.0016~0.0010)% |                                       |                |  |
|   |             |                          |                                                     | (200~1000)V           | <i>U</i> <sub>rel</sub> = (0.0016~0.0010)% |                                       |                |  |
|   |             |                          |                                                     | (10~20)mV (10Hz~40Hz) | <i>U</i> <sub>rel</sub> =(0.14~0.08)%      |                                       |                |  |
|   |             |                          |                                                     | AC Voltage            | (10~20)mV (40Hz~10kHz)                     | <i>U</i> <sub>rel</sub> =(0.08~0.04)% |                |  |
|   |             | (10~20)mV (10kHz~100kHz) |                                                     |                       | <i>U</i> <sub>rel</sub> =(0.4~0.15)%       |                                       |                |  |
|   |             | (20~100)mV(10Hz~1kHz)    |                                                     |                       | <i>U</i> <sub>rel</sub> = (0.04~0.013)%    |                                       |                |  |
|   |             | (20~100)mV(1~20)kHz      |                                                     |                       | <i>U</i> <sub>rel</sub> = (0.06~0.03)%     |                                       |                |  |
|   |             | (20~100)mV(20~50)kHz     |                                                     |                       | <i>U</i> <sub>rel</sub> =(0.08~0.05)%      |                                       |                |  |
|   |             | (20~100)mV(50~100)kHz    |                                                     |                       | <i>U</i> <sub>rel</sub> =(0.14~0.09)%      |                                       |                |  |
|   |             | (20~100)mV(100~300)kHz   |                                                     |                       | <i>U</i> <sub>rel</sub> =(0.6~0.5)%        |                                       |                |  |
|   |             | (20~100)mV(300kHz~1MHz)  |                                                     |                       | <i>U</i> <sub>rel</sub> =(1.8~1.4)%        |                                       |                |  |



No. CNAS L0730

第 48 页 共 163 页

The scope of the accreditation in Chinese remains the definitive version.

| № | Instrument | Measurand | Calibration Method | Range                   | Expanded Uncertainty<br>( $k=2$ ) | Note | Effective Date |
|---|------------|-----------|--------------------|-------------------------|-----------------------------------|------|----------------|
|   |            |           |                    | 100mV~1V(10Hz~1kHz)     | $U_{rel}=(0.06\sim0.013)\%$       |      |                |
|   |            |           |                    | 100mV~1V(1kHz~20)kHz    | $U_{rel}=(0.06\sim0.03)\%$        |      |                |
|   |            |           |                    | 100mV~1V(20~50)kHz      | $U_{rel}=(0.07\sim0.04)\%$        |      |                |
|   |            |           |                    | 100mV~1V(50~100)kHz     | $U_{rel}=(0.15\sim0.09)\%$        |      |                |
|   |            |           |                    | 100mV~1V(100~300)kHz    | $U_{rel}=(0.7\sim0.5)\%$          |      |                |
|   |            |           |                    | 100mV~1V(300kHz~1MHz)   | $U_{rel}=(2\sim1.4)\%$            |      |                |
|   |            |           |                    | (1~10)V(10Hz~1kHz)      | $U_{rel}=(0.06\sim0.013)\%$       |      |                |
|   |            |           |                    | (1~10)V(1~20)kHz        | $U_{rel}=(0.06\sim0.03)\%$        |      |                |
|   |            |           |                    | (1~10)V(20~50)kHz       | $U_{rel}=(0.07\sim0.04)\%$        |      |                |
|   |            |           |                    | (1~10)V(50~100)kHz      | $U_{rel}=(0.15\sim0.09)\%$        |      |                |
|   |            |           |                    | (1~10)V(100~300)kHz     | $U_{rel}=(0.7\sim0.5)\%$          |      |                |
|   |            |           |                    | (1~10)V(300kHz~1MHz)    | $U_{rel}=(2\sim1.4)\%$            |      |                |
|   |            |           |                    | (10~100)V(10Hz~1kHz)    | $U_{rel}=(0.09\sim0.04)\%$        |      |                |
|   |            |           |                    | (10~100)V(1kHz~20kHz)   | $U_{rel}=(0.07\sim0.04)\%$        |      |                |
|   |            |           |                    | (10~100)V(20kHz~50kHz)  | $U_{rel}=(0.09\sim0.05)\%$        |      |                |
|   |            |           |                    | (10~100)V(50kHz~100kHz) | $U_{rel}=(0.3\sim0.18)\%$         |      |                |



No. CNAS L0730

第 49 页 共 163 页

The scope of the accreditation in Chinese remains the definitive version.

| № | Instrument | Measurand  | Calibration Method                                                                                              | Range                   | Expanded Uncertainty<br>( <i>k</i> =2)    | Note | Effective Date |
|---|------------|------------|-----------------------------------------------------------------------------------------------------------------|-------------------------|-------------------------------------------|------|----------------|
|   |            |            | ilac-MEAS<br>CHINA NATIONAL ACCREDITATION SERVICE FOR CONFORMITY ASSESSMENT<br>SCHEDULE OF APPROVED CERTIFICATE | (100~1000)V(10Hz~1kHz)  | <i>U</i> <sub>rel</sub> =(0.12~0.07)%     |      |                |
|   |            |            |                                                                                                                 | (100~1000)V(1~20)kHz    | <i>U</i> <sub>rel</sub> =(0.13~0.09)%     |      |                |
|   |            |            |                                                                                                                 | (100~1000)V(20~50)kHz   | <i>U</i> <sub>rel</sub> =(0.26~0.16)%     |      |                |
|   |            |            |                                                                                                                 | (100~1000)V(50~100)kHz) | <i>U</i> <sub>rel</sub> =(0.8~0.7)%       |      |                |
|   |            | DC Current |                                                                                                                 | 20 μ A~20mA             | <i>U</i> <sub>rel</sub> =(0.0030~0.0020)% |      |                |
|   |            |            |                                                                                                                 | 20mA~20A                | <i>U</i> <sub>rel</sub> =(0.004~0.025)%   |      |                |
|   |            |            |                                                                                                                 | (20~1000)A              | <i>U</i> <sub>rel</sub> =(0.03~0.015)%    |      |                |
|   |            | AC Current |                                                                                                                 | (20~200) μ A(20Hz~5kHz) | <i>U</i> <sub>rel</sub> =(0.5~0.05)%      |      |                |
|   |            |            |                                                                                                                 | 200 μ A~2mA(20Hz~5kHz)  | <i>U</i> <sub>rel</sub> =(0.6~0.05)%      |      |                |
|   |            |            |                                                                                                                 | (2~20)mA(20Hz~10kHz)    | <i>U</i> <sub>rel</sub> =(0.6~0.05)%      |      |                |
|   |            |            |                                                                                                                 | 20 μ A~20mA(10~100)kHz  | <i>U</i> <sub>rel</sub> =(0.5~0.12)%      |      |                |
|   |            |            |                                                                                                                 | 20mA~20A(20Hz~10kHz)    | <i>U</i> <sub>rel</sub> =(0.020~0.012)%   |      |                |
|   |            |            |                                                                                                                 | 20mA~20A(10~100)kHz     | <i>U</i> <sub>rel</sub> =(0.05~0.03)%     |      |                |
|   |            |            |                                                                                                                 | (20~1000)A(45~65)Hz     | <i>U</i> <sub>rel</sub> =(0.4~0.07)%      |      |                |
|   |            | Resistance |                                                                                                                 | (0.2~20) Ω              | <i>U</i> <sub>rel</sub> =(0.004~0.0020)%  |      |                |
|   |            |            |                                                                                                                 | 20 Ω ~2M Ω              | <i>U</i> <sub>rel</sub> =(0.0020~0.003)%  |      |                |



No. CNAS L0730

第 50 页 共 163 页

The scope of the accreditation in Chinese remains the definitive version.

| № | Instrument                | Measurand  | Calibration Method                                                           | Range                          | Expanded Uncertainty<br>(k=2)    | Note          | Effective Date |
|---|---------------------------|------------|------------------------------------------------------------------------------|--------------------------------|----------------------------------|---------------|----------------|
|   |                           |            |                                                                              | (2~20)MΩ                       | U <sub>rel</sub> =(0.003~0.008)% |               |                |
|   |                           |            |                                                                              | (20~200)MΩ                     | U <sub>rel</sub> =(0.008~0.02)%  |               |                |
|   |                           | Frequency  |                                                                              | 10Hz~1MHz                      | U <sub>rel</sub> =0.0020%        |               |                |
|   |                           | Distortion |                                                                              | (0.005~30)%                    | U <sub>rel</sub> =10%            |               |                |
| 6 | *AC Power Source          | AC Voltage | C.S. for three phase variable-frequency power supply<br>JJF(Electronic) 0026 | (1~1000)V(10Hz~66Hz)           | U <sub>rel</sub> =(0.42~0.05)%   | Power≤<br>4kW |                |
|   |                           |            |                                                                              | (1~1000)V(66Hz~1kHz)           | U <sub>rel</sub> =(0.05~0.10)%   |               |                |
|   |                           | AC Current |                                                                              | (0.05~30)A(10Hz~66Hz)          | U <sub>rel</sub> =(0.42~0.05)%   |               |                |
|   |                           |            |                                                                              | (0.05~30)A(66Hz~1kHz)          | U <sub>rel</sub> =(0.05~0.10)%   |               |                |
|   |                           | AC Power   |                                                                              | (1~1000)V(0.05~30)A(10Hz~66Hz) | U <sub>rel</sub> =(0.8~0.08)%    |               |                |
|   |                           |            |                                                                              | (1~1000)V(0.05~30)A(66Hz~1kHz) | U <sub>rel</sub> =(0.08~0.15)%   |               |                |
|   |                           | Frequency  |                                                                              | 10Hz~1kHz                      | U <sub>rel</sub> =0.0020%        |               |                |
|   |                           | Distortion |                                                                              | (0.01~30)%                     | U <sub>rel</sub> =10%            |               |                |
| 7 | *DC Resistors Box         | Resistance | V.R. of D.C.Resistance Box<br>JJG 982                                        | 1mΩ~100kΩ                      | U <sub>rel</sub> =(0.14~0.002)%  |               |                |
|   |                           |            |                                                                              | 100kΩ~100MΩ                    | U <sub>rel</sub> =(0.002~0.010)% |               |                |
| 8 | *Ground Continuity Tester | Resistance | V.R. of Earth -Continuity Tester JJG 984                                     | AC(1~1000) mΩ ,<br>(45~55) Hz  | U <sub>rel</sub> =(11~0.20)%     |               |                |
|   |                           |            |                                                                              | DC(1~1000) mΩ                  | U <sub>rel</sub> =(11~0.20)%     |               |                |



No. CNAS L0730

第 51 页 共 163 页

The scope of the accreditation in Chinese remains the definitive version.

| №  | Instrument                        | Measurand                   | Calibration Method                                | Range                                 | Expanded Uncertainty<br>( <i>k</i> =2) | Note | Effective Date |
|----|-----------------------------------|-----------------------------|---------------------------------------------------|---------------------------------------|----------------------------------------|------|----------------|
|    |                                   | Current                     |                                                   | AC(1~60) A, (45~55) Hz                | <i>U</i> <sub>rel</sub> =(0.3~0.20)%   |      |                |
|    |                                   | DC(1~60) A                  |                                                   | <i>U</i> <sub>rel</sub> =(0.12~0.07)% |                                        |      |                |
|    |                                   | Voltage                     |                                                   | (1~20)V                               | <i>U</i> <sub>rel</sub> =1.5%          |      |                |
| 9  | *High Insulation Resistance Meter | Resistance                  | V.R. of High Insulation Resistance Meters JJG 690 | 100 Ω ~ 10M Ω                         | <i>U</i> <sub>rel</sub> =0.12%         |      |                |
|    |                                   |                             |                                                   | 10M Ω ~ 100G Ω                        | <i>U</i> <sub>rel</sub> =(0.6~2.3)%    |      |                |
|    |                                   |                             |                                                   | 100G Ω ~ 2T Ω                         | <i>U</i> <sub>rel</sub> =2.3%          |      |                |
|    |                                   |                             |                                                   | (2~12)T Ω                             | <i>U</i> <sub>rel</sub> =6%            |      |                |
|    |                                   | Voltage                     |                                                   | (10~1000)V                            | <i>U</i> <sub>rel</sub> =(0.11~0.02)%  |      |                |
|    |                                   |                             |                                                   |                                       |                                        |      |                |
| 10 | *Leakage Current Meter            | Test Voltage                | V.R. of Leakage Current Tester JJG 843            | (1~750)V<br>( 45~65)Hz                | <i>U</i> <sub>rel</sub> =0.3%          |      |                |
|    |                                   | AC Current                  |                                                   | 10 μ A~1A ,( 45~65)Hz                 | <i>U</i> <sub>rel</sub> =(1.0~0.10)%   |      |                |
|    |                                   | Transmission characteristic |                                                   | 10 μ A~10mA,<br>(20Hz~1MHz)           | <i>U</i> <sub>rel</sub> =(1.2~2.4)%    |      |                |
|    |                                   | Input Impedance             |                                                   | 100 Ω ~3k Ω (20Hz~1MHz)               | <i>U</i> <sub>rel</sub> =1.5%          |      |                |
|    |                                   | DC Current                  |                                                   | 10 μ A~1A                             | <i>U</i> <sub>rel</sub> =(1.0~0.10)%   |      |                |
|    |                                   | Input DC Resistance         |                                                   | 100 Ω ~3k Ω                           | <i>U</i> <sub>rel</sub> =0.5%          |      |                |
|    |                                   |                             |                                                   |                                       |                                        |      |                |
| 11 | *High Voltage Meter               | AC Voltage                  | V.R. of Digital High Voltmeter DL/T 973           | (0.1~100) kV ,( 45~65)Hz              | <i>U</i> <sub>rel</sub> =(1.5~0.2)%    |      |                |
|    |                                   | DC Voltage                  |                                                   | (0.1~100) kV                          | <i>U</i> <sub>rel</sub> =(0.2~0.12)%   |      |                |
| 12 | *Clamp ammeters                   | Current                     | C.S. of Clamp Ammeters JJF 1075                   | DCA:(0.1~200)A                        | <i>U</i> <sub>rel</sub> =(0.20~0.16)%  |      |                |



No. CNAS L0730

第 52 页 共 163 页

The scope of the accreditation in Chinese remains the definitive version.


| №  | Instrument                              | Measurand        | Calibration Method                                       | Range                          | Expanded Uncertainty<br>( $k=2$ ) | Note | Effective Date |
|----|-----------------------------------------|------------------|----------------------------------------------------------|--------------------------------|-----------------------------------|------|----------------|
| 13 | *AC Digital Power Meter                 | AC Voltage       | V.R. of AC Digital Powermeter JJG 780                    | DCA:(200~1000)A                | $U_{rel}=(0.5\sim0.26)\%$         |      |                |
|    |                                         | AC Current       |                                                          | ACA:(0.1~200)A ,( 45~65)Hz     | $U_{rel}=(0.20\sim0.16)\%$        |      |                |
|    |                                         | AC Digital Power |                                                          | ACA:(200~1000)A ,( 45~65)Hz    | $U_{rel}=(0.5\sim0.26)\%$         |      |                |
|    |                                         | Power Factor     |                                                          | -1~0~+1                        | $U=0.0006$                        |      |                |
|    |                                         | Frequency        |                                                          | (45~65)Hz                      | $U_{rel}=0.06\%$                  |      |                |
| 14 | *Electronic insulating resistance meter | Resistance       | V.R. of Electronic Insulating Resistance Meters JJG 1005 | 100 $\Omega$ ~ 10M $\Omega$    | $U_{rel}=0.12\%$                  |      |                |
|    |                                         |                  |                                                          | 10M $\Omega$ ~ 100G $\Omega$   | $U_{rel}=(0.6\sim2.3)\%$          |      |                |
|    |                                         |                  |                                                          | 100G $\Omega$ ~ 1T $\Omega$    | $U_{rel}=2.3\%$                   |      |                |
|    |                                         | Voltage          |                                                          | (10~1000)V                     | $U_{rel}=(0.11\sim0.02)\%$        |      |                |
|    |                                         |                  |                                                          | (1~10)kV                       | $U_{rel}=(2.8\sim1.1)\%$          |      |                |
| 15 | *D.C. Bridges                           | Resistance       | V.R. of D.C. Bridges JJG 125                             | (0.01~0.1) $\Omega$            | $U_{rel}=0.012\%$                 |      |                |
|    |                                         |                  |                                                          | (0.1~1) $\Omega$               | $U_{rel}=0.012\%$                 |      |                |
|    |                                         |                  |                                                          | (1 $\Omega$ ~ 100k $\Omega$ )  | $U_{rel}=0.012\%$                 |      |                |
|    |                                         |                  |                                                          | (100k $\Omega$ ~ 1M $\Omega$ ) | $U_{rel}=0.12\%$                  |      |                |



No. CNAS L0730

第 53 页 共 163 页

The scope of the accreditation in Chinese remains the definitive version.

| №  | Instrument                                       | Measurand         | Calibration Method                                                           | Range                                  | Expanded Uncertainty<br>( <i>k</i> =2) | Note                                                                                 | Effective Date |  |
|----|--------------------------------------------------|-------------------|------------------------------------------------------------------------------|----------------------------------------|----------------------------------------|--------------------------------------------------------------------------------------|----------------|--|
| 16 | *Clamp earth resistance meters                   | Resistance        | V.R. of Clamp Earth Resistance Meters JJG 1054                               | 1m Ω ~ 10k Ω                           | <i>U</i> <sub>rel</sub> =(0.6~0.12)%   |                                                                                      |                |  |
| 17 | *Surface Resistance Tester                       | Resistance        | C.S. for Surface Resistance Tester JJF 1285                                  | (10 <sup>3</sup> ~10 <sup>7</sup> ) Ω  | <i>U</i> <sub>rel</sub> =1.2%          |                                                                                      |                |  |
|    |                                                  |                   |                                                                              | (10 <sup>7</sup> ~10 <sup>9</sup> ) Ω  | <i>U</i> <sub>rel</sub> =(1.2~2.8)%    |                                                                                      |                |  |
|    |                                                  |                   |                                                                              | (10 <sup>9</sup> ~10 <sup>11</sup> ) Ω | <i>U</i> <sub>rel</sub> =(2.8~6)%      |                                                                                      |                |  |
|    |                                                  | Voltage           |                                                                              | (5~1000)V                              | <i>U</i> <sub>rel</sub> =(0.10~0.15)%  |                                                                                      |                |  |
|    |                                                  | Length            |                                                                              | (0.1~200)mm                            | <i>U</i> =0.024mm                      |                                                                                      |                |  |
|    |                                                  |                   |                                                                              | (0.1~2.5)kg                            | <i>U</i> =3g                           |                                                                                      |                |  |
|    |                                                  |                   |                                                                              | (2.5~10)kg                             | <i>U</i> =6g                           |                                                                                      |                |  |
|    |                                                  |                   |                                                                              | (10~15)kg                              | <i>U</i> =9g                           |                                                                                      |                |  |
| 18 | Transformers Turn Ratio Test Sets                | Turn Ratio        | V.R. of Transformers Turn Ratio Test Sets JJG 970                            | 1~10000                                | <i>U</i> <sub>rel</sub> =0.026%        |                                                                                      |                |  |
| 19 | *Loop Resistance Tester and DC Resistance Meters | Resistance        | V.R. of Loop Resistance Tester and DC Resistance Meters JJG 1052             | ( 200 μ Ω ~ 1.9m Ω )                   | <i>U</i> <sub>rel</sub> =0.28%         |                                                                                      |                |  |
|    |                                                  |                   |                                                                              | (2m Ω ~ 11 Ω )                         | <i>U</i> <sub>rel</sub> =(0.12~0.06)%  |                                                                                      |                |  |
|    |                                                  |                   |                                                                              | (11 Ω ~ 100k Ω )                       | <i>U</i> <sub>rel</sub> =(0.06~0.12)%  |                                                                                      |                |  |
|    |                                                  | Current           |                                                                              | (0.01~100)A                            | <i>U</i> <sub>rel</sub> =0.09%         |                                                                                      |                |  |
| 20 | *Industry Frequency Phase Meter                  | phase angle       | V.R. of Industry Frequency Single-Phase Phase Meter JJG 440                  | (0~360)°                               | <i>U</i> =0.08°                        |  |                |  |
| 21 | *Electrolytic capacitor leakage current tester   | D.C. Pole Voltage | V.R. of Electrolytic Capacitor Leakage Current Tester JJG(Electronic) 306003 | (0.1~1000)V                            | <i>U</i> <sub>rel</sub> =0.16%~0.8%    |                                                                                      |                |  |
|    |                                                  | Current           |                                                                              | 0.2 μ A~100 mA                         | <i>U</i> <sub>rel</sub> =1.0%~0.16%    |                                                                                      |                |  |



No. CNAS L0730

第 54 页 共 163 页

The scope of the accreditation in Chinese remains the definitive version.


| №  | Instrument                                     | Measurand             | Calibration Method                                      | Range                             | Expanded Uncertainty<br>( $k=2$ ) | Note | Effective Date |
|----|------------------------------------------------|-----------------------|---------------------------------------------------------|-----------------------------------|-----------------------------------|------|----------------|
|    |                                                | Time                  |                                                         | (0.5~99)s                         | $U_{rel}=1.0\%$                   |      |                |
| 22 | *A.C Electron Load                             | Voltage               | C.S.for AC Electronic Load<br>JJF(Electronic) 0002      | (1~750)V,(45~65)Hz                | $U_{rel}=(0.42\sim0.12)\%$        |      |                |
|    |                                                | Current               |                                                         | (0.05~20)A,(45~65)Hz              | $U_{rel}=(0.18\sim0.12)\%$        |      |                |
|    |                                                | power                 |                                                         | (1~750)V/(0.05~20)A<br>,(45~65)Hz | $U_{rel}=(0.17\sim0.12)\%$        |      |                |
|    |                                                | PF                    |                                                         | -1~0~+1                           | $U=0.0012$                        |      |                |
| 23 | *DC High Voltage Dividers                      | Voltage Divider Ratio | V.R. of DC High Voltage Dividers JJG 1007               | (0.1~100)kV                       | $U_{rel}=(0.15\sim0.12)\%$        |      |                |
| 24 | *High Voltage Divider at Power Frequency       | Voltage Divider Ratio | V.R. of High Voltage Divider at Power Frequency JJG 496 | (0.1~100)kV(50Hz)                 | $U_{rel}=(0.25\sim0.17)\%$        |      |                |
| 25 | *High Voltage (Electrostatic)Voltmeter (Probe) | AC Voltage            | V.R. of High Voltage Electrostratic Voltmeter JJG 494   | (0.1~100)kV,(45~65)Hz             | $U_{rel}=(0.26\sim0.14)\%$        |      |                |
|    |                                                | DC Voltage            |                                                         | (0.1~100)kV                       | $U_{rel}=(0.20\sim0.12)\%$        |      |                |
| 26 | *Relaying Protection Tester                    | AC Voltage            | V.R. of Relaying Protection JJG 1112                    | (1~750)V(45~65)Hz                 | $U_{rel}=0.012\%$                 |      |                |
|    |                                                | AC Current            |                                                         | (0.1~120)A(45~65)Hz               | $U_{rel}=(0.03\sim0.012)\%$       |      |                |
|    |                                                | AC Power              |                                                         | (1~750)V (0.1~120)A(45~65)Hz      | $U_{rel}=(0.03\sim0.012)\%$       |      |                |
|    |                                                | DC Voltage            |                                                         | (1~750)V                          | $U_{rel}=0.012\%$                 |      |                |
|    |                                                | DC Current            |                                                         | (1~100)A                          | $U_{rel}=0.03\%$                  |      |                |
|    |                                                | Time                  |                                                         | (0.001~100)s                      | $U=0.6\text{ms}$                  |      |                |
|    |                                                | Response Time         |                                                         | 10 $\mu\text{s}$ ~10ms            | $U_{rel}=2.0\%$                   |      |                |



No. CNAS L0730

第 55 页 共 163 页

The scope of the accreditation in Chinese remains the definitive version.

| №  | Instrument                                      | Measurand            | Calibration Method                                                                                             | Range                                       | Expanded Uncertainty<br>( $k=2$ ) | Note                                                                                 | Effective Date |
|----|-------------------------------------------------|----------------------|----------------------------------------------------------------------------------------------------------------|---------------------------------------------|-----------------------------------|--------------------------------------------------------------------------------------|----------------|
|    |                                                 | Phase                |                                                                                                                | $(0\sim 360)^\circ$ (45~65)Hz               | $U=0.01^\circ$                    |                                                                                      |                |
| 27 | *Double clamp<br>phase position<br>voltmeter    | AC Current           | C.S. for Digital AC Electrical<br>Parameters Meter JJF 1491                                                    | 20mA~20A(45~65)Hz                           | $U_{rel}=(0.3\sim 0.10)\%$        |                                                                                      |                |
|    |                                                 | AC Voltage           |                                                                                                                | $(2\sim 1000)V(45\sim 65)Hz$                | $U_{rel}=(0.3\sim 0.11)\%$        |                                                                                      |                |
|    |                                                 | Phase                |                                                                                                                | $(0\sim 360)^\circ$ (45~65)Hz               | $U=0.1^\circ$                     |                                                                                      |                |
| 28 | *Zinc Oxide<br>Arrester Tester                  | Testing<br>Voltage   | General Technical Conditions<br>of Resistive Current Tester<br>for Zinc Oxide Surge Arrester<br>DL/T 987 6.6.3 | $(1\sim 300)V$<br>(45~65)Hz                 | $U_{rel}=0.3\%$                   |                                                                                      |                |
|    |                                                 | AC Current           |                                                                                                                | $(0.1\sim 100)mA$<br>(45~65)Hz              | $U_{rel}=(0.5\sim 0.3)\%$         |                                                                                      |                |
|    |                                                 | AC Power             |                                                                                                                | $(1\sim 300)V(0.33\sim 100)mA$<br>(45~65)Hz | $U_{rel}=(0.4\sim 0.21)\%$        |                                                                                      |                |
|    |                                                 | phase                |                                                                                                                | $(0\sim \pm 180)^\circ$                     | $U=0.06^\circ$                    |                                                                                      |                |
| 29 | *Heavy Current<br>Tester                        | AC Current           | High Current Tester<br>Calibration Method FFD0902                                                              | $(1\sim 5000)A, (45\sim 65)Hz$              | $U_{rel}=(0.4\sim 0.2)\%$         |                                                                                      |                |
| 30 | *Electrical Energy<br>Meters                    | Electrical<br>Energy | V.R. of Electrical Meters for<br>Measuring Alternating-<br>current Electrical Energy JJG<br>596                | $(10\sim 380)V(0.1\sim 100)A$<br>(45~65)Hz  | $U_{rel}=0.06\%$                  |                                                                                      |                |
| 31 | *Electromechanical<br>Energy Meters             | Electrical<br>Energy | V.R. of Electromechanical<br>Meters for Measuring<br>Alternating-current Electrical<br>Energy JJG 307          | $(10\sim 380)V(0.1\sim 100)A$<br>(45~65)Hz  | $U_{rel}=0.24\%$                  |  |                |
| 32 | *High-voltage<br>Withstanding<br>Voltage Tester | Voltage              | V.R. of High-voltage<br>Withstanding Voltage Tester<br>JJG(Military) 18                                        | $(1\sim 100)kV, (DC, 45Hz\sim 65Hz)$        | $U_{rel}=1.1\%$                   |                                                                                      |                |
|    |                                                 | Current              |                                                                                                                | $(0.1\sim 100)mA, (DC, 45Hz\sim 65Hz)$      | $U_{rel}=1.3\%$                   |                                                                                      |                |



No. CNAS L0730

第 56 页 共 163 页

The scope of the accreditation in Chinese remains the definitive version.

| №  | Instrument                                             | Measurand  | Calibration Method                                                        | Range                                                 | Expanded Uncertainty<br>( $k=2$ )  | Note                        | Effective Date |
|----|--------------------------------------------------------|------------|---------------------------------------------------------------------------|-------------------------------------------------------|------------------------------------|-----------------------------|----------------|
|    |                                                        | Time       |                                                                           | (1~999)s                                              | $U_{\text{rel}}=0.12\%$            |                             |                |
| 33 | *Wrist Strap and Footwear Tester                       | Resistance | C.S. of Wrist Strap and Footwear Tester JJF(Electronic) 31502             | (0.1~200)M $\Omega$                                   | $U_{\text{rel}}=(1.0\sim3)\%$      |                             |                |
| 34 | *D.C. low Resistance Meters                            | Resistance | V.R. of D.C. low Resistance Meters JJG 837                                | $10^{-6}\Omega \sim 10^{-5}\Omega$                    | $U_{\text{rel}}=1.3\%$             |                             |                |
|    |                                                        |            |                                                                           | $10^{-5}\Omega \sim 10^{-4}\Omega$                    | $U_{\text{rel}}=0.6\%$             |                             |                |
|    |                                                        |            |                                                                           | $10^{-4}\Omega \sim 10^{-3}\Omega$                    | $U_{\text{rel}}=0.13\%$            |                             |                |
|    |                                                        |            |                                                                           | $10^{-3}\Omega \sim 10^{-2}\Omega$                    | $U_{\text{rel}}=0.06\%$            |                             |                |
|    |                                                        |            |                                                                           | $10^{-2}\Omega \sim 10^6\Omega$                       | $U_{\text{rel}}=0.013\%$           |                             |                |
| 35 | *AC Power Source                                       | Voltage    | V.R. for Standard AC Power Source JJG(Military) 6                         | 1mV~750V,45Hz~65Hz                                    | $U_{\text{rel}}=(0.06\sim0.012)\%$ |                             |                |
|    |                                                        | Current    |                                                                           | 1mA~120A,45Hz~65Hz                                    | $U_{\text{rel}}=(0.06\sim0.012)\%$ |                             |                |
|    |                                                        | Power      |                                                                           | 1mV~750V1mA~120A,45Hz~65Hz                            | $U_{\text{rel}}=(0.06\sim0.012)\%$ |                             |                |
|    |                                                        | Phase      |                                                                           | $0^\circ \sim 360^\circ$                              | $U=0.005^\circ$                    |                             |                |
|    |                                                        | Frequency  |                                                                           | (45~65)Hz                                             | $U_{\text{rel}}=0.001\%$           |                             |                |
| 36 | Contactless Electrostatic voltage Measuring Instrument | Voltage    | C.S. for Contactless Electrostatic Voltage Measuring Instruments JJF 1517 | (0.1~20)kV                                            | $U_{\text{rel}}=0.20\%$            | 合格评定国家认可<br>CNAS<br>认可证书专用章 |                |
| 37 | *D.C. Potentiometers                                   | Voltage    | V.R. of D.C Potentiometers JJG 123                                        | 1.0mV~100V                                            | $U_{\text{rel}}=(0.004\sim0.01)\%$ |                             |                |
| 38 | *DC Magnetoelectric Galvanometers                      | Current    | V.R. of DC Magnetoelectric Galvanometers JJG 495                          | $(1 \times 10^{-9} \sim 1 \times 10^{-1})\text{A/mm}$ | $U_{\text{rel}}=1.0\%$             |                             |                |



No. CNAS L0730

第 57 页 共 163 页

The scope of the accreditation in Chinese remains the definitive version.

| №  | Instrument                              | Measurand         | Calibration Method                                       | Range                                | Expanded Uncertainty<br>( <i>k</i> =2) | Note | Effective Date |
|----|-----------------------------------------|-------------------|----------------------------------------------------------|--------------------------------------|----------------------------------------|------|----------------|
| 39 | *Control Box                            | Voltage           | C.S.for operate box （Control Box） ZBDX-1905              | 1V～1kV                               | <i>U</i> <sub>rel</sub> =0.15%         |      |                |
|    |                                         | Current           |                                                          | 0.1mA～30A                            | <i>U</i> <sub>rel</sub> =0.15%         |      |                |
|    |                                         | Time              |                                                          | 1ms～9999s                            | <i>U</i> =3ms                          |      |                |
| 40 | *Digital AC Electrical Parameters Meter | Voltage           | C.S. for Digital AC Electrical Parameters Meter JJF 1491 | (0.01～1000)V,45Hz～1kHz               | <i>U</i> <sub>rel</sub> =(0.8～0.010)%  |      |                |
|    |                                         | Current           |                                                          | (0.001～80)A,（45Hz～1kHz)              | <i>U</i> <sub>rel</sub> =0.012%        |      |                |
|    |                                         |                   |                                                          | (0.001～120)A,（45Hz～65Hz)             | <i>U</i> <sub>rel</sub> =(0.8～0.012)%  |      |                |
|    |                                         | Power             |                                                          | (0.01～1000)V/(0.001～80)A,（45Hz～1kHz) | <i>U</i> <sub>rel</sub> =(0.8～0.012)%  |      |                |
|    |                                         |                   |                                                          | (0.01～525)V/(80～120)A,（45Hz～65Hz)    | <i>U</i> <sub>rel</sub> =0.012%        |      |                |
|    |                                         | Power Factor      |                                                          | -1～0～+1                              | <i>U</i> =0.0006                       |      |                |
|    |                                         | Frequency         |                                                          | (45～1000)Hz                          | <i>U</i> <sub>rel</sub> =0.010%        |      |                |
| 41 | *Digital DC Power Meter                 | DC Power          | V.R. of DC digital power meter JJG(YUE) 055              | 1mA～2kA/1mV～1kV                      | <i>U</i> <sub>rel</sub> =0.016%        |      |                |
| 42 | DC Standard Resistors                   | D.C. Resistance   | V.R. of DC Standard Resistors JJG 166                    | 1m Ω ～1 Ω                            | <i>U</i> <sub>rel</sub> =(0.11～0.011)% |      |                |
|    |                                         |                   |                                                          | 1 Ω ～2G Ω                            | <i>U</i> <sub>rel</sub> =(0.0011～0.4)% |      |                |
| 43 | *Charged Plate Monitor                  | Static Voltage    | C.S. for Charge Plate Monitor JJF(Electronic) 31003      | ±(10～1020)V                          | <i>U</i> <sub>rel</sub> =0.4%          |      |                |
|    |                                         | Static Decay Time |                                                          | (0.1～99.9)s                          | <i>U</i> =0.02s                        |      |                |
| 44 | Instrument Current Transformers         | Current Ratio     | V.R. of Instrument Current Transformers JJG 313          | (5～5000)A/(5A,1A)                    | <i>U</i> <sub>rel</sub> =(0.09～0.06)%  |      |                |
|    |                                         | phase error       |                                                          | -10'～+10'                            | <i>U</i> =1.1'～0.42'                   |      |                |



No. CNAS L0730

第 58 页 共 163 页

The scope of the accreditation in Chinese remains the definitive version.

| №  | Instrument                                                               | Measurand                   | Calibration Method                                    | Range                                                 | Expanded Uncertainty<br>( $k=2$ ) | Note | Effective Date |
|----|--------------------------------------------------------------------------|-----------------------------|-------------------------------------------------------|-------------------------------------------------------|-----------------------------------|------|----------------|
| 45 | Instrument<br>Voltage<br>Transformers                                    | Voltage Ratio               | V.R. of Instrument Voltage<br>Transformers JJG 314    | 10000/100 ; (10000 $\sqrt{3}$ )/<br>(100 $\sqrt{3}$ ) | $U_{rel} = (0.09 \sim 0.06)\%$    |      |                |
|    |                                                                          | phase error                 |                                                       | -10' ~ +10'                                           | $U = 2.5' \sim 1.6'$              |      |                |
| 46 | *Calibration<br>Specification of<br>Calibrators for<br>Electrical Meters | DC Voltage                  |                                                       | 10mV ~ 1000V                                          | $U_{rel} = (0.02 \sim 0.003)\%$   |      |                |
|    |                                                                          | DC Current                  |                                                       | 10 $\mu$ A ~ 100mA                                    | $U_{rel} = (0.02 \sim 0.01)\%$    |      |                |
|    |                                                                          |                             |                                                       | 100mA ~ 20A                                           | $U_{rel} = (0.05 \sim 0.01)\%$    |      |                |
|    |                                                                          | AC Voltage                  |                                                       | 10mV ~ 1000V (50 ~<br>1000)Hz                         | $U_{rel} = (0.04 \sim 0.01)\%$    |      |                |
|    |                                                                          | AC Current                  |                                                       | 10 $\mu$ A ~ 100mA (50 ~<br>1000)Hz                   | $U_{rel} = (0.20 \sim 0.04)\%$    |      |                |
|    |                                                                          |                             |                                                       | 100mA ~ 20A (50 ~<br>1000)Hz                          | $U_{rel} = (0.10 \sim 0.04)\%$    |      |                |
|    |                                                                          | Resistance                  | C.S. of Calibrators for<br>Electrical Meters JJF 1284 | 1 $\Omega$ ~ 100 M $\Omega$                           | $U_{rel} = (0.01 \sim 0.03)\%$    |      |                |
|    |                                                                          | Frequency                   |                                                       | (10 ~ 1000)Hz                                         | $U_{rel} = (0.03 \sim 0.01)\%$    |      |                |
|    |                                                                          | Ripple<br>content           |                                                       | (0.01 ~ 10)%                                          | $U = 0.2\%$                       |      |                |
|    |                                                                          | Distortion                  |                                                       | (0.03 ~ 30)% (50 ~<br>1000)Hz                         | $U_{rel} = 10\%$                  |      |                |
|    |                                                                          | Source<br>Regulate<br>Ratio |                                                       | 10mV ~ 1000V 100mA ~<br>20A                           | $U_{rel} = 0.021\%$               |      |                |
|    |                                                                          | Load<br>Regulated<br>Ratio  |                                                       | 10mV ~ 1000V 100mA ~<br>20A                           | $U_{rel} = 0.021\%$               |      |                |
| 47 | *Touch Current<br>Testers                                                | Input DC<br>Resistor        | V.R. of Touch Current Tester<br>JJG(YUE) 027          | 1 $\Omega$ ~ 100k $\Omega$                            | $U_{rel} = (0.03 \sim 0.01)\%$    |      |                |



No. CNAS L0730

第 59 页 共 163 页

The scope of the accreditation in Chinese remains the definitive version.

| №  | Instrument                                               | Measurand                  | Calibration Method                       | Range                                          | Expanded Uncertainty<br>( $k=2$ ) | Note | Effective Date |
|----|----------------------------------------------------------|----------------------------|------------------------------------------|------------------------------------------------|-----------------------------------|------|----------------|
| 48 | *Calibration<br>Specification for<br>Process Calibrators | Current                    | C.S. for Process Calibrators<br>JJF 1472 | 10 $\mu$ A $\sim$ 1A, (45Hz $\sim$ 65Hz)       | $U_{rel}=(1.0\sim 0.10)\%$        |      |                |
|    |                                                          |                            |                                          | 10 $\mu$ A $\sim$ 10mA,<br>(20Hz $\sim$ 1MHz)  | $U_{rel}=(1.2\sim 2.4)\%$         |      |                |
|    |                                                          | Input Impedance            |                                          | 1 $\Omega \sim$ 100k $\Omega$ 20Hz $\sim$ 1MHz | $U_{rel}=(0.06\sim 0.10)\%$       |      |                |
|    |                                                          | Alarm Current              |                                          | 1 $\mu$ A $\sim$ 1A                            | $U_{rel}=(1.0\sim 0.10)\%$        |      |                |
|    |                                                          | DC Voltage Output          | C.S. for Process Calibrators<br>JJF 1472 | 10mV $\sim$ 100V                               | $U_{rel}=(0.01\sim 0.003)\%$      |      |                |
|    |                                                          | DC Current Output          |                                          | 1mA $\sim$ 100mA                               | $U_{rel}=(0.01\sim 0.003)\%$      |      |                |
|    |                                                          | DC Resistance Output       |                                          | 1 $\Omega \sim$ 10k $\Omega$                   | $U_{rel}=(0.05\sim 0.01)\%$       |      |                |
|    |                                                          | Frequency Output           |                                          | 1Hz $\sim$ 50kHz                               | $U_{rel}=0.001\%$                 |      |                |
|    |                                                          | RTD Analog output          |                                          | (-200 $\sim$ 850) $^{\circ}$ C                 | $U=0.06^{\circ}$ C                |      |                |
|    |                                                          | Thermocouple Analog output |                                          | (-250 $\sim$ 1800) $^{\circ}$ C                | $U=0.06^{\circ}$ C                |      |                |
|    |                                                          | DC Voltage Measure         |                                          | 10mV $\sim$ 300V                               | $U_{rel}=(0.01\sim 0.003)\%$      |      |                |
|    |                                                          | DC Current Measure         |                                          | 1mA $\sim$ 100mA                               | $U_{rel}=(0.01\sim 0.003)\%$      |      |                |
|    |                                                          | AC Voltage Measure         |                                          | 10mV $\sim$ 300V                               | $U_{rel}=(0.3\sim 0.14)\%$        |      |                |
|    |                                                          | AC Current Measure         |                                          | 0.1mA $\sim$ 200mA                             | $U_{rel}=(0.3\sim 0.11)\%$        |      |                |



No. CNAS L0730

第 60 页 共 163 页

The scope of the accreditation in Chinese remains the definitive version.

| №  | Instrument          | Measurand             | Calibration Method            | Range                   | Expanded Uncertainty<br>( <i>k</i> =2)    | Note | Effective Date |
|----|---------------------|-----------------------|-------------------------------|-------------------------|-------------------------------------------|------|----------------|
|    |                     | DC Resistance Measure |                               | 1 Ω ~ 100k Ω            | <i>U</i> <sub>rel</sub> = (0.05~0.01)%    |      |                |
|    |                     | Frequency Measure     |                               | 1Hz~500kHz              | <i>U</i> <sub>rel</sub> =0.001%           |      |                |
|    |                     | RTD Measure           |                               | (-200~850)°C            | <i>U</i> = 0.06°C                         |      |                |
|    |                     | Thermocouple Measure  |                               | (-250~1800)°C           | <i>U</i> = 0.06°C                         |      |                |
| 49 | *Digital Multimeter | DC voltage            | C.S. for Multimeters JJF 1587 | (10~200)mV              | <i>U</i> <sub>rel</sub> =(0.2~0.0011)%    |      |                |
|    |                     |                       |                               | (0.2~2)V                | <i>U</i> <sub>rel</sub> =(0.0011~0.0006)% |      |                |
|    |                     |                       |                               | (2~10)V                 | <i>U</i> <sub>rel</sub> =(0.0006~0.0005)% |      |                |
|    |                     |                       |                               | (10~20)V                | <i>U</i> <sub>rel</sub> =0.0005%          |      |                |
|    |                     |                       |                               | (20~200)V               | <i>U</i> <sub>rel</sub> =(0.0009~0.0007)% |      |                |
|    |                     |                       |                               | (200~1020)V             | <i>U</i> <sub>rel</sub> =(0.0011~0.0009)% |      |                |
|    |                     | AC voltage            |                               | (10~20)mV,(10~40)Hz     | <i>U</i> <sub>rel</sub> =(0.23~0.03)%     |      |                |
|    |                     |                       |                               | (10~20)mV(40Hz~20kHz)   | <i>U</i> <sub>rel</sub> =(0.21~0.28)%     |      |                |
|    |                     |                       |                               | (10~20)mV,(20~100)kHz   | <i>U</i> <sub>rel</sub> =(0.23~0.05)%     |      |                |
|    |                     |                       |                               | (10~20)mV,(100~500)kHz  | <i>U</i> <sub>rel</sub> =(0.7~0.16)%      |      |                |
|    |                     |                       |                               | (10~20)mV,(500kHz~1MHz) | <i>U</i> <sub>rel</sub> =(1.4~0.4)%       |      |                |
|    |                     |                       |                               | (20~200)mV(10~40)Hz     | <i>U</i> <sub>rel</sub> =(0.09~0.013)%    |      |                |



No. CNAS L0730

第 61 页 共 163 页

The scope of the accreditation in Chinese remains the definitive version.

| № | Instrument | Measurand | Calibration Method | Range                   | Expanded Uncertainty<br>( $k=2$ ) | Note | Effective Date |
|---|------------|-----------|--------------------|-------------------------|-----------------------------------|------|----------------|
|   |            |           |                    | (20~200)mV(40Hz~20kHz)  | $U_{rel} = (0.04 \sim 0.011)\%$   |      |                |
|   |            |           |                    | (20~200)mV(20~100)kHz   | $U_{rel} = (0.06 \sim 0.025)\%$   |      |                |
|   |            |           |                    | (20~200)mV(100~500)kHz  | $U_{rel} = (0.20 \sim 0.11)\%$    |      |                |
|   |            |           |                    | (20~200)mV(500kHz~1MHz) | $U_{rel} = (0.6 \sim 0.4)\%$      |      |                |
|   |            |           |                    | 200mV~2V(10~40)Hz       | $U_{rel} = (0.05 \sim 0.010)\%$   |      |                |
|   |            |           |                    | 200mV~2V (40Hz~20kHz)   | $U_{rel} = (0.009 \sim 0.006)\%$  |      |                |
|   |            |           |                    | 200mV~2V(20~100)kHz     | $U_{rel} = (0.013 \sim 0.009)\%$  |      |                |
|   |            |           |                    | 200mV~2V(100~500)kHz    | $U_{rel} = (0.09 \sim 0.05)\%$    |      |                |
|   |            |           |                    | 200mV~2V(500kHz~1MHz)   | $U_{rel} = (0.4 \sim 0.25)\%$     |      |                |
|   |            |           |                    | (2~20)V(10~40)Hz        | $U_{rel} = (0.05 \sim 0.011)\%$   |      |                |
|   |            |           |                    | (2~20)V(40Hz~20kHz)     | $U_{rel} = (0.008 \sim 0.005)\%$  |      |                |
|   |            |           |                    | (2~20)V(20~100)kHz      | $U_{rel} = (0.013 \sim 0.009)\%$  |      |                |
|   |            |           |                    | (2~20)V(100~500)kHz     | $U_{rel} = (0.06 \sim 0.04)\%$    |      |                |
|   |            |           |                    | (2~10)V(500kHz~1MHz)    | $U_{rel} = (0.4 \sim 0.18)\%$     |      |                |
|   |            |           |                    | (20~200)V(10~40)Hz      | $U_{rel} = (0.008 \sim 0.021)\%$  |      |                |
|   |            |           |                    | (20~200)V(40Hz~20kHz)   | $U_{rel} = (0.009 \sim 0.006)\%$  |      |                |



No. CNAS L0730

第 62 页 共 163 页

The scope of the accreditation in Chinese remains the definitive version.

| № | Instrument | Measurand  | Calibration Method | Range                   | Expanded Uncertainty<br>( <i>k</i> =2)   | Note | Effective Date |
|---|------------|------------|--------------------|-------------------------|------------------------------------------|------|----------------|
|   |            |            | ilac-M             | (20~200)V(20~100)kHz    | <i>U</i> <sub>rel</sub> = (0.014~0.009)% |      |                |
|   |            |            |                    | (20~200)V(100~500)kHz   | <i>U</i> <sub>rel</sub> = (0.18~0.10)%   |      |                |
|   |            |            |                    | (200~1020)V(15~50)Hz    | <i>U</i> <sub>rel</sub> =0.04%           |      |                |
|   |            |            |                    | (200~1020)V(50Hz~1kHz)  | <i>U</i> <sub>rel</sub> = (0.010~0.008)% |      |                |
|   |            | DC current |                    | (10~20) μ A             | <i>U</i> <sub>rel</sub> =(1.0~0.04)%     |      |                |
|   |            |            |                    | (20~200) μ A            | <i>U</i> <sub>rel</sub> =(0.04~0.008)%   |      |                |
|   |            |            |                    | (0.2~2)mA               | <i>U</i> <sub>rel</sub> =(0.008~0.004)%  |      |                |
|   |            |            |                    | (2~20)mA                | <i>U</i> <sub>rel</sub> =(0.006~0.004)%  |      |                |
|   |            |            |                    | (20~200)mA              | <i>U</i> <sub>rel</sub> =(0.009~0.006)%  |      |                |
|   |            |            |                    | (0.2~2)A                | <i>U</i> <sub>rel</sub> = (0.015~0.009)% |      |                |
|   |            |            |                    | (2~11)A                 | <i>U</i> <sub>rel</sub> =(0.003~0.08)%   |      |                |
|   |            |            |                    | (11~20)A                | <i>U</i> <sub>rel</sub> =(0.003~0.01)%   |      |                |
|   |            |            |                    | (20~100)A               | <i>U</i> <sub>rel</sub> =(0.015~0.012)%  |      |                |
|   |            | AC current |                    | (20~200) μ A(10~40)Hz   | <i>U</i> <sub>rel</sub> =(0.11~0.02)%    |      |                |
|   |            |            |                    | (20~200) μ A(40Hz~1kHz) | <i>U</i> <sub>rel</sub> =(0.06~0.017)%   |      |                |
|   |            |            |                    | (20~200) μ A(1kHz~5kHz) | <i>U</i> <sub>rel</sub> =(0.10~0.04)%    |      |                |
|   |            |            |                    | (20~200) μ A(5~10)kHz   | <i>U</i> <sub>rel</sub> =(0.5~0.15)%     |      |                |



No. CNAS L0730

第 63 页 共 163 页

The scope of the accreditation in Chinese remains the definitive version.

| No | Instrument                    | Measurand  | Calibration Method                             | Range                       | Expanded Uncertainty<br>( $k=2$ ) | Note | Effective Date |
|----|-------------------------------|------------|------------------------------------------------|-----------------------------|-----------------------------------|------|----------------|
|    |                               |            |                                                | (0.2~220)mA(10~40)Hz        | $U_{rel}=(0.05\sim0.015)\%$       |      |                |
|    |                               |            |                                                | (0.2~220)mA(40Hz~1kHz)      | $U_{rel}=(0.04\sim0.015)\%$       |      |                |
|    |                               |            |                                                | (0.2~220)mA(1~5)kHz         | $U_{rel}=(0.08\sim0.027)\%$       |      |                |
|    |                               |            |                                                | (0.2~220)mA(5~10)kHz        | $U_{rel}=(0.5\sim0.10)\%$         |      |                |
|    |                               |            |                                                | (0.2~2)A(10~40)Hz           | $U_{rel}=(0.05\sim0.03)\%$        |      |                |
|    |                               |            |                                                | (0.2~2)A(40Hz~1kHz)         | $U_{rel}=(0.05\sim0.01)\%$        |      |                |
|    |                               |            |                                                | (0.2~2)A(5kHz~10kHz)        | $U_{rel}=(0.9\sim0.8)\%$          |      |                |
|    |                               |            |                                                | (2~11)A(45~65)Hz            | $U_{rel}=(0.17\sim0.09)\%$        |      |                |
|    |                               |            |                                                | (2~11)A(65~1500)Hz          | $U_{rel}=(0.21\sim0.13)\%$        |      |                |
|    |                               |            |                                                | (11~20)A(45~65)Hz           | $U_{rel}=(0.01\sim0.05)\%$        |      |                |
|    |                               |            |                                                | (20~80)A(45~65)Hz           | $U_{rel}=(0.015\sim0.012)\%$      |      |                |
|    |                               | Resistance |                                                | 0.1m $\Omega$ ~ 1m $\Omega$ | $U_{rel}=0.13\%$                  |      |                |
|    |                               |            |                                                | 1m $\Omega$ ~ 20 $\Omega$   | $U_{rel}=0.0024\%$                |      |                |
|    |                               |            |                                                | 20 $\Omega$ ~ 200k $\Omega$ | $U_{rel}=0.0011\%$                |      |                |
|    |                               |            |                                                | 200k $\Omega$ ~ 2M $\Omega$ | $U_{rel}=0.0022\%$                |      |                |
|    |                               |            |                                                | (2~20)M $\Omega$            | $U_{rel}=0.005\%$                 |      |                |
|    |                               |            |                                                | (20~100)M $\Omega$          | $U_{rel}=0.012\%$                 |      |                |
| 50 | *DC Stabilized Power Supplies | DC Voltage | C.S. for DC Stabilized Power Supplies JJF 1597 | (0.1~1000)V                 | $U_{rel}=(0.09\sim0.012)\%$       |      |                |



No. CNAS L0730

第 64 页 共 163 页

The scope of the accreditation in Chinese remains the definitive version.

| №  | Instrument                                        | Measurand                             | Calibration Method                                                                                                 | Range                                        | Expanded Uncertainty<br>( $k=2$ ) | Note                       | Effective Date |
|----|---------------------------------------------------|---------------------------------------|--------------------------------------------------------------------------------------------------------------------|----------------------------------------------|-----------------------------------|----------------------------|----------------|
|    |                                                   | DC Current                            | ilac-MRA<br>INTERNATIONAL ACCREDITATION SERVICE FOR CONFORMITY ASSESSMENT<br>SCHEDULE OF ACCREDITATION CERTIFICATE | (0.001~100)A                                 | $U_{rel}=(0.19\sim0.026)\%$       |                            |                |
|    |                                                   |                                       |                                                                                                                    | (100~240)A                                   | $U_{rel}=(0.026\sim0.08)\%$       |                            |                |
|    |                                                   | Stable Voltage<br>output cycle        |                                                                                                                    | 200 $\mu$ V~50V                              | $U_{rel}=3\%$                     |                            |                |
|    |                                                   | Stable Current<br>output cycle        |                                                                                                                    | 200 $\mu$ A~50A                              | $U_{rel}=3\%$                     |                            |                |
|    |                                                   | Load<br>Transient<br>Recovery<br>Time |                                                                                                                    | 10 $\mu$ s~1s                                | $U_{rel}=0.1\%$                   |                            |                |
| 51 | *online testers of<br>winding<br>temperature rise | Resistor                              | C.S. for online testers of<br>winding temperature rise JJF<br>1540                                                 | 0.1 $\Omega$ ~10k $\Omega$                   | $U_{rel}=0.013\%$                 |                            |                |
| 52 | *high voltage and<br>value D.C.<br>resistors      | Resistor                              | V.R. of high voltage and<br>value D.C. resistors JJG 1072                                                          | 100 $\Omega$ ~1T $\Omega$                    | $U_{rel}=0.06\%\sim2.3\%$         |                            |                |
| 53 | *VDR DC<br>parameter testers                      | DC Current                            | C.S. for VDR DC parameter<br>testers JJF(Zhe) 1088                                                                 | 10 $\mu$ A~100 mA                            | $U_{rel}=1.0\%\sim0.16\%$         |                            |                |
|    |                                                   | DC Voltage                            |                                                                                                                    | (0.1~5000)V                                  | $U_{rel}=0.16\%\sim0.8\%$         |                            |                |
|    |                                                   | ripple factor                         |                                                                                                                    | (0.1~10)%                                    | $U=0.2\%$                         |                            |                |
| 54 | DC shunts                                         | Resistor                              | V.R. of DC shunts JJG 1069                                                                                         | 1 $\mu$ $\Omega$ ~100 $\Omega$ ,(5A~<br>2kA) | $U_{rel}=(0.02\sim0.012)\%$       | 合格评定国家认可<br>委员会<br>认可证书专用章 |                |
| 55 | *micro current<br>meter                           | current                               | C.S for Digital Electrometers<br>JJF 1726                                                                          | 1pA~100mA                                    | $U_{rel}=(1.2\sim0.02)\%$         |                            |                |
|    |                                                   | Voltage                               |                                                                                                                    | (0.1~1000)V                                  | $U_{rel}=0.01\%$                  |                            |                |
|    |                                                   | Resistor                              |                                                                                                                    | 100 $\Omega$ ~12T $\Omega$                   | $U_{rel}=(0.12\sim6)\%$           |                            |                |
|    |                                                   | charge                                |                                                                                                                    | 0.1nC~2 $\mu$ C                              | $U_{rel}=0.15\%$                  |                            |                |



No. CNAS L0730

第 65 页 共 163 页

The scope of the accreditation in Chinese remains the definitive version.

| №  | Instrument                                              | Measurand         | Calibration Method                                                                  | Range                                          | Expanded Uncertainty<br>( $k=2$ ) | Note | Effective Date |
|----|---------------------------------------------------------|-------------------|-------------------------------------------------------------------------------------|------------------------------------------------|-----------------------------------|------|----------------|
| 56 | *withstanding voltage testers spot tester               | Current           | C.S. for withstanding voltage testers spot tester ZBDX-1901                         | 0.1mA~100mA,(0.1kV~20kV)                       | $U_{rel}=1.4\%$                   |      |                |
| 57 | *verification equipment for AC electrical energy meters | electric energy   | V.R. of verification equipment for AC electrical energy meters JJG 597              | (30~525)V/(0.02~120)A                          | $U_{rel}=0.024\%$                 |      |                |
| 58 | Reference Meters for Electrical Energy                  | electric energy   | V.R. of Reference Meters for Electrical Energy JJG 1085                             | (30~525)V/(0.02~120)A                          | $U_{rel}=0.016\%$                 |      |                |
| 59 | *AC shunt                                               | resistance        | C.S. for AC shunts JJF(Electronic) 0020                                             | 1m $\Omega$ ~10 $\Omega$ , (5~100)A, (45~65)Hz | $U_{rel}=(1.1\sim0.06)\%$         |      |                |
| 60 | *Welding Power Source of DC Electric Welding Machines   | Welding current   | C.S. for Welding Power Sources of DC Electric Welding Machines JJF 1985             | (1~2000)A                                      | $U_{rel}=1.6\%$                   |      |                |
|    |                                                         | No-load voltage   |                                                                                     | (20~141)V                                      | $U_{rel}=3\%$                     |      |                |
|    |                                                         | Load voltage      |                                                                                     | (10~44)V                                       | $U_{rel}=3\%$                     |      |                |
| 61 | *DC High Current Meters                                 | DC Current        | V.R. for DC High Current Meters JJG(Military) 195                                   | (100~2400)A                                    | $U_{rel}=(0.012\sim0.014)\%$      |      |                |
| 62 | *Off-board Charger for Electric Vehicles                | Electrical Energy | V.R. of Off-board Chargers for Electric Vehicles(for Trial Implementation) JJG 1149 | DCV: (30~1000)VDCA: (5~250)A                   | $U_{rel}=0.1\%$                   |      |                |
|    |                                                         | Time              |                                                                                     | (0~9999)s                                      | $U=1s$                            |      |                |
| 63 | *AC Charging Spot for Electric Vehicles                 | Electrical Energy | V.R. of AC Charge Spots for Electric Vehicles(for Trial Implementation) JJG 1148    | ACV:(30~380)VACA:(1~70)A                       | $U_{rel}=0.1\%$                   |      |                |
|    |                                                         | Time              |                                                                                     | (0~9999)s                                      | $U=1s$                            |      |                |



No. CNAS L0730

第 66 页 共 163 页

The scope of the accreditation in Chinese remains the definitive version.



| №  | Instrument                                      | Measurand                                                       | Calibration Method                                        | Range                                 | Expanded Uncertainty<br>( <i>k</i> =2) | Note | Effective Date |  |
|----|-------------------------------------------------|-----------------------------------------------------------------|-----------------------------------------------------------|---------------------------------------|----------------------------------------|------|----------------|--|
| 64 | Tesla meter/<br>Gaussian meter/<br>magnetometer | Magnetic<br>induction<br>intensity/Mag<br>netic flux<br>density | C.S.for (1 mT~2.5 T)<br>Magnetometers JJF 1832            | 1mT~2T                                | <i>U</i> <sub>rel</sub> =(3~0.5)%      |      |                |  |
| 65 | *power quality<br>analyzer                      | voltage                                                         | Verification code for power<br>quality analyzer DL/T 1028 | (1~1008)V, (16~<br>850)Hz             | <i>U</i> <sub>rel</sub> =(0.009~0.03)% |      |                |  |
|    |                                                 | frequency                                                       |                                                           | 16Hz~850Hz                            | <i>U</i> <sub>rel</sub> =0.006%        |      |                |  |
|    |                                                 | harmonic<br>voltage                                             |                                                           | (0.3~302)V, 100Hz~<br>3kHz            | <i>U</i> <sub>rel</sub> =(0.06~0.12)%  |      |                |  |
|    |                                                 | harmonic<br>current                                             |                                                           | (0.01~6)A, 100Hz~<br>3kHz             | <i>U</i> <sub>rel</sub> =0.07%~0.10%   |      |                |  |
|    |                                                 | harmonic<br>power                                               |                                                           | (0.001~1800)W,<br>100Hz~3kHz          | <i>U</i> <sub>rel</sub> =(0.12~0.15)%  |      |                |  |
|    |                                                 | flicker                                                         |                                                           | P <sub>st</sub> =1,P <sub>st</sub> =3 | <i>U</i> <sub>rel</sub> =0.3%          |      |                |  |
| 66 | *Magnetic Flux<br>Meters                        | Magnetic<br>Flux                                                | C.S.for Magnetic Flux Meters<br>JJF 1905                  | 0.1mWb~10Wb                           | <i>U</i> <sub>rel</sub> =0.07%         |      |                |  |
| 67 | *Power<br>amplifier/preampli<br>fier            | Rated output<br>power                                           | C.S.for RF & Microwave<br>Power Amplifiers JJF 1678       | 20dBm~<br>57dBm,(10kHz~1GHz)          | <i>U</i> =1.5dB                        |      |                |  |
|    |                                                 |                                                                 |                                                           | 20dBm~<br>54dBm,(1GHz~18GHz)          | <i>U</i> =1.3dB                        |      |                |  |
|    |                                                 |                                                                 |                                                           | 20dBm~<br>37dBm,(18GHz~<br>40GHz)     | <i>U</i> =1.5dB                        |      |                |  |
|    |                                                 | gain                                                            |                                                           | (0~80)dB,(10kHz~<br>40GHz)            | <i>U</i> =0.8dB~1.3dB                  |      |                |  |
|    |                                                 | Gain flatness                                                   |                                                           | (0.1~12)dB,(10kHz~<br>40GHz)          | <i>U</i> =0.8dB~1.3dB                  |      |                |  |




No. CNAS L0730

第 67 页 共 163 页

The scope of the accreditation in Chinese remains the definitive version.

| № | Instrument | Measurand                                            | Calibration Method                                                                                                                                                                           | Range                             | Expanded Uncertainty<br>( <i>k</i> =2) | Note                                                                                                                                                                        | Effective Date |
|---|------------|------------------------------------------------------|----------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|-----------------------------------|----------------------------------------|-----------------------------------------------------------------------------------------------------------------------------------------------------------------------------|----------------|
|   |            | 1dB<br>compression<br>point output<br>power          | <br>CNAS NATIONAL ACCREDITATION SERVICE FOR CONFORMITY ASSESSMENT<br>SCHEDULE OF ACCREDITATION CERTIFICATE | 20dBm~<br>57dBm,(10kHz~1GHz)      | <i>U</i> =1.5dB                        | <br> |                |
|   |            |                                                      |                                                                                                                                                                                              | 20dBm~<br>54dBm,(1GHz~18GHz)      | <i>U</i> =1.3dB                        |                                                                                                                                                                             |                |
|   |            |                                                      |                                                                                                                                                                                              | 20dBm~<br>37dBm,(18GHz~<br>40GHz) | <i>U</i> =1.5dB                        |                                                                                                                                                                             |                |
|   |            | Maximum<br>output power                              |                                                                                                                                                                                              | 20dBm~<br>57dBm,(10kHz~1GHz)      | <i>U</i> =1.5dB                        |                                                                                                                                                                             |                |
|   |            |                                                      |                                                                                                                                                                                              | 20dBm~<br>54dBm,(1GHz~18GHz)      | <i>U</i> =1.3dB                        |                                                                                                                                                                             |                |
|   |            |                                                      |                                                                                                                                                                                              | 20dBm~<br>37dBm,(18GHz~<br>40GHz) | <i>U</i> =1.5dB                        |                                                                                                                                                                             |                |
|   |            | gain<br>adjustment<br>range                          |                                                                                                                                                                                              | (0~63)dB,(10kHz~<br>1GHz)         | <i>U</i> =1.0dB                        |                                                                                                                                                                             |                |
|   |            |                                                      |                                                                                                                                                                                              | (0~55)dB,(1GHz~<br>18GHz)         | <i>U</i> =1.3dB                        |                                                                                                                                                                             |                |
|   |            |                                                      |                                                                                                                                                                                              | (0~30)dB,(18GHz~<br>40GHz)        | <i>U</i> =1.3dB                        |                                                                                                                                                                             |                |
|   |            | Harmonic<br>distortion and<br>clutter<br>suppression |                                                                                                                                                                                              | (0~140)dBc                        | <i>U</i> =1.5dB                        |                                                                                                                                                                             |                |
|   |            | Third order<br>intermodulation                       |                                                                                                                                                                                              | (0~63)dB                          | <i>U</i> =2dB                          |                                                                                                                                                                             |                |
|   |            | noise<br>coefficient                                 |                                                                                                                                                                                              | (3~<br>15)dB,(10MHz~40GHz)        | <i>U</i> =0.8dB                        |                                                                                                                                                                             |                |



No. CNAS L0730

第 68 页 共 163 页

The scope of the accreditation in Chinese remains the definitive version.

| №                                             | Instrument                          | Measurand                                | Calibration Method                                  | Range                        | Expanded Uncertainty<br>( <i>k</i> =2) | Note | Effective Date |
|-----------------------------------------------|-------------------------------------|------------------------------------------|-----------------------------------------------------|------------------------------|----------------------------------------|------|----------------|
|                                               |                                     | Input voltage<br>standing wave<br>ratio  |                                                     | (1~20),(9kHz~40GHz)          | <i>U</i> =0.05                         |      |                |
| 68                                            | *Magnetic Particle<br>Flaw Detector | Electric<br>current                      | C.S.for Magnetic Particle<br>Flaw Detector JJF 1273 | (100~2100)A                  | <i>U</i> <sub>rel</sub> =1.5%          |      |                |
| 69                                            | *Magnetic Yoke<br>Detector          | Force                                    | C.S.for Magnetic Yoke<br>Detector JJF 1458          | (20~300)N                    | <i>U</i> <sub>rel</sub> =6.0%          |      |                |
|                                               |                                     | Electric<br>current                      |                                                     | (1~1000)A                    | <i>U</i> <sub>rel</sub> =6.0%          |      |                |
| 7 Radio & Time-frequency Measuring Instrument |                                     |                                          |                                                     |                              |                                        |      |                |
| 1                                             | *Wow Flutter<br>Meter               | Wow and<br>Flutter                       | C.S. for Wow Flutter Meter<br>JJF 1683              | (0.001~3)%, (2.5~<br>3.5)kHz | <i>U</i> <sub>rel</sub> =0.8%          |      |                |
|                                               |                                     | Frequency                                |                                                     | (2.5~3.5)kHz                 | <i>U</i> <sub>rel</sub> =0.005%        |      |                |
|                                               |                                     | Voltage                                  |                                                     | (0.1~10)V, (2.5~<br>3.5)kHz  | <i>U</i> <sub>rel</sub> =(5.1~0.9)%    |      |                |
| 2                                             | *Analogue<br>Oscilloscope           | Coefficient of<br>vertical<br>deflection | V.R. of Analogue<br>Oscilloscope JJG 262            | 1mV/div~5V/div               | <i>U</i> <sub>rel</sub> =(0.7~0.5)%    |      |                |
|                                               |                                     | Scanning time<br>coefficient             |                                                     | 2ns/div~5s/div               | <i>U</i> <sub>rel</sub> =( 0.25~0.5)%  |      |                |
|                                               |                                     | Bandwidth                                |                                                     | 50kHz~500MHz                 | <i>U</i> <sub>rel</sub> =( 1.9~5.9)%   |      |                |
|                                               |                                     | Raise time                               |                                                     | 300ps~100 μ s                | <i>U</i> <sub>rel</sub> =4.6%          |      |                |
|                                               |                                     | Calibrate<br>Voltage                     |                                                     | (0.01~10)V                   | <i>U</i> <sub>rel</sub> =(1.5~0.5)%    |      |                |
|                                               |                                     | Calibrate<br>Frequency                   |                                                     | (0.01~100)kHz                | <i>U</i> <sub>rel</sub> =0.003%        |      |                |
|                                               |                                     | Resistance                               |                                                     | 50 Ω ~1M Ω                   | <i>U</i> <sub>rel</sub> =(0.16~0.22)%  |      |                |
| 3                                             | *Low Frequency                      | Frequency                                | V.R. of Low-Frequency                               | 10Hz~200kHz                  | <i>U</i> <sub>rel</sub> =(0.9~0.3)%    |      |                |



在线扫码获取验证

No. CNAS L0730

第 69 页 共 163 页

The scope of the accreditation in Chinese remains the definitive version.

| № | Instrument        | Measurand                | Calibration Method                     | Range                                                 | Expanded Uncertainty<br>( $k=2$ ) | Note | Effective Date |
|---|-------------------|--------------------------|----------------------------------------|-------------------------------------------------------|-----------------------------------|------|----------------|
|   | Signal Generator  | Voltage                  | Signal Generator JJG 602               | 1mV~300V(10Hz~1kHz)                                   | $U_{rel}=(6\sim0.20)\%$           |      |                |
|   |                   |                          |                                        | 1mV~1V(1kHz~200kHz)                                   | $U_{rel}=(6\sim0.20)\%$           |      |                |
|   |                   |                          |                                        | 1V~300V(1kHz~200kHz)                                  | $U_{rel}=(0.20\sim11)\%$          |      |                |
|   |                   | Output rating            |                                        | 1W~20W                                                | $U_{rel}=(0.4\sim0.3)\%$          |      |                |
|   |                   | Attenuation              |                                        | (0~80)dB, (20Hz~200kHz)                               | $U=(0.6\sim0.20)$ dB              |      |                |
|   |                   | Distortion               |                                        | (0.002~10)%, (20Hz~100kHz)                            | $U_{rel}=15\%$                    |      |                |
| 4 | *Signal Generator | Frequency                | C.S. for Signal Generators<br>JJF 1931 | 5kHz~50GHz                                            | $U_{rel}=3\times10^{-9}$          |      |                |
|   |                   | Power                    |                                        | (-127~+30)dBm,<br>(5kHz~50GHz)                        | $U=(0.6\sim0.21)$ dB              |      |                |
|   |                   | Harmonic                 |                                        | (-100~0)dBc, (5kHz~50GHz)                             | $U=(1.2\sim2.9)$ dB               |      |                |
|   |                   | AM                       |                                        | (5~99)%, (CW:<br>150kHz~50GHz, fm<br>: 20Hz~200kHz)   | $U_{rel}=(10\sim1.3)\%$           |      |                |
|   |                   | FM                       |                                        | (1~400)kHz,<br>(CW:150kHz~50GHz,<br>fm : 20Hz~200kHz) | $U_{rel}=(10\sim1.3)\%$           |      |                |
|   |                   | Int modulation frequency |                                        | 20Hz~100kHz                                           | $U_{rel}=1\times10^{-5}$          |      |                |
|   |                   | Distortion               |                                        | (0.01~10)%, (20Hz~100kHz)                             | $U_{rel}=15\%$                    |      |                |



No. CNAS L0730

第 70 页 共 163 页

The scope of the accreditation in Chinese remains the definitive version.

| No | Instrument | Measurand                                      | Calibration Method | Range                                                 | Expanded Uncertainty<br>( $k=2$ ) | Note | Effective Date |
|----|------------|------------------------------------------------|--------------------|-------------------------------------------------------|-----------------------------------|------|----------------|
|    |            | Voltage                                        |                    | 1mV~10V, (3Hz~1kHz)                                   | $U_{rel}=(6\sim0.20)\%$           |      |                |
|    |            |                                                |                    | 100mV~1V, (20Hz~100kHz)                               | $U_{rel}=(0.20\sim1.0)\%$         |      |                |
|    |            |                                                |                    | 1V~10V, (20Hz~100kHz)                                 | $U_{rel}=(0.20\sim11)\%$          |      |                |
|    |            | Nonharmonic<br>s                               |                    | (-100~0)dBc, (5kHz~50GHz)                             | $U=(1.2\sim2.9)$ dB               |      |                |
|    |            | Pulse<br>modulation<br>on-off ratio            |                    | (0~120)dB,(5kHz~50GHz)                                | $U=0.8$ dB                        |      |                |
|    |            | Pulse<br>modulation<br>rise / fall time        |                    | 50ps~1 $\mu$ s                                        | $U_{rel}=5.0\%$                   |      |                |
|    |            | SSB phase<br>noise                             |                    | (-30~-140)dBc/Hz,offset:10Hz~100MHz                   | $U=0.9$ dB                        |      |                |
|    |            | Carrier<br>residual<br>amplitude<br>modulation |                    | (0~10)%, (CW:150kHz~50GHz)                            | $U=0.1\%$                         |      |                |
|    |            | Carrier<br>residual<br>frequency<br>modulation |                    | (0~10)kHz, (CW:150kHz~50GHz)                          | $U=1$ Hz                          |      |                |
|    |            | Phase<br>deviation                             |                    | (1~400)rad,<br>(CW:150kHz~40GHz,<br>fm : 20Hz~200kHz) | $U_{rel}=(10\sim1.3)\%$           |      |                |



No. CNAS L0730

第 71 页 共 163 页

The scope of the accreditation in Chinese remains the definitive version.

| № | Instrument        | Measurand                                               | Calibration Method                         | Range                       | Expanded Uncertainty<br>( <i>k</i> =2) | Note | Effective Date                      |
|---|-------------------|---------------------------------------------------------|--------------------------------------------|-----------------------------|----------------------------------------|------|-------------------------------------|
|   |                   | Adjoint frequency modulation under amplitude modulation |                                            | (0~10)kHz,(CW:150kHz~50GHz) | <i>U</i> =1Hz                          |      |                                     |
|   |                   | Adjoint amplitude modulation with frequency modulation  |                                            | (0~10)%,(CW:150kHz~50GHz)   | <i>U</i> =0.1%                         |      |                                     |
| 5 | *LF Voltmeter     | AC Voltage                                              | C.S. for Low-frequency Voltmeters JJF 1925 | 1mV~330V, (1kHz)            | <i>U</i> <sub>rel</sub> =(2.5~0.5)%    |      |                                     |
|   |                   |                                                         |                                            | 1V, (10Hz~1kHz)             | <i>U</i> <sub>rel</sub> =0.5%          |      |                                     |
|   |                   |                                                         |                                            | 1V, (1kHz~500kHz)           | <i>U</i> <sub>rel</sub> =(0.5~1.0)%    |      |                                     |
| 6 | *Distortion Meter | Distortion measurement                                  | C.S.for Distortion Meters JJF 1852         | (0.01~0.03)%, (5Hz~20kHz)   | <i>U</i> <sub>rel</sub> =(6~2.6)%      |      |                                     |
|   |                   | Residual distortion                                     |                                            | (0.03~30)% , (5Hz~100kHz)   | <i>U</i> <sub>rel</sub> =(2.2~1.2)%    |      |                                     |
|   |                   |                                                         |                                            | Voltage measurement         | (0.001%~0.03%), (5Hz~200kHz)           |      | <i>U</i> <sub>rel</sub> =(2.6~6.0)% |
|   |                   | 1mV~330V, (1kHz)                                        |                                            |                             | <i>U</i> <sub>rel</sub> =(0.5~2.5)%    |      |                                     |
|   |                   | 1V, (10Hz~1kHz)                                         |                                            |                             | <i>U</i> <sub>rel</sub> =0.5%          |      |                                     |
|   |                   | 1V, (1kHz~500kHz)                                       |                                            |                             | <i>U</i> <sub>rel</sub> =(0.5~1.0)%    |      |                                     |



No. CNAS L0730

第 72 页 共 163 页

The scope of the accreditation in Chinese remains the definitive version.

| № | Instrument         | Measurand                                 | Calibration Method                     | Range                                       | Expanded Uncertainty<br>( $k=2$ )                | Note | Effective Date |
|---|--------------------|-------------------------------------------|----------------------------------------|---------------------------------------------|--------------------------------------------------|------|----------------|
|   |                    | Voltage measurement<br>Frequency response |                                        | $\pm (0.01\% \sim 20\%)$ ,<br>(1kHz~500kHz) | $U_{rel}=(0.5 \sim 2.5)\%$                       |      |                |
| 7 | *Spectrum Analyzer | Frequency                                 | C.S. for Spectrum Analyzer<br>JJF 1396 | 3Hz~50GHz                                   | $U_{rel}=1 \times 10^{-5} \sim 3 \times 10^{-9}$ |      |                |
|   |                    | Sweep width                               |                                        | 10Hz~50GHz                                  | $U_{rel}=0.01\%$                                 |      |                |
|   |                    | RBW                                       |                                        | 1Hz~10MHz                                   | $U_{rel}=0.01\%$                                 |      |                |
|   |                    | Vertical display scale                    |                                        | (-90~0)dB, (3Hz~50GHz)                      | $U=0.3\text{dB}$                                 |      |                |
|   |                    | Reference level                           |                                        | (-90~+10)dBm                                | $U=(1.0 \sim 0.27)\text{dB}$                     |      |                |
|   |                    | Input attenuator conversion effect        |                                        | (-10~10) dB, (0~100)dB                      | $U=0.30\text{dB}$                                |      |                |
|   |                    | Sweep Time                                |                                        | 1ms~1000s                                   | $U_{rel}=1.6 \times 10^{-7}$                     |      |                |
|   |                    | Resolution bandwidth conversion effects   |                                        | (-20dB~20dB),(3Hz~40GHz)                    | $U=0.2\text{dB}$                                 |      |                |
|   |                    | DANL                                      |                                        | (-160dBm~20dBm),<br>(9kHz~50GHz)            | $U=1.0\text{dB}$                                 |      |                |
|   |                    | Input frequency response                  |                                        | (-20dB~20dB),(3Hz~50GHz)                    | $U=(0.2 \sim 1.0)\text{dB}$                      |      |                |
|   |                    | Absolute level                            |                                        | (-50dBm~0dBm),(10MHz~50GHz)                 | $U=0.20\text{dB}$                                |      |                |



No. CNAS L0730

第 73 页 共 163 页

The scope of the accreditation in Chinese remains the definitive version.

| № | Instrument          | Measurand                              | Calibration Method          | Range                            | Expanded Uncertainty<br>( $k=2$ )                       | Note | Effective Date |
|---|---------------------|----------------------------------------|-----------------------------|----------------------------------|---------------------------------------------------------|------|----------------|
|   |                     | Noise Sideband                         | ilac-M                      | (-150~-20) dBc/Hz, (10MHz~50GHz) | $U=1.5\text{dB}$                                        |      |                |
|   |                     | Reference frequency                    |                             | 10MHz                            | $U_{\text{rel}}=1 \times 10^{-9}$                       |      |                |
|   |                     | Calibration signal level               |                             | (-30dBm~0dBm), (10MHz~500MHz)    | $U=0.30\text{dB}$                                       |      |                |
|   |                     | Residual Response                      |                             | (-130~-70)dBm, (100kHz~50GHz)    | $U=3\text{dB}$                                          |      |                |
|   |                     | harmonic distortion                    |                             | (-120~-20)dBc, (9kHz~25GHz)      | $U=(0.6\sim2.5)\text{dB}$                               |      |                |
|   |                     | Third-order intermodulation distortion |                             | (0~25)dBm, (10MHz~50GHz)         | $U=(0.6\sim2.5)\text{dB}$                               |      |                |
|   |                     | VSWR                                   |                             | .001~5.0, (10MHz~50GHz)          | $U_{\text{rel}}=(2.5\sim6)\%$                           |      |                |
|   |                     | image response                         |                             | (-130~-60)dBc, (10MHz~50GHz)     | $U=3\text{dB}$                                          |      |                |
|   |                     | gain compression                       |                             | (0~1)dB, (10MHz~50GHz)           | $U=0.10\text{dB}$                                       |      |                |
|   |                     | power bandwidth                        |                             | (-1~1)dB, (1Hz~10MHz)            | $U=0.10\text{dB}$                                       |      |                |
|   |                     | residual FM                            |                             | 0.2Hz~500Hz, (3Hz~50GHz)         | $U=2\text{Hz}$                                          |      |                |
|   |                     | frequency count                        |                             | 3Hz~50GHz                        | $U_{\text{rel}}=1 \times 10^{-5} \sim 3 \times 10^{-9}$ |      |                |
| 8 | *Function Generator | Frequency                              | V.R. of Function Generators | 0.1Hz~250MHz                     | $U_{\text{rel}}=1 \times 10^{-4} \sim 3 \times 10^{-9}$ |      |                |
|   |                     | ACV                                    | JJG 840                     | 5mV~55V, (1kHz)                  | $U_{\text{rel}}=(0.5\sim0.20)\%$                        |      |                |



No. CNAS L0730

第 74 页 共 163 页

The scope of the accreditation in Chinese remains the definitive version.

| №  | Instrument               | Measurand                              | Calibration Method                                     | Range                                                    | Expanded Uncertainty<br>( $k=2$ ) | Note | Effective Date |
|----|--------------------------|----------------------------------------|--------------------------------------------------------|----------------------------------------------------------|-----------------------------------|------|----------------|
|    |                          | THD                                    | ilac-M                                                 | (0.001~5)%, (20Hz~20kHz)                                 | $U_{rel}=15\%$                    |      |                |
|    |                          | Harmonic                               |                                                        | (-100~-10) dBc,<br>(100Hz~1GHz)                          | $U_{rel}=(0.4\sim3.7)\text{dB}$   |      |                |
|    |                          | Flatness of<br>sine wave<br>amplitude  |                                                        | $\pm (0.01\sim2)\text{dB}$ ,<br>(1kHz~250MHz, -60~20)dBm | $U=(0.15\sim0.6)\text{dB}$        |      |                |
|    |                          | Raise time                             |                                                        | 200ps~100 $\mu\text{s}$ , (1Hz~250MHz)                   | $U_{rel}=1.8\%$                   |      |                |
|    |                          | Pulse duration<br>ratio                |                                                        | (1~99)%, (1Hz~250MHz)                                    | $U=0.2\%$                         |      |                |
|    |                          | DCV                                    |                                                        | (-20~-0.01) V,<br>(0.01~20) V                            | $U_{rel}=0.20\%$                  |      |                |
| 9  | *Frequency<br>Sweeper    | Scope vertical<br>input<br>Sensitivity | V.R. of MSW-7124<br>AM/FM Sweeper JJG(Dianzi)<br>07003 | 1mV~1V                                                   | $U_{rel}=(4\sim2.0)\%$            |      |                |
|    |                          | CF range                               |                                                        | (0.2 ~100)MHz                                            | $U_{rel}=0.0013\%$                |      |                |
|    |                          | Sweep width                            |                                                        | (0.2~36) MHz                                             | $U_{rel}=0.0013\%$                |      |                |
|    |                          | Frequency<br>accuracy of<br>the marker |                                                        | 100kHz ~110MHz                                           | $U_{rel}=0.0013\%$                |      |                |
|    |                          | Sweep output<br>level                  |                                                        | (-87~-7)dBm                                              | $U=0.5\text{dB}$                  |      |                |
| 10 | *Digital<br>Oscilloscope | DC gain                                | V.R. for Digital Oscilloscope<br>GJB 7691              | (0.8~1.2)1mV/div~<br>50V/div                             | $U_{rel}=(0.7\sim0.3)\%$          |      |                |
|    |                          | Bandwidth                              |                                                        | 50kHz~50GHz                                              | $U_{rel}=(1.9\sim5.9)\%$          |      |                |
|    |                          | Raise time                             |                                                        | 20ps~100 $\mu\text{s}$                                   | $U_{rel}=4.6\%$                   |      |                |



No. CNAS L0730

第 75 页 共 163 页

The scope of the accreditation in Chinese remains the definitive version.

| №  | Instrument                          | Measurand                                | Calibration Method                                                                                                        | Range                             | Expanded Uncertainty<br>( $k=2$ ) | Note                     | Effective Date |
|----|-------------------------------------|------------------------------------------|---------------------------------------------------------------------------------------------------------------------------|-----------------------------------|-----------------------------------|--------------------------|----------------|
|    |                                     | Calibrate Voltage                        | ilac-MEAS CNAS<br>CHINA NATIONAL ACCREDITATION SERVICE FOR CONFORMITY ASSESSMENT<br>SCHEDULE OF ACCREDITATION CERTIFICATE | (0.01~5)V                         | $U_{rel}=(1.5\sim0.5)\%$          |                          |                |
|    |                                     | Calibrate Frequency                      |                                                                                                                           | 0.5kHz~5MHz                       | $U_{rel}=0.003\%$                 |                          |                |
|    |                                     | Resistance                               |                                                                                                                           | 50 $\Omega$ , 1M $\Omega$         | $U_{rel}=(0.16\sim0.22)\%$        |                          |                |
|    |                                     | Channel isolation                        |                                                                                                                           | 200:1~20:1                        | $U_{rel}=1\%$                     |                          |                |
|    |                                     | Input capacitance                        |                                                                                                                           | 10pF~25pF                         | $U_{rel}=2\%$                     |                          |                |
|    |                                     | Probe decay ratio                        |                                                                                                                           | 1:1~1:1000                        | $U_{rel}=0.3\%$                   |                          |                |
|    |                                     | Delay time difference between channels   |                                                                                                                           | 30ps~5ns                          | $U_{rel}=0.3\%$                   |                          |                |
|    |                                     | Time base                                |                                                                                                                           | 10MHz                             | $U_{rel}=1\times 10^{-7}$         |                          |                |
|    |                                     | Background noise                         |                                                                                                                           | 0.1mV~1V                          | $U_{rel}=2\%$                     |                          |                |
|    |                                     | Overshoot                                |                                                                                                                           | 0.5%~30%                          | $U_{rel}=2\%$                     |                          |                |
|    |                                     | Trigger sensitivity                      |                                                                                                                           | 0.1mV~1V                          | $U_{rel}=3.0\%$                   |                          |                |
|    |                                     | DC Bias                                  |                                                                                                                           | $\pm(1\text{mV}\sim 200\text{V})$ | $U_{rel}=(0.7\sim0.3)\%$          |                          |                |
| 11 | *Semiconductor Device Curve Tracers | Base voltage deflection coefficient      | C.S. for Semiconductor Device Curve Tracers JJF 1236                                                                      | 0.05V/div~2V/div                  | $U_{rel}=1.5\%$                   | 合格评定 国家认可 委员会<br>认可证书专用章 |                |
|    |                                     | collector voltage deflection coefficient |                                                                                                                           | 0.05V/div~50V/div                 | $U_{rel}=1.5\%$                   |                          |                |



在线扫码获取验证

No. CNAS L0730

第 76 页 共 163 页

The scope of the accreditation in Chinese remains the definitive version.

| №  | Instrument        | Measurand                            | Calibration Method                                                                                                 | Range                                           | Expanded Uncertainty<br>( $k=2$ )                  | Note                        | Effective Date |
|----|-------------------|--------------------------------------|--------------------------------------------------------------------------------------------------------------------|-------------------------------------------------|----------------------------------------------------|-----------------------------|----------------|
|    |                   | diode voltage deflection coefficient | ilac-MRA<br>INTERNATIONAL ACCREDITATION SERVICE FOR CONFORMITY ASSESSMENT<br>SCHEDULE OF ACCREDITATION CERTIFICATE | 100V/div ~ 500V/div                             | $U_{rel}=1.5\%$                                    |                             |                |
|    |                   | circumscribed deflection coefficient |                                                                                                                    | 0.05V/div ~ 50V/div                             | $U_{rel}=1.5\%$                                    |                             |                |
|    |                   | step voltage                         |                                                                                                                    | 0.05V/div ~ 2V/div                              | $U_{rel}=1.5\%$                                    |                             |                |
|    |                   | collector current deflection factor  |                                                                                                                    | 2 $\mu$ A/div ~ 10A/div                         | $U_{rel}=1.5\%$                                    |                             |                |
|    |                   | current deflection factor            |                                                                                                                    | 2 $\mu$ A/div ~ 10A/div                         | $U_{rel}=1.5\%$                                    |                             |                |
|    |                   | Step current                         |                                                                                                                    | 2 $\mu$ A/div ~ 10A/div                         | $U_{rel}=1.5\%$                                    |                             |                |
|    |                   | Resistance                           |                                                                                                                    | 0.1 $\Omega$ ~ 10M $\Omega$                     | $U_{rel}=1.5\%$                                    |                             |                |
| 12 | *Modulation Meter | AM                                   | C.S. for Modulation Meters<br>JJF 1111                                                                             | (5~99)% (载波: 150kHz~1.3GHz)(调制频率:20Hz~20kHz)    | $U_{rel}=(10\sim1.3)\%$                            | 合格评定国家认可<br>CNAS<br>认可证书专用章 |                |
|    |                   | FM offset                            |                                                                                                                    | (1~400)kHz (载波: 150kHz~1.3GHz)(调制频率:20Hz~20kHz) | $U_{rel}=(10\sim1.3)\%$                            |                             |                |
|    |                   | PM offset                            |                                                                                                                    | (1~400)rad (载波: 150kHz~1.3GHz)(调制频率:20Hz~20kHz) | $U_{rel}=(10\sim1.3)\%$                            |                             |                |
| 13 | *HF Q-Meter       | Frequency                            | C.S. of HF Q-Meter JJF 1073                                                                                        | 50 kHz ~ 50 MHz                                 | $U_{rel}=0.8 \times 10^{-5} \sim 4 \times 10^{-5}$ |                             |                |



No. CNAS L0730

第 77 页 共 163 页

The scope of the accreditation in Chinese remains the definitive version.

| №  | Instrument                                        | Measurand                       | Calibration Method                                                                      | Range                    | Expanded Uncertainty<br>( $k=2$ ) | Note | Effective Date |
|----|---------------------------------------------------|---------------------------------|-----------------------------------------------------------------------------------------|--------------------------|-----------------------------------|------|----------------|
|    |                                                   | Q Value                         |                                                                                         | 10~500                   | $U=3.2\sim 7.4$                   |      |                |
|    |                                                   | Capacity                        |                                                                                         | (8~500)pF                | $U=(0.09\sim 1.0)\text{ pF}$      |      |                |
| 14 | *Telephone Analyze                                | Dual tone multiple Frequency    | V.R. of Dual Tone and Multi-frequency Telephone Audio meters JJG(Youdian) 032           | (600~1700)Hz             | $U=0.20\text{Hz}$                 |      |                |
|    |                                                   | Dual tone multi frequency level |                                                                                         | (-30~0)dBm, (600~1700)Hz | $U=0.2\text{dB}$                  |      |                |
|    |                                                   | Pulse Width                     |                                                                                         | (1~150) ms               | $U=0.1\text{ms}$                  |      |                |
|    |                                                   | Pulse Ratio                     |                                                                                         | (0.1~15)                 | $U_{\text{rel}}=0.09\%$           |      |                |
| 15 | *DC and LF Characterization Tester for Transistor | DCV                             | C.S. for Semiconductor Devices DC and Low Frequency Parameters Test Equipments JJF 1895 | (0.1~1000)V              | $U_{\text{rel}}=0.4\%$            |      |                |
|    |                                                   | DCA                             |                                                                                         | 10 $\mu\text{A}$ ~1A     | $U_{\text{rel}}=0.7\%$            |      |                |
|    |                                                   | hfe                             |                                                                                         | (20~999)                 | $U_{\text{rel}}=(0.6\sim 1.1)\%$  |      |                |
| 16 | *Microwave Counter                                | Frequency                       | V.R. of Microwave Frequency Counters JJG 841                                            | 100kHz~40GHz             | $U_{\text{rel}}=6\times 10^{-10}$ |      |                |
|    |                                                   | Power-on characteristic         |                                                                                         | (1~10)MHz                | $U_{\text{rel}}=6\times 10^{-10}$ |      |                |
|    |                                                   | Daily frequency fluctuation     |                                                                                         | (1~10)MHz                | $U_{\text{rel}}=6\times 10^{-10}$ |      |                |
|    |                                                   | Daily aging rate                |                                                                                         | (1~10)MHz                | $U_{\text{rel}}=6\times 10^{-10}$ |      |                |
|    |                                                   | 1s frequency stability          |                                                                                         | (1~10)MHz                | $U_{\text{rel}}=6\times 10^{-10}$ |      |                |



No. CNAS L0730

第 78 页 共 163 页

The scope of the accreditation in Chinese remains the definitive version.

| №  | Instrument         | Measurand                   | Calibration Method                    | Range                                                 | Expanded Uncertainty<br>( $k=2$ ) | Note | Effective Date |
|----|--------------------|-----------------------------|---------------------------------------|-------------------------------------------------------|-----------------------------------|------|----------------|
|    |                    | Frequency reproducibility   |                                       | (1~10)MHz                                             | $U_{rel}=6 \times 10^{-10}$       |      |                |
|    |                    | Frequency accuracy          |                                       | (1~10)MHz                                             | $U_{rel}=6 \times 10^{-10}$       |      |                |
|    |                    | Input Sensitivity           |                                       | (-40~-10) dBm,<br>100kHz~40GHz                        | $U=0.6\text{dB}$                  |      |                |
| 17 | *Universal counter | Frequency                   | V.R. of Universal Counters<br>JJG 349 | 1Hz~18GHz                                             | $U_{rel}=6 \times 10^{-10}$       |      |                |
|    |                    | Period                      |                                       | 1ns~100s                                              | $U_{rel}=6 \times 10^{-10}$       |      |                |
|    |                    | Time interval measurement   |                                       | 100ns~4000s                                           | $U_{rel}=0.01\%$                  |      |                |
|    |                    | Power-on characteristic     |                                       | (1~10)MHz                                             | $U_{rel}=6 \times 10^{-10}$       |      |                |
|    |                    | Daily frequency fluctuation |                                       | (1~10)MHz                                             | $U_{rel}=6 \times 10^{-10}$       |      |                |
|    |                    | Daily aging rate            |                                       | (1~10)MHz                                             | $U_{rel}=6 \times 10^{-10}$       |      |                |
|    |                    | 1s frequency stability      |                                       | (1~10)MHz                                             | $U_{rel}=6 \times 10^{-10}$       |      |                |
|    |                    | Frequency reproducibility   |                                       | (1~10)MHz                                             | $U_{rel}=6 \times 10^{-10}$       |      |                |
|    |                    | Frequency accuracy          |                                       | (1~10)MHz                                             | $U_{rel}=6 \times 10^{-10}$       |      |                |
|    |                    | Input Sensitivity           |                                       | (-40~-10) dBm,<br>(1GHz~18GHz)<br>5mV~1V, (10Hz~1GHz) | $U=0.6\text{dB}$<br>$U_{rel}=5\%$ |      |                |



No. CNAS L0730

第 79 页 共 163 页

The scope of the accreditation in Chinese remains the definitive version.

| №  | Instrument                 | Measurand                             | Calibration Method                                                                           | Range                            | Expanded Uncertainty<br>( $k=2$ )  | Note | Effective Date |
|----|----------------------------|---------------------------------------|----------------------------------------------------------------------------------------------|----------------------------------|------------------------------------|------|----------------|
| 18 | *Frequency Meters          | Frequency                             | V.R. of Frequency Meters<br>JJG 603                                                          | 10 Hz~20 kHz                     | $U_{\text{rel}}=0.003\%$           |      |                |
| 19 | *Crystal Oscillator        | Frequency accuracy                    | C.S. for Crystal Oscillators<br>inside the Electrical<br>Measurement Instruments JJF<br>1984 | (1~100)MHz                       | $U_{\text{rel}}=6 \times 10^{-10}$ |      |                |
|    |                            | Power-on characteristic               |                                                                                              | (1~100)MHz                       | $U_{\text{rel}}=6 \times 10^{-10}$ |      |                |
|    |                            | Daily aging rate                      |                                                                                              | (1~100)MHz                       | $U_{\text{rel}}=6 \times 10^{-10}$ |      |                |
|    |                            | 1s frequency stability                |                                                                                              | (1~100)MHz                       | $U_{\text{rel}}=6 \times 10^{-10}$ |      |                |
|    |                            | Frequency reproducibility             |                                                                                              | (1~100)MHz                       | $U_{\text{rel}}=6 \times 10^{-10}$ |      |                |
| 20 | *Vector network<br>analyze | Internal Source Frequency             | C.S. for Vector Network<br>Analyzers JJF 1495                                                | 9kHz~50 GHz                      | $U_{\text{rel}}=1 \times 10^{-8}$  |      |                |
|    |                            | Internal Signal Source Power          |                                                                                              | (-110~20)dBm,<br>(9kHz~50 GHz)   | $U=(0.27 \sim 0.72)\text{dB}$      |      |                |
|    |                            | Dynamic Accuracy of Modulus Value     |                                                                                              | (0~90)dB, (1MHz~40 GHz)          | $U=(0.06 \sim 1.0)\text{dB}$       |      |                |
|    |                            | Modulus of Reflection                 |                                                                                              | (0~1), (0.05~18)GHz              | $U=(0.012 \sim 0.018)$             |      |                |
|    |                            | Phase Angle of reflection coefficient |                                                                                              | (-180° ~ 180° ),<br>(0.05~18)GHz | $U=(0.14 \sim 2.8)^\circ$          |      |                |
|    |                            | Scattering Parameter Modulus Sii      |                                                                                              | (0~1), (0.05~50)GHz              | $U=(0.01 \sim 0.03)$               |      |                |



No. CNAS L0730

第 80 页 共 163 页

The scope of the accreditation in Chinese remains the definitive version.

| №  | Instrument                 | Measurand                                                 | Calibration Method                          | Range                          | Expanded Uncertainty<br>( $k=2$ ) | Note | Effective Date |
|----|----------------------------|-----------------------------------------------------------|---------------------------------------------|--------------------------------|-----------------------------------|------|----------------|
|    |                            | Scattering Parameter Modulus $S_{ij}$                     | ilac-MRA                                    | (0~70) dB, (0.05~50)GHz        | $U=0.20\text{dB}$                 |      |                |
|    |                            | Measurement Accuracy Phase Angle of Scattering Parameters |                                             | (-180° ~ 180° ), (0.05~50)GHz  | $U=(0.14\sim 2.8)^\circ$          |      |                |
|    |                            | Trace Noise Magnitude                                     |                                             | (0~0.1) dB, (9kHz~50 GHz)      | $U=0.0013\text{dB}$               |      |                |
|    |                            | Trace Noise Phase                                         |                                             | (0~1) ° , (9kHz~50 GHz)        | $U=0.012^\circ$                   |      |                |
|    |                            | Noise Floor                                               |                                             | (-140~-60) dBm, (9kHz~50 GHz)  | $U=2.3\text{ dB}$                 |      |                |
|    |                            | Crosstalk                                                 |                                             | (-150~-70) dB, (9kHz~50 GHz)   | $U=1.1\text{ dB}$                 |      |                |
|    |                            | Modulus of Calibration element reflection factor          |                                             | (0~1), (0.01~50)GHz            | $U=(0.01\sim 0.03)$               |      |                |
|    |                            | phase of calibration element reflection factor            |                                             | (-180° ~ 180° ), (0.01~50)GHz  | $U=(0.4\sim 1.2)^\circ$           |      |                |
| 21 | *Artificial Mains Networks | Impedance                                                 | C.S. for Artificial Mains Networks JJF 1705 | (1~60) $\Omega$ ,(9kHz~110MHz) | $U_{\text{rel}}=7\%$              |      |                |
|    |                            | Phase                                                     |                                             | (-20° ~ 90° ) ,(9kHz~110MHz)   | $U=2.6^\circ$                     |      |                |



No. CNAS L0730

第 81 页 共 163 页

The scope of the accreditation in Chinese remains the definitive version.

| №  | Instrument               | Measurand                               | Calibration Method              | Range                             | Expanded Uncertainty<br>( $k=2$ )   | Note | Effective Date |
|----|--------------------------|-----------------------------------------|---------------------------------|-----------------------------------|-------------------------------------|------|----------------|
|    |                          | Voltage division factor                 |                                 | (-20~0)dB, (9kHz~110MHz)          | $U=0.34\text{dB}$                   |      |                |
| 22 | *Oscilloscope Calibrator | DC Calibration voltage                  | Oscilloscope Calibrator JJG 278 | $\pm(1\text{mV}\sim 200\text{V})$ | $U_{\text{rel}}=0.005\%\sim 0.05\%$ |      |                |
|    |                          | Square wave calibration voltage         |                                 | 1mV~200V(10Hz~10kHz)              | $U_{\text{rel}}=0.02\%\sim 0.2\%$   |      |                |
|    |                          | Time marker                             |                                 | 0.5ns~5s                          | $U_{\text{rel}}=5.2\times 10^{-8}$  |      |                |
|    |                          | Rise time                               |                                 | 50ps~100ns                        | $U_{\text{rel}}=5\%$                |      |                |
|    |                          | Flatness of Fixed amplitude sine signal |                                 | (-10~10) dB (9kHz~12.4GHz)        | $U=0.20\text{dB}$                   |      |                |
|    |                          | Sine wave frequency                     |                                 | 9kHz~12.4GHz                      | $U_{\text{rel}}=5.2\times 10^{-8}$  |      |                |
|    |                          | Waveform generator amplitude            |                                 | 2mV~50V (1M $\Omega$ )            | $U_{\text{rel}}=1\%$                |      |                |
|    |                          | Waveform generator amplitude            |                                 | 2mV~ 2.5V (50 $\Omega$ )          | $U_{\text{rel}}=1\%$                |      |                |
|    |                          | Waveform generator frequency            |                                 | 10Hz~10kHz                        | $U_{\text{rel}}=1\times 10^{-7}$    |      |                |
|    |                          | Waveform generator DC bias              |                                 | 2mV~50V (1M $\Omega$ )            | $U_{\text{rel}}=1\%$                |      |                |
|    |                          | Pulse generator cycle                   |                                 | 3Hz~50GHz                         | $U_{\text{rel}}=1\times 10^{-7}$    |      |                |



| №  | Instrument                       | Measurand                             | Calibration Method                                                     | Range                           | Expanded Uncertainty<br>( <i>k</i> =2)        | Note | Effective Date |
|----|----------------------------------|---------------------------------------|------------------------------------------------------------------------|---------------------------------|-----------------------------------------------|------|----------------|
|    |                                  | Pulse generator pulse width           |                                                                        | 4ns~500ns                       | <i>U</i> <sub>rel</sub> =1%                   |      |                |
|    |                                  | resistance measurement                |                                                                        | 10 Ω ~150 Ω                     | <i>U</i> <sub>rel</sub> =0.05%                |      |                |
|    |                                  | Capacitance measurement               |                                                                        | 50 kΩ ~12 MΩ                    | <i>U</i> <sub>rel</sub> =0.03%                |      |                |
|    |                                  | current output                        |                                                                        | 5pF~50pF                        | <i>U</i> <sub>rel</sub> =1.5                  |      |                |
|    |                                  |                                       |                                                                        | 88 μ A~111mA                    | <i>U</i> <sub>rel</sub> =0.1%                 |      |                |
| 23 | *RF Impedance/Material Analyzers | frequency                             | Calibration specification for RF Impedance/Material Analyzers JJF 1127 | 1MHz~3GHz                       | <i>U</i> <sub>rel</sub> =1×10 <sup>-7</sup>   |      |                |
|    |                                  | signal level                          |                                                                        | (-60~20) dBm,(1MHz~3GHz)        | <i>U</i> =(0.13~0.2) dB                       |      |                |
|    |                                  | impedance                             |                                                                        | 0.001mΩ ~50kΩ , (0.01 μ S~10S)  | <i>U</i> <sub>rel</sub> =0.4%~2%              |      |                |
|    |                                  | phase                                 |                                                                        | (-180~180) ° C, (1MHz~3GHz)     | <i>U</i> =(2~15)mrad                          |      |                |
| 24 | *Vector Signal Generator         | Internal crystal oscillator frequency | C.S. for Vector Signal Generators JJF 1174                             | 10MHz                           | <i>U</i> <sub>rel</sub> =6.7×10 <sup>-9</sup> |      |                |
|    |                                  | Frequency                             |                                                                        | 100 kHz~44GHz                   | <i>U</i> <sub>rel</sub> =6.7×10 <sup>-9</sup> |      |                |
|    |                                  | Level                                 |                                                                        | (0 ~ 20) dBm,(100 kHz ~2 GHz)   | <i>U</i> =0.17dB                              |      |                |
|    |                                  |                                       |                                                                        | (0 ~ 20) dBm,(2GHz ~ 18 GHz)    | <i>U</i> =0.27dB                              |      |                |
|    |                                  |                                       |                                                                        | (0 ~ 20) dBm,(18 GHz~ 26.5 GHz) | <i>U</i> =0.37dB                              |      |                |



No. CNAS L0730

第 83 页 共 163 页

The scope of the accreditation in Chinese remains the definitive version.

| № | Instrument | Measurand | Calibration Method | Range                               | Expanded Uncertainty<br>( $k=2$ ) | Note | Effective Date |
|---|------------|-----------|--------------------|-------------------------------------|-----------------------------------|------|----------------|
|   |            |           |                    | (0 ~ 20) dBm, (26.5 GHz ~ 33 GHz)   | $U=0.44\text{dB}$                 |      |                |
|   |            |           |                    | (0 ~ 20) dBm, (33 GHz ~ 44 GHz)     | $U=0.53\text{dB}$                 |      |                |
|   |            |           |                    | (-50 ~ 0) dBm, (100 kHz ~ 2 GHz)    | $U=0.20\text{dB}$                 |      |                |
|   |            |           |                    | (-50 ~ 0) dBm, (2 GHz ~ 18 GHz)     | $U= (0.28\sim0.31) \text{ dB}$    |      |                |
|   |            |           |                    | (-50 ~ 0) dBm (18 GHz ~ 26.5 GHz)   | $U= (0.45\sim0.48) \text{ dB}$    |      |                |
|   |            |           |                    | (-50 ~ 0) dBm (26.5 GHz ~ 33 GHz)   | $U= (0.45\sim0.48) \text{ dB}$    |      |                |
|   |            |           |                    | (-50 ~ 0) dBm (33 GHz ~ 40 GHz)     | $U= (0.54\sim0.57) \text{ dB}$    |      |                |
|   |            |           |                    | (-50 ~ 0) dBm (40 GHz ~ 44 GHz)     | $U= (0.58\sim0.61) \text{ dB}$    |      |                |
|   |            |           |                    | (-90 ~ -50) dBm (100 kHz ~ 2 GHz)   | $U= (0.12\sim0.14) \text{ dB}$    |      |                |
|   |            |           |                    | (-90 ~ -50) dBm (2 GHz ~ 18 GHz)    | $U= (0.16\sim0.18) \text{ dB}$    |      |                |
|   |            |           |                    | (-90 ~ -50) dBm (18 GHz ~ 26.5 GHz) | $U= (0.19\sim0.21) \text{ dB}$    |      |                |
|   |            |           |                    | (-90 ~ -50) dBm (26.5 GHz ~ 33 GHz) | $U= (0.23\sim0.25) \text{ dB}$    |      |                |
|   |            |           |                    | (-90 ~ -50) dBm (33 GHz ~ 44 GHz)   | $U= (0.27\sim0.29) \text{ dB}$    |      |                |
|   |            |           |                    | (-120 ~ -90) dBm (100 kHz ~ 2 GHz)  | $U= (0.17\sim0.18) \text{ dB}$    |      |                |



No. CNAS L0730

第 84 页 共 163 页

The scope of the accreditation in Chinese remains the definitive version.

| № | Instrument | Measurand                            | Calibration Method | Range                                                   | Expanded Uncertainty<br>( $k=2$ ) | Note | Effective Date |
|---|------------|--------------------------------------|--------------------|---------------------------------------------------------|-----------------------------------|------|----------------|
|   |            |                                      |                    | (-120 ~ -90)dBm(2<br>GHz ~ 18 GHz)                      | $U = (0.20 \sim 0.21) \text{ dB}$ |      |                |
|   |            |                                      |                    | (-120 ~ -90)dBm(18<br>GHz ~ 26.5 GHz)                   | $U = (0.23 \sim 0.24) \text{ dB}$ |      |                |
|   |            |                                      |                    | (-120 ~ -90)dBm(26.5<br>GHz ~ 33 GHz)                   | $U = (0.26 \sim 0.27) \text{ dB}$ |      |                |
|   |            |                                      |                    | (-120 ~ -90)dBm(33<br>GHz ~ 44 GHz)                     | $U = (0.30 \sim 0.31) \text{ dB}$ |      |                |
|   |            | Amplitude<br>Modulation<br>(AM)      |                    | (5%~99%)(CW:<br>100kHz~44GHz)(f:<br>10Hz~100kHz)        | $U_{\text{rel}}=1.2\%$            |      |                |
|   |            | Frequency<br>Modulation<br>deviation |                    | 10 Hz~400 kHz (CW:0.1<br>MHz~10 MHz;f:20 Hz~<br>40k Hz) | $U_{\text{rel}}=1.2\%$            |      |                |
|   |            | Phase<br>Modulation<br>deviation     |                    | (1~500)rad<br>(CW:0.1MHz~44<br>GHz;f:50Hz~100kHz)       | $U_{\text{rel}}=1.2\%$            |      |                |
|   |            | Signal<br>Spectral<br>Purity         |                    | (-100~-10)dBc(CW:0.1<br>MHz~44GHz)                      | $U=1.0\text{dB}$                  |      |                |
|   |            |                                      |                    | 20 MHz <CW≤3.6 GHz                                      | $U=0.41\text{dB}$                 |      |                |
|   |            |                                      |                    | 3.5 GHz <CW≤ 8.4<br>GHz                                 | $U=1.8 \text{ dB}$                |      |                |
|   |            |                                      |                    | 8.4 GHz <CW≤ 22 GHz                                     | $U=2.3\text{dB}$                  |      |                |
|   |            |                                      |                    | 22 GHz <CW≤ 34.5<br>GHz                                 | $U=2.9 \text{ dB}$                |      |                |
|   |            |                                      |                    | 34.4 GHz <CW≤50 GHz                                     | $U=3.7 \text{ dB}$                |      |                |



No. CNAS L0730

第 85 页 共 163 页

The scope of the accreditation in Chinese remains the definitive version.

| No | Instrument | Measurand                                         | Calibration Method | Range                                                   | Expanded Uncertainty<br>(k=2) | Note | Effective Date |
|----|------------|---------------------------------------------------|--------------------|---------------------------------------------------------|-------------------------------|------|----------------|
|    |            | Single Side Band Phase Noise                      |                    | (-140~-10) dBc/Hz(频率偏移:10Hz~2MHz)<br>(CW:0.1 MHz~44GHz) | $U=1.0$ dB                    |      |                |
|    |            | Error Vector Magnitude(EVM)                       |                    | 0%~25%,(100 kHz~43.5 GHz)                               | $U=0.5\%\sim 1.2\%$           |      |                |
|    |            | Frequency Error                                   |                    | -1.3 MHz ~ +1.3MHz(100 kHz≤CW≤43.5 GHz)                 | $U=16$ Hz                     |      |                |
|    |            |                                                   |                    | -1.3MHz ~ +1.3MHz(Bluetooth)                            | $U=12$ Hz                     |      |                |
|    |            |                                                   |                    | -1.3MHz ~ +1.3MHz(WLAN)                                 | $U=2$ Hz                      |      |                |
|    |            |                                                   |                    | -1.3MHz ~ +1.3MHz(LTE)                                  | $U=2$ Hz                      |      |                |
|    |            |                                                   |                    | -1.2 kHz ~ +1.2 kHz(CDMA)                               | $U=2$ Hz                      |      |                |
|    |            |                                                   |                    | -5 kHz ~ +5 kHz(WCDMA)                                  | $U=6$ Hz                      |      |                |
|    |            |                                                   |                    | -14 kHz ~ +14 kHz(TD-SCDMA)                             | $U=4$ Hz                      |      |                |
|    |            | Bluetooth Modulation Frequency Offset Measurement |                    | 100 kHz ~ 300 kHz(2.4GHz~2.5 GHz)                       | $U_{rel}=1.2\%$               |      |                |
|    |            | Phase Error Measurement                           |                    | 0° ~ 15° (100kHz~43.5GH)                                | $U=0.89^\circ$                |      |                |
|    |            | Symbol Clock Error(WLAN)                          |                    | $(-1.0\times 10^{-4}\sim 1.0\times 10^{-4})$            | $U=3.6\times 10^{-6}$         |      |                |



在线扫码获取验证

No. CNAS L0730

第 86 页 共 163 页

The scope of the accreditation in Chinese remains the definitive version.

| No | Instrument | Measurand                                     | Calibration Method | Range                              | Expanded Uncertainty<br>( $k=2$ ) | Note | Effective Date |
|----|------------|-----------------------------------------------|--------------------|------------------------------------|-----------------------------------|------|----------------|
|    |            | I/Q origin offset suppression (LTE)           |                    | (-80 ~ -10) dB<br>(0.6GHz~3.8GHz)  | $U=0.96$ dB                       |      |                |
|    |            | Waveform Quality Factor (CDMA、WCDMA、TD-SCDMA) |                    | 0.5~1.0 (0.8GHz~2.7GHz)            | $U=0.0012$                        |      |                |
|    |            | Channel Power                                 |                    | (-70 ~ 0) dBm(100 kHz ~ 18 GHz)    | $U=0.37$ dB                       |      |                |
|    |            |                                               |                    | (-70 ~ 0) dBm(18 GHz ~ 26.5 GHz)   | $U=1.7$ dB                        |      |                |
|    |            |                                               |                    | (-70 ~ 0) dBm(26.5 GHz ~ 43.5 GHz) | $U=2.5$ dB                        |      |                |
|    |            | Peak Code Domain Error                        |                    | (-50 ~ -25) dB<br>(0.7GHz~2.7GHz)  | $U=1.0$ dB                        |      |                |
|    |            |                                               |                    | (-60 ~ -50) dB<br>(0.7GHz~2.7GHz)  | $U=2.9$ dB                        |      |                |
|    |            | Occupied Bandwidth                            |                    | 70kHz~20MHz<br>(0.1GHz~44GHz)      | $U_{rel}=0.2\%$                   |      |                |
|    |            | ACPR                                          |                    | (-80~-20)dB(0.8GHz~2.7GHz)         | $U=0.87$ dB                       |      |                |
|    |            | Residual Amplitude Modulation of the Carrier  |                    | (0.01~0.5)%(250 kHz~44 GHz)        | $U_{rel}=2\%$                     |      |                |
|    |            | Residual Frequency                            |                    | (0.1Hz~50)Hz<br>100kHz~6.6GHz      | $U=1.8$ Hz                        |      |                |

No. CNAS L0730

第 87 页 共 163 页



The scope of the accreditation in Chinese remains the definitive version.

| No | Instrument        | Measurand                               | Calibration Method                      | Range                                | Expanded Uncertainty<br>(k=2)       | Note | Effective Date |
|----|-------------------|-----------------------------------------|-----------------------------------------|--------------------------------------|-------------------------------------|------|----------------|
|    |                   | Modulation of the Carrier               |                                         | (0.1Hz~50)Hz                         | $U=3.5\text{ Hz}$                   |      |                |
|    |                   |                                         |                                         | (6.6GHz~13.2 GHz)                    |                                     |      |                |
|    |                   |                                         |                                         | (0.1Hz~50)Hz (13.2GHz~31.15 GHz)     | $U=6.9\text{ Hz}$                   |      |                |
|    |                   |                                         |                                         | (0.1Hz~50)Hz (31.15GHz~44 GHz)       | $U=14\text{ Hz}$                    |      |                |
|    |                   | "Pulse modulation on-off ratio "        |                                         | (0~90)dB,(5kHz~50GHz)                | $U=0.8\text{ dB}$                   |      |                |
|    |                   | Pulse modulation rise and fall time     |                                         | 10ns~1 $\mu\text{s}$ , (250kHz~6GHz) | $U_{\text{rel}}=5.0\%$              |      |                |
|    |                   | Internal modulation generator frequency |                                         | 5Hz~200kHz                           | $U_{\text{rel}}=1 \times 10^{-6}$   |      |                |
|    |                   | Internal modulation generator amplitude |                                         | 10mV~10V(3Hz~1kHz)                   | $U_{\text{rel}}=(6\sim0.20)\%$      |      |                |
| 25 | *Bluetooth Tester | Internal crystal oscillator frequency   | C.S. for RF Bluetooth Test Set JJF 1278 | 10MHz                                | $U_{\text{rel}}=6.7 \times 10^{-9}$ |      |                |
|    |                   | Frequency                               |                                         | 2.4G Hz~2.5GHz                       | $U_{\text{rel}}=6.7 \times 10^{-9}$ |      |                |
|    |                   | Level                                   |                                         | (0 ~ 10) dBm(2.4GHz~2.5 GHz)         | $U=0.24\text{ dB}$                  |      |                |




No. CNAS L0730

第 88 页 共 163 页

The scope of the accreditation in Chinese remains the definitive version.

| № | Instrument | Measurand                                    | Calibration Method | Range                                                    | Expanded Uncertainty<br>( $k=2$ ) | Note | Effective Date |
|---|------------|----------------------------------------------|--------------------|----------------------------------------------------------|-----------------------------------|------|----------------|
|   |            |                                              |                    | (-50 ~ 0) dBm(2.4GHz ~ 2.5 GHz)                          | $U=0.30\text{dB}$                 |      |                |
|   |            |                                              |                    | (-90 ~ -50) dBm(2.4GHz~2.5 GHz)                          | $U= (0.12\sim0.14) \text{ dB}$    |      |                |
|   |            | Signal Spectral Purity                       |                    | (-100~-10)dBc(1.2GHz ~ 8GHz)                             | $U=1.0\text{dB}$                  |      |                |
|   |            | Single Side Band Phase Noise                 |                    | (-140~-10) dBc/Hz(频率偏移:10 Hz~2MHz)<br>(2.4G Hz ~ 2.5GHz) | $U=1.0 \text{ dB}$                |      |                |
|   |            | Difference Error Vector Magnitude(D EVM)     |                    | 0~25%(2.4GHz ~ 2.5 GHz)                                  | $U=1.2\%$                         |      |                |
|   |            | GFSK Modulation Frequency Offset             |                    | 100 kHz ~ 300 kHz(2.4GHz ~ 2.5 GHz)                      | $U_{\text{rel}}=0.92\%$           |      |                |
|   |            | Frequency Error                              |                    | (-1.3 MHz ~ +1.3 MHz)(2.4GHz ~ 2.5 GHz)                  | $U=12 \text{ Hz}$                 |      |                |
|   |            | Power Measurement                            |                    | (-70~ 10) dBm(2.4Hz ~ 2.5 GHz)                           | $U=0.30 \text{ dB}$               |      |                |
|   |            | GFSK Modulation Frequency Offset Measurement |                    | (100 kHz ~ 300 kHz)(2.4Hz ~ 2.5 GHz)                     | $U_{\text{rel}}=0.92\%$           |      |                |



No. CNAS L0730

第 89 页 共 163 页

The scope of the accreditation in Chinese remains the definitive version.


| №  | Instrument     | Measurand                                       | Calibration Method                    | Range                                 | Expanded Uncertainty<br>(k=2)          | Note | Effective Date |
|----|----------------|-------------------------------------------------|---------------------------------------|---------------------------------------|----------------------------------------|------|----------------|
|    |                | Differential Error Vector Amplitude Measurement |                                       | 0~25%(2.4GHz ~ 2.5 GHz)               | U=1.2%                                 |      |                |
|    |                | Frequency Error Measurement                     |                                       | (-1.3 MHz ~ +1.3MHz)(2.4Hz ~ 2.5 GHz) | U=12 Hz                                |      |                |
|    |                | Frequency Measurement                           |                                       | 2.4GHz ~ 2.5 GHz                      | U <sub>rel</sub> =1.0×10 <sup>-7</sup> |      |                |
|    |                | VSWR of RF Port                                 |                                       | (1~3), (2.4GHz ~2.5 GHz)              | U=0.02                                 |      |                |
| 26 | *WLAN Test Set | Internal crystal oscillator frequency           | C.S. for RF WLAN Test Set<br>JJF 1277 | 10MHz                                 | U <sub>rel</sub> =6.7×10 <sup>-9</sup> |      |                |
|    |                | Frequency                                       |                                       | 2.4GHz~6GHz                           | U <sub>rel</sub> =6.7×10 <sup>-9</sup> |      |                |
|    |                | Level                                           |                                       | (0 ~ 10) dBm(2.4GHz ~6 GHz)           | U=0.24 dB                              |      |                |
|    |                |                                                 |                                       | (-50 ~ 0) dBm(2.4GHz ~6 GHz)          | U=0.28 dB                              |      |                |
|    |                |                                                 |                                       | (-90 ~ -50) dBm(2.4GHz ~6 GHz)        | U=0.30 dB                              |      |                |
|    |                | Signal Spectral Purity                          |                                       | (-100~-10)dBc(2.4GHz ~6 GHz)          | U=1.0 dB                               |      |                |
|    |                | Single Side Band Phase Noise                    |                                       | (-140~-30)dBc/Hz(2.4GHz ~ 6GHz)       | U=1.0 dB                               |      |                |
|    |                | Frequency Error                                 |                                       | (-1.3MHz ~ +1.3MHz)(2.4GHz ~6GHz)     | U=10 Hz                                |      |                |

No. CNAS L0730

第 90 页 共 163 页



The scope of the accreditation in Chinese remains the definitive version.

| №  | Instrument               | Measurand                             | Calibration Method                        | Range                                                           | Expanded Uncertainty<br>( <i>k</i> =2)        | Note                                                                                 | Effective Date |
|----|--------------------------|---------------------------------------|-------------------------------------------|-----------------------------------------------------------------|-----------------------------------------------|--------------------------------------------------------------------------------------|----------------|
|    |                          | Error Vector Magnitude(E VM)          |                                           | 0~25% (2.4GHz ~ 6GHz)                                           | <i>U</i> =1.0%                                |                                                                                      |                |
|    |                          | Pilot EVM (dB)                        |                                           | 0~20%(2.4GHz ~ 6GHz)                                            | <i>U</i> =1.0%                                |                                                                                      |                |
|    |                          | Symbol Clock Error                    |                                           | 1.0×10 <sup>-4</sup> ~1.0×10 <sup>-4</sup> ), (2.4GHz ~ 6GHz)   | <i>U</i> =3.6×10 <sup>-6</sup>                |                                                                                      |                |
|    |                          | Power Measurement                     |                                           | (0 ~ 10) dBm(2.4GHz ~ 6 GHz)                                    | <i>U</i> =0.27 dB                             |                                                                                      |                |
|    |                          |                                       |                                           | (-50 ~ 0) dBm(2.4GHz ~ 6GHz)                                    | <i>U</i> =0.28dB                              |                                                                                      |                |
|    |                          |                                       |                                           | (-90 ~ -50) dBm(2.4GHz ~ 6GHz)                                  | <i>U</i> =0.30 dB                             |                                                                                      |                |
|    |                          | Error Vector Magnitude(E VM)          |                                           | 0~25% (2.4GHz ~ 6GHz)                                           | <i>U</i> =1.0%                                |                                                                                      |                |
|    |                          | Symbol Clock Error                    |                                           | (-1.0×10 <sup>-4</sup> ~1.0×10 <sup>-4</sup> ), (2.4GHz ~ 6GHz) | <i>U</i> =3.6×10 <sup>-6</sup>                |                                                                                      |                |
|    |                          | VSWR of RF Port                       |                                           | (1~3) (2.4GHz~ 6GHz)                                            | <i>U</i> =0.02                                |                                                                                      |                |
| 27 | *Vector Signal Analyzers | Internal crystal oscillator frequency | C.S. for Vector Signal Analyzers JJF 1128 | 10MHz                                                           | <i>U</i> <sub>rel</sub> =6.7×10 <sup>-9</sup> |  |                |
|    |                          | Frequency Measurement                 |                                           | 100 kHz ~ 6 GHz                                                 | <i>U</i> <sub>rel</sub> =5.0×10 <sup>-9</sup> |                                                                                      |                |
|    |                          | Power measurement                     |                                           | (-110 ~ 20) dBm(100kHz ~ 6 GHz)                                 | <i>U</i> =0.30 dB                             |                                                                                      |                |
|    |                          | Error Vector Magnitude(E VM)          |                                           | 0%~ 25%(100kHz~6GHz)                                            | <i>U</i> =0.5%~2%                             |                                                                                      |                |
|    |                          |                                       |                                           |                                                                 |                                               |                                                                                      |                |

No. CNAS L0730

第 91 页 共 163 页



在线扫码获取验证

The scope of the accreditation in Chinese remains the definitive version.

| №  | Instrument                            | Measurand                                 | Calibration Method                                     | Range                                 | Expanded Uncertainty<br>( <i>k</i> =2)        | Note | Effective Date |
|----|---------------------------------------|-------------------------------------------|--------------------------------------------------------|---------------------------------------|-----------------------------------------------|------|----------------|
|    |                                       | Frequency Error                           |                                                        | (-100 kHz ~ 100 kHz)<br>(100kHz~6GHz) | <i>U</i> =2 Hz ~ 12Hz                         |      |                |
|    |                                       | Error Vector Amplitude Measurement        |                                                        | 0%~25%(100 kHz~6GHz)                  | <i>U</i> =0.5%~2%                             |      |                |
|    |                                       | Phase Error Measurement                   |                                                        | 0° ~ 15°<br>(100kHz~6GHz)             | <i>U</i> =0.9°                                |      |                |
|    |                                       | I/Q Origin offset Suppression Measurement |                                                        | (-80 ~ -10) dB<br>(100kHz~6GHz)       | <i>U</i> =0.96 dB                             |      |                |
| 28 | *TDMA-GSM Radio Communication Testers | Internal crystal oscillator frequency     | C.S. for TDMA-GSM Radio Communication Testers JJF 1131 | 10MHz                                 | <i>U</i> <sub>rel</sub> =6.7×10 <sup>-9</sup> |      |                |
|    |                                       | Frequency                                 |                                                        | 800M Hz~2GHz                          | <i>U</i> <sub>rel</sub> =6.7×10 <sup>-9</sup> |      |                |
|    |                                       | Level                                     |                                                        | (-50 ~ 0) dBm(800MHz~2 GHz)           | <i>U</i> =0.24dB                              |      |                |
|    |                                       |                                           |                                                        | (-90 ~ -50) dBm(800MHz~2 GHz)         | <i>U</i> =0.27dB                              |      |                |
|    |                                       |                                           |                                                        | (-120 ~ -90) dBm(800MHz~2 GHz)        | <i>U</i> =0.30dB                              |      |                |
|    |                                       | Signal Spectral Purity                    |                                                        | (-100~-10) dBc<br>(400MHz~6GHz)       | <i>U</i> =1.0dB                               |      |                |
|    |                                       | Single Side Band Phase Noise              |                                                        | (-140~-30) dBc/Hz(800MHz~2GHz)        | <i>U</i> =1.0 dB                              |      |                |
|    |                                       | Phase Error                               |                                                        | (0~10)°<br>(800MHz~2GHz)              | <i>U</i> =0.3 °                               |      |                |




No. CNAS L0730

第 92 页 共 163 页

The scope of the accreditation in Chinese remains the definitive version.

| No | Instrument | Measurand                                    | Calibration Method | Range                                | Expanded Uncertainty<br>(k=2)          | Note | Effective Date |
|----|------------|----------------------------------------------|--------------------|--------------------------------------|----------------------------------------|------|----------------|
|    |            | Peak Phase Error                             |                    | (-15~15)°<br>(800MHz~2GHz)           | U=1.0°                                 |      |                |
|    |            | Frequency Error                              |                    | (-30 kHz ~ 30 kHz)<br>(800MHz~2GHz)  | U=2 Hz                                 |      |                |
|    |            | Power Measurement                            |                    | (-50 ~ 10) dBm(800MHz ~ 2 GHz)       | U=0.30 dB                              |      |                |
|    |            | Phase Error Measurement                      |                    | (0~10)°<br>(800MHz~2GHz)             | U=0.3°                                 |      |                |
|    |            | Peak Phase Error Measurement                 |                    | (-15~15)° (800MHz~2GHz)              | U=1.0°                                 |      |                |
|    |            | Frequency Error Measurement                  |                    | (-30 kHz ~ 30 kHz)<br>(800MHz~ 2GHz) | U=2 Hz                                 |      |                |
|    |            | Frequency Accuracy of Audio Signal Generator |                    | 20Hz~5kHz                            | U <sub>rel</sub> =0.49%                |      |                |
|    |            | Audio analyzer frequency measurement         |                    | 20Hz~40kHz                           | U <sub>rel</sub> =6.0×10 <sup>-5</sup> |      |                |
|    |            | Audio Signal Generator Output Voltage        |                    | (0.1~5 ) V (20 Hz~5kHz)              | U <sub>rel</sub> =0.5%                 |      |                |
|    |            | Audio Analyzer Voltage Measurement           |                    | (20mV~30V) (20 Hz~40 kHz)            | U <sub>rel</sub> =0.5%                 |      |                |



No. CNAS L0730

第 93 页 共 163 页

The scope of the accreditation in Chinese remains the definitive version.

| No | Instrument                                    | Measurand                             | Calibration Method                                             | Range                             | Expanded Uncertainty<br>(k=2) | Note | Effective Date |
|----|-----------------------------------------------|---------------------------------------|----------------------------------------------------------------|-----------------------------------|-------------------------------|------|----------------|
|    |                                               | Audio Analyzer Distortion Measurement |                                                                | 0.001%~100%(20 Hz~20 kHz)         | $U_{rel}=0.1\%\sim 5\%$       |      |                |
|    |                                               | VSWR of RF Port                       |                                                                | (1~3) 800MHz~ 2 GHz)              | $U=0.02$                      |      |                |
| 29 | *TD-SCDMA Digital Radio Communication Testers | Internal crystal oscillator frequency | C.S. for TD-SCDMA Digital Radio Communication Testers JJF 1204 | 10MHz                             | $U_{rel}=6.7\times 10^{-9}$   |      |                |
|    |                                               | Frequency                             |                                                                | 30MHz~2.7GHz                      | $U_{rel}=6.7\times 10^{-9}$   |      |                |
|    |                                               | Level                                 |                                                                | (-50 ~ 0) dBm(30MHz ~2.7 GHz)     | $U=0.24\text{dB}$             |      |                |
|    |                                               |                                       |                                                                | (-90 ~ -50) dBm(30MHz ~2.7 GHz)   | $U=0.27\text{dB}$             |      |                |
|    |                                               |                                       |                                                                | (-120 ~ -90) dBm(30MHz ~2.7 GHz)  | $U=0.30\text{dB}$             |      |                |
|    |                                               | Signal Spectral Purity                |                                                                | (-100~-10)dBc(15MHz~9GHz)         | $U=1.0\text{dB}$              |      |                |
|    |                                               | Single Side Band Phase Noise          |                                                                | (-140~-30)dBc/Hz(30MHz~2.7GHz)    | $U=1.0\text{ dB}$             |      |                |
|    |                                               | Error Vector Magnitude(EVM)           |                                                                | 0%~20%(30MHz~2.7GHz)              | $U=0.9\%$                     |      |                |
|    |                                               | Frequency Error                       |                                                                | (-14 kHz ~ +14 kHz)(30MHz~2.7GHz) | $U=4\text{Hz}$                |      |                |
|    |                                               | Waveform Quality Factor               |                                                                | (0.5~1.0)(30MHz~2.7GHz)           | $U=0.0012$                    |      |                |



No. CNAS L0730

第 94 页 共 163 页

The scope of the accreditation in Chinese remains the definitive version.

| № | Instrument | Measurand                                    | Calibration Method | Range                            | Expanded Uncertainty<br>( $k=2$ )   | Note | Effective Date |
|---|------------|----------------------------------------------|--------------------|----------------------------------|-------------------------------------|------|----------------|
|   |            | Signal Generator code Domain Power           |                    | (-50 ~ 0) dB(30MHz~2.7GHz)       | $U=0.8\text{dB}$                    |      |                |
|   |            |                                              |                    | (-60 ~ -50) dB(30MHz~2.7GHz)     | $U=2.9\text{dB}$                    |      |                |
|   |            | Power Measurement                            |                    | (-50 ~ 10) dBm(30MHz~2.7GHz)     | $U=0.27\text{ dB}$                  |      |                |
|   |            |                                              |                    | (-65 ~ -50) dBm(30MHz~2.7GHz)    | $U=0.30\text{dB}$                   |      |                |
|   |            | Error Vector Magnitude(EVM) Measurement      |                    | 0%~20% (30MHz~2.7GHz)            | $U=0.8\%$                           |      |                |
|   |            | Frequency Error Measurement                  |                    | -14 kHz ~ +14 kHz (30MHz~2.7GHz) | $U=3.5\text{Hz}$                    |      |                |
|   |            | Phase Error Measurement                      |                    | (0.01~15)° (30MHz~2.7GHz)        | $U=0.81^\circ$                      |      |                |
|   |            | Waveform Quality Factor                      |                    | 0~1.0 (30MHz~2.7GHz)             | $U=0.0012$                          |      |                |
|   |            | Frequency Accuracy of Audio Signal Generator |                    | 20Hz~20kHz                       | $U_{\text{rel}}=0.49\%$             |      |                |
|   |            | Audio analyzer frequency measurement         |                    | 20Hz~20kHz                       | $U_{\text{rel}}=6.0 \times 10^{-5}$ |      |                |
|   |            | Distortion of Audio Signal Generator         |                    | (0.001%~100%)(20Hz~20kHz)        | $U_{\text{rel}}=0.5\% \sim 5\%$     |      |                |



No. CNAS L0730

第 95 页 共 163 页

The scope of the accreditation in Chinese remains the definitive version.

| №  | Instrument                                 | Measurand                             | Calibration Method                                          | Range                           | Expanded Uncertainty<br>( <i>k</i> =2)        | Note | Effective Date |
|----|--------------------------------------------|---------------------------------------|-------------------------------------------------------------|---------------------------------|-----------------------------------------------|------|----------------|
|    |                                            | Audio Signal Generator Output Voltage |                                                             | (20 mV~5 V)(20 Hz~20 kHz)       | <i>U</i> <sub>rel</sub> =0.5%                 |      |                |
|    |                                            | Audio Analyzer Voltage Measurement    |                                                             | (20mV~5V)(20 Hz~20 kHz)         | <i>U</i> <sub>rel</sub> =0.5%                 |      |                |
|    |                                            | Audio Analyzer Distortion Measurement |                                                             | 0.001%~100%(20 Hz~20 kHz)       | <i>U</i> <sub>rel</sub> =0.1%~5%              |      |                |
|    |                                            | VSWR of RF Port                       |                                                             | (1~3)(30MHz~2.7GHz)             | <i>U</i> =0.02                                |      |                |
| 30 | *WCDMA Digital Radio Communication Testers | Internal crystal oscillator frequency | C.S. for WCDMA Digital Radio Communication Testers JJF 1276 | 10MHz                           | <i>U</i> <sub>rel</sub> =6.7×10 <sup>-9</sup> |      |                |
|    |                                            | Frequency                             |                                                             | 30MHz~2.7GHz                    | <i>U</i> <sub>rel</sub> =6.7×10 <sup>-9</sup> |      |                |
|    |                                            | Level                                 |                                                             | (-50 ~ 0) dBm (30MHz~2.7GHz)    | <i>U</i> =0.24dB                              |      |                |
|    |                                            |                                       |                                                             | (-90 ~ -50) dBm(30MHz~2.7GHz)   | <i>U</i> =0.27dB                              |      |                |
|    |                                            |                                       |                                                             | (-120 ~ -90) dBm((30MHz~2.7GHz) | <i>U</i> =0.30dB                              |      |                |
|    |                                            | Signal Spectral Purity                |                                                             | (-100~-10)dBc((30MHz~2.7GHz)    | <i>U</i> =1.0dB                               |      |                |



No. CNAS L0730

第 96 页 共 163 页

The scope of the accreditation in Chinese remains the definitive version.

| No | Instrument | Measurand                               | Calibration Method | Range                           | Expanded Uncertainty<br>( $k=2$ ) | Note | Effective Date |
|----|------------|-----------------------------------------|--------------------|---------------------------------|-----------------------------------|------|----------------|
|    |            | Single Side Band Phase Noise            |                    | (-140~-30)dBc/Hz(30MHz~2.7GHz)  | $U=1.0$ dB                        |      |                |
|    |            | Phase Error                             |                    | (0~15)° (30MHz~2.7GHz)          | $U=0.8^{\circ}$                   |      |                |
|    |            | Waveform Quality Factor                 |                    | (0~1.0)(30MHz~2.7GHz)           | $U=0.0012$                        |      |                |
|    |            | Error Vector Magnitude(EVM)             |                    | (0%~25%)(30MHz~2.7GHz)          | $U=0.9\%$                         |      |                |
|    |            | Frequency Error                         |                    | (-5 kHz ~ +5 kHz)(30MHz~2.7GHz) | $U=6$ Hz                          |      |                |
|    |            | Code Domain Power                       |                    | (-50 ~ 0) dB (30MHz~2.7GHz)     | $U=0.8$ dB                        |      |                |
|    |            |                                         |                    | (-60 ~ -50) dB (30MHz~2.7GHz)   | $U=2.9$ dB                        |      |                |
|    |            | Power Measurement                       |                    | (-50 ~ 10) dBm(30MHz~2.7GHz)    | $U=0.30$ dB                       |      |                |
|    |            | Phase Error Measurement                 |                    | (0~15)° (30MHz~2.7GHz)          | $U=0.8^{\circ}$                   |      |                |
|    |            | Error Vector Magnitude(EVM) Measurement |                    | (0%~25%)(30MHz~2.7GHz)          | $U=0.9\%$                         |      |                |
|    |            | Frequency Error Measurement             |                    | (-5 kHz ~ +5 kHz)(30MHz~2.7GHz) | $U=6$ Hz                          |      |                |
|    |            | Waveform Quality Factor                 |                    | (0.5~1.0)(30MHz~2.7GHz)         | $U=0.0012$                        |      |                |



No. CNAS L0730

第 97 页 共 163 页

The scope of the accreditation in Chinese remains the definitive version.

| №  | Instrument                               | Measurand                                    | Calibration Method                                                                                                    | Range                       | Expanded Uncertainty<br>( $k=2$ ) | Note                   | Effective Date |
|----|------------------------------------------|----------------------------------------------|-----------------------------------------------------------------------------------------------------------------------|-----------------------------|-----------------------------------|------------------------|----------------|
|    |                                          | Frequency Accuracy of Audio Signal Generator | ilac-MRA<br>CNAS<br>NATIONAL ACCREDITATION SERVICE FOR CONFORMITY ASSESSMENT<br>SCHEDULE OF ACCREDITATION CERTIFICATE | 10Hz~20kHz                  | $U_{rel}=0.49\%$                  |                        |                |
|    |                                          | Audio analyzer frequency measurement         |                                                                                                                       | 10Hz~20kHz                  | $U_{rel}=6.0 \times 10^{-5}$      |                        |                |
|    |                                          | Distortion of Audio Signal Generator         |                                                                                                                       | (0.001%~10%)(10Hz~20kHz)    | $U_{rel}=0.5\% \sim 5\%$          |                        |                |
|    |                                          | Audio Signal Generator Output Voltage        |                                                                                                                       | (10 mV~5 V)(10 Hz~20 kHz)   | $U_{rel}=0.5\%$                   |                        |                |
|    |                                          | Audio Analyzer Voltage Measurement           |                                                                                                                       | (20mV~5V) (10 Hz~20 kHz)    | $U_{rel}=0.5\%$                   |                        |                |
|    |                                          | Audio Analyzer Distortion Measurement        |                                                                                                                       | (0.001%~100%)(10 Hz~20 kHz) | $U_{rel}=0.5\% \sim 5\%$          |                        |                |
|    |                                          | VSWR of RF Port                              |                                                                                                                       | (1~3) (30 MHz~2.7 GHz)      | $U=0.02$                          |                        |                |
| 31 | *LTE Digital Radio Communication Testers | Internal crystal oscillator frequency        | C.S. for LTE Digital Radio Communication Testers JJF 1443                                                             | 10MHz                       | $U_{rel}=6.7 \times 10^{-9}$      | 合格评定国家认可委员会<br>认可证书专用章 |                |
|    |                                          | Frequency                                    |                                                                                                                       | 600MHz~3.8GHz               | $U_{rel}=6.7 \times 10^{-9}$      |                        |                |



No. CNAS L0730

第 98 页 共 163 页

The scope of the accreditation in Chinese remains the definitive version.

| No | Instrument | Measurand                     | Calibration Method | Range                               | Expanded Uncertainty<br>(k=2) | Note | Effective Date |
|----|------------|-------------------------------|--------------------|-------------------------------------|-------------------------------|------|----------------|
|    |            | Level                         | ilac-M             | (-50 ~ 0)<br>dBm(600MHz~3.8GHz)     | U=0.24dB                      |      |                |
|    |            |                               |                    | (-90 ~ -50)<br>dBm(600MHz~3.8GHz)   | U=0.27dB                      |      |                |
|    |            |                               |                    | (-120 ~ -90)<br>dBm(600MHz~3.8GHz)  | U=0.30dB                      |      |                |
|    |            | Signal Spectral Purity        |                    | (-100 ~ -10) dBc<br>(300MHz~ 12GHz) | U=1.0dB                       |      |                |
|    |            | Single Side Band Phase Noise  |                    | (-140~30)<br>dBc/Hz(600MHz~3.8GHz)  | U=1.0 dB                      |      |                |
|    |            | Error Vector Magnitude(EVM)   |                    | 0%~20%(600MHz~3.8GHz)               | U=0.5%                        |      |                |
|    |            | Frequency Error               |                    | (-30 kHz ~ +30 kHz)(600MHz~3.8GHz)  | U=2Hz                         |      |                |
|    |            | I/Q origin offset suppression |                    | (-50 ~ -15) dB                      | U=2.1 dB                      |      |                |
|    |            |                               |                    | (-60 ~ -50) dB                      | U=2.5dB                       |      |                |
|    |            | Power Measurement             |                    | (-50 ~ 10)<br>dBm(600MHz~3.8GHz)    | U=0.30 dB                     |      |                |
|    |            | Error Vector Magnitude(EVM)   |                    | 0%~20%(600MHz~3.8GHz)               | U=0.5%                        |      |                |
|    |            | Frequency Error               |                    | (-30 kHz ~ +30 kHz)(600MHz~3.8GHz)  | U=2 Hz                        |      |                |
|    |            | I/Q origin offset suppression |                    | (-50 ~ -15) dB                      | U=2.1 dB                      |      |                |



No. CNAS L0730

第 99 页 共 163 页

The scope of the accreditation in Chinese remains the definitive version.

| №  | Instrument                                              | Measurand                             | Calibration Method                                        | Range                              | Expanded Uncertainty<br>( <i>k</i> =2)        | Note | Effective Date |
|----|---------------------------------------------------------|---------------------------------------|-----------------------------------------------------------|------------------------------------|-----------------------------------------------|------|----------------|
|    |                                                         |                                       |                                                           | (-60 ~ -50) dB                     | <i>U</i> =2.5dB                               |      |                |
|    |                                                         | VSWR of RF Port                       |                                                           | (1 ~3) (600 MHz~3.8GHz)            | <i>U</i> =0.02                                |      |                |
| 32 | *CDMA Digital Mobile Communication Comprehensive Tester | Internal crystal oscillator frequency | C.S. of CDMA Digital Radio Communication Testers JJF 1177 | 10MHz                              | <i>U</i> <sub>rel</sub> =6.7×10 <sup>-9</sup> |      |                |
|    |                                                         | Frequency                             |                                                           | 100MHz~2GHz                        | <i>U</i> <sub>rel</sub> =6.7×10 <sup>-9</sup> |      |                |
|    |                                                         | Level                                 |                                                           | (-50 ~ 0) dBm(100 MHz ~2GHz)       | <i>U</i> =0.24dB                              |      |                |
|    |                                                         |                                       |                                                           | (-90 ~ -50) dBm(100 MHz ~2GHz)     | <i>U</i> =0.27dB                              |      |                |
|    |                                                         |                                       |                                                           | (-120 ~ -90) dBm(100 MHz ~2GHz)    | <i>U</i> =0.30dB                              |      |                |
|    |                                                         | Signal Spectral Purity                |                                                           | (-100~-10) dBc (100kHz~6GHz)       | <i>U</i> =1.0dB                               |      |                |
|    |                                                         | Single Side Band Phase Noise          |                                                           | (-140~-30) dBc/Hz(100MHz~2GHz)     | <i>U</i> =1.0 dB                              |      |                |
|    |                                                         | Phase Error                           |                                                           | (0~15)° (100MHz~2GHz)              | <i>U</i> =0.8°                                |      |                |
|    |                                                         | Error Vector Magnitude(EVM)           |                                                           | (0%~20%)(100MHz~2GHz)              | <i>U</i> =0.9%                                |      |                |
|    |                                                         | Frequency Error                       |                                                           | (-1.2 kHz ~ +1.2 kHz)(100MHz~2GHz) | <i>U</i> =2 Hz                                |      |                |
|    | Waveform Quality Factor                                 | (0.5~1.0)(30MHz~2.7GHz)               | <i>U</i> =0.0012                                          |                                    |                                               |      |                |



No. CNAS L0730

第 100 页 共 163 页

The scope of the accreditation in Chinese remains the definitive version.

| №  | Instrument               | Measurand                             | Calibration Method                          | Range                              | Expanded Uncertainty<br>( <i>k</i> =2)        | Note | Effective Date |
|----|--------------------------|---------------------------------------|---------------------------------------------|------------------------------------|-----------------------------------------------|------|----------------|
|    |                          | Power Measuremen                      |                                             | (-50 ~ 10) dBm(100 MHz ~ 2 GHz)    | <i>U</i> =0.30dB                              |      |                |
|    |                          | Phase Error                           |                                             | (0.1~15)° (100MHz~2GHz)            | <i>U</i> =0.8°                                |      |                |
|    |                          | Error Vector Magnitude(EVM)           |                                             | (0%~20%)(30MHz~2.7GHz)             | <i>U</i> =0.9%                                |      |                |
|    |                          | Frequency Error                       |                                             | (-1.2 kHz ~ +1.2 kHz)(100MHz~2GHz) | <i>U</i> =2Hz                                 |      |                |
|    |                          | Waveform Quality Factor               |                                             | (0.5~1.0)(30MHz~2.7GHz)            | <i>U</i> =0.0012                              |      |                |
|    |                          | VSWR of RF Port                       |                                             | (1~3) (100 MHz~2GHz)               | <i>U</i> =0.02                                |      |                |
| 33 | *RF Communication Tester | Internal crystal oscillator frequency | C.S. for RF Communication Test Set JJF 1065 | 10MHz                              | <i>U</i> <sub>rel</sub> =6.7×10 <sup>-9</sup> |      |                |
|    |                          | Frequency                             |                                             | 250kHz~1GHz                        | <i>U</i> <sub>rel</sub> =6.7×10 <sup>-9</sup> |      |                |
|    |                          | Level                                 |                                             | (0 ~ 13) dBm(100 kHz~1 GHz)        | <i>U</i> =0.24 dB                             |      |                |
|    |                          |                                       |                                             | (-50 ~ 0) dBm(100 kHz~1 GHz)       | <i>U</i> =0.25dB                              |      |                |
|    |                          |                                       |                                             | (-90 ~ -50) dBm(100 kHz ~1 GHz)    | <i>U</i> =0.27dB                              |      |                |
|    |                          |                                       |                                             | (-120 ~ -90) dBm(100 kHz~1 GHz)    | <i>U</i> =0.30dB                              |      |                |
|    |                          | Signal Spectral Purity                |                                             | (-100~-10)dBc(100 kHz~1GHz)        | <i>U</i> =1.0dB                               |      |                |



No. CNAS L0730

第 101 页 共 163 页

The scope of the accreditation in Chinese remains the definitive version.

| № | Instrument | Measurand                                    | Calibration Method | Range                                                | Expanded Uncertainty<br>( $k=2$ ) | Note | Effective Date |
|---|------------|----------------------------------------------|--------------------|------------------------------------------------------|-----------------------------------|------|----------------|
|   |            | Single Side Band Phase Noise                 |                    | (-140~-30)dBc/Hz(100kHz~1GHz)                        | $U=1.0$ dB                        |      |                |
|   |            | Amplitude Modulation (AM)                    |                    | 0%~99%<br>(CW:100kHz~1GHz;f:10Hz~50kHz)              | $U_{rel}=1.2\%$                   |      |                |
|   |            | Frequency Modulation Deviation               |                    | (1kHz~100 kHz)<br>CW: (0.1 MHz~1GHz, f:20 Hz~50k Hz) | $U_{rel}=1.2\%$                   |      |                |
|   |            | Phase Modulation Deviation                   |                    | (0.2~10) rad, (CW: 0.1MHz~1 GHz;f:20Hz~50kHz)        | $U_{rel}=1.2\%$                   |      |                |
|   |            | Power Measurement                            |                    | (-50 ~ 10) dBm(100 kHz ~1 GHz)                       | $U=0.30$ dB                       |      |                |
|   |            | Amplitude Modulation Measurement             |                    | (0~95%) (CW: 100kHz~1GHz, f: 20Hz~50kHz)             | $U_{rel}=1.2\%$                   |      |                |
|   |            | Frequency Modulation Deviation Measurement   |                    | (20Hz~100kHz)<br>(CW: 100 kHz~1GHz, f: 20Hz~50kHz)   | $U_{rel}=1.2\%$                   |      |                |
|   |            | Phase Modulation Deviation Measurement       |                    | (0.1~10) rad (CW: 100 kHz~1GHz, f: 20Hz~50kHz)       | $U_{rel}=1.2\%$                   |      |                |
|   |            | Frequency Accuracy of Audio Signal Generator |                    | 20Hz~50kHz                                           | $U_{rel}=0.49\%$                  |      |                |



No. CNAS L0730

第 102 页 共 163 页

The scope of the accreditation in Chinese remains the definitive version.

| №  | Instrument                                        | Measurand                             | Calibration Method                                                                                                           | Range                          | Expanded Uncertainty<br>( $k=2$ )                | Note                         | Effective Date |
|----|---------------------------------------------------|---------------------------------------|------------------------------------------------------------------------------------------------------------------------------|--------------------------------|--------------------------------------------------|------------------------------|----------------|
|    |                                                   | Audio analyzer frequency measurement  | ilac-MRA<br>CNAS<br>CHINA NATIONAL ACCREDITATION SERVICE FOR CONFORMANCE ASSESSMENT<br>SCHEDULE OF ACCREDITATION CERTIFICATE | 20Hz~100kHz                    | $U_{rel}=6.0 \times 10^{-5}$                     |                              |                |
|    |                                                   | Distortion of Audio Signal Generator  |                                                                                                                              | (0.001%~30%)(20Hz~20kHz)       | $U_{rel}=0.5\% \sim 5\%$                         |                              |                |
|    |                                                   | Audio Signal Generator Output Voltage |                                                                                                                              | (10 mV~5 V)(20 Hz~50 kHz)      | $U_{rel}=0.5\%$                                  |                              |                |
|    |                                                   | Audio Analyzer Voltage Measurement    |                                                                                                                              | (20mV~30V)(20 Hz~50 kHz)       | $U_{rel}=0.5\%$                                  |                              |                |
|    |                                                   | Audio Analyzer Distortion Measurement |                                                                                                                              | 0.001%~100%(20 Hz~50 kHz)      | $U_{rel}=0.5\% \sim 5\%$                         |                              |                |
|    |                                                   | VSWR of RF Port                       |                                                                                                                              | (1~3) (10 MHz~ 1 GHz)          | $U=0.02$                                         |                              |                |
| 34 | *Electro-Magnetic Interference Measuring Receiver | Frequency                             | Calibration Specification for EMI Testing Receivers JJF 1144                                                                 | 9kHz~40GHz                     | $U_{rel}=1 \times 10^{-4} \sim 1 \times 10^{-9}$ | 合格评定 国家认可 中国合格评定国家认可 认可证书专用章 |                |
|    |                                                   | Voltage                               |                                                                                                                              | (0~125)dB $\mu$ V(9kHz~40GHz)  | $U=(0.4 \sim 0.63)$ dB                           |                              |                |
|    |                                                   | 6dB bandwidth                         |                                                                                                                              | 100Hz~150kHz(9kHz~40GHz)       | $U_{rel}=0.9\% \sim 1.4\%$                       |                              |                |
|    |                                                   | Noise indicating                      |                                                                                                                              | (-30~50)dB $\mu$ V(9kHz~40GHz) | $U=0.20$ dB                                      |                              |                |
|    |                                                   | Pulse response                        |                                                                                                                              | (30~70)dB $\mu$ V(9kHz~1GHz)   | $U=0.6$ dB~1.2dB                                 |                              |                |



No. CNAS L0730

第 103 页 共 163 页

The scope of the accreditation in Chinese remains the definitive version.

| №  | Instrument                                                               | Measurand                   | Calibration Method                                                        | Range                           | Expanded Uncertainty<br>( $k=2$ )    | Note | Effective Date |
|----|--------------------------------------------------------------------------|-----------------------------|---------------------------------------------------------------------------|---------------------------------|--------------------------------------|------|----------------|
|    |                                                                          | Attenuator                  | ilac-M                                                                    | (0~80)dB(9kHz~40GHz)            | $U=0.21\text{dB} \sim 1\text{dB}$    |      |                |
|    |                                                                          | Linear indication           |                                                                           | (0~100)dB(9kHz~40GHz)           | $U=0.17\text{dB} \sim 0.65\text{dB}$ |      |                |
|    |                                                                          | Voltage standing wave ratio |                                                                           | (1~3)(10MHz~40GHz)              | $U=0.02$                             |      |                |
| 35 | *Attenuator                                                              | Attenuation                 | C.S. for Attenuators from Radio Frequency to Microwave Frequency JJF 2092 | (0 ~ 50) dB (100kHz ~ 50 GHz)   | $U=(0.036 \sim 0.082)\text{dB}$      |      |                |
|    |                                                                          |                             |                                                                           | (50 ~ 90) dB (100kHz ~ 50 GHz)  | $U=(0.26 \sim 0.28)\text{dB}$        |      |                |
|    |                                                                          |                             |                                                                           | (90 ~ 110) dB (100kHz ~ 50 GHz) | $U=(0.48 \sim 0.49)\text{dB}$        |      |                |
|    |                                                                          | Voltage standing wave ratio |                                                                           | (1 ~ 3) (100 kHz ~ 50GHz)       | $U=0.04 \sim 0.14$                   |      |                |
| 36 | *Lower power mount(Rf power sensor、Microwave power sensor、 power sensor) | Calibration factor          | C.S. for RF & Microwave Power Sensors JJF 1887                            | -20dBm~10dBm (DC~50GHz:)        | $U_{\text{rel}}=1.0\% \sim 3.7\%$    |      |                |
|    |                                                                          | Voltage standing wave ratio |                                                                           | (1 ~ 3) (10 MHz~50GHz)          | $U=0.04 \sim 0.14$                   |      |                |
| 37 | *Pulse Power meter                                                       | Calibration factor          | Verification Regulation of Pulse Power Meters JJG 1024                    | -20dBm~10dBm (DC~50GHz:)        | $U=1.0\% \sim 3.7\%$                 |      |                |
|    |                                                                          | Voltage standing wave ratio |                                                                           | (1 ~ 3) (10 MHz~50GHz)          | $U=0.04 \sim 0.14$                   |      |                |
| 38 | *Measuring Receiver                                                      | Output Reference Frequency  | C.S. for Measuring Receivers JJF 1173                                     | 10MHz                           | $U_{\text{rel}}=2 \times 10^{-9}$    |      |                |



No. CNAS L0730

第 104 页 共 163 页

The scope of the accreditation in Chinese remains the definitive version.

| № | Instrument | Measurand                         | Calibration Method | Range                                                            | Expanded Uncertainty<br>( $k=2$ ) | Note | Effective Date |
|---|------------|-----------------------------------|--------------------|------------------------------------------------------------------|-----------------------------------|------|----------------|
|   |            | Frequency Measurement             |                    | 100kHz~50GHz                                                     | $U_{rel}=1 \times 10^{-8}$        |      |                |
|   |            | Tuning Level                      |                    | (-120~0)dB(100kHz~50GHz)                                         | $U=(0.008 \sim 0.4)$ dB           |      |                |
|   |            | Frequency Modulation Deviation    |                    | 0.1Hz~500kHz(carrier frequency :100kHz~17.1GHz, MF:10Hz~400kHz)  | $U_{rel}=1.0\%$                   |      |                |
|   |            |                                   |                    | 0.1Hz~500kHz(carrier frequency :17.1GHz~34.5GHz, MF:10Hz~200kHz) | $U_{rel}=1.2\%$                   |      |                |
|   |            |                                   |                    | 0.1Hz~500kHz(carrier frequency :34.5GHz~50GHz, MF:10Hz~200kHz)   | $U_{rel}=1.4\%$                   |      |                |
|   |            | Phase Modulation                  |                    | (0.2~500)rad(carrier frequency :100kHz~50GHz, MF:200Hz~100kHz)   | $U_{rel}=1.0\%$                   |      |                |
|   |            | Amplitude Demodulation Distortion |                    | 0.001%~100%(frequency:20Hz~20kHz)                                | $U_{rel}=15\%$                    |      |                |
|   |            |                                   |                    | 0.001%~100%(frequency:20kHz~100kHz)                              | $U_{rel}=26\%$                    |      |                |
|   |            | Residential Frequency Modulation  |                    | 0.1Hz(rms)~10Hz(rms)                                             | $U=0.9$ Hz(rms)                   |      |                |



No. CNAS L0730

第 105 页 共 163 页

The scope of the accreditation in Chinese remains the definitive version.


| №  | Instrument          | Measurand                         | Calibration Method                   | Range                                                                    | Expanded Uncertainty<br>( $k=2$ ) | Note                   | Effective Date |
|----|---------------------|-----------------------------------|--------------------------------------|--------------------------------------------------------------------------|-----------------------------------|------------------------|----------------|
|    |                     | Amplitude Modulation Rejection    | ilac-MRA                             | 0.1Hz~100Hz                                                              | $U=0.9\text{Hz}$                  |                        |                |
|    |                     | Amplitude Modulation Depth        |                                      | 5%~99%(Carrier Frequency:100kHz~50GHz, Modulation Frequency:10Hz~100kHz) | $U_{\text{rel}}=1.0\%$            |                        |                |
|    |                     | Frequency Demodulation Distortion |                                      | 0.001%~100%(Frequency:20Hz~20kHz)                                        | $U_{\text{rel}}=15\%$             |                        |                |
|    |                     | Residual Amplitude Modulation     |                                      | 0.001%~100%(Frequency:20kHz~100kHz)                                      | $U_{\text{rel}}=26\%$             |                        |                |
|    |                     | Frequency Modulation Rejection    |                                      | 0.001%(rms)~1%(rms)                                                      | $U=0.0024\%(\text{rms})$          |                        |                |
|    |                     | Frequency Modulation Rejection    |                                      | 0.001%~2%                                                                | $U=0.0024\%$                      |                        |                |
| 39 | *Waveform Recorders | Direct Voltage                    | C.S. for Waveform Recorders JJF 1876 | $\pm(5\text{mV}\sim1000\text{V})$                                        | $U_{\text{rel}}=0.04\%\sim0.12\%$ | 合格评定国家认可委员会<br>认可证书专用章 |                |
|    |                     | Alternating Voltage               |                                      | 5mV~1000V                                                                | $U_{\text{rel}}=0.07\%\sim0.17\%$ |                        |                |
|    |                     | Time Base                         |                                      | 0.5ns~10s                                                                | $U_{\text{rel}}=3\times10^{-7}$   |                        |                |
|    |                     | Frequency Bandwidth               |                                      | 50kHz~100MHz                                                             | $U_{\text{rel}}=2\%$              |                        |                |
|    |                     | Input Resistance                  |                                      | 40 $\Omega$ ~ 300M $\Omega$                                              | $U_{\text{rel}}=0.24\%$           |                        |                |



No. CNAS L0730

第 106 页 共 163 页

The scope of the accreditation in Chinese remains the definitive version.

| №  | Instrument                         | Measurand                          | Calibration Method                                        | Range                              | Expanded Uncertainty<br>( <i>k</i> =2)        | Note                                                                                 | Effective Date |
|----|------------------------------------|------------------------------------|-----------------------------------------------------------|------------------------------------|-----------------------------------------------|--------------------------------------------------------------------------------------|----------------|
|    |                                    | Direct Current Gain                |                                                           | 0.8~1.2（Voltage: ±<br>(5mV~1000V)） | <i>U</i> <sub>rel</sub> =0.1%                 |                                                                                      |                |
| 40 | *Distortion Meter Calibrator       | Frequency                          | V.R.of Distortion Meter Calibrators JJG 802               | 5Hz~200 kHz                        | <i>U</i> <sub>rel</sub> =1.8×10 <sup>-5</sup> |                                                                                      |                |
|    |                                    | Voltage                            |                                                           | 90 μ<br>V~10V(5Hz~400kHz)          | <i>U</i> <sub>rel</sub> =0.02%~1.2%           |                                                                                      |                |
|    |                                    | distortion                         |                                                           | 0.001%~1.0% (10<br>Hz~20 kHz)      | <i>U</i> <sub>rel</sub> =6%                   |                                                                                      |                |
|    |                                    |                                    |                                                           | 0.001%~2.0% (20<br>kHz~400 kHz)    | <i>U</i> <sub>rel</sub> =7%                   |                                                                                      |                |
| 41 | *Range Calibrator                  | Direct Voltage                     | V.R.for HP 11683A power range calibrator JJG(Dianzi)10005 | 100 μ V~200mV                      | <i>U</i> <sub>rel</sub> =0.02%                |                                                                                      |                |
|    |                                    | Resistance                         |                                                           | 10 Ω ~5000 Ω                       | <i>U</i> <sub>rel</sub> =0.02%                |                                                                                      |                |
| 42 | *Power Meters                      | Zero Point                         | C.S.of Power Meters JJF 1757                              | 1nW~10 μ W                         | <i>U</i> =6nW                                 | accredited only for Amplified                                                        |                |
|    |                                    | Zero Drift                         |                                                           | 1nW~10 μ W                         | <i>U</i> =10nW~20nW                           |                                                                                      |                |
|    |                                    | Power Range                        |                                                           | 3uW~100mW                          | <i>U</i> <sub>rel</sub> =(0.30%~0.65%)        |                                                                                      |                |
|    |                                    | Calibration Factor                 |                                                           | 85%~100%                           | <i>U</i> <sub>rel</sub> =1.0%                 |                                                                                      |                |
|    |                                    | Output Power of Calibration Source |                                                           | 0.9mW~1.1mW, (10MHz~1GHz)          | <i>U</i> <sub>rel</sub> =1.3%                 |                                                                                      |                |
| 43 | *Vibrating Wire Frequency Readouts | Frequency                          | C.S. for Vibrating Wire Frequency Readouts JJF 1401       | (300~6000) Hz                      | <i>U</i> =0.11 Hz                             |  |                |
| 44 | *Audio Analyzer/Audio meter        | AC voltage measurement             | C.S. for Audio Analyzer JJF 1395                          | 1mV~300V, (10Hz~200kHz)            | <i>U</i> <sub>rel</sub> =0.1%~0.4%            |                                                                                      |                |
|    |                                    | Distortion measurement             |                                                           | (0.0001~30)%, (10Hz~100kHz)        | <i>U</i> <sub>rel</sub> =1.2%                 |                                                                                      |                |



No. CNAS L0730

第 107 页 共 163 页

The scope of the accreditation in Chinese remains the definitive version.

| №                        | Instrument                                | Measurand                              | Calibration Method | Range                       | Expanded Uncertainty<br>( <i>k</i> =2)      | Note | Effective Date |
|--------------------------|-------------------------------------------|----------------------------------------|--------------------|-----------------------------|---------------------------------------------|------|----------------|
|                          |                                           | Frequency measurement                  |                    | 10Hz~200kHz                 | <i>U</i> <sub>rel</sub> =0.0007%            |      |                |
|                          |                                           | Voltage measurement inherent noise     |                    | ≤300 μ V, (20kHz BW)        | <i>U</i> <sub>rel</sub> =5.6%               |      |                |
|                          |                                           | Distortion measure residual distortion |                    | (0.001~0.05)%,(20Hz~100kHz) | <i>U</i> <sub>rel</sub> =2%                 |      |                |
|                          |                                           | Output frequency                       |                    | 10Hz~200kHz                 | <i>U</i> <sub>rel</sub> =1×10 <sup>-6</sup> |      |                |
|                          |                                           | Output voltage                         |                    | 1mV~20V, (10Hz~200kHz)      | <i>U</i> <sub>rel</sub> =0.1%~0.4%          |      |                |
|                          |                                           | Output sine wave waveform distortion   |                    | "0.001%~10% (5Hz~200kHz) "  | <i>U</i> <sub>rel</sub> =15%                |      |                |
|                          |                                           | 45                                     |                    | *Dynamic Signal Analyzer    | Amplitude                                   |      |                |
| (10~33)mV, (10Hz~200kHz) | <i>U</i> <sub>rel</sub> =0.24%            |                                        |                    |                             |                                             |      |                |
| 33mV~10V, (10Hz~200kHz)  | <i>U</i> <sub>rel</sub> =0.12%            |                                        |                    |                             |                                             |      |                |
| Dynamic range            | 30dB~130dB, (10Hz~200kHz)                 |                                        | <i>U</i> =0.6dB    |                             |                                             |      |                |
| Channel consistency      | phase difference: -5° ~5° , (10Hz~200kHz) |                                        | <i>U</i> =0.08°    |                             |                                             |      |                |
|                          | amplitude ratio: -1dB~1dB, (10Hz~200kHz)  |                                        | <i>U</i> =0.02dB   |                             |                                             |      |                |



No. CNAS L0730

第 108 页 共 163 页

The scope of the accreditation in Chinese remains the definitive version.

| №  | Instrument                           | Measurand                              | Calibration Method                                   | Range                                  | Expanded Uncertainty<br>( <i>k</i> =2) | Note | Effective Date |
|----|--------------------------------------|----------------------------------------|------------------------------------------------------|----------------------------------------|----------------------------------------|------|----------------|
|    |                                      | frequency                              |                                                      | 0.1Hz~200kHz                           | <i>U</i> <sub>rel</sub> =0.003%        |      |                |
| 46 | *Network cable analyzers             | DC Loop Resistance                     | C.S.for Network cable analyzers JJF 1494             | (0.1~100) Ω                            | <i>U</i> <sub>rel</sub> =(0.22~3)%     |      |                |
|    |                                      | Propagation delay                      |                                                      | (1~500)ns                              | <i>U</i> <sub>rel</sub> =1.0%          |      |                |
|    |                                      | Insertion loss                         |                                                      | (0~40)dB , (1MHz~250MHz)               | <i>U</i> =0.4dB                        |      |                |
|    |                                      | Near-end crosstalk                     |                                                      | (20~70)dB , (1MHz~250MHz)              | <i>U</i> =0.4dB                        |      |                |
|    |                                      | Equivalent far-end crosstalk           |                                                      | (10~60)dB , (1MHz~250MHz)              | <i>U</i> =0.6dB                        |      |                |
|    |                                      | length                                 |                                                      | (0.1~150)m                             | <i>U</i> <sub>rel</sub> =(0.5~1)%      |      |                |
|    |                                      | Return loss                            |                                                      | (8~26)dB , (1MHz~250MHz)               | <i>U</i> =0.8dB                        |      |                |
| 47 | *RF Voltmeters                       | Voltage measurement                    | V.R. of RF Voltmeters JJG 308                        | 10mV~100mV, (10kHz~100kHz)             | <i>U</i> <sub>rel</sub> =(0.08~1.0)%   |      |                |
|    |                                      |                                        |                                                      | 100mV~100V, (10kHz~100kHz)             | <i>U</i> <sub>rel</sub> =0.08%         |      |                |
|    |                                      | Voltage measurement frequency response |                                                      | ± (0~30) %, (100mV~2V, 10kHz~200MHz)   | <i>U</i> =(0.7~3)%                     |      |                |
|    |                                      |                                        |                                                      | ± (0~30) %。 (100mV~2V, 200MHz~2.5G Hz) | <i>U</i> =3%                           |      |                |
| 48 | *Directional Coupler and SWR Bridges | Return Loss                            | C.S.for Directional Coupler and SWR Bridges JJF 1680 | (15~60) dB, (300kHz~50GHz)             | <i>U</i> =0.2dB                        |      |                |



No. CNAS L0730

第 109 页 共 163 页

The scope of the accreditation in Chinese remains the definitive version.

| №  | Instrument                    | Measurand                          | Calibration Method                            | Range                          | Expanded Uncertainty<br>( $k=2$ ) | Note | Effective Date |
|----|-------------------------------|------------------------------------|-----------------------------------------------|--------------------------------|-----------------------------------|------|----------------|
|    |                               | Insertion loss                     |                                               | (0.01~10)dB,<br>(300kHz~50GHz) | $U=0.12\text{dB}$                 |      |                |
|    |                               | reflection coefficient measurement |                                               | 0.001~1.0,<br>(300kHz~50GHz)   | $U=0.01$                          |      |                |
|    |                               | Coupling factor                    |                                               | (10~50)dB,<br>(300kHz~50GHz)   | $U=0.2\text{dB}$                  |      |                |
|    |                               | Directivity                        |                                               | (20~60)dB,<br>(300kHz~50GHz)   | $U=0.3\text{dB}$                  |      |                |
| 49 | *RF/ microwave switch         | VSWR                               | C.S.for RF/ microwave switch JJF(电子) 30503    | 1.0~4.0<br>(10MHz~50GHz)       | $U=0.05$                          |      |                |
|    |                               | Insertion loss                     |                                               | (0.01~10) dB<br>(10MHz~50GHz)  | $U=0.12\text{dB}$                 |      |                |
|    |                               | Isolation degree                   |                                               | (30~120) dB<br>(10MHz~50GHz)   | $U=(0.5\sim 2)\text{dB}$          |      |                |
| 50 | *Cable and Antenna Analyzers  | frequency                          | C.S.for Cable and Antenna Analyzers JJF 1740  | 2MHz~18GHz                     | $U_{\text{rel}}=1 \times 10^{-7}$ |      |                |
|    |                               | output power                       |                                               | (-50~20) dBm , (2 MHz~18 GHz)  | $U=0.25\text{dB}$                 |      |                |
|    |                               | VSWR                               |                                               | 1.0~4.0, (2 MHz~18 GHz)        | $U_{\text{rel}}=2\%$              |      |                |
|    |                               | Attenuation                        |                                               | (0~30) dB , (2 MHz~18 GHz)     | $U=0.3\text{dB}$                  |      |                |
|    |                               | RF power                           |                                               | (-50~20) dBm , (2 MHz~18 GHz)  | $U=0.25\text{dB}$                 |      |                |
|    |                               | Distance-to-fault                  |                                               | (1~15) m , (2 MHz~18 GHz)      | $U_{\text{rel}}=2\%$              |      |                |
| 51 | *Digital Storage Oscilloscope | bandwidth                          | C.S.for Digital Storage Oscilloscope JJF 1057 | 30kHz~50GHz                    | $U_{\text{rel}}=2\%$              |      |                |



No. CNAS L0730

第 110 页 共 163 页

The scope of the accreditation in Chinese remains the definitive version.

| №  | Instrument                      | Measurand                                              | Calibration Method                                  | Range                              | Expanded Uncertainty<br>( $k=2$ ) | Note | Effective Date |
|----|---------------------------------|--------------------------------------------------------|-----------------------------------------------------|------------------------------------|-----------------------------------|------|----------------|
|    |                                 | rise time                                              |                                                     | 20ps~100 $\mu$ s                   | $U_{rel}=3\%$                     |      |                |
|    |                                 | overshoot                                              |                                                     | 0.1%~30%                           | $U_{rel}=2\%$                     |      |                |
|    |                                 | Time measurement                                       |                                                     | 2ns/div~5s/div                     | $U_{rel}=1\%$                     |      |                |
|    |                                 | Amplitude measurement                                  |                                                     | 5mV/div~10V/div                    | $U_{rel}=(0.7\sim0.3)\%$          |      |                |
|    |                                 | DC gain accuracy                                       |                                                     | $\pm(1\text{mV}\sim200\text{V})$   | $U_{rel}=(0.7\sim0.3)\%$          |      |                |
|    |                                 | Square wave calibration signal frequency               |                                                     | 0.5kHz~1MHz                        | $U_{rel}=0.003\%$                 |      |                |
|    |                                 | Dynamic Significant Digits                             |                                                     | (4~14) bits                        | $U=0.1\text{bits}$                |      |                |
|    |                                 | acquisition rate                                       |                                                     | 100MS/s~500GS/s                    | $U_{rel}=1\times10^{-6}$          |      |                |
|    |                                 | Scan time factor                                       |                                                     | $1\times10^{-5}\sim1\times10^{-8}$ | $U_{rel}=1\times10^{-6}$          |      |                |
|    |                                 | Impedance                                              |                                                     | 50 $\Omega$ 、 1M $\Omega$          | $U_{rel}=0.3\%$                   |      |                |
|    |                                 | Calibrate signal amplitude                             |                                                     | (0.01~5)V                          | $U_{rel}=(1.5\sim0.5)\%$          |      |                |
| 52 | *5G Radio Communication Testers | Reference crystal oscillator output frequency accuracy | C.S.for 5G Radio Communication Testers JJF (通信) 052 | 10 MHz                             | $U_{rel}=6.7\times10^{-9}$        |      |                |




No. CNAS L0730

第 111 页 共 163 页

The scope of the accreditation in Chinese remains the definitive version.

| № | Instrument | Measurand                                                  | Calibration Method | Range                                    | Expanded Uncertainty<br>( $k=2$ )   | Note | Effective Date |
|---|------------|------------------------------------------------------------|--------------------|------------------------------------------|-------------------------------------|------|----------------|
|   |            | Frequency accuracy of output signal of RF signal generator |                    | 450 MHz~50 GHz                           | $U_{\text{rel}}=6.7 \times 10^{-9}$ |      |                |
|   |            | Rf signal generator output level                           |                    | (-120~0) dBm<br>(CW:450 MHz~50 GHz)      | $U=0.30 \text{ dB}$                 |      |                |
|   |            | Frequency spectrum purity of RF signal generator           |                    | (-120~-10) dBc<br>(CW:450 MHz~50 GHz)    | $U=1.0 \text{ dB}$                  |      |                |
|   |            | Rf signal generator single sideband phase noise            |                    | (-140~-10) dBc/Hz<br>(CW:450 MHz~50 GHz) | $U=1.0 \text{ dB}$                  |      |                |
|   |            | Digital modulation quality of 5G signal generator (EVM)    |                    | (0~8)% (CW:450 MHz~50 GHz)               | $U=0.6\%$                           |      |                |
|   |            | 5G signal generators occupy bandwidth                      |                    | (5~400) MHz<br>(CW:450 MHz~50 GHz)       | $U_{\text{rel}}=1.0\%$              |      |                |



No. CNAS L0730

第 112 页 共 163 页

The scope of the accreditation in Chinese remains the definitive version.

在线扫码获取验证

| No | Instrument | Measurand                                                              | Calibration Method | Range                            | Expanded Uncertainty<br>( $k=2$ ) | Note                              | Effective Date |
|----|------------|------------------------------------------------------------------------|--------------------|----------------------------------|-----------------------------------|-----------------------------------|----------------|
|    |            | Adjacent channel power ratio of 5G signal generator                    |                    | (0~-90)dBc, CW:450 MHz~50 GHz)   | $U=0.9$ dB                        |                                   |                |
|    |            | Rf analyzer                                                            |                    | (-50~30) (CW:450 MHz~44 GHz)     | $U=0.30$ dB                       |                                   |                |
|    |            | RF power analysis                                                      |                    | (-50~10) dBm (CW:6Hz~44 GHz)     | $U=0.50$ dB                       |                                   |                |
|    |            | Quality Parameter Analysis of Digital Modulation of 5G NR Signal (EVM) |                    | (0~17.5)%, (CW:450 MHz~44 GHz)   | $U=0.6\%$                         |                                   |                |
|    |            | 5G NR signal analyzer occupies bandwidth                               |                    | (5~400) MHz, (CW:450 MHz~44 GHz) | $U_{rel}=1.0\%$                   |                                   |                |
|    |            | Adjacent channel power ratio of 5G NR signal analyzer                  |                    | (0~-90)dBc, CW:450 MHz~50 GHz)   | $U=0.9$ dB                        |                                   |                |
|    |            | VSWR of RF port                                                        |                    | (1.001~1.5) (CW:450 MHz~50 GHz)  | $U=0.02$                          |                                   |                |
|    |            | shielding effectiveness                                                |                    | 5 dB~120 dB, (CW:14 kHz~18GHz)   | $U=3.0$ dB                        | The frequency range from 14kHz to |                |



No. CNAS L0730

第 113 页 共 163 页

The scope of the accreditation in Chinese remains the definitive version.

| No | Instrument                                             | Measurand                                    | Calibration Method                                                     | Range                        | Expanded Uncertainty<br>(k=2) | Note                                                                                                                                      | Effective Date |
|----|--------------------------------------------------------|----------------------------------------------|------------------------------------------------------------------------|------------------------------|-------------------------------|-------------------------------------------------------------------------------------------------------------------------------------------|----------------|
|    |                                                        | VSWR of RF<br>Input/Output<br>port           |                                                                        | (1~5), (CW:14 kHz~18<br>GHz) | $U=0.05$                      | 200kHz is<br>accredited<br>only for at<br>50cm ×<br>30cm ×<br>10cm Box<br>size above,<br>400MHz to<br>1GHz is<br>25cm ×<br>25cm ×<br>5cm。 |                |
| 54 | *Calibration<br>Device for Electric<br>Cardiac Monitor | Amplitude<br>(voltage)                       | V.R. of Calibration Device<br>for Electric Cardiac Monitor<br>JJG 1016 | 0.2 mV~2m V                  | $U_{rel}=0.3\%$               |                                                                                                                                           |                |
|    |                                                        | period                                       |                                                                        | 100 ms~20s                   | $U_{rel}=0.1\%$               |                                                                                                                                           |                |
|    |                                                        | frequency                                    |                                                                        | 0.1 Hz~100 Hz                | $U_{rel}=0.1\%$               |                                                                                                                                           |                |
|    |                                                        | distortion                                   |                                                                        | 0.01%~30% (20Hz~<br>100Hz)   | $U_{rel}=5.0\%$               |                                                                                                                                           |                |
|    |                                                        | polarization<br>voltage                      |                                                                        | -10V~10V                     | $U=0.01mV$                    |                                                                                                                                           |                |
|    |                                                        | Standard heart<br>rate                       |                                                                        | 27 次/min~300 次/min           | $U_{rel}=0.1\%$               |                                                                                                                                           |                |
|    |                                                        | Standard heart<br>rate signal<br>spike width |                                                                        | (50~150)ms                   | $U_{rel}=0.1\%$               |                                                                                                                                           |                |
|    |                                                        | Standard heart<br>rate signal<br>amplitude   |                                                                        | ± (0.2 mV~5mV)               | $U_{rel}=0.3\%$               |                                                                                                                                           |                |



No. CNAS L0730

第 114 页 共 163 页

The scope of the accreditation in Chinese remains the definitive version.

| №  | Instrument                          | Measurand                              | Calibration Method                                       | Range                                                 | Expanded Uncertainty<br>( $k=2$ ) | Note                                | Effective Date |
|----|-------------------------------------|----------------------------------------|----------------------------------------------------------|-------------------------------------------------------|-----------------------------------|-------------------------------------|----------------|
|    |                                     | Input loop current sampling resistance |                                                          | $(5\sim 50)\text{k}\Omega$                            | $U_{\text{rel}}=0.5\%$            |                                     |                |
| 55 | *Medium power meter                 | direct-current power                   | C.S.for Medium Power Meters JJF 1386                     | 0.1W~9W                                               | $U_{\text{rel}}=1.2\%$            | Except for transfer Standard Method |                |
|    |                                     | calibration factor/RF power            |                                                          | 10%~200%,<br>(1MHz~18GHz)                             | $U_{\text{rel}}=4.0\%$            |                                     |                |
|    |                                     | voltage standing wave ratio            |                                                          | (1~2), (1MHz~18GHz)                                   | $U=0.05$                          |                                     |                |
| 56 | *Broadband Coaxial Noise Generators | excess noise ratio                     | C.S.for Broadband Coaxial Noise Generators JJF 1442      | (1~30)dB,<br>(10MHz~40GHz)                            | $U=0.26\text{dB}$                 |                                     |                |
|    |                                     | voltage standing wave ratio            |                                                          | (1~1.5),<br>(10MHz~40GHz)                             | $U=0.05$                          |                                     |                |
| 57 | *Plus generator of EMI Calibration  | pulse repetition rate                  | C.S.of Pulse Generator for EMI Calibration JJF(电子) 30814 | 0.1Hz~1000Hz                                          | $U_{\text{rel}}=0.3\%$            |                                     |                |
|    |                                     | pulse area                             |                                                          | $0.01\text{ }\mu\text{Vs}\sim 20\text{ }\mu\text{Vs}$ | $U_{\text{rel}}=2.0\%$            |                                     |                |
|    |                                     | impulse amplitude                      |                                                          | (0~80)dB,<br>(9kHz~1GHz)                              | $U=0.3\text{dB}$                  |                                     |                |
|    |                                     | Flatness of the pulse spectrum         |                                                          | $\pm(0\sim 5)\text{dB}$ , (9kHz~1GHz)                 | $U=0.35\text{dB}$                 |                                     |                |
|    |                                     | pulse width                            |                                                          | 0.1ns~10 $\mu\text{s}$ , (9kHz~1GHz)                  | $U_{\text{rel}}=2.0\%$            |                                     |                |
| 58 | *GNSS Signal Simulators             | integrator carrier frequency deviation | C.S.for GNSS Signal Simulators JJF 1471                  | 1000MHz~3000MHz                                       | $U_{\text{rel}}=3\times 10^{-9}$  |                                     |                |

No. CNAS L0730

第 115 页 共 163 页



The scope of the accreditation in Chinese remains the definitive version.

| №  | Instrument   | Measurand                       | Calibration Method       | Range                               | Expanded Uncertainty<br>( $k=2$ )  | Note | Effective Date |
|----|--------------|---------------------------------|--------------------------|-------------------------------------|------------------------------------|------|----------------|
|    |              | power range                     |                          | (-150~-20)dBm, 1000MHz~3000MHz      | $U=0.20\text{dB} \sim 1\text{dB}$  |      |                |
|    |              | velocity dynamic range          |                          | (0~36000)m/s                        | $U=1\text{m/s}$                    |      |                |
|    |              | acceleration dynamic range      |                          | (0~2000)m/s <sup>2</sup>            | $U=0.3\text{m/s}^2$                |      |                |
|    |              | plus acceleration dynamic range |                          | (0~2000)m/s <sup>3</sup>            | $U=0.6\text{m/s}^3$                |      |                |
|    |              | internal channel delay          |                          | (0~15)ns, (1000MHz~3000MHz)         | $U=0.14\text{ns}$                  |      |                |
|    |              | harmonic suppression            |                          | (-60~-20)dB, (1GHz~10GHz)           | $U=1.2\text{dB}$                   |      |                |
|    |              | non-harmonic suppression        |                          | (-80~-30)dB, (10kHz~10GHz)          | $U=1.2\text{dB}$                   |      |                |
|    |              | phase noise                     |                          | (-60~-130)dBc/Hz, (1000MHz~3000MHz) | $U=1.6\text{dB}$                   |      |                |
|    |              | time-base accuracy              |                          | 10MHz                               | $U_{\text{rel}}=3 \times 10^{-9}$  |      |                |
|    |              | frequency stability             |                          | 10MHz                               | $U_{\text{rel}}=6 \times 10^{-11}$ |      |                |
|    |              | amplitude of error vector       |                          | 0.5%~20%, (1000MHz~3000MHz)         | $U=0.7\%$                          |      |                |
| 59 | *Power Meter | calibration                     | C.S.for RF and Microwave | 30%~130%(9kHz~6GHz)                 | $U_{\text{rel}}=1.8\%$             |      |                |

No. CNAS L0730

第 116 页 共 163 页



在线扫码获取验证

The scope of the accreditation in Chinese remains the definitive version.

| №  | Instrument                                                                | Measurand                   | Calibration Method                                                                       | Range                              | Expanded Uncertainty<br>( $k=2$ ) | Note | Effective Date |
|----|---------------------------------------------------------------------------|-----------------------------|------------------------------------------------------------------------------------------|------------------------------------|-----------------------------------|------|----------------|
|    |                                                                           | factor                      | Power Meters JJF 1885                                                                    | 30%~130%(6GHz~18GHz)               | $U_{rel}=2.2\%$                   |      |                |
|    |                                                                           |                             |                                                                                          | 30%~130%(>18GHz~26GHz)             | $U_{rel}=3.4\%$                   |      |                |
|    |                                                                           |                             |                                                                                          | 30%~130%(>26GHz~40GHz)             | $U_{rel}=4.2\%$                   |      |                |
|    |                                                                           |                             |                                                                                          | 30%~130%(>40GHz~50GHz)             | $U_{rel}=5.0\%$                   |      |                |
|    |                                                                           | reference source power      |                                                                                          | 1mW, (50MHz)                       | $U_{rel}=0.8\%$                   |      |                |
|    |                                                                           | Linearity                   |                                                                                          | (-2~2)dB, (-20dBm~10dBm) (50MHz)   | $U=0.10\text{dB}$                 |      |                |
|    |                                                                           | voltage standing wave ratio |                                                                                          | 1.0~2, (9kHz~50GHz)                | $U=0.05$                          |      |                |
| 60 | *Coupling Decoupling Networks                                             | Common Impedance            | C.S.for Coupling Decoupling Networks JJF(电子)0034                                         | (30~400) $\Omega$<br>(0.15~300)MHz | $U_{rel}=5\%$                     |      |                |
|    |                                                                           | Coupling Factor             |                                                                                          | (-20dB~0dB),<br>(0.15~300)MHz      | $U=1.2\text{dB}$                  |      |                |
| 61 | *Verification Instrument for Electrocardiograph and Electroencephalograph | Amplitude                   | V.R. of Verification Instrument for Electrocardiograph and Electroencephalograph JJG 749 | 0.1mV~10V                          | $U_{rel}=0.1\%$                   |      |                |
|    |                                                                           | sine wave Frequency         |                                                                                          | (0.1~1000)Hz                       | $U_{rel}=0.1\%$                   |      |                |
|    |                                                                           | square Period               |                                                                                          | 1ms~50s                            | $U_{rel}=0.1\%$                   |      |                |
|    |                                                                           | Polarization voltage        |                                                                                          | $\pm(0.1\sim1)\text{V}$            | $U_{rel}=0.1\%$                   |      |                |
|    |                                                                           | sine wave Distortion        |                                                                                          | 0.01%~10%, (20Hz~200Hz)            | $U_{rel}=5\%$                     |      |                |
|    |                                                                           | Attenuation ratio           |                                                                                          | 1/100~1/2000                       | $U_{rel}=0.1\%$                   |      |                |



No. CNAS L0730

第 117 页 共 163 页

The scope of the accreditation in Chinese remains the definitive version.

| No | Instrument           | Measurand    | Calibration Method                             | Range                         | Expanded Uncertainty<br>(k=2)    | Note | Effective Date |
|----|----------------------|--------------|------------------------------------------------|-------------------------------|----------------------------------|------|----------------|
| 62 | *LCR Meter           | Inductance   | V.R. for wide range digital RLC meter GJB 8817 | 10 μ H~100mH, (1kHz)          | U <sub>rel</sub> =(2.3~0.06)%    |      |                |
|    |                      |              |                                                | 1H (1kHz) , 10H (100Hz)       | U <sub>rel</sub> =0.12%          |      |                |
|    |                      | Capacitance  |                                                | 1pF~1nF (1kHz~13MHz)          | U <sub>rel</sub> =(0.04~1.5)%    |      |                |
|    |                      |              |                                                | 1nF~1 μ F, (1kHz)             | U <sub>rel</sub> =0.06%          |      |                |
|    |                      |              |                                                | 1 μ F~1mF, ( 100Hz)           | U <sub>rel</sub> = (0.06~0.09) % |      |                |
|    |                      | Resistance   |                                                | 10m Ω ~100k Ω , (100Hz~10MHz) | U <sub>rel</sub> =(0.05~1.5)%    |      |                |
|    |                      |              |                                                | 1 Ω ~100k Ω , (1kHz)          | U <sub>rel</sub> =0.024%         |      |                |
|    |                      |              |                                                | 100k Ω ~1.1M Ω , (1kHz)       | U <sub>rel</sub> =0.12%          |      |                |
|    |                      | Frequency    |                                                | 10Hz~13MHz                    | U <sub>rel</sub> =0.002%         |      |                |
|    |                      | Voltage      |                                                | 1mV~20V, (10Hz~1kHz)          | U <sub>rel</sub> =(6~0.20)%      |      |                |
|    |                      |              |                                                | 1mV~1V, (300kHz~1kHz)         | U <sub>rel</sub> =(6~0.20)%      |      |                |
|    |                      |              |                                                | 1V~20V, (1kHz~300kHz)         | U <sub>rel</sub> =(0.20~6)%      |      |                |
|    |                      |              |                                                | 10mV~2V, (300kHz~10MHz)       | U <sub>rel</sub> =(2.8~1.5)%     |      |                |
|    |                      | Depreciation |                                                | 0.0001~1, (1kHz)              | U <sub>rel</sub> =0.24%          |      |                |
|    |                      |              |                                                | 0.0001~0.01, (1kHz~10MHz)     | U=0.0003                         |      |                |
| 63 | *Standard Capacitors | Capacitance  | V.R. of Standard Capacitor JJG 183             | 1pF~100 μ F, (20Hz~1MHz)      | U <sub>rel</sub> =(2.1~0.06)%    |      |                |



No. CNAS L0730

第 118 页 共 163 页

The scope of the accreditation in Chinese remains the definitive version.

| №  | Instrument                                                                     | Measurand                            | Calibration Method                                           | Range                                      | Expanded Uncertainty<br>( $k=2$ ) | Note | Effective Date |
|----|--------------------------------------------------------------------------------|--------------------------------------|--------------------------------------------------------------|--------------------------------------------|-----------------------------------|------|----------------|
|    |                                                                                |                                      |                                                              | 100 $\mu$ F $\sim$ 100mF(20Hz $\sim$ 1kHz) | $U_{rel}=(0.06\sim 0.29)\%$       |      |                |
|    |                                                                                | Depreciation                         |                                                              | 0.0001 $\sim$ 1(20Hz $\sim$ 1MHz)          | $U=0.0006\sim 0.004$              |      |                |
| 64 | *Standard Inductors                                                            | Inductance                           | V.R. of Standard Inductors JJG 726                           | 10 $\mu$ H $\sim$ 1H(1kHz)                 | $U_{rel}=(0.12\sim 0.06)\%$       |      |                |
|    |                                                                                |                                      |                                                              | 1H $\sim$ 10H(100Hz $\sim$ 120Hz)          | $U_{rel}=0.12\%$                  |      |                |
|    |                                                                                | DC Resistor                          |                                                              | 1m $\Omega \sim$ 100k $\Omega$             | $U_{rel}=(1.2\sim 0.12)\%$        |      |                |
| 65 | AC Resistors Box                                                               | AC Resistance                        | C.S. for A.C. Resistance Boxes JJF 1636                      | 10m $\Omega \sim$ 1M $\Omega$              | $U_{rel}=(0.06\sim 14)\%$         |      |                |
|    |                                                                                | time constant                        |                                                              | 0.022ns $\sim$ 1ms                         | $U_{rel}=(6\sim 0.28)\%$          |      |                |
|    |                                                                                | residual resistance                  |                                                              | 1m $\Omega \sim$ 1 $\Omega$                | $U_{rel}=(0.4\sim 0.03)\%$        |      |                |
|    |                                                                                | Residual inductance                  |                                                              | 0.1 $\mu$ H $\sim$ 100 $\mu$ H(1kHz)       | $U=(0.05\sim 0.08) \mu$ H         |      |                |
| 66 | *Verification Regulation of Calibrator for Transistor Specificity Oscilloscope | DCV                                  | C.S. for Calibrators of Semiconductor Curve Tracers JJF 1894 | (0.01 $\sim$ 1000)V                        | $U_{rel}=(0.11\sim 0.04)\%$       |      |                |
|    |                                                                                | resistance                           |                                                              | 0.1 $\Omega \sim$ 10M $\Omega$             | $U_{rel}=(0.06\sim 0.10)\%$       |      |                |
|    |                                                                                | DCI                                  |                                                              | 10nA $\sim$ 10A                            | $U_{rel}=(0.5\sim 0.15)\%$        |      |                |
| 67 | *Stopwatch                                                                     | Time interval (Electronic Stopwatch) | V.R. of Stopwatches JJG 237                                  | (1 $\sim$ 99999)s                          | $U=(5\sim 6)$ ms                  |      |                |
|    |                                                                                | Time interval (Mechanical Stopwatch) |                                                              | (1 $\sim$ 99999)s                          | $U=(0.05\sim 0.08)$ s             |      |                |



No. CNAS L0730

第 119 页 共 163 页

The scope of the accreditation in Chinese remains the definitive version.


| №  | Instrument                                         | Measurand                                     | Calibration Method                                                   | Range          | Expanded Uncertainty<br>( $k=2$ )                | Note | Effective Date |
|----|----------------------------------------------------|-----------------------------------------------|----------------------------------------------------------------------|----------------|--------------------------------------------------|------|----------------|
|    |                                                    | Time interval<br>(electrical stopwatches)     |                                                                      | 0.1ms~9999.9s  | $U_{rel}=2 \times 10^{-3} \sim 2 \times 10^{-7}$ |      |                |
|    |                                                    | Time error in a day<br>(electronic stopwatch) |                                                                      | (-99~99)s      | $U=0.023s$                                       |      |                |
| 68 | *Instantaneous daily clock time difference testers | Instantaneous diurnal difference              | V.R. of instantaneous daily clock time difference testers<br>JJG 488 | (-99~+99)s     | $U=0.008s$                                       |      |                |
|    |                                                    | Instantaneous monthly difference              |                                                                      | (-1000~+1000)s | $U=0.24s$                                        |      |                |
|    |                                                    | Internal crystal oscillator frequency         |                                                                      | 1MHz~10MHz     | $U_{rel}=5 \times 10^{-9}$                       |      |                |
| 69 | *Time interval meter                               | Warm up                                       | V.R. of time interval meters<br>JJG 238                              | 1MHz~10MHz     | $U_{rel}=6 \times 10^{-8}$                       |      |                |
|    |                                                    | Daily frequency fluctuations                  |                                                                      | 1MHz~10MHz     | $U_{rel}=6 \times 10^{-8}$                       |      |                |
|    |                                                    | 1s frequency stability                        |                                                                      | 1MHz~10MHz     | $U_{rel}=6 \times 10^{-8}$                       |      |                |
|    |                                                    | Frequency Reproducibility                     |                                                                      | 1MHz~10MHz     | $U_{rel}=6 \times 10^{-8}$                       |      |                |
|    |                                                    | Relative frequency deviation                  |                                                                      | 1MHz~10MHz     | $U_{rel}=6 \times 10^{-8}$                       |      |                |



No. CNAS L0730

第 120 页 共 163 页

The scope of the accreditation in Chinese remains the definitive version.

| No                            | Instrument                                                                  | Measurand                    | Calibration Method                                                                                                      | Range               | Expanded Uncertainty<br>(k=2)          | Note                                                                                 | Effective Date |
|-------------------------------|-----------------------------------------------------------------------------|------------------------------|-------------------------------------------------------------------------------------------------------------------------|---------------------|----------------------------------------|--------------------------------------------------------------------------------------|----------------|
|                               |                                                                             | time interval measurement    |                                                                                                                         | 10 μ s~10000 s      | U=(0.006~0.3)ms                        |                                                                                      |                |
| 70                            | *Electronic time relay                                                      | Time                         | C.S. for Electronic Time Relay JJF 1282                                                                                 | 1ms~9999s           | U=0.1ms                                |                                                                                      |                |
| 71                            | *C.S. for Time Relay Testers                                                | Time measurement             | C.S. for Time Relay Testers JJF 1400                                                                                    | 1ms~9999s           | U=0.03 ms                              |                                                                                      |                |
|                               |                                                                             | Warm up                      |                                                                                                                         | (1~10)MHz           | U <sub>rel</sub> =1×10 <sup>-8</sup>   |                                                                                      |                |
|                               |                                                                             | Frequency accuracy           |                                                                                                                         | (1~100)MHz          | U <sub>rel</sub> =1×10 <sup>-8</sup>   |                                                                                      |                |
|                               |                                                                             | DC voltage output            |                                                                                                                         | (10~400) V          | U <sub>rel</sub> =0.5%                 |                                                                                      |                |
|                               |                                                                             | AC voltage output            |                                                                                                                         | (10~400) V          | U <sub>rel</sub> =0.5%                 |                                                                                      |                |
| 72                            | *Residual Current Operated Protective Device Operated Characteristic Tester | Cut Timer                    | C.S. for Residual Current Operated Protective Device Operated Protective Device Operated Characteristic Tester JJF 1283 | 10ms~5s             | U=1ms                                  |                                                                                      |                |
|                               |                                                                             | Residual Current             |                                                                                                                         | (3~1500)mA          | U <sub>rel</sub> =1.1%                 |                                                                                      |                |
| 73                            | *Calibration Specification for Pulse Counters                               | Pulse Counters               | C.S. for Pulse Counters JJF1686                                                                                         | 1~1×10 <sup>6</sup> | U <sub>rel</sub> =1.6×10 <sup>-6</sup> |                                                                                      |                |
|                               |                                                                             | Built-in time base frequency |                                                                                                                         | (1~10)MHz           | U <sub>rel</sub> =5×10 <sup>-9</sup>   |                                                                                      |                |
| 8 Optics Measuring Instrument |                                                                             |                              |                                                                                                                         |                     |                                        |                                                                                      |                |
| 1                             | *Specular Gloss Meter                                                       | Specular Gloss               | V.R. of Specular Gloss Meters JJG 696                                                                                   | (0~100)GU           | U=1.3GU                                |  |                |
| 2                             | *Colorimeters and Color Difference Meters                                   | Chroma                       | V.R. of Colorimeters and Color Difference Meters JJG 595                                                                | Y: (0.1~100)        | U=2.1(Y)                               |                                                                                      |                |
|                               |                                                                             |                              |                                                                                                                         | x,y:(0~1)           | U=0.007(x) U=0.007(y)                  |                                                                                      |                |



在线扫码获取验证

No. CNAS L0730

第 121 页 共 163 页

The scope of the accreditation in Chinese remains the definitive version.

| №  | Instrument                    | Measurand         | Calibration Method                                | Range                                       | Expanded Uncertainty<br>( $k=2$ ) | Note | Effective Date |
|----|-------------------------------|-------------------|---------------------------------------------------|---------------------------------------------|-----------------------------------|------|----------------|
| 3  | Illuminance Meter             | Illuminance       | V.R. of Illuminance Meter JJG 245                 | (1~3000) lx                                 | $U_{\text{rel}}=1.3\%$            |      |                |
| 4  | *Whiteness Meter              | Whiteness         | V.R. of Whiteness Meter JJG 512                   | 0~100                                       | $U=2.0$                           |      |                |
| 5  | *Standard Light Sources Boxes | Color Temperature | C.S. for Standard Light Sources Boxes JJF(EZ) 055 | (2000~2856)K                                | $U=31\text{K}$                    |      |                |
|    |                               |                   |                                                   | (2857~5500)K                                | $U=0.11\text{kK}$                 |      |                |
|    |                               |                   |                                                   | (5501~6500)K                                | $U=0.14\text{kK}$                 |      |                |
| 6  | *Clarity Test Equipment       | Illumination      | C.S. for Clarity Test Equipment JJF1287           | (50~3000)lx                                 | $U_{\text{rel}}=11\%$             |      |                |
|    |                               | time              |                                                   | (1~60)s                                     | $U=0.2\text{s}$                   |      |                |
| 7  | *Hazemeter                    | Haze              | C.S. for Hazemeter JJF 1303                       | 0~30                                        | $U=0.41$                          |      |                |
|    |                               | Transmittance     |                                                   | 0~100%                                      | $U=1.1\%$                         |      |                |
| 8  | *Trial Case Lenses            | Spherical Degree  | V.R. of Trial Case Lenses JJG 579                 | $(-20 \sim +20) \text{ m}^{-1}$             | $U=0.03 \text{ m}^{-1}$           |      |                |
|    |                               | Cylinder Degree   |                                                   | $(-1.5 \sim +1.5) \text{ m}^{-1}$           | $U=0.03 \text{ m}^{-1}$           |      |                |
|    |                               | Prismatic Power   |                                                   | $(2 \sim 20) \text{ cm/m}$                  | $U=0.013 \text{ cm/m}$            |      |                |
| 9  | *Focimeters                   | Spherical Degree  | V.R. of Focimeters JJG 580                        | $(-25 \sim +25) \text{ m}^{-1}$             | $U=0.03 \text{ m}^{-1}$           |      |                |
|    |                               | Cylinder Degree   |                                                   | $(-1.5 \sim +1.5) \text{ m}^{-1}$           | $U=0.03 \text{ m}^{-1}$           |      |                |
|    |                               | Prismatic Degree  |                                                   | $(2 \sim 10) \text{ cm/m}$                  | $U=0.03 \text{ cm/m}$             |      |                |
| 10 | *Eye Refractometers           | Vertex Power      | V.R. of Eye Refractometers JJG 892                | objective: $(-20 \sim +20) \text{ m}^{-1}$  | $U=0.08 \text{ m}^{-1}$           |      |                |
|    |                               |                   |                                                   | subjective: $(-15 \sim +15) \text{ m}^{-1}$ | $U=0.08 \text{ m}^{-1}$           |      |                |



No. CNAS L0730

第 122 页 共 163 页

The scope of the accreditation in Chinese remains the definitive version.

| №  | Instrument                                                      | Measurand                    | Calibration Method                                        | Range                     | Expanded Uncertainty<br>( $k=2$ ) | Note | Effective Date |
|----|-----------------------------------------------------------------|------------------------------|-----------------------------------------------------------|---------------------------|-----------------------------------|------|----------------|
|    |                                                                 | Pupil Distance               |                                                           | (55~75)mm                 | $U=0.3\text{mm}$                  |      |                |
|    |                                                                 | Curvature radius             |                                                           | (6.5~9.4)mm               | $U=0.004\text{ mm}$               |      |                |
|    |                                                                 | Axis                         |                                                           | (0~180)°                  | $U=1.2^\circ$                     |      |                |
| 11 | *Pupil Distance Meter                                           | Pupil Distance               | V.R. of Pupil Distance Meters JJG 952                     | (55~75) mm                | $U=0.2\text{mm}$                  |      |                |
| 12 | *Ophthalmometers                                                | Curvature radius             | V.R. of Ophthalmometers JJG 1011                          | (6.5~9.4)mm               | $U=0.004\text{ mm}$               |      |                |
|    |                                                                 | Axis                         |                                                           | (0~180)°                  | $U=1.2^\circ$                     |      |                |
| 13 | *Phoropters (Eye Chart Included)                                | Sphericity                   | V.R. of Phoropters (Eye Chart Included) JJG 1097          | (-20~+20) $\text{m}^{-1}$ | $U_S=(0.05\sim0.08)\text{m}^{-1}$ |      |                |
|    |                                                                 | Columnar mirroriness         |                                                           | (-20~+20) $\text{m}^{-1}$ | $U_S=(0.05\sim0.08)\text{m}^{-1}$ |      |                |
|    |                                                                 | Optical center               |                                                           | (0~1.0)cm/m               | $U_p=0.05\text{cm/m}$             |      |                |
|    |                                                                 | Column mirror axial position |                                                           | (0~180)°                  | $U=1^\circ$                       |      |                |
|    |                                                                 | Visual scale size            |                                                           | (2~80)mm                  | $U_{\text{rel}}=3.0\%$            |      |                |
| 14 | *Abbe Refractometer                                             | Refractive index             | V.R. of Abbe Refractometer JJG 625                        | $n_D: 1.300\sim1.700$     | $U=1\times10^{-4}$                |      |                |
| 15 | *Optical Power Meters in Telecommunication/Optical Power Meters | Power                        | V.R. of Optical Power Meters in Telecommunication JJG 965 | -60dBm~+10 dBm            | $U=(0.20\sim0.50)\text{dB}$       |      |                |
| 16 | *Fiber Optical Power Meters                                     | Power                        | V.R. of Fiber Optical Power Meters JJG 813                | -60dBm~+10 dBm            | $U=(0.20\sim0.50)\text{dB}$       |      |                |



No. CNAS L0730

第 123 页 共 163 页

The scope of the accreditation in Chinese remains the definitive version.

| №  | Instrument                                                        | Measurand                        | Calibration Method                                              | Range                             | Expanded Uncertainty<br>( $k=2$ )   | Note | Effective Date |
|----|-------------------------------------------------------------------|----------------------------------|-----------------------------------------------------------------|-----------------------------------|-------------------------------------|------|----------------|
| 17 | *Optical Spectrum Analyzers                                       | Wavelength                       | V.R.of Optical Spectrum Analyzers in Telecommunication JJG 1035 | 900nm~1700nm                      | $U_{\text{rel}}=2 \times 10^{-6}$   |      |                |
|    |                                                                   | Power                            |                                                                 | -60dBm~+10 dBm                    | $U=0.20\text{dB}$                   |      |                |
|    |                                                                   | Resolution bandwidth             |                                                                 | (0.01~5)nm                        | $U_{\text{rel}}=0.7\%$              |      |                |
| 18 | *Optical Wavelength Counters                                      | Wavelength                       | V.R.of Optical Wavelength Counters in Telecommunication JJG 963 | 900nm~1200nm                      | $U_{\text{rel}}=2 \times 10^{-6}$   |      |                |
|    |                                                                   |                                  |                                                                 | 1200nm~1700nm                     | $U_{\text{rel}}=0.5 \times 10^{-6}$ |      |                |
|    |                                                                   | power                            |                                                                 | -60dBm~+10 dBm                    | $U=(0.20\sim0.50)\text{dB}$         |      |                |
| 19 | *Stabilized Laser Sources for Optical Transmit/ Laser Sources     | Wavelength                       | V.R.of Stabilized Laser Sources for Optical Transmit JJG 958    | 900nm~1700nm                      | $U=0.05\text{nm}$                   |      |                |
|    |                                                                   | Output power                     |                                                                 | -60dBm~+10 dBm                    | $U=0.3\text{dB}$                    |      |                |
|    |                                                                   | Output power stability           |                                                                 | -60dBm~+10 dBm                    | $U=0.005\text{dB}$                  |      |                |
| 20 | *Tunable Laser Source for Telecommunications/Tunable Laser Source | Optical wavelength               | C.S.of Tunable Laser Source for Telecommunications JJF 1198     | 1260nm~1640nm                     | $U=0.5\text{pm}$                    |      |                |
|    |                                                                   | Optical wavelength repeatability |                                                                 | 1260nm~1640nm                     | $U=0.5\text{pm}$                    |      |                |
|    |                                                                   | Optical wavelength stability     |                                                                 | 1260nm~1640nm                     | $U=0.5\text{pm}$                    |      |                |
|    |                                                                   | Optical output power             |                                                                 | -60dBm~+10 dBm                    | $U=0.3\text{dB}$                    |      |                |
|    |                                                                   | Optical output power stability   |                                                                 | -1.000dB~1.000dB (-60dBm~+10 dBm) | $U=0.005\text{dB}$                  |      |                |
|    |                                                                   |                                  |                                                                 |                                   |                                     |      |                |
| 21 | *Optical Attenuator for Telecommunications                        | Attenuation                      | C.S.of Optical Attenuator for Telecommunications JJF 1199       | (0~65) dB                         | $U=0.03\text{ dB}$                  |      |                |



No. CNAS L0730

第 124 页 共 163 页

The scope of the accreditation in Chinese remains the definitive version.

| №  | Instrument                         | Measurand                   | Calibration Method                                                | Range                                                 | Expanded Uncertainty<br>( $k=2$ )         | Note                                                                                                 | Effective Date |
|----|------------------------------------|-----------------------------|-------------------------------------------------------------------|-------------------------------------------------------|-------------------------------------------|------------------------------------------------------------------------------------------------------|----------------|
|    | ns/Optical Attenuator              | Insertion loss              |                                                                   | (0~5) dB                                              | $U=0.04\text{dB}$                         |                                                                                                      |                |
| 22 | *Optical Return Loss Meters        | Optical Return Loss         | C.S.for Optical Return Loss Meters for Telecommunication JJF 1325 | (0~60) dB                                             | $U=(0.6\sim 2.0)\text{ dB}$               |                                                                                                      |                |
| 23 | *optical time-domain reflectometer | Optical wavelength          | V.R. of Optical Time Domain Reflectometer JJG 959                 | (600~1700)nm                                          | $U=0.05\text{nm}$                         |                                                                                                      |                |
|    |                                    | Distance scale coefficient  |                                                                   | 0.1~2                                                 | $U_{\text{rel}}=2\times 10^{-5}$          |                                                                                                      |                |
|    |                                    | Distance/Position deviation |                                                                   | (0~40)km                                              | $U=0.3\text{m}+1.5\times 10^{-5}\text{L}$ |                                                                                                      |                |
|    |                                    | Loss scale factor           |                                                                   | (0~10) dB/dB                                          | $U=0.04\text{dB/dB}$                      |                                                                                                      |                |
|    |                                    | Loss characteristics        |                                                                   | (0~60) dB/km                                          | $U=0.04\text{dB/km}$                      |                                                                                                      |                |
| 24 | LED                                | Total Light Flux            | C.S. for Single Low Power LED JJF1501                             | (0.5~200)lm                                           | $U_{\text{rel}}=4.6\%$                    | Do average luminous intensity calibration by standard detector measurement method on CIE B condition |                |
|    |                                    | Chromaticity coordinates    |                                                                   | x, y: (0~1)                                           | $U=0.008(x)$ $U=0.008(y)$                 |                                                                                                      |                |
|    |                                    | Average Luminous Intensity  |                                                                   | (0.1~100) cd                                          | $U_{\text{rel}}=4.5\%$                    |                                                                                                      |                |
| 25 | Instantaneous Spectral Instrument  | Spectral Radiant Intensity  | C.S.for Instantaneous Spectral Instruments JJF1329                | (250~1700)nm:<br>(0.001~10) W/m <sup>2</sup> /nm      | $U_{\text{rel}}=8.7\%$                    |                                                                                                      |                |
| 26 | Pyranometer Sensor                 | Irradiance                  | C.S.for Pyranometer Sensors FFJ1409                               | (350~1300) W/m <sup>2</sup>                           | $U_{\text{rel}}=3.7\%$                    |                                                                                                      |                |
|    |                                    | Sensitivity                 |                                                                   | (6~15) $\mu\text{V}\cdot\text{W}^{-1}\cdot\text{m}^2$ | $U_{\text{rel}}=3.7\%$                    |                                                                                                      |                |



No. CNAS L0730

第 125 页 共 163 页

The scope of the accreditation in Chinese remains the definitive version.

| №                                | Instrument                        | Measurand               | Calibration Method                                                                        | Range                                                             | Expanded Uncertainty<br>( $k=2$ ) | Note                  | Effective Date |
|----------------------------------|-----------------------------------|-------------------------|-------------------------------------------------------------------------------------------|-------------------------------------------------------------------|-----------------------------------|-----------------------|----------------|
| 27                               | Solar Cell                        | Open Circuit Voltage    | C.S. for Solar Cells:<br>Photoelectric Properties<br>JJF1622                              | (0.001~5) V                                                       | $U_{rel}=1.2\%$                   |                       |                |
|                                  |                                   | Short Circuit Current   |                                                                                           | (0.001~10) A                                                      | $U_{rel}=2.7\%$                   |                       |                |
|                                  |                                   | Maximum Power           |                                                                                           | (0.01~50) W                                                       | $U_{rel}=3.0\%$                   |                       |                |
|                                  |                                   | Optimal Working Current |                                                                                           | (0.001~10) A                                                      | $U_{rel}=2.7\%$                   |                       |                |
|                                  |                                   | Optimal Working Voltage |                                                                                           | (0.001~5) V                                                       | $U_{rel}=1.2\%$                   |                       |                |
|                                  |                                   | Conversion Efficiency   |                                                                                           | (0.01~99.9)%                                                      | $U_{rel}=3.4\%$                   |                       |                |
| 28                               | Solar Cell Photoelectric Detector | Short Circuit Current   | C.S. for Reference Solar Cell<br>FFJ1903                                                  | (0.1~10000) mA                                                    | $U_{rel}=2.8\%$                   |                       |                |
|                                  |                                   | Sensitivity             |                                                                                           | (20~10000) mV·kW <sup>-1</sup> ·m <sup>2</sup>                    | $U_{rel}=2.8\%$                   |                       |                |
| 29                               | Spectral Radiance Lamp            | Spectral Radiance       | V.R. of Spectral Radiance Standard Lamp<br>JJG383, Spectral Radiance Lamp ZYZ226-2021     | (250~2500) nm:<br>(0.001~100) μW/cm <sup>2</sup> /nm/sr           | $U_{rel}=4.1\%~6.9\%$             |                       |                |
| 30                               | Spectral Irradiance Lamp          | Spectral Irradiance     | V.R. of Spectral Irradiance Standard Lamp<br>JJG384, Spectral Irradiance Lamp ZYZ225-2021 | (250~2500) nm:<br>(0.01~100) μW·cm <sup>2</sup> ·nm <sup>-1</sup> | $U_{rel}=2.7\%~5.3\%$             |                       |                |
| 31                               | Solar Simulator Irradiance Meter  | Irradiance              | C.S. for Solar Simulator Irradiance Meter FFJ1604                                         | (350~1300) W/m <sup>2</sup>                                       | $U_{rel}=2.8\%$                   |                       |                |
| 9 Chemistry Measuring Instrument |                                   |                         |                                                                                           |                                                                   |                                   |                       |                |
| 1                                | *Gas Chromatog-                   | Sensitivity             | V.R. of Gas Chromatographs                                                                | TCD: ≥800 mV·mL/mg                                                | $U_{rel}=4\%$                     | Accredited except for |                |

No. CNAS L0730

第 126 页 共 163 页



The scope of the accreditation in Chinese remains the definitive version.


| No | Instrument                            | Measurand                            | Calibration Method                                   | Range                                                 | Expanded Uncertainty<br>(k=2) | Note             | Effective Date |
|----|---------------------------------------|--------------------------------------|------------------------------------------------------|-------------------------------------------------------|-------------------------------|------------------|----------------|
|    |                                       | Detection limit                      |                                                      | FID: $\leq 0.5\text{ng/s}$                            | $U_{\text{rel}}=6\%$          | gas injection GC |                |
|    |                                       |                                      |                                                      | NPD: $\leq 5\text{pg/s(N)}$                           | $U_{\text{rel}}=6\%$          |                  |                |
|    |                                       |                                      |                                                      | NPD: $\leq 10\text{pg/s (P)}$                         | $U_{\text{rel}}=6\%$          |                  |                |
|    |                                       |                                      |                                                      | FPD: $\leq 0.5\text{ng/s(S)}$                         | $U_{\text{rel}}=5\%$          |                  |                |
|    |                                       |                                      |                                                      | FPD: $\leq 0.1\text{ng/s (P)}$                        | $U_{\text{rel}}=5\%$          |                  |                |
|    |                                       |                                      |                                                      | ECD: $\leq 5\text{pg/mL}$                             | $U_{\text{rel}}=6\%$          |                  |                |
| 2  | *Liquid Chromatograph                 | The minimum detectable concentration | V.R.of Liquid Chromatographs JJG 705                 | UVD: Naphthalene: $\leq 5 \times 10^{-8}\text{g/mL}$  | $U_{\text{rel}}=8\%$          |                  |                |
|    |                                       |                                      |                                                      | PDAD:Naphthalene : $\leq 5 \times 10^{-8}\text{g/mL}$ | $U_{\text{rel}}=8\%$          |                  |                |
|    |                                       |                                      |                                                      | FD:Naphthalene : $\leq 5 \times 10^{-9}\text{g/mL}$   | $U_{\text{rel}}=8\%$          |                  |                |
|    |                                       |                                      |                                                      | RID: Cholesterol: $\leq 5 \times 10^{-6}\text{g/mL}$  | $U_{\text{rel}}=8\%$          |                  |                |
|    |                                       |                                      |                                                      | ELSD:Cholesterol: $\leq 5 \times 10^{-6}\text{g/mL}$  | $U_{\text{rel}}=8\%$          |                  |                |
| 3  | *Atomic Absorption Spectrophotometers | Detection limit                      | V.R.for Atomic Absorption Spectrophotometers JJG 694 | Flame Atomic(Cu) : $\leq 0.02 \mu\text{g/mL}$         | $U=0.005 \mu\text{g/mL}$      |                  |                |
|    |                                       |                                      |                                                      | Graphite Furnace Atomic(Cd) : $\leq 4\text{pg}$       | $U=0.22\text{pg}$             |                  |                |
| 4  | *Electrolytic Conductivity Meter      | Electrolytic conductivity            | V.R.of Electrolytic Conductivity Meters JJG 376      | electric: $(0.05 \sim 1 \times 10^5) \mu\text{S/cm}$  | $U=0.2\%\text{FS}$            |                  |                |
|    |                                       |                                      |                                                      | Instrument: $147.8 \mu\text{S/cm}$                    | $U=0.4\%\text{FS}$            |                  |                |
|    |                                       |                                      |                                                      | Instrument: $1410 \mu\text{S/cm}$                     | $U=0.4\%\text{FS}$            |                  |                |



No. CNAS L0730

第 127 页 共 163 页

The scope of the accreditation in Chinese remains the definitive version.

| №  | Instrument                                                 | Measurand        | Calibration Method                                                        | Range                           | Expanded Uncertainty<br>( <i>k</i> =2) | Note                                                                                 | Effective Date |
|----|------------------------------------------------------------|------------------|---------------------------------------------------------------------------|---------------------------------|----------------------------------------|--------------------------------------------------------------------------------------|----------------|
| 5  | *Laboratory pH Meters                                      | pH               | V.R.of Laboratory pH Meters JIG 119                                       | Electric pH: 0~14               | <i>U</i> =0.002                        | Accreditation except for grade 0.001pH meters                                        |                |
|    |                                                            |                  |                                                                           | Instrument pH: 4.00~9.18        | <i>U</i> =0.015                        |                                                                                      |                |
|    |                                                            | Voltage          |                                                                           | (-1999~1999)mV                  | <i>U</i> =0.01%FS                      |                                                                                      |                |
| 6  | Rotational Viscometers                                     | Viscosity        | V.R.of Rotational Viscometers JIG 1002                                    | (2~1.2×10 <sup>5</sup> )mPa • s | <i>U</i> <sub>rel</sub> =2.4%          |                                                                                      |                |
| 7  | *Turbidimeters                                             | Turbidity        | V.R.of Turbidimeters JIG 880                                              | (1~10)NTU                       | <i>U</i> <sub>rel</sub> =10%           |                                                                                      |                |
|    |                                                            |                  |                                                                           | 10~20)NTU                       | <i>U</i> <sub>rel</sub> =8%            |                                                                                      |                |
|    |                                                            |                  |                                                                           | (20~400)NTU                     | <i>U</i> <sub>rel</sub> =4%            |                                                                                      |                |
| 8  | *Ultraviolet, Visible Spectrophotometer                    | Wavelength       | V.R.of Ultraviolet, Visible , Near-Infrared Spectrophotometer JIG 178     | (200~ 900)nm                    | <i>U</i> =0.4nm                        |                                                                                      |                |
|    |                                                            | Transmittance    |                                                                           | (8~32)%                         | <i>U</i> =0.4%                         |                                                                                      |                |
| 9  | *Polarimeter and Polarimetric Saccharimeters               | Polarization     | V.R.of Polarimeter and Polarimetric Saccharimeters JIG 536                | -72° ~+72°                      | <i>U</i> =0.005°                       |                                                                                      |                |
|    |                                                            | Saccharin        |                                                                           | -20° Z~+105° Z                  | <i>U</i> =0.015° Z                     |                                                                                      |                |
| 10 | *Hand Saccharinmeter (Content-meter)and Hand Refractometer | Refractive index | V.R.of Hand Saccharinmeter (Content-meter )and Hand Refractometer JIG 820 | n <sub>D</sub> :1.3300~1.6000   | <i>U</i> =2×10 <sup>-4</sup>           |  |                |
|    |                                                            | Saccharization   |                                                                           | scale division: 0.1%: (0.1~15)% | <i>U</i> =0.3%                         |                                                                                      |                |
|    |                                                            |                  |                                                                           | scale division: 0.2%: (0.2~30)% | <i>U</i> =0.5%                         |                                                                                      |                |
|    |                                                            |                  |                                                                           | scale division: 0.5%: (0.5~50)% | <i>U</i> =1.1%                         |                                                                                      |                |
|    |                                                            |                  |                                                                           | scale division: 1.0%: (1.0~60)% | <i>U</i> =1.8%                         |                                                                                      |                |



在线扫码获取验证

No. CNAS L0730

第 128 页 共 163 页

The scope of the accreditation in Chinese remains the definitive version.

| №  | Instrument                                                | Measurand               | Calibration Method                                                        | Range                                                                                | Expanded Uncertainty<br>( $k=2$ ) | Note                                                     | Effective Date |
|----|-----------------------------------------------------------|-------------------------|---------------------------------------------------------------------------|--------------------------------------------------------------------------------------|-----------------------------------|----------------------------------------------------------|----------------|
|    |                                                           |                         |                                                                           | scale division: 2.0%:<br>(2.0~60)%                                                   | $U=2.8\%$                         |                                                          |                |
| 11 | *On-line Automatic Determinator of Chemical Oxygen Demand | Concentration           | V.R. of On-line Automatic Determinator of Chemical Oxygen Demand JJG 1012 | (50~1000)mg/L                                                                        | $U_{rel}=4\%$                     |                                                          |                |
| 12 | *Ammonia-Nitrogen Automatic Analyzers                     | Concentration           | V.R. of Ammonia-Nitrogen Automatic Analyzers JJG 631                      | Type A: (1~2)mg/L                                                                    | $U=0.05\text{mg/L}$               |                                                          |                |
|    |                                                           |                         |                                                                           | Type A: (2~100)mg/L                                                                  | $U_{rel}=4.2\%$                   |                                                          |                |
|    |                                                           |                         |                                                                           | Type B: (1~100)mg/L                                                                  | $U_{rel}=3.6\%$                   |                                                          |                |
| 13 | *Alarmer Detectors of Combustible Gas                     | Concentration           | V.R. of Alarmer Detectors of Combustible Gas JJG 693                      | (5~65)%LEL                                                                           | $U=2.2\%\text{FS}$                | Accreditation only for Methane and Isobutane Gas Alarmer |                |
| 14 | *Gas Chromatography-Mass Spectrometers                    | Signal- to- Noise ratio | C.S. for Gas Chromatography-Mass Spectrometers JJF 1164                   | EI <sup>+</sup> (Ion trap, single quadrupole, triple quadrupole) S/N: $\geq 10:1$    | $U_{rel}=9\%$                     |                                                          |                |
|    |                                                           |                         |                                                                           | EI <sup>+</sup> (Time of flight, electrostatic field, orbital trap) S/N: $\geq 50:1$ | $U_{rel}=9\%$                     |                                                          |                |
|    |                                                           |                         |                                                                           | CI <sup>+</sup> (Ion trap, single quadrupole, triple quadrupole) S/N: $\geq 10:1$    | $U_{rel}=9\%$                     |                                                          |                |
|    |                                                           |                         |                                                                           | CI <sup>+</sup> (Ion trap, single quadrupole) S/N: $\geq 10:1$                       | $U_{rel}=8\%$                     |                                                          |                |
| 15 | *Laboratory Ion Meters                                    | pX                      | V.R. of Laboratory Ion Meters JJG 757                                     | Electric pX : 0~14                                                                   | $U=0.01$                          |                                                          |                |



No. CNAS L0730

第 129 页 共 163 页

The scope of the accreditation in Chinese remains the definitive version.

| №  | Instrument                                                              | Measurand     | Calibration Method                                                                        | Range                           | Expanded Uncertainty<br>( $k=2$ ) | Note | Effective Date |
|----|-------------------------------------------------------------------------|---------------|-------------------------------------------------------------------------------------------|---------------------------------|-----------------------------------|------|----------------|
|    |                                                                         |               |                                                                                           | Instrument: pX: F-(2~4)         | $U=0.02$                          |      |                |
|    |                                                                         | Voltage       |                                                                                           | (-1999~1999)mV                  | $U=0.03\%FS$                      |      |                |
| 16 | *Water Quality On-line Analyzers of Total Phosphorus and Total Nitrogen | Concentration | V.R. of Water Quality On-line Analyzers of Total Phosphorus and Total Nitrogen JJG 1094   | Total Phosphorus: (0.1~0.5)mg/L | $U=0.02mg/L$                      |      |                |
|    |                                                                         |               |                                                                                           | Total Phosphorus: (0.5~500)mg/L | $U_{rel}=4.0\%$                   |      |                |
|    |                                                                         | Concentration |                                                                                           | Total Nitrogen: (0.1~2)mg/L     | $U=0.08mg/L$                      |      |                |
|    |                                                                         |               |                                                                                           | Total Nitrogen: (2~500)mg/L     | $U_{rel}=3.5\%$                   |      |                |
| 17 | *Energy Dispersive X-Ray Fluorescence Spectrometers                     | Concentration | Calibration Specification for Energy Dispersive X-Ray Fluorescence Spectrometers JJF 2024 | Cd:(8.7~107)mg/kg               | $U_{rel}=8\%$                     |      |                |
|    |                                                                         |               |                                                                                           | Cr:(97.3~1122)mg/kg             | $U_{rel}=10\%$                    |      |                |
|    |                                                                         |               |                                                                                           | Hg:(91.5~1096)mg/kg             | $U_{rel}=9\%$                     |      |                |
|    |                                                                         |               |                                                                                           | Pb:(93.1~1122)mg/kg             | $U_{rel}=9\%$                     |      |                |
| 18 | *On-line pH Meters                                                      | pH            | C.S. for On-line pH Meters JJF 1547                                                       | Electric Meter pH: 0.~14        | $U=0.012$                         |      |                |
|    |                                                                         |               |                                                                                           | Instrument pH: 4.00~9.18        | $U=0.02$                          |      |                |
|    |                                                                         | Voltage       |                                                                                           | (-2000~2000)mV                  | $U=0.7mV$                         |      |                |
| 19 | *Automatic Potentiometric Titrators                                     | Voltage       | V.R. of Automatic Potentiometric Titrators JJG 814                                        | (-2000~2000)mV                  | $U=0.13mV$                        |      |                |
|    |                                                                         | Capacity      |                                                                                           | (2~100)mL                       | $U=0.03mL$                        |      |                |
|    |                                                                         | Concentration |                                                                                           | NaOH:0.1mol/L                   | $U_{rel}=0.3\%$                   |      |                |



No. CNAS L0730

第 130 页 共 163 页

The scope of the accreditation in Chinese remains the definitive version.

| №  | Instrument                                       | Measurand                      | Calibration Method                                                            | Range                                                | Expanded Uncertainty<br>( $k=2$ ) | Note | Effective Date |
|----|--------------------------------------------------|--------------------------------|-------------------------------------------------------------------------------|------------------------------------------------------|-----------------------------------|------|----------------|
| 20 | Dissolved Oxygen Meter                           | Dissolved oxygen Concentration | V.R. of Dissolved Oxygen Meters JJG 291                                       | (6~12) mg/L                                          | $U=0.16\text{mg/L}$               |      |                |
|    |                                                  | Temperature                    |                                                                               | (5~50) °C                                            | $U=0.1\text{ °C}$                 |      |                |
|    |                                                  | Time                           |                                                                               | (1~60)s                                              | $U=0.7\text{s}$                   |      |                |
| 21 | *Air Samplers                                    | Flow                           | V.R. of Air Samplers JJG 956                                                  | (0.01~6)L/min                                        | $U_{\text{rel}}=1.8\%$            |      |                |
| 22 | *Electrochemical Oxygen Meter                    | Concentration                  | Verification Regulation of Electrochemical Oxygen Meter JJG 365               | (5~25)%                                              | $U_{\text{rel}}=3\%$              |      |                |
|    |                                                  |                                |                                                                               | (25~80)%                                             | $U_{\text{rel}}=1.2\%$            |      |                |
| 23 | *Mercury Analyzers                               | Detection limit                | Verification Regulation of Mercury Analyzers JJG 548                          | Absorption class: $\leq 1.0\text{ng}$                | $U=0.3\text{ng}$                  |      |                |
|    |                                                  |                                |                                                                               | Fluorescence Spectrophotometers: $\leq 0.1\text{ng}$ | $U=0.04\text{ng}$                 |      |                |
| 24 | *Atomic Fluorescence Spectrophotometers          | Detection limit                | Verification Regulation for Atomic Fluorescence Spectrophotometers JJG 939    | As: $\leq 0.4\text{ng}$                              | $U=0.015\text{ng}$                |      |                |
|    |                                                  |                                |                                                                               | Sb: $\leq 0.4\text{ng}$                              | $U=0.05\text{ng}$                 |      |                |
| 25 | *Dust Sampler                                    | Flow                           | Verification Regulation of Dust Sampler JJG 520                               | (0.1~1200)L/min                                      | $U_{\text{rel}}=1.9\%$            |      |                |
| 26 | *Water Hardness Meters                           | Water Hardness                 | C. S. for Water Hardness Meters JJF 1949                                      | (10~1000) mg/L                                       | $U_{\text{rel}}=3.8\%$            |      |                |
| 27 | *Flame Photometer                                | Detection limit                | Verification Regulation of Flame Photometer JJG 630                           | K: $\leq 0.004\text{mmol/L}$                         | $U=0.001\text{mmol/L}$            |      |                |
|    |                                                  |                                |                                                                               | Na: $\leq 0.008\text{mmol/L}$                        | $U=0.002\text{mmol/L}$            |      |                |
| 28 | *Water Quality On-line Analyzers of Heavy Metals | Concentration                  | Calibration Specification for Water Quality On-line Analyzers of Heavy Metals | Cr <sup>6+</sup> : (0.01~100) mg/L                   | $U_{\text{rel}}=4\%$              |      |                |
|    |                                                  |                                |                                                                               | Cu: (0.01~500) mg/L                                  | $U_{\text{rel}}=4\%$              |      |                |



No. CNAS L0730

第 131 页 共 163 页

The scope of the accreditation in Chinese remains the definitive version.


| №  | Instrument                          | Measurand                       | Calibration Method                                              | Range                                                | Expanded Uncertainty<br>( $k=2$ ) | Note | Effective Date |
|----|-------------------------------------|---------------------------------|-----------------------------------------------------------------|------------------------------------------------------|-----------------------------------|------|----------------|
|    |                                     |                                 | JJF1565                                                         | Ni: (0.02~100) mg/L                                  | $U_{rel}=4\%$                     |      |                |
|    |                                     |                                 |                                                                 | Pb: (0.01~1000) mg/L                                 | $U_{rel}=4\%$                     |      |                |
|    |                                     |                                 |                                                                 | Cd: (0.001~10) mg/L                                  | $U_{rel}=4\%$                     |      |                |
|    |                                     |                                 |                                                                 | Hg: (0.001~100) mg/L                                 | $U_{rel}=4\%$                     |      |                |
|    |                                     |                                 |                                                                 | As: (0.01~1000) mg/L                                 | $U_{rel}=4\%$                     |      |                |
|    |                                     |                                 |                                                                 | Fe: (0.02~1000) mg/L                                 | $U_{rel}=4\%$                     |      |                |
|    |                                     |                                 |                                                                 | Mn: (0.01~1000) mg/L                                 | $U_{rel}=4\%$                     |      |                |
|    |                                     |                                 |                                                                 | Zn: (0.02~1000) mg/L                                 | $U_{rel}=4\%$                     |      |                |
|    |                                     |                                 |                                                                 | Cr: (0.01~500) mg/L                                  | $U_{rel}=4\%$                     |      |                |
| 29 | *Analyzers for oil content in water | Concentration                   | V.R.of Analyzers for oil content in water JJG 950               | (0.1~10) mg/L                                        | $U=0.2\text{mg/L}$                |      |                |
|    |                                     |                                 |                                                                 | (>10~1000) mg/L                                      | $U_{rel}=2.4\%$                   |      |                |
| 30 | *Total Organic Carbon Analyzer      | Concentration                   | Verification Regulation of Total Organic Carbon Analyzer JJG821 | OC:(0.1~1000)mg/L                                    | $U_{rel}=2.6\%$                   |      |                |
|    |                                     |                                 |                                                                 | IC:(0.1~1000)mg/L                                    | $U_{rel}=2.6\%$                   |      |                |
| 31 | *Ion Chromatographs                 | Minimum detection concentration | Verification Regulation of Ion Chromatograph JJG823             | Conductivity detector:<br>$\leq 0.02 \mu\text{g/mL}$ | $U_{rel}=6\%$                     |      |                |
|    |                                     |                                 |                                                                 | UV visible detector: $\leq 0.02 \mu\text{g/mL}$      | $U_{rel}=9\%$                     |      |                |
|    |                                     |                                 |                                                                 | Electrochemical detector: $\leq 0.02 \mu\text{g/mL}$ | $U_{rel}=5\%$                     |      |                |



No. CNAS L0730

第 132 页 共 163 页

The scope of the accreditation in Chinese remains the definitive version.

| №  | Instrument                    | Measurand       | Calibration Method                                                 | Range                                          | Expanded Uncertainty<br>( $k=2$ ) | Note                                                                                 | Effective Date |
|----|-------------------------------|-----------------|--------------------------------------------------------------------|------------------------------------------------|-----------------------------------|--------------------------------------------------------------------------------------|----------------|
| 32 | *Ammonia Gas Detectors        | Concentration   | Verification Regulation of Ammonia Gas Detectors JJG1105           | (0.1~300) $\mu$ mol/mol                        | $U_{rel}=3.6\%$                   | Do not make instruments with maximum allowable error less than $\pm 10\%$            |                |
| 33 | *Alarmer Detectors of Benzene | Concentration   | Calibration Specification for Alarmer Detectors of Benzene JJF1674 | (1~100) $\mu$ mol/mol                          | $U_{rel}=3.6\%$                   |                                                                                      |                |
| 34 | *Emission Spectrometer        | Detection limit | Verification Regulation of Emission Spectrometer JJG768            | ICP Spectrometer Zn: $\leq 0.003\text{mg/L}$   | $U=0.0012\text{mg/L}$             |  |                |
|    |                               |                 |                                                                    | ICP Spectrometer Ni: $\leq 0.01\text{mg/L}$    | $U=0.004\text{mg/L}$              |                                                                                      |                |
|    |                               |                 |                                                                    | ICP Spectrometer Mn: $\leq 0.002\text{mg/L}$   | $U=0.0008\text{mg/L}$             |                                                                                      |                |
|    |                               |                 |                                                                    | ICP Spectrometer Cr: $\leq 0.007\text{mg/L}$   | $U=0.0028\text{mg/L}$             |                                                                                      |                |
|    |                               |                 |                                                                    | ICP Spectrometer Cu: $\leq 0.007\text{mg/L}$   | $U=0.0028\text{mg/L}$             |                                                                                      |                |
|    |                               |                 |                                                                    | ICP Spectrometer Ba: $\leq 0.001\text{mg/L}$   | $U=0.0004\text{mg/L}$             |                                                                                      |                |
|    |                               |                 |                                                                    | Direct reading spectrometer C: $\leq 0.005\%$  | $U=0.0019\%$                      |                                                                                      |                |
|    |                               |                 |                                                                    | Direct reading spectrometer Si: $\leq 0.005\%$ | $U=0.0020\%$                      |                                                                                      |                |



No. CNAS L0730

第 133 页 共 163 页

The scope of the accreditation in Chinese remains the definitive version.


| №  | Instrument                                                | Measurand     | Calibration Method                                                        | Range                                                            | Expanded Uncertainty<br>( $k=2$ ) | Note | Effective Date |
|----|-----------------------------------------------------------|---------------|---------------------------------------------------------------------------|------------------------------------------------------------------|-----------------------------------|------|----------------|
|    |                                                           |               |                                                                           | Direct reading spectrometer Mn: $\leq 0.003\%$                   | $U=0.0012\%$                      |      |                |
|    |                                                           |               |                                                                           | Direct reading spectrometer Cr: $\leq 0.003\%$                   | $U=0.0013\%$                      |      |                |
|    |                                                           |               |                                                                           | Direct reading spectrometer Ni: $\leq 0.005\%$                   | $U=0.0020\%$                      |      |                |
| 35 | *Total Susended Particulates Samler                       | Flow          | Verification Regulation of Total Susended Particulates Samler JJG943      | (80~120)L/min                                                    | $U_{rel}=1.3\%$                   |      |                |
|    |                                                           |               |                                                                           | (800~1200)L/min                                                  | $U_{rel}=1.3\%$                   |      |                |
| 36 | *Polymerase Chain Reaction Analyzers                      | Temperature   |                                                                           | (30~95)°C                                                        | $U=0.6^{\circ}\text{C}$           |      |                |
|    |                                                           | Concentration | Calibration Specification for Polymerase Chain Reaction Analyzers JJF1527 | ( $10 \sim 1.0 \times 10^3$ ) copies/ $\mu\text{L}$              | $U_{rel}=15\%$                    |      |                |
|    |                                                           |               |                                                                           | ( $1.0 \times 10^3 \sim 1.2 \times 10^7$ ) copies/ $\mu\text{L}$ | $U_{rel}=10\%$                    |      |                |
| 37 | *Melting-point Measurement                                | temperature   | V.R.of Melting-point Measurement Instruments JJG 701                      | (50~300)°C                                                       | $U=0.3^{\circ}\text{C}$           |      |                |
| 38 | *Carbon Monoxide and Carbon Dioxide Infrared Gas Analyzer | concentration | V.R.of Carbon Monoxide and Carbon Dioxide Infrared Gas Analyzer JJG 635   | CO <sub>2</sub> : (0.01~20)%                                     | $U_{rel}=1.5\%$                   |      |                |
|    |                                                           |               |                                                                           | CO: (1~1000) $\mu\text{mol/mol}$                                 | $U_{rel}=1.5\%$                   |      |                |
| 39 | *Sulfur Hydrogen Gas Detectors                            | concentration | V.R.of Sulfur Hydrogen Gas Detectors JJG 695                              | (0.1~200) $\mu\text{mol/mol}$                                    | $U_{rel}=2.5\%$                   |      |                |
| 40 | *Kjeldahl method                                          | Content       | C.S.for Elemental Analyzers JJF 1321                                      | 20%~50%                                                          | $U_{rel}=1.8\%$                   |      |                |



No. CNAS L0730

第 134 页 共 163 页

The scope of the accreditation in Chinese remains the definitive version.

| №  | Instrument                                                         | Measurand          | Calibration Method                                                     | Range                                      | Expanded Uncertainty<br>( $k=2$ ) | Note                                                                                 | Effective Date |
|----|--------------------------------------------------------------------|--------------------|------------------------------------------------------------------------|--------------------------------------------|-----------------------------------|--------------------------------------------------------------------------------------|----------------|
| 41 | *Instrument for KF Coulometry Titration                            | water content      | V.R.of Instrument for KF Coulometry Titration JJG 1044                 | (10~5000) $\mu$ g                          | $U_{rel}=2.4\%$                   |                                                                                      |                |
| 42 | *Karl Fischer Volumetric Titrators for Water Content               | water content      | V.R.of Karl Fischer Volumetric Titrators for Water Content JJG 1154    | (1~20)mg                                   | $U_{rel}=1.9\%$                   |                                                                                      |                |
| 43 | *Residual Chlorine Meters                                          | concentration      | C.S.for Residual Chlorine Meters JJF 1609                              | Residual Chlorine:<br>(0.01~10) mg/L       | $U_{rel}=4\%$                     |                                                                                      |                |
|    |                                                                    |                    |                                                                        | Total Chlorine: (0.01~50) mg/L             | $U_{rel}=4\%$                     |                                                                                      |                |
| 44 | *Differential Scanning Calorimeters                                | temperature        | V.R.of Differential Scanning Calorimeters JJG 936                      | (120~600)°C                                | $U=0.8\text{ }^{\circ}\text{C}$   |                                                                                      |                |
|    |                                                                    | heat quantity      |                                                                        | (25~100)J/g                                | $U_{rel}=1.8\%$                   |                                                                                      |                |
| 45 | *Sulfur Dioxide Gas Detectors                                      | concentration      | V.R.of Sulfur Dioxide Gas Detectors JJG 551                            | (1~2000) $\mu$ mol/mol                     | $U_{rel}=2.3\%$                   |                                                                                      |                |
| 46 | *Liquid Chromatography-Mass Spectrometers                          | signal-noise ratio | C.S.for Liquid Chromatography-Mass Spectrometers JJF 1317              | Reserpine $\geq 10: 1$                     | $U_{rel}=10\%$                    |                                                                                      |                |
| 47 | *Calibration Specification for Open/Closed Cup Flash Point Testers | temperature        | C.S.for Open/Closed Cup Flash Point Testers JJF 1384                   | Closed Cup Flash Point Testers :(70~200)°C | $U=6^{\circ}\text{C}$             |                                                                                      |                |
|    |                                                                    |                    |                                                                        | Open Cup Flash Point Testers :(110~250)°C  | $U=8^{\circ}\text{C}$             |                                                                                      |                |
| 48 | *Flue Gas Samplers                                                 | flow               | V.R.of Flue Gas Samplers JJG 1169                                      | (0.01~2.0)L/min                            | $U_{rel}=1.2\%$                   |  |                |
| 49 | *Volatile Organic Compounds Photo Ionization Detectors             | concentration      | C.S.for Volatile Organic Compounds Photo Ionization Detectors JJF 1172 | (0.1~2000) $\times 10^{-6}$ mol/mol        | $U_{rel}=2.4\%$                   |                                                                                      |                |



No. CNAS L0730

第 135 页 共 163 页

The scope of the accreditation in Chinese remains the definitive version.

| №  | Instrument                                              | Measurand                  | Calibration Method                                                         | Range                                                      | Expanded Uncertainty<br>( $k=2$ ) | Note | Effective Date |
|----|---------------------------------------------------------|----------------------------|----------------------------------------------------------------------------|------------------------------------------------------------|-----------------------------------|------|----------------|
| 50 | *Flue Gas Analyzers                                     | concentration              | V.R. of Flue Gas Analyzers<br>JJG 968                                      | NO: $(0.1 \sim 1000) \times 10^{-6}$ mol/mol               | $U_{\text{rel}}=2.4\%$            |      |                |
|    |                                                         |                            |                                                                            | SO <sub>2</sub> : $(0.1 \sim 2000) \times 10^{-6}$ mol/mol | $U_{\text{rel}}=3.0\%$            |      |                |
|    |                                                         |                            |                                                                            | CO: $(0.1 \sim 1000) \times 10^{-6}$ mol/mol               | $U_{\text{rel}}=2.4\%$            |      |                |
|    |                                                         |                            |                                                                            | O <sub>2</sub> : $(0.1 \sim 25)\%$ mol/mol                 | $U_{\text{rel}}=2.4\%$            |      |                |
| 51 | *Liquid Chromatograph-Atomic Fluorescence Spectrometers | detection limit            | V.R. of Liquid Chromatograph-Atomic Fluorescence Spectrometers<br>JJG 1151 | As(V): $<1.0\text{ng}$                                     | $U_{\text{rel}}=11\%$             |      |                |
|    |                                                         |                            |                                                                            | MMA: $<0.7\text{ng}$                                       | $U_{\text{rel}}=11\%$             |      |                |
|    |                                                         |                            |                                                                            | DMA: $<0.7\text{ng}$                                       | $U_{\text{rel}}=10\%$             |      |                |
| 52 | *Fourier Transform Infrared Spectrometer                | wave number                | C.S. for Transform Infrared Spectrometers JJF 1319                         | $(4000 \sim 400)\text{cm}^{-1}$                            | $U=0.4\text{cm}^{-1}$             |      |                |
| 53 | *Pesticide Residue Detectors                            | Wavelength                 | C.S. for Pesticide Residue Detectors JJF 1729                              | $(330 \sim 450)\text{nm}$                                  | $U=3.5\text{nm}$                  |      |                |
|    |                                                         | Transmittance              |                                                                            | $(7 \sim 33)\%$                                            | $U=0.3\%$                         |      |                |
|    |                                                         | Suppression ratio          |                                                                            | $(50 \sim 100)\%$                                          | $U=5.2\%$                         |      |                |
| 54 | *Water Colorimeters                                     | Water quality chromaticity | C.S. for Water Colorimeters JJF 1689                                       | $(0.1 \sim 500)\text{PCU}$                                 | $U_{\text{rel}}=2.0\%$            |      |                |
| 55 | *Flow Cup Viscometers                                   | Viscosity                  | V.R. of Flow Cup Viscometers JJG 743                                       | $(2 \sim 1000)\text{mm}^2/\text{s}$                        | $U_{\text{rel}}=5\%$              |      |                |
| 56 | Wood Moisture Content Measuring Meters                  | Moisture Content           | V.R. of Wood Moisture Content Measuring Meters JJG 986                     | $(6 \sim 28)\%$                                            | $U=0.6\%$                         |      |                |



No. CNAS L0730

第 136 页 共 163 页

The scope of the accreditation in Chinese remains the definitive version.

| №                                          | Instrument                                                   | Measurand             | Calibration Method                                                           | Range                        | Expanded Uncertainty<br>( <i>k</i> =2) | Note                            | Effective Date |
|--------------------------------------------|--------------------------------------------------------------|-----------------------|------------------------------------------------------------------------------|------------------------------|----------------------------------------|---------------------------------|----------------|
| 57                                         | *Automatic Closed Luminescence Immunoassay Analyzers         | Concentration         | C.S.for Automatic Closed Luminescence Immunoassay Analyzers JJF1752          | AFP: (0 ~300) IU/mL          | <i>U</i> =1.3 IU/mL                    |                                 |                |
| 58                                         | *Glycohemoglobin Analyzers                                   | Concentration         | C.S.for Glycohemoglobin Analyzers JJF1841                                    | 4%~11%                       | <i>U</i> =3.3 %                        |                                 |                |
| 59                                         | Dust Concentration Measuring Instruments                     | Concentration         | V.R.of Dust Concentration Measuring Instruments JJG846                       | (0.1~1000)mg/m <sup>3</sup>  | <i>U</i> <sub>rel</sub> =9.6%          |                                 |                |
| 60                                         | PM <sub>2.5</sub> Mass Concentration Measurement Instruments | Concentration         | C.M.for PM <sub>2.5</sub> Mass Concentration Measurement Instruments JJF1659 | (10~1000) μ g/m <sup>3</sup> | <i>U</i> <sub>rel</sub> =9%            |                                 |                |
| 61                                         | *Particle Filtrig Efficiency Testers for Mask                | Flow                  | V.R.of Particle Filtrig Efficiency Testers for Mask JJG(Yue) 061             | (15~100) L/min               | <i>U</i> <sub>rel</sub> =1.3%          | Standard photometer method only |                |
|                                            |                                                              | Resistance            |                                                                              | (50~1000)Pa                  | <i>U</i> <sub>rel</sub> =1.0%          |                                 |                |
|                                            |                                                              | Filtration efficiency |                                                                              | 80%~90%                      | <i>U</i> =0.7%                         |                                 |                |
|                                            |                                                              |                       |                                                                              | ≥90%~99.999%                 | <i>U</i> =1.3%                         |                                 |                |
| 62                                         | *Mask Synthetic Blood Penetration Testers                    | Pressure              | C.M.for Mask Synthetic Blood Penetration Testers FFH2117                     | 10.6kPa、16.0kPa、21.3kPa      | <i>U</i> <sub>rel</sub> =0.4%          |                                 |                |
|                                            |                                                              | Length                |                                                                              | (1~15)mm                     | <i>U</i> =40 μ m                       |                                 |                |
|                                            |                                                              | Volume                |                                                                              | (1~5)mL                      | <i>U</i> <sub>rel</sub> =1.4%          |                                 |                |
| 63                                         | *Time-resolved Fluorescence Immunoassay Analyzers            | Concentration         | C.M.for Time-resolved Fluorescence Immunoassay                               | AFP: (1 ~300) IU/mL          | <i>U</i> <sub>rel</sub> =7.6%          |                                 |                |
|                                            |                                                              | Temperature           | Analzyers FFH2128                                                            | (20~50)℃                     | <i>U</i> =0.5℃                         |                                 |                |
| 10 Ionizing radiation Measuring Instrument |                                                              |                       |                                                                              |                              |                                        |                                 |                |



No. CNAS L0730

第 137 页 共 163 页

The scope of the accreditation in Chinese remains the definitive version.

| No | Instrument                                                                                            | Measurand                                           | Calibration Method                                                                                             | Range              | Expanded Uncertainty<br>( $k=2$ ) | Note | Effective Date |
|----|-------------------------------------------------------------------------------------------------------|-----------------------------------------------------|----------------------------------------------------------------------------------------------------------------|--------------------|-----------------------------------|------|----------------|
| 1  | *Medical<br>Diagnostical X-ray<br>Radiation Source<br>for Spiral<br>Computed<br>Tomography(CT)        | Dose index                                          | V.R.of Medical Diagnostic<br>X-ray Radiation Source for<br>Spiral Computed<br>Tomography(CT) JJG961            | (0.1~999.9)mGy/min | $U_{rel}=5\%$                     |      |                |
| 2  | *Medical<br>Diagnostical X-ray<br>Radiation Source                                                    | Air kerma rate                                      | V.R.of Medical Diagnostical<br>X-ray Radiation Source<br>JJG744                                                | (0.1~999.9)mGy/min | $U_{rel}=7\%(k=2)$                |      |                |
|    |                                                                                                       | Tube Voltage                                        |                                                                                                                | (50~120) kV        | $U_{rel}=4\%$                     |      |                |
| 3  | *Medical<br>Diagnostic X-ray<br>Source for Dental<br>Panorama                                         | Air kerma rate                                      | V.R.of Medical Diagnostic<br>X-ray Source for Dental<br>Panorama JJG1101                                       | (0.1~999.9)mGy/min | $U_{rel}=6\%$                     |      |                |
|    |                                                                                                       | Tube Voltage                                        |                                                                                                                | (60~120) kV        | $U_{rel}=2.2\%$                   |      |                |
| 4  | *X-ray Flaw<br>Detectors                                                                              | Kerma Rate                                          | Verification Regulation of X-<br>ray Flaw Detectors JJG40                                                      | (0.01~10) Gy/min   | $U_{rel}=5\%$                     |      |                |
| 5  | *Medical<br>Diagnostic X-ray<br>Radiation Source<br>for Medical Digital<br>Subtraction<br>Angiography | Air Kerma                                           | V.R.of Medical Diagnostic<br>X-ray Radiation Source for<br>Medical Digital Subtraction<br>Angiography JJG 1067 | (0.001~10)mGy/s    | $U_{rel}=7\%$                     |      |                |
|    |                                                                                                       | Tube Voltage                                        |                                                                                                                | (60~120)kV         | $U_{rel}=4\%$                     |      |                |
|    |                                                                                                       | Simulate the<br>minimum size<br>of blood<br>vessels |                                                                                                                | (0.25~2)mm         | $U=0.11\text{mm}$                 |      |                |
|    |                                                                                                       | spatial<br>resolution                               |                                                                                                                | (6~50)Lp/cm        | $U=1\text{Lp/cm}$                 |      |                |
|    |                                                                                                       | low contrast<br>resolution                          |                                                                                                                | (0.5~4.0)mm        | $U=0.11\text{mm}$                 |      |                |
|    |                                                                                                       | Subtraction<br>performance<br>impact                |                                                                                                                | 2mm                | $U=0.11\text{mm}$                 |      |                |
|    |                                                                                                       |                                                     |                                                                                                                |                    |                                   |      |                |



No. CNAS L0730

第 138 页 共 163 页

The scope of the accreditation in Chinese remains the definitive version.

| No                                                        | Instrument                                       | Measurand                           | Calibration Method                                                          | Range                                             | Expanded Uncertainty<br>(k=2) | Note | Effective Date |
|-----------------------------------------------------------|--------------------------------------------------|-------------------------------------|-----------------------------------------------------------------------------|---------------------------------------------------|-------------------------------|------|----------------|
| 6                                                         | Radon Measuring Instruments                      | Volume Activity                     | V.R.of the Radon Measuring Instruments JJG825                               | (100~10000) Bq/m <sup>3</sup>                     | U <sub>rel</sub> =6%          |      |                |
| 7                                                         | *Low Background Alpha/Beta Measuring Instruments | Surface Emission Rate of α Particle | V.R.of the Low Background Alpha/Beta Measuring Instruments JJG853           | (1~1×10 <sup>6</sup> ) (min•2 π sr) <sup>-1</sup> | U <sub>rel</sub> =7%          |      |                |
|                                                           |                                                  | Surface Emission Rate of β Particle |                                                                             | (1~1×10 <sup>6</sup> ) (min•2 π sr) <sup>-1</sup> | U <sub>rel</sub> =4%          |      |                |
| 8                                                         | *Radioactivity Meter                             | Activity                            | V.R.of Radioactivity Meter JJG377                                           | (3.7×10 <sup>5</sup> ~3.7×10 <sup>10</sup> ) Bq   | U <sub>rel</sub> =3.0%        |      |                |
| 9                                                         | *Germanium Gamma-ray Spectrometers               | Fully-energy peak efficiency        | C. S. for Germanium Gamma-ray Spectrometers JJF1850                         | (0.01~100)%                                       | U <sub>rel</sub> =7%          |      |                |
| 10                                                        | *Alpha Beta Surface Contamination Instruments    | Surface Emission Rate of α Particle | V.R.of Alpha Beta Surface Contamination Instruments JJG478                  | (1~1×10 <sup>6</sup> ) (min•2 π sr) <sup>-1</sup> | U <sub>rel</sub> =4%          |      |                |
|                                                           |                                                  | Surface Emission Rate of β Particle |                                                                             | (1~1×10 <sup>6</sup> ) (min•2 π sr) <sup>-1</sup> | U <sub>rel</sub> =7%          |      |                |
| 11 Specialized Measuring Instrument for Textile & Leather |                                                  |                                     |                                                                             |                                                   |                               |      |                |
| 1                                                         | *Perspiration Fastness Instrumints               | Force value                         | Calibration Specification for Perspiration Fastness Instrumints JJF(纺织) 028 | (10~100) N                                        | U <sub>rel</sub> =0.11%       |      |                |
|                                                           |                                                  | Size                                |                                                                             | (0~150) mm                                        | U=0.04mm                      |      |                |
| 2                                                         | *Fabric Shrinkage Testers                        | Temperature                         | C.S.for Fabric Shrinkage Testers JJF(纺织) 052                                | (0~100) °C                                        | U=0.6℃                        |      |                |



No. CNAS L0730

第 139 页 共 163 页

The scope of the accreditation in Chinese remains the definitive version.

| №                                                     | Instrument                                              | Measurand          | Calibration Method                                                          | Range                    | Expanded Uncertainty<br>( $k=2$ ) | Note | Effective Date |
|-------------------------------------------------------|---------------------------------------------------------|--------------------|-----------------------------------------------------------------------------|--------------------------|-----------------------------------|------|----------------|
|                                                       |                                                         | Washing speed      |                                                                             | (30~100) r/min           | $U=0.18\text{r/min}$              |      |                |
|                                                       |                                                         | Dehydration speed  |                                                                             | (100~1000) r/min         | $U=1.3\text{r/min}$               |      |                |
|                                                       |                                                         | Time               |                                                                             | (0~30) min               | $U=0.3\text{s}$                   |      |                |
| 12 Specialized Measuring Instrument for Motor Vehicle |                                                         |                    |                                                                             |                          |                                   |      |                |
| 1                                                     | *tire tread depth gauge                                 | Length             | C.S for tire tread depth gauge JJF 1477                                     | (0~50)mm                 | $U=0.01\text{mm}$                 |      |                |
| 2                                                     | *portable braking performance tester for motor vehicles | Acceleration       | C.S for portable braking performance tester for motor vehicles JJF 1168     | (0~9.81)m/s <sup>2</sup> | $U=0.05\text{m/s}^2$              |      |                |
| 3                                                     | *steering force and steering angle tester for motor car | Angle              | C.S. of Motor vehicle tester for steering force and steering angle JJF 1196 | (0~1080)°                | $U=1^\circ$                       |      |                |
|                                                       |                                                         | Force value        |                                                                             | (10~500)N                | $U_{\text{rel}}=0.8\%$            |      |                |
| 4                                                     | *Automobile Side Slip Testers                           | side slip distance | V.R of Automobile Side Slip Testers JJG 908                                 | (15~+15)m/km             | $U=0.04\text{ m/km}$              |      |                |
| 5                                                     | *roller type speedometer tester                         | speed              | V.R of roller type speedometer tester JJG 909                               | (0.1~80)km/h             | $U_{\text{rel}}=1\%$              |      |                |
| 6                                                     | *turning angle testers                                  | angle              | C.S for turning angle testers JJF 1141                                      | (-50~+50)°               | $U=0.32^\circ$                    |      |                |
| 7                                                     | *Four-wheel aligners                                    | Angle              | C.S. for Four-wheel aligners JJF 1154                                       | Toe-in: (-3~+3) °        | $U=1.4'$                          |      |                |
|                                                       |                                                         |                    |                                                                             | Camber: (-10~+10) °      | $U=4'$                            |      |                |
| 8                                                     | *Vehicle Contour Dimensions Testers                     | length             | C.S.for Vehicle Contour Dimensions Testers JJF 1749                         | (0.001~16)m              | $U_{\text{rel}}=0.3\%$            |      |                |
| 9                                                     | *Motor Vehicle Engine Speed Measuring Instrument        | speed              | C.S. for Motor Vehicle Engine Speed Measuring Instruments JJF 1375          | (500~6000)r/min          | $U_{\text{rel}}=0.03\%$           |      |                |
|                                                       |                                                         | Time               |                                                                             | (0~60)s                  | $U=0.30\text{ s}$                 |      |                |



No. CNAS L0730

第 140 页 共 163 页

The scope of the accreditation in Chinese remains the definitive version.

| №  | Instrument                                       | Measurand       | Calibration Method                                               | Range                                        | Expanded Uncertainty<br>(k=2) | Note | Effective Date |
|----|--------------------------------------------------|-----------------|------------------------------------------------------------------|----------------------------------------------|-------------------------------|------|----------------|
| 10 | *opacimeters                                     | smoke intensity | V.R. of opacimeters JJG 976                                      | (15~80)%                                     | $U_{rel}=1.0\%$               |      |                |
| 11 | *vehicle exhaust emissions measuring instruments | Concentration   | V.R. of vehicle exhaust emissions measuring instruments JJG 688  | CO:(0.5~5.0)×10 <sup>-2</sup>                | $U_{rel}=1.4\%$               |      |                |
|    |                                                  |                 |                                                                  | CO <sub>2</sub> :(2.0~15.0)×10 <sup>-2</sup> | $U_{rel}=1.4\%$               |      |                |
|    |                                                  |                 |                                                                  | NO:(300~3200)×10 <sup>-6</sup>               | $U_{rel}=2.2\%$               |      |                |
|    |                                                  |                 |                                                                  | O <sub>2</sub> :(0.5~30.0)×10 <sup>-2</sup>  | $U_{rel}=2.2\%$               |      |                |
|    |                                                  |                 |                                                                  | HC:(200~3200)×10 <sup>-6</sup>               | $U_{rel}=1.5\%$               |      |                |
| 12 | *headlamp testers for motor vehicle              | optical angle   | V.R. of headlamp testers for motor vehicle JJG 745               | -3° ~+3°                                     | $U=3'$                        |      |                |
|    |                                                  | light intensity |                                                                  | (8~60)kcd                                    | $U_{rel}=0.47\%$              |      |                |
|    |                                                  | Length          |                                                                  | (0.1~2000)mm                                 | $U_{rel}=0.3\%$               |      |                |
| 13 | *Diesel Vehicle Nitrogen Oxides(NOx) Measuring   | concentration   | C.S. for Diesel Vehicle Nitrogen Oxides(NOx) Measuring JJF 1873  | CO <sub>2</sub> :(0.1~18)×10 <sup>-2</sup>   | $U_{rel}=2\%$                 |      |                |
|    |                                                  |                 |                                                                  | NO:(10~4000)×10 <sup>-6</sup>                | $U_{rel}=2\%$                 |      |                |
|    |                                                  |                 |                                                                  | NO <sub>2</sub> :(5~1000)×10 <sup>-6</sup>   | $U_{rel}=3\%$                 |      |                |
| 14 | *Transmittance Meter                             | Transmittance   | C.S. for Transmittance Meter of Automobile JJF 1225              | 50%~90%                                      | $U=0.8\%$                     |      |                |
| 15 | *Manipulating force tester for automotive brake  | Force value     | C.S. for Manipulating force tester for automotive brake JJF 1169 | (0.1~1000)N                                  | $U_{rel}=0.8\%$               |      |                |
| 16 | *automotive suspension tester                    | Mass            | C.S. for automotive suspension tester JJF 1192                   | (1~3000)kg                                   | $U_{rel}=1\%$                 |      |                |
| 17 | *Equipment of power measuring                    | Torque          | V.R. of Equipment of Power Measuring JJG 653                     | (1~5000)Nm                                   | $U_{rel}=1\%$                 |      |                |
|    |                                                  | rotate speed    |                                                                  | (1~9999)r/min                                | $U_{rel}=0.3\%$               |      |                |



No. CNAS L0730

第 141 页 共 163 页

The scope of the accreditation in Chinese remains the definitive version.


| №  | Instrument                                                                 | Measurand                      | Calibration Method                                                                          | Range                                                          | Expanded Uncertainty<br>( $k=2$ ) | Note | Effective Date |
|----|----------------------------------------------------------------------------|--------------------------------|---------------------------------------------------------------------------------------------|----------------------------------------------------------------|-----------------------------------|------|----------------|
|    |                                                                            | Length                         |                                                                                             | (0.1~600)mm                                                    | $U=0.04\text{mm}$                 |      |                |
| 18 | *wheel dynamic balancers                                                   | Mass                           | C.S. for wheel dynamic balancers JJF 1151                                                   | (80~140) g                                                     | $U=4.3\text{g}$                   |      |                |
| 19 | *special axle(wheel) load scale for motor                                  | Mass                           | V.R. of special axle(wheel) load scale for motor JJG 1014                                   | (0.01~15)t                                                     | $U_{\text{rel}}=0.8\%$            |      |                |
| 20 | *roller opposite force type brake testers                                  | Force value                    | V.R. of roller opposite force type brake testers JJG 906                                    | (0.1~30)kN                                                     | $U_{\text{rel}}=1.1\%$            |      |                |
|    |                                                                            | Velocity ratio                 |                                                                                             | (5~40)%                                                        | $U=2\%$                           |      |                |
| 21 | *Retroreflection Coefficient Meters for Motor Vehicle's Reflecting Marking | coefficient of retroreflection | C.S. for Retroreflection Coefficient Meters for Motor Vehicle's Reflecting Marking JJF 1747 | (0.1~400) $\text{cd} \cdot \text{lx}^{-1} \cdot \text{m}^{-2}$ | $U_{\text{rel}}=4.0\%$            |      |                |
| 22 | *Platform Brake Testers                                                    | force                          | V.R. of Platform Brake Testers JJG 1020                                                     | Brake force: (0.1~30)kN                                        | $U_{\text{rel}}=1.1\%$            |      |                |
|    |                                                                            |                                |                                                                                             | axle load: (20~4000)kg                                         | $U_{\text{rel}}=0.8\%$            |      |                |
|    |                                                                            | Length                         |                                                                                             | (0.1~100)mm                                                    | $U=0.5\text{mm}$                  |      |                |
| 23 | *Loading Method Automobile Brake Testers                                   | force                          | V.R. of Loading Method Automobile Brake Testers JJG 1160                                    | Brake force: (0.1~30)kN                                        | $U_{\text{rel}}=1.1\%$            |      |                |
|    |                                                                            |                                |                                                                                             | axle load: (20~30000)kg                                        | $U_{\text{rel}}=4.0\%$            |      |                |
|    |                                                                            | Length                         |                                                                                             | (0.1~300)mm                                                    | $U=0.1\text{mm}$                  |      |                |
|    |                                                                            | Velocity ratio                 |                                                                                             | (5~40)%                                                        | $U=2\%$                           |      |                |
| 24 | *Chassis Dynamometers for Automobile                                       | force                          | C.S. for Chassis Dynamometers for Automobile Emissions                                      | torque: (0~16)kN                                               | $U_{\text{rel}}=0.6\%$            |      |                |



No. CNAS L0730

第 142 页 共 163 页

The scope of the accreditation in Chinese remains the definitive version.

| №                                                                  | Instrument                                    | Measurand                              | Calibration Method                                           | Range                                             | Expanded Uncertainty<br>(k=2) | Note                                                                                 | Effective Date |
|--------------------------------------------------------------------|-----------------------------------------------|----------------------------------------|--------------------------------------------------------------|---------------------------------------------------|-------------------------------|--------------------------------------------------------------------------------------|----------------|
|                                                                    | Emissions Testing                             | speed                                  | Testing JJF 1221                                             | (0~100)km/h                                       | $U_{rel}=0.26\%$              |                                                                                      |                |
|                                                                    |                                               | Length                                 |                                                              | (0.1~600)mm                                       | $U=0.04\text{mm}$             |                                                                                      |                |
|                                                                    |                                               | Time                                   |                                                              | (0.01~3600)s                                      | $U_{rel}=1\%$                 |                                                                                      |                |
| 25                                                                 | *Simulation transportation vibration test bed | Frequency                              | C.M.for Simulated Transport Vibration Test Bed ZBDX 1954     | 1Hz~5Hz                                           | $U_{rel}=3.7\%$               |                                                                                      |                |
|                                                                    |                                               |                                        |                                                              | 5Hz~2000Hz                                        | $U_{rel}=0.5\%$               |                                                                                      |                |
|                                                                    |                                               | Total root-mean-square of acceleration |                                                              | (0.1~100) m/s <sup>2</sup>                        | $U_{rel}=1.8\%$               |                                                                                      |                |
|                                                                    |                                               | Displacement                           |                                                              | (0.01~30) mm                                      | $U_{rel}=2.8\%$               |                                                                                      |                |
|                                                                    |                                               | Distortion                             |                                                              | (0.01~30)%                                        | $U_{rel}=10\%$                |                                                                                      |                |
| 26                                                                 | *Simulation Test-bed for Road Transportation  | Frequency                              | C.S.for Simulation Test-bed for Road Transportation JJF 1271 | (5~400) Hz                                        | $U_{rel}=0.16\%$              |                                                                                      |                |
|                                                                    |                                               | Acceleration                           |                                                              | (0.1~100)m/s <sup>2</sup>                         | $U_{rel}=3.4\%$               |                                                                                      |                |
|                                                                    |                                               | Acceleration power spectral density    |                                                              | (0.01~10000) (m/s <sup>2</sup> ) <sup>2</sup> /Hz | $U_{rel}=3.4\%$               |                                                                                      |                |
| 13 Specialized Measuring Instrument for Meteorology & Marine Field |                                               |                                        |                                                              |                                                   |                               |                                                                                      |                |
| 1                                                                  | Meteorological Bimetallic of Thermograph      | Temperature                            | V.R.of Meteorological Bimetallic of Thermograph JJG 287      | (-50~+50)°C                                       | $U=0.3^{\circ}\text{C}$       |  |                |
| 2                                                                  | Bucket Thermometers                           | Temperature                            | V.R.of Bucket Thermometers JJG 289                           | (-5~+40)°C                                        | $U=0.09$                      |                                                                                      |                |
| 3                                                                  | Pitot tubes                                   | coefficient of Pitot tubes             | V.R.of Pitot Tubes JJG518                                    | L Type(0.95~1.05)、S Type(0.75~0.90)               | $U_{rel}=0.4\%$               |                                                                                      |                |

No. CNAS L0730

第 143 页 共 163 页



在线扫码获取验证

The scope of the accreditation in Chinese remains the definitive version.


| №                                                                     | Instrument                                   | Measurand      | Calibration Method                                          | Range                      | Expanded Uncertainty<br>( $k=2$ ) | Note | Effective Date |
|-----------------------------------------------------------------------|----------------------------------------------|----------------|-------------------------------------------------------------|----------------------------|-----------------------------------|------|----------------|
| 4                                                                     | *Portable Induction Anemometer               | Wing speed     | V.R.of Portable Induction Anemometer JJG515                 | (0.2~50) m/s               | $U=(0.1\sim1.0)$ m/s              |      |                |
| 5                                                                     | *Portable 3-cup Anemometer                   | Wing speed     | V.R.of Portable 3-cup Anemometers JJG431                    | (0.2~50) m/s               | $U=(0.1\sim1.0)$ m/s              |      |                |
| 6                                                                     | *Contact Anemometer                          | Wing speed     | V.R.of Contact Anemometer JJG613                            | (0.2~50) m/s               | $U=(0.1\sim1.0)$ m/s              |      |                |
| 7                                                                     | *Hot Ball Shaped Anemometer                  | Wing speed     | V.R.of Hot Ball Shaped Anemometer JJG(建设)0001               | (0.2~50) m/s               | $U=(0.2\sim1.5)\%$ FS             |      |                |
| 8                                                                     | Magnetoelectricity Wind Sensor for Wind Farm | Wind speed     | C.S.of Magnetoelectricity Wind Sensor for Wind Farm JJF1431 | (0.2~50)m/s                | $U=(0.1\sim1)$ m/s                |      |                |
|                                                                       |                                              | Frequency      |                                                             | (0.1~1000)Hz               | $U_{rel}=1.0\%$                   |      |                |
| 9                                                                     | Anemometers Used in Marine Field             | Air velocity   | V.R.of Anemometers Used in Marine Field JJG1167             | (0.2~50)m/s                | $U=0.1$ m/s                       |      |                |
|                                                                       |                                              | Wind Directino |                                                             | (0~360)°                   | $U=1.6^\circ$                     |      |                |
| 10                                                                    | Fanning Mill Anemometer                      | Wind Velocity  | C.S.for Fanning Mill Anemometer JJF1971                     | (1~30)m/s                  | $U=(0.10\sim0.40)$ m/s            |      |                |
| 11                                                                    | *Exhaust Air Rate Instrument                 | Air volume     | Calibration Method of Exhaust Air Rate Instrument FFN1001   | (60~6000)m <sup>3</sup> /h | $U_{rel}=(0.5\sim2.2)\%$          |      |                |
| 14 Specialized Measuring Instrument for Papermaking & Paper and Board |                                              |                |                                                             |                            |                                   |      |                |
| 1                                                                     | *Grammage Tester for Paper and Board         | Mass           | V.R. of Grammage Tester for Paper and Board JJG (轻工) 54.2   | (1~50) g                   | $U_{rel}=0.02\%\sim0.3\%$         |      |                |
| 2                                                                     | *MIT Type Folding Endurance Tester           | Force          | V.R. of MIT Type Folding Endurance Tester JJG (轻工) 59       | (4.41~15.19) N             | $U=0.1$ N                         |      |                |
|                                                                       |                                              | Angle          |                                                             | (0~320) °                  | $U=1'$                            |      |                |



No. CNAS L0730

第 144 页 共 163 页

The scope of the accreditation in Chinese remains the definitive version.

| №                                                      | Instrument                                  | Measurand      | Calibration Method                                                               | Range                                  | Expanded Uncertainty<br>(k=2) | Note                                                                                 | Effective Date |
|--------------------------------------------------------|---------------------------------------------|----------------|----------------------------------------------------------------------------------|----------------------------------------|-------------------------------|--------------------------------------------------------------------------------------|----------------|
|                                                        |                                             | Speed          |                                                                                  | (10~250) min <sup>-1</sup>             | U=0.6min <sup>-1</sup>        |                                                                                      |                |
| 3                                                      | *Absorption Tester for Paper and Board      | Length         | V.R. of Absorption Tester for Paper and Board JJG(轻工) 555                        | (0~200) mm                             | U=0.03 mm                     |                                                                                      |                |
|                                                        |                                             | Mass           |                                                                                  | (10~15) kg                             | U=6 g                         |                                                                                      |                |
|                                                        |                                             | Area           |                                                                                  | (0.1~110) cm <sup>2</sup>              | U=0.06 cm <sup>2</sup>        |                                                                                      |                |
| 4                                                      | *Bursting Strength Testers for Paper(Board) | Pressure       | Calibration Specification for Bursting Strength Testers for Paper(Board) JJF1811 | (0~6) MPa                              | U=0.07%FS                     |                                                                                      |                |
| 5                                                      | *Compression Strength Tester for Board      | Force value    | V.R.of Compression Strength Tester for Board JJG(粤) 018                          | (0.2~5)kN                              | U <sub>rel</sub> =0.15%       |                                                                                      |                |
|                                                        |                                             | Test speed     |                                                                                  | (5~20)mm/min                           | U=0.5mm/min                   |                                                                                      |                |
| 6                                                      | *Carton Box Tester of Crush Resistance      | Force value    | V.R.of Carton Box Tester of Crush Resistance JJG(粤) 043                          | (1~50)kN                               | U <sub>rel</sub> =0.34%       |                                                                                      |                |
|                                                        |                                             | Pressing speed |                                                                                  | Constant speed type:<br>(5~15)mm/min   | U=0.5mm/min                   |                                                                                      |                |
|                                                        |                                             |                |                                                                                  | Adjustable speed type:<br>(1~65)mm/min | U <sub>rel</sub> =1%          |                                                                                      |                |
|                                                        |                                             | Deformation    |                                                                                  | (0~500)mm                              | U=0.06mm                      |                                                                                      |                |
| 15 Specialized Measuring Instrument for Medical Device |                                             |                |                                                                                  |                                        |                               |                                                                                      |                |
| 1                                                      | *Electrocardiograph                         | Voltage        | V.R.of Electrocardiograph JJG543                                                 | 0.1mV~1V                               | U <sub>rel</sub> =1.6%        |  |                |
| 2                                                      | *Digital Electrocardiograph                 | Voltage        | V.R.of Digital Electrocardiograph JJG1041                                        | 0.03V~10V                              | U <sub>rel</sub> =1.6%        |                                                                                      |                |
|                                                        |                                             | Heart Rate     |                                                                                  | (10~200)bpm                            | U=1bpm                        |                                                                                      |                |
| 3                                                      | *Electro cardiac monitor                    | Voltage        | V.R.of Electro Cardiac Monitor JJG 760                                           | (0.1~1000)mV                           | U <sub>rel</sub> =2.2%        |                                                                                      |                |
|                                                        |                                             | Heart rate     |                                                                                  | (10~200)bpm                            | U=1bpm                        |                                                                                      |                |



No. CNAS L0730

第 145 页 共 163 页

The scope of the accreditation in Chinese remains the definitive version.

| №  | Instrument                                    | Measurand            | Calibration Method                                           | Range                       | Expanded Uncertainty<br>( $k=2$ ) | Note | Effective Date |
|----|-----------------------------------------------|----------------------|--------------------------------------------------------------|-----------------------------|-----------------------------------|------|----------------|
| 4  | *Multifunction Patient Monitoring Instruments | Voltage              | V.R.of Multifunction Patient Monitoring Instruments JJG 1163 | (1~1000)mV                  | $U_{rel}=2.2\%$                   |      |                |
|    |                                               | Static pressure      |                                                              | (0~40)kPa                   | $U=0.2\text{kPa}$                 |      |                |
|    |                                               | Oxygen saturation    |                                                              | 35%~100%                    | $U=2\%$                           |      |                |
| 5  | *Medical Ultrasonic Diagnostic Equipment      | Ultrasonic intensity | V.R.of Medical Ultrasonic Diagnostic Equipment JJG639        | (0.1~100)mW/cm <sup>2</sup> | $U_{rel}=10\%$                    |      |                |
| 6  | *ELISA Analytical Instruments                 | Wavelength           | V.R.of ELISA Analytical Instruments JJG861                   | I、II: (360~700) nm          | $U=1.0\text{nm}$                  |      |                |
|    |                                               |                      |                                                              | III: (360~700) nm           | $U=1.4\text{nm}$                  |      |                |
|    |                                               | Absorbance           |                                                              | 0.10~1.70                   | $U=0.02$                          |      |                |
| 7  | *Sphygmomanometer                             | Pressure             | V.R.of Sphygmomanometer JJG 270                              | (0~40) kPa                  | $U=0.3\text{kPa}$                 |      |                |
| 8  | *Non-invasive automated sphygmomanometers     | Pressure             | V.R.of Non-invasive automated sphygmomanometers JJG 692      | (0~40) kPa                  | $U=0.3\text{kPa}$                 |      |                |
| 9  | *Electroencephalographs                       | Voltage              | V.R.of Electroencephalographs JJG 1043                       | 0.005mV~5V                  | $U_{rel}=2.1\%$                   |      |                |
|    |                                               | Time                 |                                                              | 0.05s~10s                   | $U_{rel}=1.0\%$                   |      |                |
| 10 | *Cardiac defibrillators-monitor               | Energy               | C.S.for Cardiac Defibrillators JJF 1149                      | (0.1~99.9)J                 | $U=2.2\text{J}$                   |      |                |
|    |                                               |                      |                                                              | (100~360)J                  | $U_{rel}=6\%$                     |      |                |
|    |                                               | Heart rate           |                                                              | (10~200)bpm                 | $U=1\text{bpm}$                   |      |                |
| 11 | *Baby Incubator                               | Temperature          | C.S.for Baby Incubator JJF1260                               | (10~40)°C                   | $U=0.22^{\circ}\text{C}$          |      |                |



No. CNAS L0730

第 146 页 共 163 页

The scope of the accreditation in Chinese remains the definitive version.

| №  | Instrument                       | Measurand                      | Calibration Method                                                     | Range              | Expanded Uncertainty<br>( $k=2$ ) | Note | Effective Date |
|----|----------------------------------|--------------------------------|------------------------------------------------------------------------|--------------------|-----------------------------------|------|----------------|
| 12 | *Hemodialysis Equipment          | Noise                          | C.S.for Hemodialysis Equipment JJF1353                                 | (30~90)dB          | $U=1.4\text{dB}$                  |      |                |
|    |                                  | Humidity                       |                                                                        | 30 RH%~90 RH%      | $U=2.2\%\text{RH}$                |      |                |
|    |                                  | Oxygen Concentration           |                                                                        | (30~40) %          | $U=1.8\%$                         |      |                |
|    |                                  | Conductivity                   |                                                                        | (12.5~15.5)mS/cm   | $U=0.14\text{mS/cm}$              |      |                |
|    |                                  | Temperature                    |                                                                        | (35~40)°C          | $U=0.2^\circ\text{C}$             |      |                |
|    |                                  | Pressure                       |                                                                        | (-40~60) kPa       | $U=0.3\text{kPa}$                 |      |                |
|    |                                  | pH                             |                                                                        | 0.01~14            | $U=0.04$                          |      |                |
| 13 | *Urine Analyzers                 | Voltage                        | Calibration Specification of Urine Analyzers JJF1129                   | (0.1~2000) mL/min  | $U_{\text{rel}}=2.0\%$            |      |                |
|    |                                  | pH                             |                                                                        | 5.5~7.5            | $U=0.4$                           |      |                |
|    |                                  | PRO                            |                                                                        | (0~2.0)g/L         | $U=0.4\text{g/L}$                 |      |                |
|    |                                  | GLU                            |                                                                        | (0~43)mmol/L       | $U=1.1\text{mmol/L}$              |      |                |
| 14 | *Syringe Pumps and Infusion Pump | Flow rate                      | Calibration Specification for Syringe Pumps and Infusion Pumps JJF1259 | (5.00~19.99)mL/h   | $U_{\text{rel}}=2.5\%$            |      |                |
|    |                                  |                                |                                                                        | (20~200)mL/h       | $U_{\text{rel}}=1.3\%$            |      |                |
|    |                                  |                                |                                                                        | (200.1~1000.0)mL/h | $U_{\text{rel}}=2.4\%$            |      |                |
| 15 | *Electrosurgical Generator       | High-frequency leakage current | Calibration Specification for Electrosurgical Generator JJF1217        | (0.001~0.5)A       | $U_{\text{rel}}=2.8\%$            |      |                |
|    |                                  | output power                   |                                                                        | (1~500)W           | $U_{\text{rel}}=6\%$              |      |                |



No. CNAS L0730

第 147 页 共 163 页

The scope of the accreditation in Chinese remains the definitive version.


| №  | Instrument                                       | Measurand              | Calibration Method                                                     | Range                                     | Expanded Uncertainty<br>( <i>k</i> =2) | Note | Effective Date |
|----|--------------------------------------------------|------------------------|------------------------------------------------------------------------|-------------------------------------------|----------------------------------------|------|----------------|
| 16 | *Ventilators                                     | tidal volume           | Calibration Specification<br>for Ventilators JJF1234                   | (0.1～2000) mL                             | <i>U</i> <sub>rel</sub> =3.6%          |      |                |
|    |                                                  | frequency              |                                                                        | (1～80) BPM                                | <i>U</i> =1BPM                         |      |                |
|    |                                                  | pressure               |                                                                        | (0～10) kPa                                | <i>U</i> =0.14kPa                      |      |                |
|    |                                                  | oxygen                 |                                                                        | 21%～100%                                  | <i>U</i> =2.4%                         |      |                |
| 17 | *Ultrasound Bone<br>Sonometers                   | speed of<br>sound      | Calibration Specification<br>for Ultrasound Bone<br>Sonometers JJF1649 | (1400～1700)m/s,<br>(2500～3000) m/s        | <i>U</i> <sub>rel</sub> =2.1%          |      |                |
| 18 | Pulmonary<br>Function<br>Measuring<br>Instrument | Capacity               | C.S.for the Pulmonary<br>Function Measuring<br>Instrument JJF 1213     | (0～10)L                                   | <i>U</i> <sub>rel</sub> =1.4%          |      |                |
|    |                                                  | Flow                   |                                                                        | (2～14)L/s                                 | <i>U</i> <sub>rel</sub> =2.5%          |      |                |
|    |                                                  | Concentration          |                                                                        | (0～30)%                                   | <i>U</i> <sub>rel</sub> =1.7%          |      |                |
|    |                                                  |                        |                                                                        | (0～30)%                                   | <i>U</i> <sub>rel</sub> =1.8%          |      |                |
| 19 | *Electrolyte<br>Analyzer                         | Ionic<br>Concentration | V.R.of Electrolyte Analyzers<br>JJG 1051                               | K <sup>+</sup> : (1.50～7.50)<br>mmol/L    | <i>U</i> <sub>rel</sub> =4%            |      |                |
|    |                                                  |                        |                                                                        | Na <sup>+</sup> : (100.0～180.0)<br>mmol/L | <i>U</i> <sub>rel</sub> =3%            |      |                |
|    |                                                  |                        |                                                                        | Cl <sup>-</sup> : (80.0～160.0)<br>mmol/L  | <i>U</i> <sub>rel</sub> =4%            |      |                |
|    |                                                  |                        |                                                                        | Li <sup>+</sup> : (0.40～2.00)<br>mmol/L   | <i>U</i> <sub>rel</sub> =5%            |      |                |
|    |                                                  |                        |                                                                        | iCa <sup>2+</sup> : (0.50～2.50)<br>mmol/L | <i>U</i> <sub>rel</sub> =6%            |      |                |
| 20 | *Digital<br>Electroencephalogr<br>aphs           | Voltage                | V.R.of Digital<br>Electroencephalographs JJG<br>954                    | 2 μ V～2mV                                 | <i>U</i> <sub>rel</sub> =4%            |      |                |
|    |                                                  | Period                 |                                                                        | (0.01～10)s                                | <i>U</i> <sub>rel</sub> =2.2%          |      |                |
| 21 | *Ambulatory                                      | Voltage                | V.R.of Ambulatory                                                      | (0.05～6)mV                                | <i>U</i> <sub>rel</sub> =4%            |      |                |



No. CNAS L0730

第 148 页 共 163 页

The scope of the accreditation in Chinese remains the definitive version.

| №                                                                 | Instrument                               | Measurand        | Calibration Method                                            | Range          | Expanded Uncertainty<br>( <i>k</i> =2) | Note                                                                                 | Effective Date |
|-------------------------------------------------------------------|------------------------------------------|------------------|---------------------------------------------------------------|----------------|----------------------------------------|--------------------------------------------------------------------------------------|----------------|
|                                                                   | Electrocardiographs                      | Frequency        | Electrocardiographs JJG 1042                                  | (0.5～60)Hz     | <i>U</i> <sub>rel</sub> =1.2%          |                                                                                      |                |
| 22                                                                | *Exercise Treadmills                     | Voltage          | C.S.for Exercise Treadmills JJF 1722                          | (0.05～10)mV    | <i>U</i> <sub>rel</sub> =4%            |                                                                                      |                |
|                                                                   |                                          | Heart Rate       |                                                               | (30～300)次/分    | <i>U</i> <sub>rel</sub> =1.6%          |                                                                                      |                |
|                                                                   |                                          | Velocity         |                                                               | (5～10)km/h     | <i>U</i> <sub>rel</sub> =7%            |                                                                                      |                |
| 16 Specialized Measuring Instrument for Building & Transportation |                                          |                  |                                                               |                |                                        |                                                                                      |                |
| 1                                                                 | Rebound Test Hammer                      | Length           | V.R.of the Rebound Test Hammer JJG 817                        | (0～150) mm     | <i>U</i> =0.04 mm                      |                                                                                      |                |
|                                                                   |                                          | Friction force   |                                                               | (0.1～1) N      | <i>U</i> =0.07 N                       |                                                                                      |                |
|                                                                   |                                          | Spring rigidity  |                                                               | (65～1045) N/m  | <i>U</i> =8 N/m                        |                                                                                      |                |
| 2                                                                 | *Vibrator for Compacting Mortar Specimen | Swing            | C.S.for Cement Mortar Vibration Compaction Equipment JJF 1867 | (0～1) mm       | <i>U</i> =0.014mm                      |                                                                                      |                |
|                                                                   |                                          | Frequency        |                                                               | (10～100) Hz    | <i>U</i> =0.6Hz                        |                                                                                      |                |
|                                                                   |                                          | Time             |                                                               | (0～5)min       | <i>U</i> =0.4s                         |                                                                                      |                |
|                                                                   |                                          | Mass             |                                                               | (0～15)kg       | <i>U</i> =6g                           |                                                                                      |                |
|                                                                   |                                          | Size             |                                                               | (0～200)mm      | <i>U</i> =0.04mm                       |                                                                                      |                |
| 3                                                                 | *Coment Mortars Mixer                    | Displacement     | V.R.for Cement Mortars Mixer JJG(建材) 102                      | (1～200) mm     | <i>U</i> =0.08 mm                      |  |                |
|                                                                   |                                          | Rotational speed |                                                               | (10～300) r/min | <i>U</i> =0.7 r/min                    |                                                                                      |                |
|                                                                   |                                          | Time             |                                                               | (10～200) s     | <i>U</i> =0.4 s                        |                                                                                      |                |
| 4                                                                 | *Mixer for cement paste                  | Diameter         | C.S.for Mixer for Cement Paste JJF(JianCai) 104               | (0.5～160) mm   | <i>U</i> =0.05mm                       |                                                                                      |                |
|                                                                   |                                          | Rotational speed |                                                               | (10～300) r/min | <i>U</i> =0.6r/min                     |                                                                                      |                |



No. CNAS L0730

第 149 页 共 163 页

The scope of the accreditation in Chinese remains the definitive version.

| №  | Instrument                                     | Measurand                 | Calibration Method                                                                       | Range           | Expanded Uncertainty<br>( $k=2$ ) | Note | Effective Date |
|----|------------------------------------------------|---------------------------|------------------------------------------------------------------------------------------|-----------------|-----------------------------------|------|----------------|
|    |                                                | Time                      |                                                                                          | (1~3600) s      | $U=0.3$ s                         |      |                |
| 5  | *Mixer for Mixing Mortars                      | Width                     | C.S.for Mixer for Mixing Mortars JJF(JianCai) 123                                        | (0.5~200) mm    | $U=0.05$ mm                       |      |                |
|    |                                                | Rotational speed          |                                                                                          | (10~300) r/min  | $U=0.6$ r/min                     |      |                |
|    |                                                | Working time              |                                                                                          | (1~3600) s      | $U=0.3$ s                         |      |                |
| 6  | *Apparatus of fluidity of cement mortar        | Length                    | V.R.for Inspecting specification for apparatus of fluidity of cement mortar JJG(建材) 126  | (10~300) mm     | $U=0.06$ mm                       |      |                |
|    |                                                | Mass                      |                                                                                          | 200 g~34 kg     | $U=0.18$ g                        |      |                |
|    |                                                | Time                      |                                                                                          | (15~120) s      | $U=0.4$ s                         |      |                |
| 7  | *Jolting Table for Compacting Mortars Specimen | Working Time              | C.S.for Jolting Table for Compacting Mortars Specimen JJF(JianCai) 124                   | (1~3600) s      | $U=0.3$ s                         |      |                |
|    |                                                | Displacement              |                                                                                          | (10~1000) mm    | $U=0.07$ mm                       |      |                |
|    |                                                | Mass                      |                                                                                          | (0~15)kg        | $U=0.07$ kg                       |      |                |
| 8  | *Wearing Machine for Cement Mortar             | Force                     | V.R.of Wearing Machine for Cement Mortar JJG(建材) 125                                     | (20~1000) N     | $U_{rel}=0.37\%$                  |      |                |
|    |                                                | Displacement              |                                                                                          | (0.5~200) mm    | $U=0.06$ mm                       |      |                |
|    |                                                | Rotational speed          |                                                                                          | (10~1000) r/min | $U=0.4$ r/min                     |      |                |
| 9  | *Expansion meter of Le-chatelier apparatus     | Staff Scale               | C.S.for the Expansion Tester for Cement Le Chatelier Needles JJF (建材) 110                | (2~10) mm       | $U_{rel}=0.6\%$                   |      |                |
|    |                                                | Mass                      |                                                                                          | (2~1200) g      | $U=0.06$ g                        |      |                |
| 10 | *Compacting Instrument                         | Mass of Compacting Hammer | V.R.of Marshall Asphalt Compactor JJG (交通) 065,V.R.of Standard Compactor JJD 1024,V.R.of | (20~15000) g    | $U=0.6$ g                         |      |                |
|    |                                                | Length                    |                                                                                          | H:(0~500) mm    | $U=0.1$ mm                        |      |                |



No. CNAS L0730

第 150 页 共 163 页

The scope of the accreditation in Chinese remains the definitive version.

| №  | Instrument                                                                | Measurand         | Calibration Method                                                                                           | Range               | Expanded Uncertainty<br>(k=2) | Note | Effective Date |
|----|---------------------------------------------------------------------------|-------------------|--------------------------------------------------------------------------------------------------------------|---------------------|-------------------------------|------|----------------|
|    |                                                                           |                   | Compacting Instrument of Soil JJG (JIAOTONG) 058                                                             | Dimmter: (0~150) mm | U=0.03mm                      |      |                |
| 11 | *Air Entrainment Meter of Freshly Mixed Concrete by The Volumetric Method | Presure           | V.R.of Air Entrainment Meter of Freshly Mixed Concrete by The Volumetric Method JJG(交通) 094                  | (0~0.25)MPa         | U=0.3kPa                      |      |                |
|    |                                                                           | Volume            |                                                                                                              | (1~8) L             | U=2mL                         |      |                |
|    |                                                                           | Air content       |                                                                                                              | 0%~10%              | U=0.008%~0.08%                |      |                |
| 12 | *Apparatus to Measure Water Permeability of Concrete                      | Presure value     | C.S. of Apparatus to Measure Water Permeability of Concrete JJF 1812                                         | (0~6) MPa           | U=0.01 MPa                    |      |                |
| 13 | *Flow Table for Determine Cement Mortar Fluidity                          | Length            | V.R. of Metallic Webster Hardness Testing Machines JJG(交通) 096                                               | (10~300) mm         | U=0.06 mm                     |      |                |
|    |                                                                           | Mass              |                                                                                                              | (0.2~34)kg          | U=18 g                        |      |                |
|    |                                                                           | Time              |                                                                                                              | (15~120) s          | U=0.4 s                       |      |                |
| 14 | Instrument of Testing Mortar-strength by Penetration Resistance Method    | Force value       | Calibration Specification for Instrument of Testing Mortar-strength by Penetration Resistance Mtghod JJF1372 | (100~1000)N         | U=3.7N                        |      |                |
|    |                                                                           | Length            |                                                                                                              | (0~150)mm           | U=0.03mm                      |      |                |
| 15 | *Asphalt Mixture's Mixing Machine                                         | Volume            | Verification Regulation of Asphalt Mixture's Mixing Machine JJG(交通) 064                                      | (5~22)L             | U <sub>rel</sub> =0.1%        |      |                |
|    |                                                                           | Revolution Speed  |                                                                                                              | (10~200)r/min       | U <sub>rel</sub> =0.1%        |      |                |
|    |                                                                           | Time              |                                                                                                              | (0~4)min            | U=0.03min                     |      |                |
|    |                                                                           | Temperature       |                                                                                                              | (0~300)°C           | U=0.3°C                       |      |                |
| 16 | *Los Angeles Testing Machine                                              | Revolution Speed  | Verification Regulation of Los Angeles Testing Machine JJG(交通) 108                                           | (10~50)r/min        | U=0.2r/min                    |      |                |
|    |                                                                           | Internal diameter |                                                                                                              | (700~720)mm         | U=1mm                         |      |                |



No. CNAS L0730

第 151 页 共 163 页

The scope of the accreditation in Chinese remains the definitive version.

| №  | Instrument                                                                   | Measurand         | Calibration Method                                                                                                 | Range                  | Expanded Uncertainty<br>( $k=2$ ) | Note | Effective Date |
|----|------------------------------------------------------------------------------|-------------------|--------------------------------------------------------------------------------------------------------------------|------------------------|-----------------------------------|------|----------------|
|    |                                                                              | Length            |                                                                                                                    | (500~520)mm            | $U=1\text{mm}$                    |      |                |
|    |                                                                              | Diameter          |                                                                                                                    | (40~50)mm              | $U=0.06\text{mm}$                 |      |                |
|    |                                                                              | Mass              |                                                                                                                    | (2000~6000)g           | $U=0.6\text{g}$                   |      |                |
| 17 | *Asphalt Centrifugal Extractor                                               | Revolution Speed  | Verification Regulation of Asphalt Centrifugal Extractor JJG(交通) 132                                               | (100~20000)r/min       | $U_{\text{rel}}=0.4\%\sim 0.1\%$  |      |                |
|    |                                                                              | Mesh size         |                                                                                                                    | (0~1)mm                | $U=3.7\text{ }\mu\text{m}$        |      |                |
|    |                                                                              | Filament diameter |                                                                                                                    | (0~1)mm                | $U=3.6\text{ }\mu\text{m}$        |      |                |
| 18 | *Apparatus Time of Setting of Concrete Mixture by Penetration Resistance     | Force value       | Verification Regulation of Apparatus for Time of Setting of Concrete Mixture by Penetration Resistance JJG(交通) 095 | (100~3000)N            | $U_{\text{rel}}=0.36\%$           |      |                |
|    |                                                                              | Diameter          |                                                                                                                    | (0~30)mm               | $U=0.004\text{mm}$                |      |                |
|    |                                                                              | Size              |                                                                                                                    | (0~300)mm              | $U=0.08\text{mm}$                 |      |                |
| 19 | *Cement Fineness Negative Pressure Screen Analyzers                          | Pressure          | Calibration Specification for Cement Fineness Negative Pressure Screen Analyzers JJF1827                           | (-100~0) hPa           | $U=0.1\text{hPa}$                 |      |                |
| 20 | *Test Apparatus for Theoretical Maximum Specific Gravity of Asphalt Mixtures | Pressure          | V.R.of Test Apparants for Theoretical Maximum Specific Gravity of Asphalt Mixtures JJG(交通) 105                     | (0~120) kPa            | $U=0.09\text{kPa}$                |      |                |
|    |                                                                              | Vacuum            |                                                                                                                    | (-100~0) kPa           | $U=0.5\text{kPa}$                 |      |                |
|    |                                                                              | Time              |                                                                                                                    | (0~24) h               | $U=3\text{s}$                     |      |                |
| 21 | *Nonmetal Building Materials Plastic Limit Measuring Instruments             | Length            | C.S.for Nonmetal Building Materials Plastic Limit Measuring Instruments JJF 1090                                   | (0~145)mm              | $U=0.04\text{mm}$                 |      |                |
|    |                                                                              | Angle             |                                                                                                                    | (0~50)°                | $U=6'$                            |      |                |
|    |                                                                              | Mass              |                                                                                                                    | (0~300)g               | $U=0.04\text{g}$                  |      |                |
| 22 | construction quality tester sets                                             | Length            | C.S of construction quality tester sets JJF 1110                                                                   | slope feeler: (0~35)mm | $U=0.02\text{mm}$                 |      |                |



No. CNAS L0730

第 152 页 共 163 页

The scope of the accreditation in Chinese remains the definitive version.

| №                                                                         | Instrument              | Measurand     | Calibration Method                        | Range                                     | Expanded Uncertainty<br>( $k=2$ )   | Note | Effective Date |
|---------------------------------------------------------------------------|-------------------------|---------------|-------------------------------------------|-------------------------------------------|-------------------------------------|------|----------------|
|                                                                           |                         | angle         |                                           | diagonal tester: -50 mm~+50 mm            | $U=0.1\text{mm}$                    |      |                |
|                                                                           |                         |               |                                           | grid : (50~500)mm                         | $U_{\text{rel}}=0.1\%$              |      |                |
|                                                                           |                         |               |                                           | gradient tester: -10 mm/m~+10 mm/m        | $U=0.05\text{mm/m}$                 |      |                |
|                                                                           |                         |               |                                           | vertical tester: -15 mm/2 m~+15 mm/2 m    | $U=0.05\text{mm/2m}$                |      |                |
|                                                                           |                         |               |                                           | angle tester: -10 mm/150 mm~+10 mm/150 mm | $U=0.05\text{mm/150mm}$             |      |                |
| 23                                                                        | *needle and flake gages | Length        | C.S for needle and flake gages JJF1593    | (0~100)mm                                 | $U=0.02\text{mm}$                   |      |                |
| 17 Specialized Measuring Instrument for Electrical & Electronic Appliance |                         |               |                                           |                                           |                                     |      |                |
| 1                                                                         | *Spark Tester           | AC Voltage    | C.S. of Wire Spark Tester JJF(LU) 63      | (1~30) kV,(45~65)Hz                       | $U_{\text{rel}}=(5\sim1.2)\%$       |      |                |
|                                                                           |                         | DC Voltage    |                                           | (1~30) kV                                 | $U_{\text{rel}}=(1.5\sim0.7)\%$     |      |                |
|                                                                           |                         | length        |                                           | (0.1~200)mm                               | $U=0.024\text{mm}$                  |      |                |
|                                                                           |                         |               |                                           | 200mm~5m                                  | $U=0.9\text{mm}$                    |      |                |
|                                                                           |                         | ripple factor |                                           | (0.1~10)%                                 | $U=0.2\%$                           |      |                |
| 2                                                                         | *Pulse Hi-pot tester    | Voltage       | C.S. of Pulse High Voltage Tester DJF 009 | (0.5~30) kV                               | $U_{\text{rel}}=(4\sim1.2)\%$       |      |                |
| 3                                                                         | *DC Electronic Load     | DC Voltage    | C.S. for DC Electronic Loads JJF 1462     | (0.1~1000)V                               | $U_{\text{rel}}=(0.014\sim0.011)\%$ |      |                |
|                                                                           |                         | DC Current    |                                           | (0.1~1000)A                               | $U_{\text{rel}}=(0.04\sim0.03)\%$   |      |                |
|                                                                           |                         | Power         |                                           | 0.1W~1kW                                  | $U_{\text{rel}}=(0.06\sim0.03)\%$   |      |                |
|                                                                           |                         | Current Speed |                                           | 0.1mA/ $\mu\text{s}$ ~10A/ $\mu\text{s}$  | $U_{\text{rel}}=1.3\%$              |      |                |



No. CNAS L0730

第 153 页 共 163 页

The scope of the accreditation in Chinese remains the definitive version.


| № | Instrument                            | Measurand                               | Calibration Method                                              | Range                                     | Expanded Uncertainty<br>( $k=2$ ) | Note | Effective Date |
|---|---------------------------------------|-----------------------------------------|-----------------------------------------------------------------|-------------------------------------------|-----------------------------------|------|----------------|
|   |                                       | Current Load<br>Time and<br>Unload time |                                                                 | 10 $\mu$ s ~ 10s                          | $U_{rel}=0.13\%$                  |      |                |
|   |                                       | Resistance                              |                                                                 | 0.1 $\Omega$ ~ 100k $\Omega$              | $U_{rel}=0.13\%$                  |      |                |
| 4 | *Cable Testers                        | Resistance                              | C.S. for Cable Testers JJF 1457                                 | 1 $\Omega$ ~ 100k $\Omega$                | $U_{rel}=(2.4\sim1.0)\%$          |      |                |
|   |                                       | Insulation Resistance                   |                                                                 | 100 $\Omega$ ~ 100M $\Omega$              | $U_{rel}=(2.4\sim1.0)\%$          |      |                |
|   |                                       |                                         |                                                                 | 100M $\Omega$ ~ 10G $\Omega$              | $U_{rel}=(1.5\sim2.4)\%$          |      |                |
|   |                                       | AC/DC voltage                           |                                                                 | (4~2000)V,(DC,45Hz~65Hz)                  | $U_{rel}=1.2\%$                   |      |                |
|   |                                       | Current                                 |                                                                 | (0.1~100)mA,(DC,45Hz~65Hz)                | $U_{rel}=1.2\%$                   |      |                |
|   |                                       | Capacitance                             |                                                                 | 100pF~100 $\mu$ F,1kHz                    | $U_{rel}=0.5\%$                   |      |                |
|   |                                       | Distortion                              |                                                                 | (0.01~100)%(20Hz~100kHz)                  | $U_{rel}=15\%$                    |      |                |
| 5 | *Turns Ratio Meter                    | Turns                                   | C.S. of Turns Ratio Meter DJF 016                               | 1:1~1:0.001(100Hz~100kHz)                 | $U_{rel}=0.12\%$                  |      |                |
| 6 | *Coil number testing instructing      | Coil Number                             | C.S. of coil number meters JJF(Electronic) 0041                 | 1 n~11110 n                               | $U_{rel}=0.23\%$                  |      |                |
| 7 | *Battery Internal Resistance Testers  | Voltage                                 | C.S. for Battery Internal Resistance Testers JJF1620            | 0.1V~1kV                                  | $U_{rel}=(0.01\sim1.5)\%$         |      |                |
|   |                                       | Internal Resistance                     |                                                                 | 1m $\Omega$ ~ 100k $\Omega$ (100Hz~10kHz) | $U_{rel}=(0.05\sim1.5)\%$         |      |                |
| 8 | *Charge & Discharge of Battery Tester | Voltage                                 | C.S. for Charge & Discharge of Battery Tester JJF(Military) 108 | (1~10)V                                   | $U_{rel}=(0.008\sim0.004)\%$      |      |                |
|   |                                       |                                         |                                                                 | (10~1000)V                                | $U_{rel}=(0.004\sim0.006)\%$      |      |                |
|   |                                       | Current                                 |                                                                 | (10~100)mA                                | $U_{rel}=(0.09\sim0.07)\%$        |      |                |



No. CNAS L0730

第 154 页 共 163 页

The scope of the accreditation in Chinese remains the definitive version.

| №                   | Instrument                                                         | Measurand                            | Calibration Method                                                                                                  | Range                           | Expanded Uncertainty<br>( <i>k</i> =2) | Note                                                                                 | Effective Date |
|---------------------|--------------------------------------------------------------------|--------------------------------------|---------------------------------------------------------------------------------------------------------------------|---------------------------------|----------------------------------------|--------------------------------------------------------------------------------------|----------------|
|                     |                                                                    | Resistance                           | ilac-MRA<br>CHINA NATIONAL ACCREDITATION SERVICE FOR CONFORMITY ASSESSMENT<br>SCHEDULE OF ACCREDITATION CERTIFICATE | 100mA~1kA                       | <i>U</i> <sub>rel</sub> =(0.03~0.09)%  |                                                                                      |                |
|                     |                                                                    |                                      |                                                                                                                     | 100m Ω ~10 Ω                    | <i>U</i> <sub>rel</sub> = (0.6~0.08)%  |                                                                                      |                |
|                     |                                                                    |                                      |                                                                                                                     | 10 Ω ~1k Ω                      | <i>U</i> <sub>rel</sub> = (0.08~0.09)% |                                                                                      |                |
|                     |                                                                    |                                      |                                                                                                                     | (1~10)k Ω                       | <i>U</i> <sub>rel</sub> =0.05%         |                                                                                      |                |
|                     |                                                                    | Power                                |                                                                                                                     | (10~100)mW                      | <i>U</i> <sub>rel</sub> = (0.18~0.07)% |                                                                                      |                |
|                     |                                                                    |                                      |                                                                                                                     | 100mW~100W                      | <i>U</i> <sub>rel</sub> = (0.09~0.04)% |                                                                                      |                |
|                     |                                                                    |                                      |                                                                                                                     | 100W~10kW                       | <i>U</i> <sub>rel</sub> = (0.04~0.07)% |                                                                                      |                |
|                     |                                                                    | Capacity                             |                                                                                                                     | 0.1Ah~10Ah                      | <i>U</i> <sub>rel</sub> = (0.09~0.04)% |                                                                                      |                |
|                     |                                                                    |                                      |                                                                                                                     | 10Ah~1kAh                       | <i>U</i> <sub>rel</sub> = (0.03~0.04)% |                                                                                      |                |
| Current<br>Risetime | 10 μ s~10ms                                                        | <i>U</i> <sub>rel</sub> = (0.3~0.2)% |                                                                                                                     |                                 |                                        |                                                                                      |                |
| 9                   | *Reference<br>Ballasts                                             | Voltage                              | C.S. for Reference Ballasts<br>JJF 1502                                                                             | (10~400)V/(0.1~20)A,50Hz,60Hz   | <i>U</i> <sub>rel</sub> =0.27%         |                                                                                      |                |
|                     |                                                                    | Current ratio                        |                                                                                                                     | (10~400)V/(0.1~20)A,20kHz~50kHz | <i>U</i> <sub>rel</sub> =0.3%          |                                                                                      |                |
|                     |                                                                    | Power Factor                         |                                                                                                                     | 0.1~1                           | <i>U</i> =0.006                        |                                                                                      |                |
| 10                  | *impulse voltage<br>testers for winding<br>interturn insulation    | peak pulse<br>voltage                | C.S. for impulse voltage<br>testers for winding interturn<br>insulation JJF 1691                                    | (0.5~30) kV                     | <i>U</i> <sub>rel</sub> =(4~1.2)%      |  |                |
|                     |                                                                    | wave front<br>time                   |                                                                                                                     | 2ns~1s                          | <i>U</i> <sub>rel</sub> =0.1%          |                                                                                      |                |
| 11                  | *Mechanical<br>Characteristic<br>Tester for High<br>Voltage Switch | Close Time                           | V.R. of High Voltage Switch<br>Operation Characteristic<br>Testers JJG 1120                                         | (1.0~1000.0)ms                  | <i>U</i> =(0.03~0.1)ms                 |                                                                                      |                |
|                     |                                                                    | Open time                            |                                                                                                                     | (1.0~1000.0)ms                  | <i>U</i> =(0.03~0.1)ms                 |                                                                                      |                |
|                     |                                                                    | Spring time                          |                                                                                                                     | (1.0~10.0)ms                    | <i>U</i> =0.03ms                       |                                                                                      |                |



No. CNAS L0730

第 155 页 共 163 页

The scope of the accreditation in Chinese remains the definitive version.


| №  | Instrument                                               | Measurand                          | Calibration Method                                           | Range                         | Expanded Uncertainty<br>( $k=2$ ) | Note | Effective Date |
|----|----------------------------------------------------------|------------------------------------|--------------------------------------------------------------|-------------------------------|-----------------------------------|------|----------------|
| 12 | *Calibration Specification for Pulse Spark Leak Detector | Pulse Peak voltage                 | C.S. for Pulsed Spark Leak Detector JJF(xin) 32              | (0.5~40) kV                   | $U_{rel}=3\%$                     |      |                |
| 13 | *Oscilloscope Voltage Probe                              | DC Voltage attenuation ratio       | C.S. for Oscilloscope Voltage Probes JJF 1437                | 1~1000                        | $U_{rel}=0.2\%$                   |      |                |
|    |                                                          | Frequency response                 |                                                              | (0~-3)dB(50kHz~6GHz)          | $U=0.9\text{dB}$                  |      |                |
|    |                                                          | Raise time                         |                                                              | 150ps~100 $\mu\text{s}$       | $U_{rel}=6.4\%$                   |      |                |
|    |                                                          | Input resistance                   |                                                              | 50 $\Omega$ ~ 100M $\Omega$   | $U_{rel}=0.6\%$                   |      |                |
| 14 | *Electrical Fast Transient/Burst Generator               | voltage                            | C.S. for Electrical fast Transient/Burst Simulators JJF 1672 | (0.1~4)kV                     | $U_{rel}=3\%$                     |      |                |
|    |                                                          | Pulse Rise Time                    |                                                              | (3~6) ns                      | $U_{rel}=6\%$                     |      |                |
|    |                                                          | Pulse group duration               |                                                              | 0.5ms~20ms                    | $U_{rel}=3\%$                     |      |                |
|    |                                                          | Pulse Repetition Frequency         |                                                              | 1kHz~100kHz                   | $U_{rel}=0.3\%$                   |      |                |
|    |                                                          | Pulse group period                 |                                                              | 250ms~350ms                   | $U_{rel}=0.3\%$                   |      |                |
|    |                                                          | Pulse width                        |                                                              | (35~150) ns                   | $U_{rel}=6\%$                     |      |                |
| 15 | *Current Probes of Oscilloscope, Current Probe           | DC current attenuation coefficient | C.S. for Current Probes of Oscilloscope JJF(Electronic)0036  | 1~100                         | $U_{rel}=0.7\%$                   |      |                |
|    |                                                          | AC current                         |                                                              | (0.001A~1000A)<br>(50Hz~1kHz) | $U_{rel}=0.7\%$                   |      |                |
|    |                                                          | DC current                         |                                                              | (0.001A~1000A)                | $U_{rel}=0.7\%$                   |      |                |



No. CNAS L0730

第 156 页 共 163 页

The scope of the accreditation in Chinese remains the definitive version.

| №  | Instrument                                       | Measurand                            | Calibration Method                                               | Range                         | Expanded Uncertainty<br>( <i>k</i> =2) | Note                                                                                 | Effective Date |
|----|--------------------------------------------------|--------------------------------------|------------------------------------------------------------------|-------------------------------|----------------------------------------|--------------------------------------------------------------------------------------|----------------|
|    |                                                  | Bandwidth                            |                                                                  | 1kHz~200MHz                   | <i>U</i> <sub>rel</sub> =4%            |                                                                                      |                |
|    |                                                  | Rise Time                            |                                                                  | 3ns~10ms                      | <i>U</i> <sub>rel</sub> =6.4%          |                                                                                      |                |
| 16 | Electrostatic Discharge Simulators               | Output voltage                       | C.S. for Electrostatic Discharge Simulators JJF 1397             | ±(0.1~40)kV                   | <i>U</i> <sub>rel</sub> =1.0%          |                                                                                      |                |
|    |                                                  | Peak current of discharge            |                                                                  | (0.1~150)A                    | <i>U</i> <sub>rel</sub> =5.6%          |                                                                                      |                |
|    |                                                  | 30ns/60ns Current of discharge       |                                                                  | (0.1~100)A                    | <i>U</i> <sub>rel</sub> =6%            |                                                                                      |                |
|    |                                                  | Rise time                            |                                                                  | (0.1~10)ns                    | <i>U</i> <sub>rel</sub> =4.5%          |                                                                                      |                |
| 17 | *Absorbing Clamp in the Range of 30MHz to 1.0GHz | Insertion loss                       | C.S.for Absorbing Clamp in the Range of 30MHz to 1.0GHz JJF 1155 | (10~30) dB,<br>(30MHz~1.0GHz) | <i>U</i> =1.7dB                        |                                                                                      |                |
| 18 | *Power frequency Magnetic Field Generator        | Current                              | C.S.for power frequency magnetic field simulators JJF 1737       | (0.1~600) A                   | <i>U</i> <sub>rel</sub> =2.5%          |                                                                                      |                |
|    |                                                  | Magnetic field intensity             |                                                                  | (0.1~15) A/m                  | <i>U</i> <sub>rel</sub> =6%            |                                                                                      |                |
|    |                                                  |                                      |                                                                  | (15~150) A/m                  | <i>U</i> <sub>rel</sub> =6.3%          |                                                                                      |                |
|    |                                                  |                                      |                                                                  | (150~1000) A/m                | <i>U</i> <sub>rel</sub> =12%           |                                                                                      |                |
| 19 | *Surge (impact) simulator                        | Peak-Voltage of Open Circuit         | C.S.for surge (impact) simulator JJF 1741                        | ±(0.05~15)kV                  | <i>U</i> <sub>rel</sub> =3.4%          |  |                |
|    |                                                  | Open circuit voltage wave front time |                                                                  | (0.1~100) μs                  | <i>U</i> <sub>rel</sub> =3.2%          |                                                                                      |                |
|    |                                                  | Open circuit voltage duration        |                                                                  | (1~1000) μs                   | <i>U</i> <sub>rel</sub> =3%            |                                                                                      |                |



No. CNAS L0730

第 157 页 共 163 页

The scope of the accreditation in Chinese remains the definitive version.

| №  | Instrument                                                                | Measurand                             | Calibration Method                                                                                            | Range                         | Expanded Uncertainty<br>(k=2) | Note                   | Effective Date |
|----|---------------------------------------------------------------------------|---------------------------------------|---------------------------------------------------------------------------------------------------------------|-------------------------------|-------------------------------|------------------------|----------------|
|    |                                                                           | Peak-Current of Short Circuit         | ilac-MRA<br>NATIONAL ACCREDITATION SERVICE FOR CONFORMITY ASSESSMENT<br>SCHEDULE OF ACCREDITATION CERTIFICATE | $\pm(0.001\sim 20)\text{kA}$  | $U_{\text{rel}}=3\%$          |                        |                |
|    |                                                                           | Short circuit current wave front time |                                                                                                               | $(0.01\sim 100)\ \mu\text{s}$ | $U_{\text{rel}}=3.2\%$        |                        |                |
|    |                                                                           | Short circuit current duration        |                                                                                                               | $(1\sim 1000)\ \mu\text{s}$   | $U_{\text{rel}}=3\%$          |                        |                |
|    |                                                                           | phase                                 |                                                                                                               | $(0\sim 360)^\circ$           | $U=3^\circ$                   |                        |                |
|    |                                                                           | Open circuit undershoot               |                                                                                                               | $(-50\sim -0.1)\ \%$          | $U_{\text{rel}}=3\%$          |                        |                |
|    |                                                                           | Short circuit undershoot              |                                                                                                               | $(-50\sim -0.1)\ \%$          | $U_{\text{rel}}=3\%$          |                        |                |
|    |                                                                           |                                       |                                                                                                               |                               |                               |                        |                |
| 20 | *Voltage Dips, Short Interruptions and Voltage Variations Test Generators | output voltage                        | C.S. for Voltage Dips, Short Interruptions and Voltage Variations Test Generators<br>JJF 1673                 | $(0.1\sim 360)\text{V}$       | $U_{\text{rel}}=1.0\%$        | 合格评定国家认可委员会<br>认可证书专用章 |                |
|    |                                                                           | load regulation                       |                                                                                                               | $0.01\%\sim 50\%$             | $U_{\text{rel}}=1.0\%$        |                        |                |
|    |                                                                           | voltage rise time and fall time       |                                                                                                               | $(1\sim 60)\ \mu\text{s}$     | $U_{\text{rel}}=7.7\%$        |                        |                |
|    |                                                                           | Voltage overshoot and undershoot      |                                                                                                               | $0.01\%\sim 20\%$             | $U_{\text{rel}}=5.0\%$        |                        |                |
|    |                                                                           | phase angle                           |                                                                                                               | $0^\circ\sim 360^\circ$       | $U=2^\circ$                   |                        |                |
|    |                                                                           | time of duration                      |                                                                                                               | $1\text{ms}\sim 60\text{s}$   | $U_{\text{rel}}=2.0\%$        |                        |                |
|    |                                                                           | interval time                         |                                                                                                               | $1\text{ms}\sim 60\text{s}$   | $U_{\text{rel}}=2.0\%$        |                        |                |
|    |                                                                           | Peak impulse current                  |                                                                                                               | $40\text{A}\sim 2\text{kA}$   | $U_{\text{rel}}=3.0\%$        |                        |                |



No. CNAS L0730

第 158 页 共 163 页


The scope of the accreditation in Chinese remains the definitive version.

| No | Instrument                              | Measurand                          | Calibration Method                                       | Range                  | Expanded Uncertainty<br>( $k=2$ ) | Note | Effective Date |
|----|-----------------------------------------|------------------------------------|----------------------------------------------------------|------------------------|-----------------------------------|------|----------------|
| 21 | *Harmonious and Flicker Analysis System | Current and linearity              | C.S. for Harmonious and Flicker Analysis System JJF 1205 | 0.01A ~ 1A             | $U_{rel}=0.1\%$                   |      |                |
|    |                                         |                                    |                                                          | 1A ~ 5A                | $U_{rel}=0.06\%$                  |      |                |
|    |                                         |                                    |                                                          | 5A ~ 60A               | $U_{rel}=0.1\%$                   |      |                |
|    |                                         | voltage amplitude                  |                                                          | (110~240)V             | $U_{rel}=(0.02\sim0.05)\%$        |      |                |
|    |                                         | power                              |                                                          | (1~12000)W, PF=1.0     | $U_{rel}=(0.08\sim0.10)\%$        |      |                |
|    |                                         |                                    |                                                          | (1~12000)W, PF=0.5     | $U_{rel}=(0.08\sim0.10)\%$        |      |                |
|    |                                         | harmonic current                   |                                                          | (0.01~6)A, 100Hz~3kHz  | $U_{rel}=(0.02\sim0.3)\%$         |      |                |
|    |                                         | The transient harmonic current     |                                                          | (1~10) A               | $U_{rel}=(0.32\sim0.7)\%$         |      |                |
|    |                                         | flicker                            |                                                          | Pst=1,Pst=5            | $U_{rel}=1.6\%$                   |      |                |
|    |                                         | inter-harmonics                    |                                                          | (0.003~6)A, 100Hz~2kHz | $U_{rel}=1.2\%\sim4.7\%$          |      |                |
|    |                                         | Harmonic content of supply voltage |                                                          | 0.01%~10%              | $U_{rel}=1.0\%$                   |      |                |
|    |                                         | Voltage stability with frequency   |                                                          | 0.1%~20%               | $U_{rel}=0.7\%$                   |      |                |
|    |                                         | Voltage with load stability        |                                                          | 0.1%~20%               | $U_{rel}=0.7\%$                   |      |                |



No. CNAS L0730

The scope of the accreditation in Chinese remains the definitive version.

| №  | Instrument                              | Measurand                                             | Calibration Method                                    | Range                           | Expanded Uncertainty<br>( $k=2$ ) | Note                                                                                 | Effective Date |
|----|-----------------------------------------|-------------------------------------------------------|-------------------------------------------------------|---------------------------------|-----------------------------------|--------------------------------------------------------------------------------------|----------------|
|    |                                         | Stability of voltage with load<br>frequency with load |                                                       | 0.1%~20%                        | $U_{rel}=0.2\%$                   |                                                                                      |                |
| 22 | *Discontinuous Interference Analyzers   | frequency                                             | C.S.for Discontinuous Interference Analyzers JJF 1845 | (0.15~30)MHz                    | $U_{rel}=0.30\%$                  |                                                                                      |                |
|    |                                         | level                                                 |                                                       | (45~110)dB $\mu$ V,(0.15~30)MHz | $U=0.3$ dB                        |                                                                                      |                |
|    |                                         | IF bandwidth                                          |                                                       | 9kHz                            | $U=0.09$ kHz                      |                                                                                      |                |
|    |                                         | impulse response                                      |                                                       | (0~30)dB, (0.15~30)MHz          | $U=0.7$ dB                        |                                                                                      |                |
| 23 | *Electrostatic Discharge Current Target | Insertion loss                                        | C.S. of ESD Current Targets JJF(Electronic)0040       | (-70dB~-30dB), (9kHz~1GHz)      | $U=0.20$ dB                       |                                                                                      |                |
|    |                                         | Transfer impedance                                    |                                                       | (-70dB~-30dB), (1GHz~4GHz)      | $U=0.30$ dB                       |                                                                                      |                |
|    |                                         | Input impedance                                       |                                                       | (0.02~1.0)V/A                   | $U_{rel}=0.12\%$                  |                                                                                      |                |
|    |                                         |                                                       |                                                       | (0.1~10) $\Omega$               | $U_{rel}=0.1\%$                   |                                                                                      |                |
| 24 | *Ring wave generator                    | Pulse voltage output                                  | C.S.for Oscillatory wave generator JJF(Zhe) 1059      | $\pm (0.05\sim4)$ kV            | $U_{rel}=3.4\%$                   |  |                |
|    |                                         | First peak voltage rise time                          |                                                       | 10ns~10 $\mu$ s                 | $U_{rel}=5\%$                     |                                                                                      |                |
|    |                                         | oscillation frequency                                 |                                                       | (1~2000)kHz                     | $U_{rel}=2.0\%$                   |                                                                                      |                |
|    |                                         | peak short circuit current                            |                                                       | $\pm (1\sim1000)$ A             | $U_{rel}=3.4\%$                   |                                                                                      |                |



No. CNAS L0730

第 160 页 共 163 页

The scope of the accreditation in Chinese remains the definitive version.

| №  | Instrument                       | Measurand                                  | Calibration Method                                          | Range                                 | Expanded Uncertainty<br>( $k=2$ ) | Note | Effective Date |
|----|----------------------------------|--------------------------------------------|-------------------------------------------------------------|---------------------------------------|-----------------------------------|------|----------------|
|    |                                  | First peak current rise time               |                                                             | 10ns~10 $\mu$ s                       | $U_{rel}=2.6\%$                   |      |                |
|    |                                  | output impedance                           |                                                             | (1~500) $\Omega$                      | $U_{rel}=4.8\%$                   |      |                |
|    |                                  | attenuation                                |                                                             | (10~150)%                             | $U_{rel}=3.4\%$                   |      |                |
|    |                                  | Pulse group duration                       |                                                             | 0.1s~100s                             | $U_{rel}=2.6\%$                   |      |                |
| 25 | *Oscillatory wave generator      | Open-circuit peak voltage                  | C.S.for Damped Oscillatory Wave Simulators JJF 2016         | $\pm$ (0.05~4)kV                      | $U_{rel}=3.4\%$                   |      |                |
|    |                                  | open circuit voltage rise time             |                                                             | 10ns~10 $\mu$ s                       | $U_{rel}=5\%$                     |      |                |
|    |                                  | oscillation frequency                      |                                                             | 1kHz~30MHz                            | $U_{rel}=4\%$                     |      |                |
|    |                                  | peak short circuit current                 |                                                             | $\pm$ (1~1000)A                       | $U_{rel}=3.4\%$                   |      |                |
|    |                                  | repetition frequency                       |                                                             | (1~5000)/s                            | $U_{rel}=2.0\%$                   |      |                |
|    |                                  | Open-circuit voltage oscillation frequency |                                                             | 1kHz~30MHz                            | $U_{rel}=2.0\%$                   |      |                |
|    |                                  | Durationtime                               |                                                             | 0.1s~100s                             | $U_{rel}=2.6\%$                   |      |                |
|    |                                  | attenuation                                |                                                             | (10~100)%                             | $U_{rel}=3.4\%$                   |      |                |
| 26 | *Triple-Loop antenna             | validation factor                          | C.S.for Triple-Loop Antenna Systems JJF(Electronic)0035     | (20~100)dB<br>$\Omega$ , 300kHz~30MHz | $U=1.5$ dB                        |      |                |
| 27 | *Pulse magnetic field generators | Peak output current                        | C.S.for Pulse magnetic field generators JJF(Electronic)0037 | (0.1~5)kA                             | $U_{rel}=3.0\%$                   |      |                |

No. CNAS L0730

第 161 页 共 163 页



The scope of the accreditation in Chinese remains the definitive version.

| №  | Instrument                           | Measurand                   | Calibration Method                                                | Range                       | Expanded Uncertainty<br>( $k=2$ ) | Note | Effective Date |
|----|--------------------------------------|-----------------------------|-------------------------------------------------------------------|-----------------------------|-----------------------------------|------|----------------|
|    |                                      | Front time                  |                                                                   | (5~10) $\mu$ s              | $U_{rel}=3.8\%$                   |      |                |
|    |                                      | Duration Time               |                                                                   | (10~30) $\mu$ s             | $U_{rel}=2.6\%$                   |      |                |
|    |                                      | Induction Coil Factor       |                                                                   | (0.5~110) 1/m               | $U_{rel}=5\%$                     |      |                |
| 28 | *High Frequency Noise Simulator      | Pulse voltage amplitude     | C.S.for High Frequency Noise Simulator JJF(Electronic) 30810      | (1~4000)V                   | $U_{rel}=3\%$                     |      |                |
|    |                                      | Pulse voltage width         |                                                                   | (50~1000)ns                 | $U_{rel}=3\%$                     |      |                |
|    |                                      | Pulse voltage rise time     |                                                                   | (0.3~1)ns                   | $U_{rel}=6\%$                     |      |                |
|    |                                      | Pulse Repetition Frequency  |                                                                   | (1~100)Hz                   | $U_{rel}=0.3\%$                   |      |                |
| 29 | *Current Injection Probes            | Insertion Loss              | C.S.for Current Probes and Injection Probes JJF(communication)030 | (0~60)dB,(300kHz~1GHz)      | $U=0.7$ dB                        |      |                |
|    |                                      | Transfer Impedance          |                                                                   | (-60~30)dB $\Omega$         | $U=0.7$ dB                        |      |                |
|    |                                      | Transfer Admittance         |                                                                   | (-30~60)dBS                 | $U=0.7$ dB                        |      |                |
| 30 | *Differential Probes of Oscilloscope | Common-mode rejection ratio | C.S.for Differential Probes of Oscilloscope JJF(Electronic)30306  | (18~90)dB                   | $U=1.5$ dB                        |      |                |
|    |                                      | DC attenuation coefficient  |                                                                   | 1~1000,(1V~10kV)            | $U_{rel}=0.2\%$                   |      |                |
|    |                                      | Band width                  |                                                                   | 1kHz~1GHz                   | $U_{rel}=4\%$                     |      |                |
|    |                                      | RiSe time                   |                                                                   | 150ps~100 $\mu$ s           | $U_{rel}=6.4\%$                   |      |                |
|    |                                      | Input Impedance             |                                                                   | 50 $\Omega$ ~ 100M $\Omega$ | $U_{rel}=0.6\%$                   |      |                |

No. CNAS L0730

第 162 页 共 163 页



The scope of the accreditation in Chinese remains the definitive version.

| №  | Instrument                                 | Measurand              | Calibration Method                                                      | Range                      | Expanded Uncertainty<br>( $k=2$ ) | Note | Effective Date |
|----|--------------------------------------------|------------------------|-------------------------------------------------------------------------|----------------------------|-----------------------------------|------|----------------|
| 31 | *High Voltage<br>Probes of<br>Oscilloscope | Voltage decay<br>ratio | C.S.for High Voltage Probes<br>of Oscilloscope<br>JJF(Electronic )30304 | 1~1000,(1kV~100kV)         | $U_{rel}=0.2\%$                   |      |                |
|    |                                            | Band width             |                                                                         | 1kHz~1GHz                  | $U_{rel}=4\%$                     |      |                |
|    |                                            | Input<br>capacitance   |                                                                         | 1pF~100pF, 1kHz            | $U_{rel}=0.1\%$                   |      |                |
|    |                                            | Input<br>Impedance     |                                                                         | 50 $\Omega$ ~100M $\Omega$ | $U_{rel}=0.6\%$                   |      |                |



No. CNAS L0730

第 163 页 共 163 页

The scope of the accreditation in Chinese remains the definitive version.

在线扫码获取验证